

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040424\
 Data File : BN030641.D
 Acq On : 04 Apr 2024 00:02
 Operator : MA/JU
 Sample : P1849-03MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0LY5MSD

Quant Time: Apr 04 00:50:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Mar 31 12:24:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.693	152	281160	20.000	ng/u1	0.00
20) Naphthalene-d8	10.481	136	1226314	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.339	164	783613	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.086	188	1519000	20.000	ng/u1	0.00
79) Chrysene-d12	21.327	240	888883	20.000	ng/u1	0.00
88) Perylene-d12	24.274	264	975565	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.199	96	12603	2.102	ng/uL	0.00
4) Pyridine-d5	3.605	84	73340	4.079	ng/u1	0.00
7) Phenol-d5	6.863	99	292176	12.846	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.034	67	174634	12.431	ng/u1	0.00
11) 2-Chlorophenol-d4	7.228	132	246247	14.272	ng/u1	0.00
15) 4-Methylphenol-d8	8.399	113	224305	11.767	ng/u1	0.00
21) Nitrobenzene-d5	8.852	128	124714	14.660	ng/u1	0.00
24) 2-Nitrophenol-d4	9.563	143	132302	17.001	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.099	165	265888	15.606	ng/u1	0.00
31) 4-Chloroaniline-d4	10.616	131	16050	0.572	ng/u1	0.00
46) Dimethylphthalate-d6	13.751	166	786005	15.023	ng/u1	0.00
49) Acenaphthylene-d8	14.028	160	893945	14.861	ng/u1	0.00
54) 4-Nitrophenol-d4	14.545	143	131205	13.111	ng/u1	0.00
60) Fluorene-d10	15.334	176	693886	15.156	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.457	200	77334	11.534	ng/u1	0.00
73) Anthracene-d10	17.186	188	1013064	15.479	ng/u1	0.00
81) Pyrene-d10	19.486	212	969098	20.278	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.074	264	764404	17.005	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.228	88	44141	6.871	ng/uL	99
5) Pyridine	3.622	79	225236	12.415	ng/u1	97
6) Benzaldehyde	6.846	77	249678	23.573	ng/u1	97
8) Phenol	6.887	94	468860	19.788	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.128	93	374414	19.378	ng/u1	98
12) 2-Chlorophenol	7.258	128	386171	20.864	ng/u1	98
13) 2-Methylphenol	8.134	108	319220	17.484	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.216	45	508525	18.691	ng/u1	99
16) Acetophenone	8.516	105	597566	19.855	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.499	70	284206	18.749	ng/u1	97
18) 4-Methylphenol	8.463	108	364935	17.854	ng/u1	99
19) Hexachloroethane	8.763	117	165618	21.342	ng/u1	99
22) Nitrobenzene	8.893	77	438260	21.104	ng/u1	100
23) Isophorone	9.416	82	842460	20.342	ng/u1	99
25) 2-Nitrophenol	9.599	139	217724	23.805	ng/u1	98
26) 2,4-Dimethylphenol	9.657	107	136784	6.729	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.899	93	529318	20.102	ng/u1	99
29) 2,4-Dichlorophenol	10.128	162	387554	22.286	ng/u1	99
30) Naphthalene	10.528	128	1326414	21.089	ng/u1	100
32) 4-Chloroaniline	10.646	127	254492	9.415	ng/u1	99
33) Hexachlorobutadiene	10.810	225	248929	22.742	ng/u1	93
34) Caprolactam	11.422	113	126778	19.515	ng/u1	93
35) 4-Chloro-3-methylphenol	11.769	107	413940	21.134	ng/u1	99
36) 2-Methylnaphthalene	12.146	142	924332	21.044	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040424\
 Data File : BN030641.D
 Acq On : 04 Apr 2024 00:02
 Operator : MA/JU
 Sample : P1849-03MSD
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E0LY5MSD

Quant Time: Apr 04 00:50:23 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Mar 31 12:24:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.369	142	918461	20.599	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.516	216	469447	23.214	ng/ul	97
40) Hexachlorocyclopentadiene	12.493	237	71044	6.186	ng/ul	98
41) 2,4,6-Trichlorophenol	12.763	196	314772	24.909	ng/ul	99
42) 2,4,5-Trichlorophenol	12.834	196	345696	24.223	ng/ul	95
43) 1,1'-Biphenyl	13.169	154	1195452	22.382	ng/ul	97
44) 2-Chloronaphthalene	13.210	162	946312	22.507	ng/ul	99
45) 2-Nitroaniline	13.416	65	263150	23.518	ng/ul	100
47) Dimethylphthalate	13.804	163	1196843	22.145	ng/ul	99
48) 2,6-Dinitrotoluene	13.922	165	240362	23.988	ng/ul	97
50) Acenaphthylene	14.057	152	1515648	21.807	ng/ul	100
51) 3-Nitroaniline	14.245	138	199264	18.736	ng/ul	97
52) Acenaphthene	14.404	153	1062794	22.946	ng/ul	99
53) 2,4-Dinitrophenol	14.451	184	75831	16.477	ng/ul	91
55) 4-Nitrophenol	14.557	109	162585	20.506	ng/ul	98
56) Dibenzofuran	14.739	168	1439276	22.355	ng/ul	98
57) 2,4-Dinitrotoluene	14.710	165	360328	24.631	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	14.963	232	289588	24.432	ng/ul	96
59) Diethylphthalate	15.169	149	1251185	22.789	ng/ul	99
61) Fluorene	15.392	166	1165105	22.767	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.387	204	554367	22.080	ng/ul	97
63) 4-Nitroaniline	15.416	138	218745	21.156	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.469	198	143222	19.880	ng/ul	99
67) N-Nitrosodiphenylamine	15.604	169	997471	24.112	ng/ul	100
68) 4-Bromophenyl-phenylether	16.281	248	367434	25.474	ng/ul	90
69) Hexachlorobenzene	16.386	284	394766	23.242	ng/ul	98
70) Atrazine	16.563	200	360044	25.111	ng/ul	99
71) Pentachlorophenol	16.733	266	225762	22.544	ng/ul	100
72) Phenanthrene	17.133	178	2457405	31.488	ng/ul	99
74) Anthracene	17.222	178	1919748	24.333	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.128	216	475261	25.467	ng/ul	96
76) Pentachlorobenzene	14.651	250	458962	24.213	ng/ul	93
77) Carbazole	17.492	167	1680758	24.293	ng/ul	100
78) Di-n-butylphthalate	18.063	149	2251364	26.674	ng/ul	100
80) Fluoranthene	19.151	202	3032092	53.372	ng/ul	99
82) Pyrene	19.516	202	2815466	46.926	ng/ul	98
83) Butylbenzylphthalate	20.427	149	823350	35.935	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.239	252	207222	12.687	ng/ul#	97
85) Benzo(a)anthracene	21.310	228	1840374	31.023	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.245	149	1720228	50.910	ng/ul	99
87) Chrysene	21.374	228	1803010	32.201	ng/ul	100
89) Di-n-octyl phthalate	22.363	149	1731600	28.029	ng/ul	100
90) Benzo(b)fluoranthene	23.351	252	1926211	33.550	ng/ul	97
91) Benzo(k)fluoranthene	23.410	252	1613960	27.680	ng/ul	97
93) Benzo(a)pyrene	24.139	252	1683306	31.037	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.598	276	1931845	32.951	ng/ul	97
95) Dibenzo(a,h)anthracene	27.662	278	1419150	28.754	ng/ul	97
96) Benzo(g,h,i)perylene	28.639	276	1523579	33.103	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 6.871 ng/uL
 RT: 3.228 min Scan# 41
 Delta R.T. -0.006 min
 Lab File: BN030641.D
 Acq: 04 Apr 2024 00:02




Instrument :
 BNA_N
 ClientSampleId :
 E0LY5MSD


Tgt Ion: 88	Resp: 44141
Ion Ratio	Lower Upper
88	100
43	29.8 23.8 35.8
58	69.8 56.6 84.8



#5
 Pyridine
 Concen: 12.415 ng/uL
 RT: 3.622 min Scan# 108
 Delta R.T. -0.000 min
 Lab File: BN030641.D
 Acq: 04 Apr 2024 00:02

Tgt Ion: 79	Resp: 225236
Ion Ratio	Lower Upper
79	100
52	66.7 56.2 84.4
51	34.6 27.0 40.4









#6
Benzaldehyde
Concen: 23.573 ng/ul
RT: 6.846 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion: 77 Resp: 249678
Ion Ratio Lower Upper
77 100
105 100.2 78.1 117.1
106 96.9 75.6 113.4



#8
Phenol
Concen: 19.788 ng/ul
RT: 6.887 min Scan# 663
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion: 94 Resp: 468860
Ion Ratio Lower Upper
94 100
65 26.8 21.4 32.0
66 35.8 28.5 42.7









#10
Bis(2-Chloroethyl)ether
Concen: 19.378 ng/ul
RT: 7.128 min Scan# 703
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion: 93 Resp: 374414
Ion Ratio Lower Upper
93 100
63 70.7 58.2 87.4
95 33.3 26.0 39.0



#12
2-Chlorophenol
Concen: 20.864 ng/ul
RT: 7.258 min Scan# 726
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion: 128 Resp: 386171
Ion Ratio Lower Upper
128 100
64 43.6 37.0 55.4
130 32.8 26.3 39.5









#13
2-Methylphenol
Concen: 17.484 ng/ul
RT: 8.134 min Scan# 81
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:108 Resp: 319220
Ion Ratio Lower Upper
108 100
107 87.0 71.8 107.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.691 ng/ul
RT: 8.216 min Scan# 889
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion: 45 Resp: 508525
Ion Ratio Lower Upper
45 100
77 14.8 12.1 18.1
79 11.6 8.6 13.0









#16
Acetophenone
Concen: 19.855 ng/ul
RT: 8.516 min Scan# 940
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:105 Resp: 597566
Ion Ratio Lower Upper
105 100
77 75.3 61.2 91.8
51 25.0 21.0 31.4



#17
N-Nitroso-di-n-propylamine
Concen: 18.749 ng/ul
RT: 8.499 min Scan# 937
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion: 70 Resp: 284206
Ion Ratio Lower Upper
70 100
42 55.8 47.4 71.0
101 10.2 7.9 11.9
130 24.5 19.0 28.4









#18
4-Methylphenol
Concen: 17.854 ng/ul
RT: 8.463 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:108 Resp: 364935
Ion Ratio Lower Upper
108 100
107 111.0 88.1 132.1



#19
Hexachloroethane
Concen: 21.342 ng/ul
RT: 8.763 min Scan# 982
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:117 Resp: 165618
Ion Ratio Lower Upper
117 100
201 95.1 76.5 114.7
199 59.6 48.5 72.7









#22
Nitrobenzene
Concen: 21.104 ng/ul
RT: 8.893 min Scan# 1004
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion: 77 Resp: 438260
Ion Ratio Lower Upper
77 100
123 46.9 37.5 56.3
65 13.2 10.6 15.8



#23
Isophorone
Concen: 20.342 ng/ul
RT: 9.416 min Scan# 1093
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion: 82 Resp: 842460
Ion Ratio Lower Upper
82 100
95 7.6 5.7 8.5
138 18.8 14.6 21.8









#25
2-Nitrophenol
Concen: 23.805 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:139 Resp: 217724
Ion Ratio Lower Upper
139 100
65 48.7 39.5 59.3
109 31.5 23.8 35.8



#26
2,4-Dimethylphenol
Concen: 6.729 ng/ul
RT: 9.657 min Scan# 1134
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:107 Resp: 136784
Ion Ratio Lower Upper
107 100
121 50.6 41.6 62.4
122 87.0 71.3 106.9





#27

Bis(2-Chloroethoxy)methane

Concen: 20.102 ng/ul

RT: 9.899 min Scan# 1175

Delta R.T. -0.006 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Instrument :

BNA_N

ClientSampleId :

E0LY5MSD





Tgt Ion: 93 Resp: 529318

Ion Ratio Lower Upper

93 100

95 32.1 25.4 38.0

123 11.6 9.4 14.0



#29

2,4-Dichlorophenol

Concen: 22.286 ng/ul

RT: 10.128 min Scan# 1214

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02




Tgt Ion:162 Resp: 387554


Ion Ratio Lower Upper

162 100

164 65.7 51.8 77.8

98 35.4 28.0 42.0









#30
Naphthalene
Concen: 21.089 ng/ul
RT: 10.528 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
EOLY5MSD


Tgt Ion:128 Resp: 1326414
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 9.415 ng/ul
RT: 10.646 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02


Tgt Ion:127 Resp: 254492
Ion Ratio Lower Upper
127 100
129 32.1 25.3 37.9








#33
Hexachlorobutadiene
Concen: 22.742 ng/ul
RT: 10.810 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02


Instrument :
BNA_N
ClientSampleId :
E0LY5MSD





Tgt Ion:225 Resp: 248929
Ion Ratio Lower Upper
225 100
223 58.1 50.9 76.3
227 59.1 51.6 77.4




#34
Caprolactam
Concen: 19.515 ng/ul
RT: 11.422 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02



Tgt Ion:113 Resp: 126778
Ion Ratio Lower Upper
113 100
55 145.3 124.7 187.1
56 108.2 90.6 135.8









#35
4-Chloro-3-methylphenol
Concen: 21.134 ng/ul
RT: 11.769 min Scan# 1492
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:107 Resp: 413940
Ion Ratio Lower Upper
107 100
144 26.9 21.4 32.0
142 81.2 65.9 98.9



#36
2-Methylnaphthalene
Concen: 21.044 ng/ul
RT: 12.146 min Scan# 1557
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:142 Resp: 924332
Ion Ratio Lower Upper
142 100
141 87.8 70.8 106.2









#37
1-Methylnaphthalene
Concen: 20.599 ng/ul
RT: 12.369 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:142 Resp: 918461
Ion Ratio Lower Upper
142 100
141 92.0 73.0 109.4
116 3.8 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 23.214 ng/ul
RT: 12.516 min Scan# 1620
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:216 Resp: 469447
Ion Ratio Lower Upper
216 100
214 73.5 60.9 91.3
179 20.7 17.4 26.0
108 16.8 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 6.186 ng/ul
RT: 12.493 min Scan# 1616
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:237 Resp: 71044
Ion Ratio Lower Upper
237 100
235 61.6 50.6 76.0
272 13.7 10.2 15.4



#41
2,4,6-Trichlorophenol
Concen: 24.909 ng/ul
RT: 12.763 min Scan# 1662
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:196 Resp: 314772
Ion Ratio Lower Upper
196 100
198 94.6 74.6 111.8
200 30.8 24.8 37.2









#42
2,4,5-Trichlorophenol
Concen: 24.223 ng/ul
RT: 12.834 min Scan# 1674
Delta R.T. 0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:196 Resp: 345696
Ion Ratio Lower Upper
196 100
198 97.6 73.0 109.6
200 31.3 25.2 37.8



#43
1,1'-Biphenyl
Concen: 22.382 ng/ul
RT: 13.169 min Scan# 1731
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:154 Resp: 1195452
Ion Ratio Lower Upper
154 100
153 41.6 31.5 47.3
76 13.1 9.8 14.6









#44
2-Chloronaphthalene
Concen: 22.507 ng/ul
RT: 13.210 min Scan# 1738
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:162 Resp: 946312
Ion Ratio Lower Upper
162 100
164 32.7 25.7 38.5
127 40.1 32.2 48.2



#45
2-Nitroaniline
Concen: 23.518 ng/ul
RT: 13.416 min Scan# 1773
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion: 65 Resp: 263150
Ion Ratio Lower Upper
65 100
92 72.5 57.7 86.5
138 112.8 90.6 135.8





#47

Dimethylphthalate

Concen: 22.145 ng/ul

RT: 13.804 min Scan# 1839

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Instrument :

BNA_N

ClientSampleId :

E0LY5MSD





Tgt Ion:163 Resp: 1196843

Ion Ratio Lower Upper

163 100

194 4.2 3.3 4.9

164 9.9 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 23.988 ng/ul

RT: 13.922 min Scan# 1859

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02




Tgt Ion:165 Resp: 240362


Ion Ratio Lower Upper

165 100

89 56.5 43.5 65.3

121 21.6 15.7 23.5









#50
Acenaphthylene
Concen: 21.807 ng/ul
RT: 14.057 min Scan# 1882
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:152 Resp: 1515648
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 13.5 10.7 16.1



#51
3-Nitroaniline
Concen: 18.736 ng/ul
RT: 14.245 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:138 Resp: 199264
Ion Ratio Lower Upper
138 100
108 11.9 9.6 14.4
92 114.9 89.6 134.4









#52
 Acenaphthene
 Concen: 22.946 ng/ul
 RT: 14.404 min Scan# 1941
 Delta R.T. -0.000 min
 Lab File: BN030641.D
 Acq: 04 Apr 2024 00:02

Instrument :
 BNA_N
 ClientSampleId :
 EOLY5MSD

Tgt Ion:153 Resp: 1062794




Ion	Ratio	Lower	Upper
153	100		
152	47.2	37.6	56.4
154	88.7	70.3	105.5




#53
 2,4-Dinitrophenol
 Concen: 16.477 ng/ul
 RT: 14.451 min Scan# 1949
 Delta R.T. -0.000 min
 Lab File: BN030641.D
 Acq: 04 Apr 2024 00:02

Tgt Ion:184 Resp: 75831

Ion	Ratio	Lower	Upper
184	100		
63	67.4	49.9	74.9
154	70.1	49.5	74.3









#55
4-Nitrophenol
Concen: 20.506 ng/ul
RT: 14.557 min Scan# 1967
Delta R.T. 0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:109 Resp: 162585
Ion Ratio Lower Upper
109 100
139 134.9 111.8 167.6
65 122.9 98.7 148.1



#56
Dibenzofuran
Concen: 22.355 ng/ul
RT: 14.739 min Scan# 1998
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:168 Resp: 1439276
Ion Ratio Lower Upper
168 100
139 38.9 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 24.631 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:165 Resp: 360328
Ion Ratio Lower Upper
165 100
63 39.9 34.7 52.1
182 2.9 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 24.432 ng/ul
RT: 14.963 min Scan# 2036
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:232 Resp: 289588
Ion Ratio Lower Upper
232 100
131 49.9 36.6 55.0
130 2.8 2.2 3.2
166 33.2 26.0 39.0





#59

Diethylphthalate

Concen: 22.789 ng/ul

RT: 15.169 min Scan# 2071

Delta R.T. -0.006 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Instrument :

BNA_N

ClientSampleId :

EOLY5MSD





Tgt Ion:149 Resp: 1251185

Ion Ratio Lower Upper

149 100

177 21.5 17.4 26.2

150 12.5 9.8 14.8



#61

Fluorene

Concen: 22.767 ng/ul

RT: 15.392 min Scan# 2109

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02




Tgt Ion:166 Resp: 1165105


Ion Ratio Lower Upper

166 100

165 99.9 78.2 117.4

167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 22.080 ng/ul
RT: 15.387 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:204 Resp: 554367
Ion Ratio Lower Upper
204 100
206 32.8 25.8 38.6
141 64.5 49.1 73.7



#63
4-Nitroaniline
Concen: 21.156 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:138 Resp: 218745
Ion Ratio Lower Upper
138 100
92 49.9 39.7 59.5
108 76.7 62.6 94.0









#66
4,6-Dinitro-2-methylphenol
Concen: 19.880 ng/ul
RT: 15.469 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:198 Resp: 143222
Ion Ratio Lower Upper
198 100
182 5.0 3.6 5.4
77 28.3 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 24.112 ng/ul
RT: 15.604 min Scan# 2145
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:169 Resp: 997471
Ion Ratio Lower Upper
169 100
168 65.0 51.9 77.9
167 35.9 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 25.474 ng/ul
RT: 16.281 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Instrument :
BNA_N
ClientSampleId :
E0LY5MSD

Tgt Ion:248 Resp: 367434




Ion	Ratio	Lower	Upper
248	100		
250	86.8	80.0	120.0
141	80.4	60.5	90.7




#69
Hexachlorobenzene
Concen: 23.242 ng/ul
RT: 16.386 min Scan# 2278
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:284 Resp: 394766

Ion	Ratio	Lower	Upper
284	100		
142	37.6	28.6	43.0
249	31.5	25.0	37.4









#70
Atrazine
Concen: 25.111 ng/ul
RT: 16.563 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:200 Resp: 360044
Ion Ratio Lower Upper
200 100
173 26.3 21.5 32.3
215 46.2 37.8 56.8



#71
Pentachlorophenol
Concen: 22.544 ng/ul
RT: 16.733 min Scan# 2337
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:266 Resp: 225762
Ion Ratio Lower Upper
266 100
264 64.2 51.0 76.6
268 65.8 52.5 78.7









#72
Phenanthrene
Concen: 31.488 ng/ul
RT: 17.133 min Scan# 2405
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:178 Resp: 2457405
Ion Ratio Lower Upper
178 100
179 15.3 12.4 18.6
176 19.7 15.0 22.6



#74
Anthracene
Concen: 24.333 ng/ul
RT: 17.222 min Scan# 2420
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:178 Resp: 1919748
Ion Ratio Lower Upper
178 100
179 15.4 12.0 18.0
176 18.4 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 25.467 ng/uL
RT: 13.128 min Scan# 1723
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:216 Resp: 475261
Ion Ratio Lower Upper
216 100
214 74.9 63.1 94.7
218 47.7 39.5 59.3



#76
Pentachlorobenzene
Concen: 24.213 ng/uL
RT: 14.651 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:250 Resp: 458962
Ion Ratio Lower Upper
250 100
248 65.2 45.8 68.6
252 63.4 48.7 73.1





#77

Carbazole

Concen: 24.293 ng/ul

RT: 17.492 min Scan# 2466

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Instrument :

BNA_N

ClientSampleId :

EOLY5MSD





Tgt Ion:167 Resp: 1680758

Ion Ratio Lower Upper

167 100

166 20.9 16.6 24.8

139 13.1 10.5 15.7



#78

Di-n-butylphthalate

Concen: 26.674 ng/ul

RT: 18.063 min Scan# 2563

Delta R.T. -0.006 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02




Tgt Ion:149 Resp: 2251364


Ion Ratio Lower Upper

149 100

150 9.0 7.2 10.8

104 5.1 4.1 6.1









#80
Fluoranthene
Concen: 53.372 ng/ul
RT: 19.151 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:202 Resp: 3032092
Ion Ratio Lower Upper
202 100
101 11.9 9.0 13.6
100 8.8 6.8 10.2



#82
Pyrene
Concen: 46.926 ng/ul
RT: 19.516 min Scan# 2810
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:202 Resp: 2815466
Ion Ratio Lower Upper
202 100
101 13.4 11.3 16.9
100 10.8 9.2 13.8









#83
 Butylbenzylphthalate
 Concen: 35.935 ng/ul
 RT: 20.427 min Scan# 2965
 Delta R.T. -0.006 min
 Lab File: BN030641.D
 Acq: 04 Apr 2024 00:02




Instrument :
 BNA_N
 ClientSampleId :
 EOLY5MSD


Tgt Ion:149 Resp: 823350
 Ion Ratio Lower Upper
 149 100
 91 69.3 58.2 87.2
 206 20.9 15.4 23.0



#84
 3,3'-Dichlorobenzidine
 Concen: 12.687 ng/ul
 RT: 21.239 min Scan# 3103
 Delta R.T. -0.000 min
 Lab File: BN030641.D
 Acq: 04 Apr 2024 00:02

Tgt Ion:252 Resp: 207222
 Ion Ratio Lower Upper
 252 100
 254 62.9 51.1 76.7
 126 15.5 9.3 13.9#









#85
Benzo(a)anthracene
Concen: 31.023 ng/ul
RT: 21.310 min Scan# 3115
Delta R.T. -0.000 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:228 Resp: 1840374
Ion Ratio Lower Upper
228 100
229 20.0 15.8 23.6
226 26.7 20.5 30.7



#86
Bis(2-ethylhexyl)phthalate
Concen: 50.910 ng/ul
RT: 21.245 min Scan# 3104
Delta R.T. -0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:149 Resp: 1720228
Ion Ratio Lower Upper
149 100
167 27.4 21.4 32.0
279 4.5 3.5 5.3





#87

Chrysene

Concen: 32.201 ng/ul

RT: 21.374 min Scan# 3125

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Instrument :

BNA_N

ClientSampleId :

EOLY5MSD





Tgt Ion:228 Resp: 1803010

Ion Ratio Lower Upper

228 100

226 30.4 24.3 36.5

229 19.7 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 28.029 ng/ul




RT: 22.363 min Scan# 3294


Delta R.T. -0.012 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Tgt Ion:149 Resp: 1731600









#90
Benzo(b)fluoranthene
Concen: 33.550 ng/ul
RT: 23.351 min Scan# 3459
Delta R.T. 0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02




Instrument :
BNA_N
ClientSampleId :
E0LY5MSD


Tgt Ion:252 Resp: 1926211
Ion Ratio Lower Upper
252 100
253 21.8 16.7 25.1
125 12.2 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 27.680 ng/ul
RT: 23.410 min Scan# 3472
Delta R.T. 0.006 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:252 Resp: 1613960
Ion Ratio Lower Upper
252 100
253 21.9 16.9 25.3
125 12.7 8.5 12.7





#93

Benzo(a)pyrene

Concen: 31.037 ng/ul

RT: 24.139 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02

Instrument :

BNA_N

ClientSampleId :

E0LY5MSD





Tgt Ion:252 Resp: 1683306

Ion Ratio Lower Upper

252 100

253 21.9 16.7 25.1

125 13.3 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 32.951 ng/ul

RT: 27.598 min Scan# 4184

Delta R.T. 0.012 min

Lab File: BN030641.D

Acq: 04 Apr 2024 00:02




Tgt Ion:276 Resp: 1931845


Ion Ratio Lower Upper

276 100

138 22.4 17.8 26.8

277 25.4 18.4 27.6









#95
Dibenzo(a,h)anthracene
Concen: 28.754 ng/ul
RT: 27.662 min Scan# 4195
Delta R.T. 0.012 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Instrument :
BNA_N
ClientSampleId :
E0LY5MSD

Tgt Ion:278 Resp: 1419150

Ion	Ratio	Lower	Upper
278	100		
139	18.0	13.2	19.8
279	23.4	17.4	26.2



#96
Benzo(g,h,i)perylene
Concen: 33.103 ng/ul
RT: 28.639 min Scan# 4361
Delta R.T. 0.023 min
Lab File: BN030641.D
Acq: 04 Apr 2024 00:02

Tgt Ion:276 Resp: 1523579

Ion	Ratio	Lower	Upper
276	100		
138	22.6	19.3	28.9
277	23.5	20.1	30.1

