




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040522\
Data File : BN019347.D
Acq On : 05 Apr 2022 11:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 06 03:22:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019347.D\data.ms

(70) Pentachlorophenol (C)




16.857min (-0.015) 0.00 ng

response 324872

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.76
264.00	61.60	64.85
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040522\
Data File : BN019347.D
Acq On : 05 Apr 2022 11:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 06 03:22:42 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 23 16:56:32 2022
Response via : Initial Calibration



TIC: BN019347.D\data.ms

(77) Benzidine**19.445min (-0.021) 744033.50 ng****response 1261431**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.24
183.00	13.20	11.81
0.00	0.00	0.00