

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040621\
 Data File : BN014207.D
 Acq On : 05 Apr 2021 17:14
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 05 19:00:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 05 18:58:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.644	152	5937	0.40	ng	0.00
7) Naphthalene-d8	10.429	136	23219	0.40	ng	0.00
13) Acenaphthene-d10	14.283	164	14211	0.40	ng	0.00
19) Phenanthrene-d10	17.031	188	31356	0.40	ng	0.00
29) Chrysene-d12	21.226	240	31651	0.40	ng	# 0.00
36) Perylene-d12	23.425	264	33104	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	1634	0.11	ng	0.00
5) Phenol-d6	6.814	99	2095	0.11	ng	0.00
8) Nitrobenzene-d5	8.794	82	1664	0.10	ng	0.00
11) 2-Methylnaphthalene-d10	12.017	152	5239	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.780	330	492	0.11	ng	0.00
15) 2-Fluorobiphenyl	12.900	172	5419	0.10	ng	0.00
27) Fluoranthene-d10	19.066	212	11905	0.10	ng	0.00
31) Terphenyl-d14	19.672	244	7576	0.10	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.174	88	904	0.12	ng	92
3) n-Nitrosodimethylamine	3.521	42	422	0.08	ng	# 90
6) bis(2-Chloroethyl)ether	7.080	93	1757	0.11	ng	100
9) Naphthalene	10.467	128	6268	0.10	ng	97
10) Hexachlorobutadiene	10.758	225	1158	0.10	ng	# 100
12) 2-Methylnaphthalene	12.093	142	4388	0.11	ng	99
16) Acenaphthylene	14.004	152	5256	0.09	ng	99
17) Acenaphthene	14.346	154	4113	0.10	ng	99
18) Fluorene	15.336	166	5163	0.10	ng	99
20) 4,6-Dinitro-2-methylph...	15.412	198	354	0.09	ng	# 21
21) 4-Bromophenyl-phenylether	16.227	248	1729	0.10	ng	98
22) Hexachlorobenzene	16.337	284	1917	0.10	ng	99
23) Atrazine	16.495	200	1084	0.10	ng	94
24) Pentachlorophenol	16.690	266	421	0.12	ng	96
25) Phenanthrene	17.067	178	9038	0.10	ng	99
26) Anthracene	17.164	178	7176	0.09	ng	99
28) Fluoranthene	19.096	202	10049	0.10	ng	100
30) Pyrene	19.459	202	10553	0.10	ng	100
32) Benzo(a)anthracene	21.205	228	9630	0.10	ng	99
33) Chrysene	21.258	228	10682	0.10	ng	98
34) Bis(2-ethylhexyl)phtha...	21.141	149	3809	0.13	ng	99
35) Indeno(1,2,3-cd)pyrene	25.629	276	13215	0.11	ng	99
37) Benzo(b)fluoranthene	22.809	252	11577	0.09	ng	96
38) Benzo(k)fluoranthene	22.809	252	11577	0.09	ng	# 93
39) Benzo(a)pyrene	23.329	252	10738	0.10	ng	# 75
40) Dibenzo(a,h)anthracene	25.639	278	10894	0.09	ng	94
41) Benzo(g,h,i)perylene	26.303	276	11415	0.09	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040621\
Data File : BN014207.D
Acq On : 05 Apr 2021 17:14
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: Apr 05 19:00:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 05 18:58:13 2021
Response via : Initial Calibration





Abundance Scan 604 (7.644 min): BN014209.D\data.ms (-599) (#1)




1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.644 min Scan# 604
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:152 Resp: 5937
Ion Ratio Lower Upper
152 100
150 157.5 127.4 191.0
115 61.0 50.3 75.5



Abundance Scan 604 (7.644 min): BN014207.D\data.ms (-592) (-)





Abundance Scan 49 (3.174 min): BN014209.D\data.ms (-45) (#2)





1,4-Dioxane
Concen: 0.12 ng
RT: 3.174 min Scan# 49
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion: 88 Resp: 904
Ion Ratio Lower Upper
88 100
58 66.0 59.4 89.0
43 30.0 25.8 38.8



Abundance Scan 49 (3.174 min): BN014207.D\data.ms (-37) (-)









n-Nitrosodimethylamine
Concen: 0.08 ng
RT: 3.521 min Scan# 92
Delta R.T. 0.008 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 42 Resp: 422
Ion Ratio Lower Upper
42 100
74 128.2 95.1 142.7
44 21.6 4.2 6.2#



2-Fluorophenol
Concen: 0.11 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion: 112 Resp: 1634
Ion Ratio Lower Upper
112 100
64 61.9 48.2 72.4
63 32.9 25.5 38.3









Phenol-d6
 Concen: 0.11 ng
 RT: 6.814 min Scan# 501
 Delta R.T. 0.000 min
 Lab File: BN014207.D
 Acq: 05 Apr 2021 17:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:	99	Resp:	2095
Ion	Ratio	Lower	Upper
99	100		
42	23.1	18.2	27.2
71	34.5	26.9	40.3



bis(2-Chloroethyl)ether
 Concen: 0.11 ng
 RT: 7.080 min Scan# 534
 Delta R.T. 0.000 min
 Lab File: BN014207.D
 Acq: 05 Apr 2021 17:14

Tgt Ion:	93	Resp:	1757
Ion	Ratio	Lower	Upper
93	100		
63	81.4	65.0	97.6
95	33.1	26.8	40.2









Naphthalene-d8
 Concen: 0.40 ng
 RT: 10.429 min Scan# 854
 Delta R.T. 0.000 min
 Lab File: BN014207.D
 Acq: 05 Apr 2021 17:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	136	Resp:	23219
Ion	Ratio	Lower	Upper
136	100		
137	11.5	9.2	13.8
54	5.1	4.3	6.5
68	4.2	3.6	5.4




Nitrobenzene-d5
 Concen: 0.10 ng
 RT: 8.794 min Scan# 725
 Delta R.T. 0.000 min
 Lab File: BN014207.D
 Acq: 05 Apr 2021 17:14

Tgt Ion:	82	Resp:	1664
Ion	Ratio	Lower	Upper
82	100		
128	46.8	34.1	51.1
54	51.3	39.5	59.3





Abundance Scan 857 (10.467 min): BN014209.D\data.ms (-857) (-)




Naphthalene
Concen: 0.10 ng
RT: 10.467 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion:	128	Resp:	6268
Ion	Ratio	Lower	Upper
128	100		
129	12.1	8.9	13.3
127	14.6	10.7	16.1



Abundance Scan 857 (10.467 min): BN014207.D\data.ms (-849) (-)





Abundance Scan 880 (10.758 min): BN014209.D\data.ms (-875) (-)




Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.758 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Tgt Ion:	225	Resp:	1158
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.2	51.2	76.8



Abundance Scan 880 (10.758 min): BN014207.D\data.ms (-872) (-)






Abundance Scan 1054 (12.017 min): BN014209.D\data.ms (-1039) (-)




2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.017 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Tgt Ion:152 Resp: 5239
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2






Abundance Scan 1071 (12.093 min): BN014209.D\data.ms (-1039) (-)




2-Methylnaphthalene
Concen: 0.11 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:142 Resp: 4388
Ion Ratio Lower Upper
142 100
141 88.1 70.5 105.7
115 36.8 28.3 42.5






Abundance Scan 1304 (14.283 min): BN014209.D\data.ms (-1288) (-)




Acenaphthene-d10
Concen: 0.40 ng
RT: 14.283 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:164 Resp: 14211
Ion Ratio Lower Upper
164 100
162 101.7 81.2 121.8
160 46.6 37.5 56.3






Abundance Scan 1422 (15.780 min): BN014209.D\data.ms (-1414) (-)




2,4,6-Tribromophenol
Concen: 0.11 ng
RT: 15.780 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:330 Resp: 492
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.2 30.5 45.7






Abundance Scan 1195 (12.900 min): BN014209.D\data.ms (-1187) (-)




2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.900 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	172	Resp:	5419
Ion	Ratio	Lower	Upper
172	100		
171	37.0	28.0	42.0
170	25.0	19.1	28.7






Abundance Scan 1282 (14.004 min): BN014209.D\data.ms (-1266) (-)



Acenaphthylene
Concen: 0.09 ng
RT: 14.004 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Tgt Ion:	152	Resp:	5256
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.5	23.3
153	13.2	10.2	15.4






Abundance Scan 1309 (14.346 min): BN014209.D\data.ms (-1303) (-)

Acenaphthene
Concen: 0.10 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	154	Resp:	4113
Ion	Ratio	Lower	Upper
154	100		
153	110.9	89.4	134.2
152	54.4	43.9	65.9






Abundance Scan 1387 (15.336 min): BN014209.D\data.ms (-1315) (-)

Fluorene
Concen: 0.10 ng
RT: 15.336 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:	166	Resp:	5163
Ion	Ratio	Lower	Upper
166	100		
165	98.9	78.1	117.1
167	13.6	10.8	16.2






Abundance Scan 1524 (17.031 min): BN014209.D\data.ms (-1519) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.031 min Scan# 1524
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1






Tgt Ion:	188	Resp:	31356
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	8.1	6.9	10.3




Abundance Scan 1393 (15.412 min): BN014209.D\data.ms (-1329) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.09 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	198	Resp:	354
Ion	Ratio	Lower	Upper
198	100		
51	149.8	51.7	77.5#
105	73.4	38.2	57.2#





Abundance Scan 1458 (16.228 min): BN014209.D\data.ms (-1423) (-)




4-Bromophenyl-phenylether
Concen: 0.10 ng
RT: 16.227 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	248	Resp:	1729
Ion	Ratio	Lower	Upper
248	100		
250	95.6	74.6	111.8
141	62.9	49.4	74.0



Abundance Scan 1458 (16.227 min): BN014207.D\data.ms (-1450) (-)





Abundance Scan 1467 (16.337 min): BN014209.D\data.ms (-1423) (-)




Hexachlorobenzene
Concen: 0.10 ng
RT: 16.337 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Tgt Ion:	284	Resp:	1917
Ion	Ratio	Lower	Upper
284	100		
142	36.9	29.0	43.4
249	31.4	25.2	37.8



Abundance Scan 1467 (16.337 min): BN014207.D\data.ms (-1459) (-)





Abundance Scan 1480 (16.495 min): BN014209.D\data.ms (-1425) (-)




Atrazine
Concen: 0.10 ng
RT: 16.495 min Scan# 1480
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	200	Resp:	1084
Ion	Ratio	Lower	Upper
200	100		
173	32.5	22.9	34.3
215	50.8	37.7	56.5



Abundance Scan 1480 (16.495 min): BN014207.D\data.ms (-1439) (-)





Abundance Scan 1496 (16.690 min): BN014209.D\data.ms (-1426) (-)





Pentachlorophenol
Concen: 0.12 ng
RT: 16.690 min Scan# 1496
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	266	Resp:	421
Ion	Ratio	Lower	Upper
266	100		
264	65.8	49.0	73.4
268	64.6	50.3	75.5



Abundance Scan 1496 (16.690 min): BN014207.D\data.ms (-1488) (-)









Phenanthrene
Concen: 0.10 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	178	Resp:	9038
Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.1	22.7
179	15.8	12.4	18.6




Anthracene
Concen: 0.09 ng
RT: 17.164 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	178	Resp:	7176
Ion	Ratio	Lower	Upper
178	100		
176	18.9	14.6	21.8
179	15.6	12.2	18.2





Abundance Scan 1816 (19.064 min): BN014209.D\data.ms (-1827) (-)




Fluoranthene-d10
Concen: 0.10 ng
RT: 19.066 min Scan# 1818
Delta R.T. 0.002 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 11905
Ion Ratio Lower Upper
212 100
106 12.6 10.0 15.0
104 7.1 5.6 8.4



Abundance Scan 1818 (19.066 min): BN014207.D\data.ms (-1777) (-)





Abundance Scan 1822 (19.094 min): BN014209.D\data.ms (-1828) (-)




Fluoranthene
Concen: 0.10 ng
RT: 19.096 min Scan# 1824
Delta R.T. 0.002 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Tgt Ion:202 Resp: 10049
Ion Ratio Lower Upper
202 100
101 11.0 8.7 13.1
203 17.0 13.8 20.6



Abundance Scan 1824 (19.096 min): BN014207.D\data.ms (-1783) (-)






Abundance Scan 2120 (21.218 min): BN014209.D\data.ms (-2120) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.226 min Scan# 2122
Delta R.T. 0.008 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	240	Resp:	31651
Ion	Ratio	Lower	Upper
240	100		
120	5.5	10.3	15.5#
236	25.2	21.4	32.0






Abundance Scan 1896 (19.461 min): BN014209.D\data.ms (-1896) (-)



Pyrene
Concen: 0.10 ng
RT: 19.459 min Scan# 1897
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Tgt Ion:	202	Resp:	10553
Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.2	24.4
203	17.5	13.9	20.9






Abundance Scan 1938 (19.670 min): BN014209.D\data.ms (-1938) (-)

Terphenyl-d14
Concen: 0.10 ng
RT: 19.672 min Scan# 1940
Delta R.T. 0.002 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1







Tgt Ion:	244	Resp:	7576
Ion	Ratio	Lower	Upper
244	100		
212	7.9	6.0	9.0
122	10.6	9.4	14.0




Abundance Scan 2119 (21.207 min): BN014209.D\data.ms (-2119) (-)

Benzo(a)anthracene
Concen: 0.10 ng
RT: 21.205 min Scan# 2120
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	228	Resp:	9630
Ion	Ratio	Lower	Upper
228	100		
226	27.2	21.1	31.7
229	20.1	15.7	23.5






Abundance Scan 2124 (21.261 min): BN014209.D\data.ms (-2116) (-)




Chrysene
Concen: 0.10 ng
RT: 21.258 min Scan# 2125
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	228	Resp:	10682
Ion	Ratio	Lower	Upper
228	100		
226	30.0	23.0	34.6
229	20.0	15.8	23.8






Abundance Scan 2113 (21.144 min): BN014209.D\data.ms (-2116) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.13 ng
RT: 21.141 min Scan# 2114
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	149	Resp:	3809
Ion	Ratio	Lower	Upper
149	100		
167	27.3	21.5	32.3
279	4.6	3.1	4.7




Abundance Scan 3174 (25.632 min): BN014209.D\data.ms (-3135) (-)





Indeno(1,2,3-cd)pyrene
Concen: 0.11 ng
RT: 25.629 min Scan# 3175
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.8	18.6	28.0
277	24.6	19.8	29.8





Abundance Scan 2530 (23.427 min): BN014209.D\data.ms (-2536) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.425 min Scan# 2531
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.4	29.2
265	25.8	20.7	31.1





Abundance Scan 2337 (22.767 min): BN014209.D\data.ms (-227) (-)




Benzo(b)fluoranthene
Concen: 0.09 ng
RT: 22.809 min Scan# 2351
Delta R.T. 0.042 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 11577
Ion Ratio Lower Upper
252 100
253 26.7 18.9 28.3
125 12.2 10.1 15.1



Abundance Scan 2351 (22.809 min): BN014207.D\data.ms (-2280) (-)





Abundance Scan 2350 (22.811 min): BN014209.D\data.ms (-228) (-)




Benzo(k)fluoranthene
Concen: 0.09 ng
RT: 22.809 min Scan# 2351
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:252 Resp: 11577
Ion Ratio Lower Upper
252 100
253 26.7 18.5 27.7
125 12.2 7.9 11.9#




Abundance Scan 2351 (22.809 min): BN014207.D\data.ms (-2264) (-)





Abundance Scan 2501 (23.328 min): BN014209.D\data.ms (-2439) (-)

Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.329 min Scan# 2503
Delta R.T. 0.001 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	252	Resp:	10738
Ion	Ratio	Lower	Upper
252	100		
253	32.1	19.4	29.0#
125	35.4	15.0	22.6#





Abundance Scan 2503 (23.329 min): BN014207.D\data.ms (-2444) (-)




Abundance Scan 3177 (25.642 min): BN014209.D\data.ms (-3142) (-)

Dibenzo(a,h)anthracene
Concen: 0.09 ng
RT: 25.639 min Scan# 3178
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14


Tgt Ion:	278	Resp:	10894
Ion	Ratio	Lower	Upper
278	100		
139	17.8	12.7	19.1
279	27.4	19.4	29.2



Abundance Scan 3178 (25.639 min): BN014207.D\data.ms (-3155) (-)




Abundance Scan 3371 (26.306 min): BN014209.D\data.ms (-3340) (-)




Benzo(g,h,i)perylene
Concen: 0.09 ng
RT: 26.303 min Scan# 3372
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Abundance Scan 3372 (26.303 min): BN014207.D\data.ms



Tgt Ion:	276	Resp:	11415
Ion	Ratio	Lower	Upper
276	100		
277	24.9	19.2	28.8
138	22.8	16.8	25.2



Abundance Scan 3372 (26.303 min): BN014207.D\data.ms (-3314) (-)

