

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040621\
 Data File : BN014208.D
 Acq On : 05 Apr 2021 17:50
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Apr 05 18:59:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 05 18:58:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.644	152	5165	0.40	ng	0.00
7) Naphthalene-d8	10.416	136	20190	0.40	ng	#-0.01
13) Acenaphthene-d10	14.283	164	12284	0.40	ng	0.00
19) Phenanthrene-d10	17.031	188	26850	0.40	ng	0.00
29) Chrysene-d12	21.218	240	27610	0.40	ng	0.00
36) Perylene-d12	23.428	264	29817	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	2784	0.22	ng	0.00
5) Phenol-d6	6.814	99	3567	0.22	ng	0.00
8) Nitrobenzene-d5	8.794	82	2807	0.20	ng	0.00
11) 2-Methylnaphthalene-d10	12.017	152	9082	0.21	ng	0.00
14) 2,4,6-Tribromophenol	15.780	330	812	0.21	ng	0.00
15) 2-Fluorobiphenyl	12.900	172	9368	0.20	ng	0.00
27) Fluoranthene-d10	19.064	212	20713	0.20	ng	0.00
31) Terphenyl-d14	19.670	244	13095	0.20	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.174	88	1347	0.21	ng	97
3) n-Nitrosodimethylamine	3.529	42	802	0.17	ng	# 96
6) bis(2-Chloroethyl)ether	7.080	93	3043	0.21	ng	99
9) Naphthalene	10.467	128	10921	0.20	ng	99
10) Hexachlorobutadiene	10.758	225	2005	0.20	ng	# 98
12) 2-Methylnaphthalene	12.093	142	7514	0.21	ng	99
16) Acenaphthylene	14.004	152	9193	0.18	ng	99
17) Acenaphthene	14.346	154	7111	0.20	ng	100
18) Fluorene	15.336	166	9015	0.20	ng	99
20) 4,6-Dinitro-2-methylph...	15.412	198	600	0.18	ng	# 65
21) 4-Bromophenyl-phenylether	16.228	248	2952	0.19	ng	99
22) Hexachlorobenzene	16.337	284	3323	0.20	ng	99
23) Atrazine	16.496	200	1932	0.21	ng	99
24) Pentachlorophenol	16.690	266	635	0.21	ng	96
25) Phenanthrene	17.068	178	15694	0.20	ng	100
26) Anthracene	17.165	178	12451	0.19	ng	100
28) Fluoranthene	19.094	202	17561	0.20	ng	100
30) Pyrene	19.461	202	18050	0.20	ng	100
32) Benzo(a)anthracene	21.208	228	16237	0.19	ng	99
33) Chrysene	21.261	228	19603	0.21	ng	100
34) Bis(2-ethylhexyl)phtha...	21.144	149	5612	0.22	ng	98
35) Indeno(1,2,3-cd)pyrene	25.632	276	24882	0.23	ng	100
37) Benzo(b)fluoranthene	22.808	252	21547	0.19	ng	97
38) Benzo(k)fluoranthene	22.808	252	21547	0.19	ng	98
39) Benzo(a)pyrene	23.328	252	19351	0.20	ng	# 90
40) Dibenzo(a,h)anthracene	25.646	278	20858	0.20	ng	98
41) Benzo(g,h,i)perylene	26.303	276	21457	0.19	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040621\
Data File : BN014208.D
Acq On : 05 Apr 2021 17:50
Operator : CG/JU
Sample : SSTDIC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Quant Time: Apr 05 18:59:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 05 18:58:13 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.644 min Scan# 604
 Delta R.T. -0.000 min
 Lab File: BN014208.D
 Acq: 05 Apr 2021 17:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	152	Resp:	5165
Ion	Ratio	Lower	Upper
152	100		
150	158.8	127.4	191.0
115	62.2	50.3	75.5



1,4-Dioxane
 Concen: 0.21 ng
 RT: 3.174 min Scan# 49
 Delta R.T. 0.000 min
 Lab File: BN014208.D
 Acq: 05 Apr 2021 17:50

Tgt Ion:	88	Resp:	1347
Ion	Ratio	Lower	Upper
88	100		
58	77.4	59.4	89.0
43	33.1	25.8	38.8









#3
n-Nitrosodimethylamine
Concen: 0.17 ng
RT: 3.529 min Scan# 93
Delta R.T. 0.016 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	42	Resp:	802
Ion	Ratio	Lower	Upper
42	100		
74	115.3	95.1	142.7
44	9.1	4.2	6.2#



#4
2-Fluorophenol
Concen: 0.22 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	112	Resp:	2784
Ion	Ratio	Lower	Upper
112	100		
64	59.9	48.2	72.4
63	32.1	25.5	38.3









Phenol-d6
 Concen: 0.22 ng
 RT: 6.814 min Scan# 501
 Delta R.T. 0.000 min
 Lab File: BN014208.D
 Acq: 05 Apr 2021 17:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:	99	Resp:	3567
Ion	Ratio	Lower	Upper
99	100		
42	22.2	18.2	27.2
71	34.5	26.9	40.3



bis(2-Chloroethyl)ether
 Concen: 0.21 ng
 RT: 7.080 min Scan# 534
 Delta R.T. 0.000 min
 Lab File: BN014208.D
 Acq: 05 Apr 2021 17:50

Tgt Ion:	93	Resp:	3043
Ion	Ratio	Lower	Upper
93	100		
63	81.2	65.0	97.6
95	34.5	26.8	40.2









Naphthalene-d8
 Concen: 0.40 ng
 RT: 10.416 min Scan# 853
 Delta R.T. -0.013 min
 Lab File: BN014208.D
 Acq: 05 Apr 2021 17:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Tgt Ion:	136	Resp:	20190
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.2	13.8
54	11.0	4.3	6.5#
68	7.8	3.6	5.4#




Nitrobenzene-d5
 Concen: 0.20 ng
 RT: 8.794 min Scan# 725
 Delta R.T. 0.000 min
 Lab File: BN014208.D
 Acq: 05 Apr 2021 17:50

Tgt Ion:	82	Resp:	2807
Ion	Ratio	Lower	Upper
82	100		
128	43.3	34.1	51.1
54	51.3	39.5	59.3






Abundance Scan 857 (10.467 min): BN014209.D\data.ms (-857) (-)




Naphthalene
Concen: 0.20 ng
RT: 10.467 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	128	Resp:	10921
Ion	Ratio	Lower	Upper
128	100		
129	11.6	8.9	13.3
127	13.8	10.7	16.1






Abundance Scan 880 (10.758 min): BN014209.D\data.ms (-872) (-)



Hexachlorobutadiene
Concen: 0.20 ng
RT: 10.758 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Tgt Ion:	225	Resp:	2005
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	65.6	51.2	76.8





Abundance Scan 1054 (12.017 min): BN014209.D\data.ms (-1039) (-)

2-Methylnaphthalene-d10
Concen: 0.21 ng
RT: 12.017 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Instrument :
BNA_N
ClientSampled :
SSTDICC0.2



Tgt Ion:152 Resp: 9082
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2






Abundance Scan 1054 (12.017 min): BN014208.D\data.ms (-1009) (-)




Abundance Scan 1071 (12.093 min): BN014209.D\data.ms (-1039) (-)

2-Methylnaphthalene
Concen: 0.21 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:142 Resp: 7514
Ion Ratio Lower Upper
142 100
141 87.0 70.5 105.7
115 35.4 28.3 42.5




Abundance Scan 1071 (12.093 min): BN014208.D\data.ms (-1048) (-)






Abundance Scan 1304 (14.283 min): BN014209.D\data.ms (-1288) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.283 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2







Tgt Ion:	164	Resp:	12284
Ion	Ratio	Lower	Upper
164	100		
162	101.9	81.2	121.8
160	46.5	37.5	56.3




Abundance Scan 1422 (15.780 min): BN014209.D\data.ms (-1414) (-)

2,4,6-Tribromophenol
Concen: 0.21 ng
RT: 15.780 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	330	Resp:	812
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	37.9	30.5	45.7





Abundance Scan 1195 (12.900 min): BN014209.D\data.ms (-1187) (-)




2-Fluorobiphenyl
Concen: 0.20 ng
RT: 12.900 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	172	Resp:	9368
Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.0	42.0
170	24.6	19.1	28.7



Abundance Scan 1195 (12.900 min): BN014208.D\data.ms (-1187) (-)





Abundance Scan 1282 (14.004 min): BN014209.D\data.ms (-1266) (-)




Acenaphthylene
Concen: 0.18 ng
RT: 14.004 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	152	Resp:	9193
Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.5	23.3
153	13.1	10.2	15.4




Abundance Scan 1282 (14.004 min): BN014208.D\data.ms (-1266) (-)






Abundance Scan 1309 (14.346 min): BN014209.D\data.ms (-1309) (-)

Acenaphthene
Concen: 0.20 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:	154	Resp:	7111
Ion	Ratio	Lower	Upper
154	100		
153	112.2	89.4	134.2
152	54.8	43.9	65.9






Abundance Scan 1387 (15.336 min): BN014209.D\data.ms (-1387) (-)

Fluorene
Concen: 0.20 ng
RT: 15.336 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50




Tgt Ion:	166	Resp:	9015
Ion	Ratio	Lower	Upper
166	100		
165	98.8	78.1	117.1
167	13.4	10.8	16.2





Abundance Scan 1524 (17.031 min): BN014209.D\data.ms (-1510) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.031 min Scan# 1524
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	188	Resp:	26850
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	8.4	6.9	10.3






Abundance Scan 1524 (17.031 min): BN014208.D\data.ms (-1508) (-)




Abundance Scan 1393 (15.412 min): BN014209.D\data.ms (-1329) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.18 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	198	Resp:	600
Ion	Ratio	Lower	Upper
198	100		
51	102.0	51.7	77.5#
105	60.4	38.2	57.2#



Abundance Scan 1393 (15.412 min): BN014208.D\data.ms (-1385) (-)



Abundance Scan 1458 (16.228 min): BN014209.D\data.ms (-1423) (-)

4-Bromophenyl-phenylether
Concen: 0.19 ng
RT: 16.228 min Scan# 1458
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Ref 50 141.0 173.0 248.0 330.0

m/z-->

Tgt Ion:248 Resp: 2952
Ion Ratio Lower Upper
248 100
250 94.0 74.6 111.8
141 61.0 49.4 74.0

Abundance Scan 1458 (16.228 min): BN014208.D\data.ms

Raw 50 141.0 173.0 215.0 248.0 284.0 330.0

m/z-->

Abundance

16.228

2000

1500

1000

500

0

16.20 16.30

Time-->

Abundance Scan 1458 (16.228 min): BN014208.D\data.ms (-1450) (-)

Sub 50 141.0 173.0 215.0 248.0

m/z-->

Abundance Scan 1467 (16.337 min): BN014209.D\data.ms (-1423) (-)

Hexachlorobenzene
Concen: 0.20 ng
RT: 16.337 min Scan# 1467
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Ref 50 142.0 179.0 215.0 249.0 284.0 330.0

m/z-->

Tgt Ion:284 Resp: 3323
Ion Ratio Lower Upper
284 100
142 36.0 29.0 43.4
249 30.7 25.2 37.8

Abundance Scan 1467 (16.337 min): BN014208.D\data.ms

Raw 50 142.0 179.0 215.0 249.0 284.0 330.0

m/z-->

Abundance

16.337

2000

1500

1000

500

0

16.30 16.40

Time-->

Abundance Scan 1467 (16.337 min): BN014208.D\data.ms (-1459) (-)


Sub 50 142.0 179.0 215.0 249.0 284.0

m/z-->



Abundance Scan 1480 (16.495 min): BN014209.D\data.ms (-1425) (-)

Atrazine
Concen: 0.21 ng
RT: 16.496 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2



Tgt Ion:	200	Resp:	1932
Ion	Ratio	Lower	Upper
200	100		
173	29.1	22.9	34.3
215	46.8	37.7	56.5






Abundance Scan 1480 (16.496 min): BN014208.D\data.ms (-1439) (-)




Abundance Scan 1496 (16.690 min): BN014209.D\data.ms (-1426) (-)


Pentachlorophenol
Concen: 0.21 ng
RT: 16.690 min Scan# 1496
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	266	Resp:	635
Ion	Ratio	Lower	Upper
266	100		
264	64.6	49.0	73.4
268	65.4	50.3	75.5



Abundance Scan 1496 (16.690 min): BN014208.D\data.ms (-1488) (-)









Phenanthrene
Concen: 0.20 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	178	Resp:	15694
Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.1	22.7
179	15.4	12.4	18.6




Anthracene
Concen: 0.19 ng
RT: 17.165 min Scan# 1535
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	178	Resp:	12451
Ion	Ratio	Lower	Upper
178	100		
176	18.1	14.6	21.8
179	15.4	12.2	18.2





Abundance Scan 1816 (19.064 min): BN014209.D\data.ms (-1827) (-)




Fluoranthene-d10
Concen: 0.20 ng
RT: 19.064 min Scan# 1816
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:	212	Resp:	20713
Ion	Ratio	Lower	Upper
212	100		
106	12.4	10.0	15.0
104	6.9	5.6	8.4



Abundance Scan 1816 (19.064 min): BN014208.D\data.ms (-1776) (-)





Abundance Scan 1822 (19.094 min): BN014209.D\data.ms (-1825) (-)




Fluoranthene
Concen: 0.20 ng
RT: 19.094 min Scan# 1822
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Tgt Ion:	202	Resp:	17561
Ion	Ratio	Lower	Upper
202	100		
101	11.0	8.7	13.1
203	17.2	13.8	20.6



Abundance Scan 1822 (19.094 min): BN014208.D\data.ms (-1782) (-)






Abundance Scan 2120 (21.218 min): BN014209.D\data.ms (-2129) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.218 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	240	Resp:	27610
Ion	Ratio	Lower	Upper
240	100		
120	13.4	10.3	15.5
236	27.1	21.4	32.0






Abundance Scan 1896 (19.461 min): BN014209.D\data.ms (-1836) (-)



Pyrene
Concen: 0.20 ng
RT: 19.461 min Scan# 1896
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Tgt Ion:	202	Resp:	18050
Ion	Ratio	Lower	Upper
202	100		
200	20.2	16.2	24.4
203	17.3	13.9	20.9






Abundance Scan 1938 (19.670 min): BN014209.D\data.ms (-1938) (-)

Terphenyl-d14
Concen: 0.20 ng
RT: 19.670 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2







Tgt Ion:	244	Resp:	13095
Ion	Ratio	Lower	Upper
244	100		
212	7.9	6.0	9.0
122	11.7	9.4	14.0




Abundance Scan 2119 (21.207 min): BN014209.D\data.ms (-2119) (-)

Benzo(a)anthracene
Concen: 0.19 ng
RT: 21.208 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	228	Resp:	16237
Ion	Ratio	Lower	Upper
228	100		
226	26.6	21.1	31.7
229	19.9	15.7	23.5






Abundance Scan 2124 (21.261 min): BN014209.D\data.ms (-2115) (-)




Chrysene
Concen: 0.21 ng
RT: 21.261 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2

Tgt Ion:	228	Resp:	19603
Ion	Ratio	Lower	Upper
228	100		
226	28.9	23.0	34.6
229	19.8	15.8	23.8






Abundance Scan 2113 (21.144 min): BN014209.D\data.ms (-2104) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.22 ng
RT: 21.144 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	149	Resp:	5612
Ion	Ratio	Lower	Upper
149	100		
167	26.0	21.5	32.3
279	4.3	3.1	4.7




Abundance Scan 3174 (25.632 min): BN014209.D\data.ms (-3145) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.23 ng
RT: 25.632 min Scan# 3174
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion	Ratio	Lower	Upper
276	100		
138	23.1	18.6	28.0
277	24.8	19.8	29.8




Abundance Scan 3174 (25.632 min): BN014208.D\data.ms (-3145) (-)




Abundance Scan 2530 (23.427 min): BN014209.D\data.ms (-2486) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.428 min Scan# 2530
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50




Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.7	19.4	29.2
265	26.5	20.7	31.1



Abundance Scan 2530 (23.428 min): BN014208.D\data.ms (-2486) (-)



Abundance Scan 2337 (22.767 min): BN014209.D\data.ms (-227) (-)





Benzo(b)fluoranthene
Concen: 0.19 ng
RT: 22.808 min Scan# 2349
Delta R.T. 0.041 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50


Tgt Ion:	252	Resp:	21547
Ion	Ratio	Lower	Upper
252	100		
253	24.3	18.9	28.3
125	10.6	10.1	15.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Abundance Scan 2349 (22.808 min): BN014208.D\data.ms



Abundance Scan 2349 (22.808 min): BN014208.D\data.ms (-2279) (-)





Abundance Scan 2350 (22.811 min): BN014209.D\data.ms (-228) (-)




Benzo(k)fluoranthene
Concen: 0.19 ng
RT: 22.808 min Scan# 2349
Delta R.T. -0.003 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Tgt Ion:	252	Resp:	21547
Ion	Ratio	Lower	Upper
252	100		
253	24.3	18.5	27.7
125	10.6	7.9	11.9

Abundance Scan 2349 (22.808 min): BN014208.D\data.ms




Abundance Scan 2349 (22.808 min): BN014208.D\data.ms (-2262) (-)






Abundance Scan 2501 (23.328 min): BN014209.D\data.ms (-2443) (-)

Benzo(a)pyrene
Concen: 0.20 ng
RT: 23.328 min Scan# 2501
Delta R.T. 0.000 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2




Tgt Ion:	252	Resp:	19351
Ion	Ratio	Lower	Upper
252	100		
253	27.1	19.4	29.0
125	25.4	15.0	22.6#






Abundance Scan 3177 (25.642 min): BN014209.D\data.ms (-3154) (-)


Dibenzo(a,h)anthracene
Concen: 0.20 ng
RT: 25.646 min Scan# 3178
Delta R.T. 0.004 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50



Tgt Ion:	278	Resp:	20858
Ion	Ratio	Lower	Upper
278	100		
139	16.5	12.7	19.1
279	25.7	19.4	29.2




Abundance Scan 3371 (26.306 min): BN014209.D\data.ms (-3371) (-)




Benzo(g,h,i)perylene
Concen: 0.19 ng
RT: 26.303 min Scan# 3370
Delta R.T. -0.003 min
Lab File: BN014208.D
Acq: 05 Apr 2021 17:50

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Abundance Scan 3370 (26.303 min): BN014208.D\data.ms



Tgt Ion:	276	Resp:	21457
Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.2	28.8
138	21.2	16.8	25.2



Abundance Scan 3370 (26.303 min): BN014208.D\data.ms (-3313) (-)

