

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040621\
 Data File : BN014207.D
 Acq On : 05 Apr 2021 17:14
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 4/6/2021 3:24:34 PM

Quant Time: Apr 05 19:00:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 05 18:58:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.644	152	5937	0.40	ng	0.00
7) Naphthalene-d8	10.429	136	23219	0.40	ng	0.00
13) Acenaphthene-d10	14.283	164	14211	0.40	ng	0.00
19) Phenanthrene-d10	17.031	188	31356	0.40	ng	0.00
29) Chrysene-d12	21.226	240	31651	0.40	ng	# 0.00
36) Perylene-d12	23.425	264	33104	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	1634	0.11	ng	0.00
5) Phenol-d6	6.814	99	2095	0.11	ng	0.00
8) Nitrobenzene-d5	8.794	82	1664	0.10	ng	0.00
11) 2-Methylnaphthalene-d10	12.017	152	5239	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.780	330	492	0.11	ng	0.00
15) 2-Fluorobiphenyl	12.900	172	5419	0.10	ng	0.00
27) Fluoranthene-d10	19.066	212	11905	0.10	ng	0.00
31) Terphenyl-d14	19.672	244	7576	0.10	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.174	88	904	0.12	ng	92
6) bis(2-Chloroethyl)ether	7.080	93	1757	0.11	ng	100
9) Naphthalene	10.467	128	6268	0.10	ng	97
10) Hexachlorobutadiene	10.758	225	1158	0.10	ng	# 100
12) 2-Methylnaphthalene	12.093	142	4388	0.11	ng	99
16) Acenaphthylene	14.004	152	5256	0.09	ng	99
17) Acenaphthene	14.346	154	4113	0.10	ng	99
18) Fluorene	15.336	166	5163	0.10	ng	99
20) 4,6-Dinitro-2-methylph...	15.412	198	354	0.09	ng	# 21
21) 4-Bromophenyl-phenylether	16.227	248	1729	0.10	ng	98
22) Hexachlorobenzene	16.337	284	1917	0.10	ng	99
23) Atrazine	16.495	200	1084	0.10	ng	94
25) Phenanthrene	17.067	178	9038	0.10	ng	99
26) Anthracene	17.164	178	7176	0.09	ng	99
28) Fluoranthene	19.096	202	10049	0.10	ng	100
30) Pyrene	19.459	202	10553	0.10	ng	100
32) Benzo(a)anthracene	21.205	228	9630	0.10	ng	99
33) Chrysene	21.258	228	10682	0.10	ng	98
34) Bis(2-ethylhexyl)phtha...	21.141	149	3809	0.13	ng	99
35) Indeno(1,2,3-cd)pyrene	25.629	276	13215	0.11	ng	99
37) Benzo(b)fluoranthene	22.764	252	12359m	0.10	ng	
38) Benzo(k)fluoranthene	22.809	252	11577	0.09	ng	# 93
39) Benzo(a)pyrene	23.329	252	10738	0.10	ng	# 75
40) Dibenzo(a,h)anthracene	25.639	278	10894	0.09	ng	94
41) Benzo(g,h,i)perylene	26.303	276	11415	0.09	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040621\
Data File : BN014207.D
Acq On : 05 Apr 2021 17:14
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM


Quant Time: Apr 05 19:00:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 05 18:58:13 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.644 min Scan# 604
 Delta R.T. -0.000 min
 Lab File: BN014207.D
 Acq: 05 Apr 2021 17:14




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1




Tgt Ion:152 Resp: 5937
 Ion Ratio Lower Upper
 152 100
 150 157.5 127.4 191.0
 115 61.0 50.3 75.5

Manual Integrations
 APPROVED



mohammad
 4/6/2021 3:24:34 PM




#2
 1,4-Dioxane
 Concen: 0.12 ng
 RT: 3.174 min Scan# 49
 Delta R.T. 0.000 min
 Lab File: BN014207.D
 Acq: 05 Apr 2021 17:14



Tgt Ion: 88 Resp: 904
 Ion Ratio Lower Upper
 88 100
 58 66.0 59.4 89.0
 43 30.0 25.8 38.8









#4
2-Fluorophenol
Concen: 0.11 ng
RT: 5.236 min Scan# 305
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	112	Resp:	1634
Ion	Ratio	Lower	Upper
112	100		
64	61.9	48.2	72.4
63	32.9	25.5	38.3

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM



#5
Phenol-d6
Concen: 0.11 ng
RT: 6.814 min Scan# 501
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	99	Resp:	2095
Ion	Ratio	Lower	Upper
99	100		
42	23.1	18.2	27.2
71	34.5	26.9	40.3









#6
bis(2-Chloroethyl)ether
Concen: 0.11 ng
RT: 7.080 min Scan# 534
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion: 93 Resp: 1757
Ion Ratio Lower Upper
93 100
63 81.4 65.0 97.6
95 33.1 26.8 40.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.429 min Scan# 854
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion: 136 Resp: 23219
Ion Ratio Lower Upper
136 100
137 11.5 9.2 13.8
54 5.1 4.3 6.5
68 4.2 3.6 5.4









#8
Nitrobenzene-d5
Concen: 0.10 ng
RT: 8.794 min Scan# 725
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	82	Resp:	1664
Ion	Ratio	Lower	Upper
82	100		
128	46.8	34.1	51.1
54	51.3	39.5	59.3

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM




#9
Naphthalene
Concen: 0.10 ng
RT: 10.467 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	128	Resp:	6268
Ion	Ratio	Lower	Upper
128	100		
129	12.1	8.9	13.3
127	14.6	10.7	16.1









#10
Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.758 min Scan# 880
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion:225 Resp: 1158
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.2 51.2 76.8



#11
2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.017 min Scan# 1054
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:152 Resp: 5239
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM



#12
2-Methylnaphthalene
Concen: 0.11 ng
RT: 12.093 min Scan# 1071
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	142	Resp:	4388
Ion Ratio	Lower	Upper	
142	100		
141	88.1	70.5	105.7
115	36.8	28.3	42.5

Manual Integrations
APPROVED



mohammad
4/6/2021 3:24:34 PM




#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.283 min Scan# 1304
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14



Tgt Ion:	164	Resp:	14211
Ion Ratio	Lower	Upper	
164	100		
162	101.7	81.2	121.8
160	46.6	37.5	56.3









#14
2,4,6-Tribromophenol
Concen: 0.11 ng
RT: 15.780 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion: 330 Resp: 492
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.2 30.5 45.7

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM




#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.900 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion: 172 Resp: 5419
Ion Ratio Lower Upper
172 100
171 37.0 28.0 42.0
170 25.0 19.1 28.7









#16
Acenaphthylene
Concen: 0.09 ng
RT: 14.004 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion:152 Resp: 5256
Ion Ratio Lower Upper
152 100
151 19.1 15.5 23.3
153 13.2 10.2 15.4



#17
Acenaphthene
Concen: 0.10 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:154 Resp: 4113
Ion Ratio Lower Upper
154 100
153 110.9 89.4 134.2
152 54.4 43.9 65.9



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM



#18
Fluorene
Concen: 0.10 ng
RT: 15.336 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:166 Resp: 5163
Ion Ratio Lower Upper
166 100
165 98.9 78.1 117.1
167 13.6 10.8 16.2

Manual Integrations
APPROVED



mohammad
4/6/2021 3:24:34 PM




#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.031 min Scan# 1524
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14



Tgt Ion:188 Resp: 31356
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 8.1 6.9 10.3









#20
4,6-Dinitro-2-methylphenol
Concen: 0.09 ng
RT: 15.412 min Scan# 1393
Delta R.T. 0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion	Ratio	Lower	Upper
198	100		
51	149.8	51.7	77.5#
105	73.4	38.2	57.2#

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM




#21
4-Bromophenyl-phenylether
Concen: 0.10 ng
RT: 16.227 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.6	74.6	111.8
141	62.9	49.4	74.0









#22
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.337 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion:284 Resp: 1917
Ion Ratio Lower Upper
284 100
142 36.9 29.0 43.4
249 31.4 25.2 37.8



#23
Atrazine
Concen: 0.10 ng
RT: 16.495 min Scan# 1480
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:200 Resp: 1084
Ion Ratio Lower Upper
200 100
173 32.5 22.9 34.3
215 50.8 37.7 56.5




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED




mohammad
4/6/2021 3:24:34 PM




#25
Phenanthrene
Concen: 0.10 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion:178 Resp: 9038
Ion Ratio Lower Upper
178 100
176 18.5 15.1 22.7
179 15.8 12.4 18.6



#26
Anthracene
Concen: 0.09 ng
RT: 17.164 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:178 Resp: 7176
Ion Ratio Lower Upper
178 100
176 18.9 14.6 21.8
179 15.6 12.2 18.2



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM



#27
Fluoranthene-d10
Concen: 0.10 ng
RT: 19.066 min Scan# 1818
Delta R.T. 0.002 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:212 Resp: 11905
Ion Ratio Lower Upper
212 100
106 12.6 10.0 15.0
104 7.1 5.6 8.4

Manual Integrations
APPROVED



mohammad
4/6/2021 3:24:34 PM




#28
Fluoranthene
Concen: 0.10 ng
RT: 19.096 min Scan# 1824
Delta R.T. 0.002 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:202 Resp: 10049
Ion Ratio Lower Upper
202 100
101 11.0 8.7 13.1
203 17.0 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.226 min Scan# 2122
Delta R.T. 0.008 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion:240 Resp: 31651
Ion Ratio Lower Upper
240 100
120 5.5 10.3 15.5#
236 25.2 21.4 32.0



#30
Pyrene
Concen: 0.10 ng
RT: 19.459 min Scan# 1897
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:202 Resp: 10553
Ion Ratio Lower Upper
202 100
200 20.2 16.2 24.4
203 17.5 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM



#31
Terphenyl-d14
Concen: 0.10 ng
RT: 19.672 min Scan# 1940
Delta R.T. 0.002 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:244 Resp: 7576
Ion Ratio Lower Upper
244 100
212 7.9 6.0 9.0
122 10.6 9.4 14.0




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM



#32
Benzo(a)anthracene
Concen: 0.10 ng
RT: 21.205 min Scan# 2120
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:228 Resp: 9630
Ion Ratio Lower Upper
228 100
226 27.2 21.1 31.7
229 20.1 15.7 23.5









#33
Chrysene
Concen: 0.10 ng
RT: 21.258 min Scan# 2125
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	228	Resp:	10682
Ion Ratio	Lower	Upper	
228	100		
226	30.0	23.0	34.6
229	20.0	15.8	23.8

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.13 ng
RT: 21.141 min Scan# 2114
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	149	Resp:	3809
Ion Ratio	Lower	Upper	
149	100		
167	27.3	21.5	32.3
279	4.6	3.1	4.7









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.11 ng
RT: 25.629 min Scan# 3175
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion: 276 Resp: 13215
Ion Ratio Lower Upper
276 100
138 22.8 18.6 28.0
277 24.6 19.8 29.8



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.425 min Scan# 2531
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion: 264 Resp: 33104
Ion Ratio Lower Upper
264 100
260 24.6 19.4 29.2
265 25.8 20.7 31.1




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED




mohammad
4/6/2021 3:24:34 PM




#37
Benzo(b)fluoranthene
Concen: 0.10 ng m
RT: 22.764 min Scan# 2338
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14





Tgt Ion:252 Resp: 12359
Ion Ratio Lower Upper
252 100
253 26.9 18.9 28.3
125 19.2 10.1 15.1#



#38
Benzo(k)fluoranthene
Concen: 0.09 ng
RT: 22.809 min Scan# 2351
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14




Tgt Ion:252 Resp: 11577
Ion Ratio Lower Upper
252 100
253 26.7 18.5 27.7
125 12.2 7.9 11.9#



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM







#39
Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.329 min Scan# 2503
Delta R.T. 0.001 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion	Ratio	Lower	Upper
252	100		
253	32.1	19.4	29.0#
125	35.4	15.0	22.6#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/6/2021 3:24:34 PM



#40
Dibenzo(a,h)anthracene
Concen: 0.09 ng
RT: 25.639 min Scan# 3178
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion	Ratio	Lower	Upper
278	100		
139	17.8	12.7	19.1
279	27.4	19.4	29.2





#41
Benzo(g,h,i)perylene
Concen: 0.09 ng
RT: 26.303 min Scan# 3372
Delta R.T. -0.003 min
Lab File: BN014207.D
Acq: 05 Apr 2021 17:14

Tgt Ion:	276	Resp:	11415
Ion Ratio	Lower	Upper	
276	100		
277	24.9	19.2	28.8
138	22.8	16.8	25.2

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/6/2021 3:24:34 PM

