

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040624\
 Data File : BN030701.D
 Acq On : 06 Apr 2024 09:59
 Operator : MA/JU
 Sample : P1893-03MS 10X
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E07D1MS

Quant Time: Apr 07 12:06:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Mar 31 12:24:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.687	152	103117	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.475	136	510159	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.334	164	369686	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.080	188	819405	20.000	ng/u1	0.00
79) Chrysene-d12	21.327	240	778588	20.000	ng/u1	0.00
88) Perylene-d12	24.262	264	877582	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	9324	4.239	ng/uL	0.00
4) Pyridine-d5	3.599	84	140898	21.367	ng/u1	0.00
7) Phenol-d5	6.857	99	121367	14.549	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.028	67	313002	60.751	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.222	132	328367	51.891	ng/u1	0.00
15) 4-Methylphenol-d8	8.399	113	235787	33.726	ng/u1	0.00
21) Nitrobenzene-d5	8.846	128	234755	66.335	ng/u1	0.00
24) 2-Nitrophenol-d4	9.563	143	243658	75.262	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.093	165	423422	59.741	ng/u1	0.00
31) 4-Chloroaniline-d4	10.616	131	609408	52.227	ng/u1	0.00
46) Dimethylphthalate-d6	13.751	166	1742850	70.611	ng/u1	0.00
49) Acenaphthylene-d8	14.028	160	1852788	65.288	ng/u1	0.00
54) 4-Nitrophenol-d4	14.539	143	82621	17.500	ng/u1	0.00
60) Fluorene-d10	15.328	176	1454638	67.348	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.457	200	291704	80.652	ng/u1	0.00
73) Anthracene-d10	17.186	188	2401342	68.017	ng/u1	0.00
81) Pyrene-d10	19.486	212	2916701	69.677	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.068	264	3128052	77.356	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.228	88	22119	9.389	ng/uL#	1
5) Pyridine	3.616	79	142990	21.489	ng/u1	96
6) Benzaldehyde	6.840	77	290934	74.893	ng/u1	98
8) Phenol	6.887	94	128510	14.788	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.122	93	433106	61.118	ng/u1	97
12) 2-Chlorophenol	7.252	128	333062	49.064	ng/u1	96
13) 2-Methylphenol	8.134	108	259063	38.687	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.210	45	562990	56.422	ng/u1	98
16) Acetophenone	8.510	105	715267	64.799	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.499	70	354579	63.780	ng/u1	97
18) 4-Methylphenol	8.457	108	254532	33.954	ng/u1	99
19) Hexachloroethane	8.757	117	164018	57.630	ng/u1	97
22) Nitrobenzene	8.887	77	514583	59.565	ng/u1	99
23) Isophorone	9.410	82	1077336	62.531	ng/u1	100
25) 2-Nitrophenol	9.593	139	254628	66.921	ng/u1	96
26) 2,4-Dimethylphenol	9.657	107	366017	43.281	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.893	93	642548	58.658	ng/u1	99
29) 2,4-Dichlorophenol	10.122	162	412156	56.972	ng/u1	97
30) Naphthalene	10.522	128	1525587	58.306	ng/u1	100
32) 4-Chloroaniline	10.640	127	546673	48.614	ng/u1	100
33) Hexachlorobutadiene	10.804	225	256678	56.368	ng/u1	97
34) Caprolactam	11.422	113	31780	11.759	ng/u1	89
35) 4-Chloro-3-methylphenol	11.763	107	468406	57.486	ng/u1	98
36) 2-Methylnaphthalene	12.140	142	1117187	61.140	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040624\
 Data File : BN030701.D
 Acq On : 06 Apr 2024 09:59
 Operator : MA/JU
 Sample : P1893-03MS 10X
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E07D1MS

Quant Time: Apr 07 12:06:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Mar 31 12:24:07 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.363	142	1121877	60.481	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.510	216	560366	58.736	ng/ul	97
40) Hexachlorocyclopentadiene	12.487	237	261314	48.230	ng/ul	96
41) 2,4,6-Trichlorophenol	12.757	196	403466	67.676	ng/ul	100
42) 2,4,5-Trichlorophenol	12.828	196	435746	64.719	ng/ul	98
43) 1,1'-Biphenyl	13.163	154	1513003	60.045	ng/ul	97
44) 2-Chloronaphthalene	13.204	162	1194069	60.198	ng/ul	99
45) 2-Nitroaniline	13.416	65	387152	73.341	ng/ul	100
47) Dimethylphthalate	13.798	163	1721811	67.531	ng/ul	100
48) 2,6-Dinitrotoluene	13.922	165	369161	78.092	ng/ul	99
50) Acenaphthylene	14.051	152	2074927	63.280	ng/ul	100
51) 3-Nitroaniline	14.245	138	364067	72.559	ng/ul	100
52) Acenaphthene	14.398	153	1366328	62.529	ng/ul	97
53) 2,4-Dinitrophenol	14.451	184	172865	79.615	ng/ul	95
55) 4-Nitrophenol	14.551	109	64219	17.169	ng/ul	98
56) Dibenzofuran	14.734	168	1929153	63.513	ng/ul	100
57) 2,4-Dinitrotoluene	14.704	165	540848	78.365	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.963	232	405460	72.511	ng/ul	99
59) Diethylphthalate	15.169	149	1851342	71.474	ng/ul	99
61) Fluorene	15.386	166	1559703	64.602	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.381	204	751109	63.412	ng/ul	98
63) 4-Nitroaniline	15.416	138	400190	82.039	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.469	198	299032	76.946	ng/ul	96
67) N-Nitrosodiphenylamine	15.598	169	1445639	64.782	ng/ul	98
68) 4-Bromophenyl-phenylether	16.275	248	515550	66.259	ng/ul	92
69) Hexachlorobenzene	16.381	284	597281	65.190	ng/ul	98
70) Atrazine	16.557	200	502478	64.965	ng/ul	95
71) Pentachlorophenol	16.728	266	357215	66.126	ng/ul	98
72) Phenanthrene	17.128	178	2749129	65.302	ng/ul	99
74) Anthracene	17.216	178	2752023	64.665	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.122	216	586276	58.239	ng/ul	99
76) Pentachlorobenzene	14.645	250	623711	60.998	ng/ul	99
77) Carbazole	17.492	167	2661049	71.299	ng/ul	99
78) Di-n-butylphthalate	18.057	149	3437473	75.498	ng/ul	100
80) Fluoranthene	19.145	202	3327372	66.867	ng/ul	99
82) Pyrene	19.510	202	3396272	64.626	ng/ul	99
83) Butylbenzylphthalate	20.421	149	1625880	81.013	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.239	252	1013583	70.849	ng/ul	99
85) Benzo(a)anthracene	21.304	228	3595981	69.205	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.239	149	2307366	77.960	ng/ul	99
87) Chrysene	21.369	228	3298997	67.264	ng/ul	97
89) Di-n-octyl phthalate	22.357	149	4380180	78.816	ng/ul	100
90) Benzo(b)fluoranthene	23.339	252	3796529	73.509	ng/ul	99
91) Benzo(k)fluoranthene	23.404	252	3739557	71.296	ng/ul	99
93) Benzo(a)pyrene	24.133	252	3455803	70.833	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	27.580	276	4085736	77.471	ng/ul	98
95) Dibenzo(a,h)anthracene	27.645	278	3423945	77.119	ng/ul	99
96) Benzo(g,h,i)perylene	28.615	276	3248851	78.468	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#2
1,4-Dioxane
Concen: 9.389 ng/uL
RT: 3.228 min Scan# 41
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Instrument :
BNA_N
ClientSampleId :
E07D1MS







Tgt Ion: 88 Resp: 22119
Ion Ratio Lower Upper
88 100
43 926.7 23.8 35.8#
58 69.3 56.6 84.8



#5
Pyridine
Concen: 21.489 ng/uL
RT: 3.616 min Scan# 107
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion: 79 Resp: 142990
Ion Ratio Lower Upper
79 100
52 74.4 56.2 84.4
51 33.2 27.0 40.4









#6
Benzaldehyde
Concen: 74.893 ng/ul
RT: 6.840 min Scan# 61
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion: 77 Resp: 290934
Ion Ratio Lower Upper
77 100
105 98.7 78.1 117.1
106 97.9 75.6 113.4



#8
Phenol
Concen: 14.788 ng/ul
RT: 6.887 min Scan# 663
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion: 94 Resp: 128510
Ion Ratio Lower Upper
94 100
65 27.2 21.4 32.0
66 35.1 28.5 42.7









#10
Bis(2-Chloroethyl)ether
Concen: 61.118 ng/ul
RT: 7.122 min Scan# 703
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion: 93 Resp: 433106
Ion Ratio Lower Upper
93 100
63 70.1 58.2 87.4
95 31.9 26.0 39.0



#12
2-Chlorophenol
Concen: 49.064 ng/ul
RT: 7.252 min Scan# 725
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:128 Resp: 333062
Ion Ratio Lower Upper
128 100
64 43.6 37.0 55.4
130 31.2 26.3 39.5









#13
2-Methylphenol
Concen: 38.687 ng/ul
RT: 8.134 min Scan# 81
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:108 Resp: 259063
Ion Ratio Lower Upper
108 100
107 84.5 71.8 107.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 56.422 ng/ul
RT: 8.210 min Scan# 888
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion: 45 Resp: 562990
Ion Ratio Lower Upper
45 100
77 15.9 12.1 18.1
79 11.4 8.6 13.0









#16
Acetophenone
Concen: 64.799 ng/ul
RT: 8.510 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:105 Resp: 715267
Ion Ratio Lower Upper
105 100
77 75.7 61.2 91.8
51 24.5 21.0 31.4



#17
N-Nitroso-di-n-propylamine
Concen: 63.780 ng/ul
RT: 8.499 min Scan# 937
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion: 70 Resp: 354579
Ion Ratio Lower Upper
70 100
42 56.1 47.4 71.0
101 10.4 7.9 11.9
130 23.7 19.0 28.4









#18
4-Methylphenol
Concen: 33.954 ng/ul
RT: 8.457 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:108 Resp: 254532
Ion Ratio Lower Upper
108 100
107 111.4 88.1 132.1



#19
Hexachloroethane
Concen: 57.630 ng/ul
RT: 8.757 min Scan# 981
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:117 Resp: 164018
Ion Ratio Lower Upper
117 100
201 99.7 76.5 114.7
199 60.6 48.5 72.7





#22

Nitrobenzene

Concen: 59.565 ng/ul

RT: 8.887 min Scan# 1003

Delta R.T. -0.006 min

Lab File: BN030701.D

Acq: 06 Apr 2024 09:59

Instrument :

BNA_N

ClientSampleId :

E07D1MS





Tgt Ion: 77 Resp: 514583

Ion Ratio Lower Upper

77 100

123 47.2 37.5 56.3

65 13.8 10.6 15.8



#23

Isophorone

Concen: 62.531 ng/ul

RT: 9.410 min Scan# 1092

Delta R.T. -0.006 min

Lab File: BN030701.D

Acq: 06 Apr 2024 09:59




Tgt Ion: 82 Resp: 1077336


Ion Ratio Lower Upper

82 100

95 7.0 5.7 8.5

138 18.1 14.6 21.8





#25

2-Nitrophenol

Concen: 66.921 ng/ul

RT: 9.593 min Scan# 1123

Delta R.T. -0.006 min

Lab File: BN030701.D

Acq: 06 Apr 2024 09:59

Instrument :

BNA_N

ClientSampleId :

E07D1MS





Tgt Ion:139 Resp: 254628

Ion Ratio Lower Upper

139 100

65 47.3 39.5 59.3

109 32.2 23.8 35.8



#26

2,4-Dimethylphenol

Concen: 43.281 ng/ul

RT: 9.657 min Scan# 1134

Delta R.T. -0.006 min

Lab File: BN030701.D

Acq: 06 Apr 2024 09:59




Tgt Ion:107 Resp: 366017


Ion Ratio Lower Upper

107 100

121 50.2 41.6 62.4

122 89.5 71.3 106.9









#27
 Bis(2-Chloroethoxy)methane
 Concen: 58.658 ng/ul
 RT: 9.893 min Scan# 1174
 Delta R.T. -0.012 min
 Lab File: BN030701.D
 Acq: 06 Apr 2024 09:59

Instrument :
 BNA_N
 ClientSampleId :
 E07D1MS

Tgt Ion: 93 Resp: 642548




Ion	Ratio	Lower	Upper
93	100		
95	32.6	25.4	38.0
123	11.8	9.4	14.0




#29
 2,4-Dichlorophenol
 Concen: 56.972 ng/ul
 RT: 10.122 min Scan# 1213
 Delta R.T. -0.006 min
 Lab File: BN030701.D
 Acq: 06 Apr 2024 09:59

Tgt Ion: 162 Resp: 412156

Ion	Ratio	Lower	Upper
162	100		
164	62.3	51.8	77.8
98	34.0	28.0	42.0









#30
Naphthalene
Concen: 58.306 ng/ul
RT: 10.522 min Scan# 1281
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:128 Resp: 1525587
Ion Ratio Lower Upper
128 100
129 11.1 8.7 13.1
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 48.614 ng/ul
RT: 10.640 min Scan# 1301
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:127 Resp: 546673
Ion Ratio Lower Upper
127 100
129 31.9 25.3 37.9









#33
Hexachlorobutadiene
Concen: 56.368 ng/ul
RT: 10.804 min Scan# 11
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:225 Resp: 256678
Ion Ratio Lower Upper
225 100
223 65.0 50.9 76.3
227 68.5 51.6 77.4



#34
Caprolactam
Concen: 11.759 ng/ul
RT: 11.422 min Scan# 1434
Delta R.T. 0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:113 Resp: 31780
Ion Ratio Lower Upper
113 100
55 138.3 124.7 187.1
56 106.2 90.6 135.8









#35
4-Chloro-3-methylphenol
Concen: 57.486 ng/ul
RT: 11.763 min Scan# 1492
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:107 Resp: 468406
Ion Ratio Lower Upper
107 100
144 25.1 21.4 32.0
142 81.4 65.9 98.9



#36
2-Methylnaphthalene
Concen: 61.140 ng/ul
RT: 12.140 min Scan# 1556
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:142 Resp: 1117187
Ion Ratio Lower Upper
142 100
141 86.3 70.8 106.2









#37
1-Methylnaphthalene
Concen: 60.481 ng/ul
RT: 12.363 min Scan# 1594
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:142 Resp: 1121877
Ion Ratio Lower Upper
142 100
141 93.1 73.0 109.4
116 3.7 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 58.736 ng/ul
RT: 12.510 min Scan# 1619
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:216 Resp: 560366
Ion Ratio Lower Upper
216 100
214 79.4 60.9 91.3
179 23.4 17.4 26.0
108 18.4 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 48.230 ng/ul
RT: 12.487 min Scan# 1615
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:237 Resp: 261314
Ion Ratio Lower Upper
237 100
235 59.4 50.6 76.0
272 13.4 10.2 15.4



#41
2,4,6-Trichlorophenol
Concen: 67.676 ng/ul
RT: 12.757 min Scan# 1661
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59


Tgt Ion:196 Resp: 403466
Ion Ratio Lower Upper
196 100
198 93.2 74.6 111.8
200 31.2 24.8 37.2








#42
2,4,5-Trichlorophenol
Concen: 64.719 ng/u1
RT: 12.828 min Scan# 1673
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Instrument :
BNA_N
ClientSampleId :
E07D1MS







Tgt Ion:196 Resp: 435746
Ion Ratio Lower Upper
196 100
198 92.6 73.0 109.6
200 30.0 25.2 37.8



#43
1,1'-Biphenyl
Concen: 60.045 ng/u1
RT: 13.163 min Scan# 1730
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:154 Resp: 1513003
Ion Ratio Lower Upper
154 100
153 41.5 31.5 47.3
76 13.2 9.8 14.6









#44
2-Chloronaphthalene
Concen: 60.198 ng/ul
RT: 13.204 min Scan# 1737
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:162 Resp: 1194069
Ion Ratio Lower Upper
162 100
164 32.7 25.7 38.5
127 40.7 32.2 48.2



#45
2-Nitroaniline
Concen: 73.341 ng/ul
RT: 13.416 min Scan# 1773
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion: 65 Resp: 387152
Ion Ratio Lower Upper
65 100
92 72.7 57.7 86.5
138 113.2 90.6 135.8









#47
Dimethylphthalate
Concen: 67.531 ng/ul
RT: 13.798 min Scan# 1838
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:163 Resp: 1721811
Ion Ratio Lower Upper
163 100
194 4.3 3.3 4.9
164 10.4 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 78.092 ng/ul
RT: 13.922 min Scan# 1859
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:165 Resp: 369161
Ion Ratio Lower Upper
165 100
89 54.2 43.5 65.3
121 20.3 15.7 23.5









#50
Acenaphthylene
Concen: 63.280 ng/ul
RT: 14.051 min Scan# 1882
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:152 Resp: 2074927
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.3 10.7 16.1



#51
3-Nitroaniline
Concen: 72.559 ng/ul
RT: 14.245 min Scan# 1914
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59


Tgt Ion:138 Resp: 364067
Ion Ratio Lower Upper
138 100
108 11.3 9.6 14.4
92 112.2 89.6 134.4








#52
Acenaphthene
Concen: 62.529 ng/ul
RT: 14.398 min Scan# 1940
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Instrument :
BNA_N
ClientSampleId :
E07D1MS







Tgt Ion:153 Resp: 1366328
Ion Ratio Lower Upper
153 100
152 48.1 37.6 56.4
154 91.3 70.3 105.5



#53
2,4-Dinitrophenol
Concen: 79.615 ng/ul
RT: 14.451 min Scan# 1949
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:184 Resp: 172865
Ion Ratio Lower Upper
184 100
63 60.4 49.9 74.9
154 66.9 49.5 74.3









#55
4-Nitrophenol
Concen: 17.169 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:109 Resp: 64219
Ion Ratio Lower Upper
109 100
139 136.5 111.8 167.6
65 125.6 98.7 148.1



#56
Dibenzofuran
Concen: 63.513 ng/ul
RT: 14.734 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:168 Resp: 1929153
Ion Ratio Lower Upper
168 100
139 38.0 30.2 45.4









#57
2,4-Dinitrotoluene
Concen: 78.365 ng/ul
RT: 14.704 min Scan# 1992
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:165 Resp: 540848
Ion Ratio Lower Upper
165 100
63 41.9 34.7 52.1
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 72.511 ng/ul
RT: 14.963 min Scan# 2036
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:232 Resp: 405460
Ion Ratio Lower Upper
232 100
131 46.9 36.6 55.0
130 2.3 2.2 3.2
166 32.9 26.0 39.0









#59
Diethylphthalate
Concen: 71.474 ng/ul
RT: 15.169 min Scan# 2071
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:149 Resp: 1851342
Ion Ratio Lower Upper
149 100
177 21.6 17.4 26.2
150 12.0 9.8 14.8



#61
Fluorene
Concen: 64.602 ng/ul
RT: 15.386 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:166 Resp: 1559703
Ion Ratio Lower Upper
166 100
165 100.6 78.2 117.4
167 14.2 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 63.412 ng/ul
RT: 15.381 min Scan# 2107
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:204 Resp: 751109
Ion Ratio Lower Upper
204 100
206 31.3 25.8 38.6
141 63.3 49.1 73.7



#63
4-Nitroaniline
Concen: 82.039 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:138 Resp: 400190
Ion Ratio Lower Upper
138 100
92 49.2 39.7 59.5
108 79.6 62.6 94.0









#66
4,6-Dinitro-2-methylphenol
Concen: 76.946 ng/ul
RT: 15.469 min Scan# 2121
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:198 Resp: 299032
Ion Ratio Lower Upper
198 100
182 4.1 3.6 5.4
77 25.7 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 64.782 ng/ul
RT: 15.598 min Scan# 2144
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:169 Resp: 1445639
Ion Ratio Lower Upper
169 100
168 66.2 51.9 77.9
167 34.9 29.0 43.4









#68
4-Bromophenyl-phenylether
Concen: 66.259 ng/ul
RT: 16.275 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:248 Resp: 515550
Ion Ratio Lower Upper
248 100
250 92.0 80.0 120.0
141 82.9 60.5 90.7



#69
Hexachlorobenzene
Concen: 65.190 ng/ul
RT: 16.381 min Scan# 2277
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:284 Resp: 597281
Ion Ratio Lower Upper
284 100
142 37.7 28.6 43.0
249 30.8 25.0 37.4









#70
Atrazine
Concen: 64.965 ng/ul
RT: 16.557 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:200 Resp: 502478
Ion Ratio Lower Upper
200 100
173 28.1 21.5 32.3
215 51.6 37.8 56.8



#71
Pentachlorophenol
Concen: 66.126 ng/ul
RT: 16.728 min Scan# 2336
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:266 Resp: 357215
Ion Ratio Lower Upper
266 100
264 65.0 51.0 76.6
268 64.1 52.5 78.7









#72
Phenanthrene
Concen: 65.302 ng/u1
RT: 17.128 min Scan# 2404
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:178 Resp: 2749129
Ion Ratio Lower Upper
178 100
179 15.8 12.4 18.6
176 19.3 15.0 22.6



#74
Anthracene
Concen: 64.665 ng/u1
RT: 17.216 min Scan# 2419
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:178 Resp: 2752023
Ion Ratio Lower Upper
178 100
179 16.4 12.0 18.0
176 19.2 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 58.239 ng/uL
RT: 13.122 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:216 Resp: 586276
Ion Ratio Lower Upper
216 100
214 79.2 63.1 94.7
218 47.2 39.5 59.3



#76
Pentachlorobenzene
Concen: 60.998 ng/uL
RT: 14.645 min Scan# 1982
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59


Tgt Ion:250 Resp: 623711
Ion Ratio Lower Upper
250 100
248 58.3 45.8 68.6
252 59.9 48.7 73.1








#77
Carbazole
Concen: 71.299 ng/ul
RT: 17.492 min Scan# 2466
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Instrument :
BNA_N
ClientSampleId :
E07D1MS







Tgt Ion:167 Resp: 2661049
Ion Ratio Lower Upper
167 100
166 21.1 16.6 24.8
139 13.1 10.5 15.7



#78
Di-n-butylphthalate
Concen: 75.498 ng/ul
RT: 18.057 min Scan# 2562
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:149 Resp: 3437473
Ion Ratio Lower Upper
149 100
150 9.0 7.2 10.8
104 5.1 4.1 6.1









#80
Fluoranthene
Concen: 66.867 ng/ul
RT: 19.145 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:202 Resp: 3327372
Ion Ratio Lower Upper
202 100
101 11.6 9.0 13.6
100 9.1 6.8 10.2



#82
Pyrene
Concen: 64.626 ng/ul
RT: 19.510 min Scan# 2809
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:202 Resp: 3396272
Ion Ratio Lower Upper
202 100
101 14.5 11.3 16.9
100 11.6 9.2 13.8









#83
Butylbenzylphthalate
Concen: 81.013 ng/ul
RT: 20.421 min Scan# 2964
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:149 Resp: 1625880
Ion Ratio Lower Upper
149 100
91 72.7 58.2 87.2
206 21.4 15.4 23.0



#84
3,3'-Dichlorobenzidine
Concen: 70.849 ng/ul
RT: 21.239 min Scan# 3103
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:252 Resp: 1013583
Ion Ratio Lower Upper
252 100
254 65.0 51.1 76.7
126 11.3 9.3 13.9









#85
Benzo(a)anthracene
Concen: 69.205 ng/u1
RT: 21.304 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:228 Resp: 3595981
Ion Ratio Lower Upper
228 100
229 19.6 15.8 23.6
226 27.8 20.5 30.7



#86
Bis(2-ethylhexyl)phthalate
Concen: 77.960 ng/u1
RT: 21.239 min Scan# 3103
Delta R.T. -0.012 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:149 Resp: 2307366
Ion Ratio Lower Upper
149 100
167 27.1 21.4 32.0
279 4.3 3.5 5.3





#87

Chrysene

Concen: 67.264 ng/ul

RT: 21.369 min Scan# 3124

Delta R.T. -0.006 min

Lab File: BN030701.D

Acq: 06 Apr 2024 09:59

Instrument :

BNA_N

ClientSampleId :

E07D1MS





Tgt Ion:228 Resp: 3298997

Ion Ratio Lower Upper

228 100

226 32.2 24.3 36.5

229 21.3 15.9 23.9



#89

Di-n-octyl phthalate

Concen: 78.816 ng/ul




RT: 22.357 min Scan# 3293


Delta R.T. -0.018 min

Lab File: BN030701.D

Acq: 06 Apr 2024 09:59

Tgt Ion:149 Resp: 4380180









#90
Benzo(b)fluoranthene
Concen: 73.509 ng/u1
RT: 23.339 min Scan# 3459
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:252 Resp: 3796529
Ion Ratio Lower Upper
252 100
253 21.9 16.7 25.1
125 10.6 8.5 12.7



#91
Benzo(k)fluoranthene
Concen: 71.296 ng/u1
RT: 23.404 min Scan# 3471
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:252 Resp: 3739557
Ion Ratio Lower Upper
252 100
253 21.4 16.9 25.3
125 10.3 8.5 12.7









#93
Benzo(a)pyrene
Concen: 70.833 ng/u1
RT: 24.133 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E07D1MS


Tgt Ion:252 Resp: 3455803
Ion Ratio Lower Upper
252 100
253 23.2 16.7 25.1
125 12.4 9.6 14.4



#94
Indeno(1,2,3-cd)pyrene
Concen: 77.471 ng/u1
RT: 27.580 min Scan# 4181
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:276 Resp: 4085736
Ion Ratio Lower Upper
276 100
138 21.8 17.8 26.8
277 24.6 18.4 27.6









#95
Dibenzo(a,h)anthracene
Concen: 77.119 ng/ul
RT: 27.645 min Scan# 4189
Delta R.T. -0.006 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Instrument :
BNA_N
ClientSampleId :
E07D1MS

Tgt Ion:278 Resp: 3423945
Ion Ratio Lower Upper
278 100
139 17.5 13.2 19.8
279 22.2 17.4 26.2



#96
Benzo(g,h,i)perylene
Concen: 78.468 ng/ul
RT: 28.615 min Scan# 4357
Delta R.T. -0.000 min
Lab File: BN030701.D
Acq: 06 Apr 2024 09:59

Tgt Ion:276 Resp: 3248851
Ion Ratio Lower Upper
276 100
138 23.5 19.3 28.9
277 24.3 20.1 30.1

