


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040624\
Data File : BN030667.D
Acq On : 05 Apr 2024 08:57
Operator : MA/JU
Sample : DFTPP313
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP313

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Title : SVOA CALIBRATION
Last Update : Sun Mar 31 12:24:07 2024



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.1	157474	PASS
68	69	0.00	2	2.0	3277	PASS
69	198	0.00	100	35.1	166963	PASS
70	69	0.00	2	0.7	1108	PASS
127	198	10	80	46.3	219904	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	475221	PASS
199	198	5	9	6.5	30781	PASS
275	198	10	60	25.7	122307	PASS
365	198	1	100	3.6	17054	PASS
441	198	0.01	100	12.0	57261	PASS
442	198	50	100	73.0	346944	PASS
443	442	15	24	21.2	73400	PASS

m/z	Abundance
39.00	270.0
40.00	2792.0
41.00	426.0
43.00	309.0
44.00	4961.0
45.00	260.0
48.90	293.0
51.00	257.0
54.80	252.0
56.90	181.0
65.90	327.0
72.90	631.0
76.90	327.0
77.90	158.0
83.90	343.0
85.70	174.0
91.00	194.0
95.90	354.0
96.80	154.0
104.80	154.0
132.90	419.0
191.00	352.0
193.00	248.0
207.00	2610.0
207.90	550.0
209.10	350.0
210.00	151.0
266.90	154.0
281.00	659.0
282.00	152.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

m/z	Abundance
37.00	904.0
38.00	2225.0
39.00	12363.0
40.00	2360.0
41.00	582.0
42.90	412.0
44.00	3835.0
45.00	428.0
49.00	1616.0
50.00	44768.0
51.00	167552.0
52.00	9065.0
53.00	569.0
55.00	1113.0
55.90	4969.0
57.00	10842.0
57.90	593.0
58.90	182.0
59.90	212.0
60.90	2268.0
62.00	2206.0
63.00	7025.0
63.90	1189.0
65.00	3475.0
66.00	459.0
66.90	453.0
68.00	3641.0
69.00	176512.0
70.00	1108.0
73.00	1806.0
74.00	19080.0
75.00	28576.0
76.00	9847.0
77.00	209408.0
78.00	14570.0
79.00	15339.0
79.90	9973.0
81.00	14657.0
82.00	3584.0
83.00	2970.0
83.90	638.0
85.00	2319.0
86.00	4643.0
87.00	1927.0
87.90	989.0
88.90	323.0
91.00	3274.0
92.00	3860.0
93.00	24472.0
94.00	1621.0
94.90	386.0
96.00	1194.0
96.90	524.0
98.00	19800.0
98.90	15487.0
100.00	1372.0
101.00	8488.0
101.80	565.0
102.90	2968.0
103.90	5068.0
105.00	5672.0
106.00	1703.0
107.00	64992.0
108.00	9840.0
109.10	2160.0
110.00	118536.0
111.00	17688.0
111.90	2211.0
112.90	795.0
114.90	425.0
116.00	3910.0
117.00	59600.0
118.00	4398.0
119.00	495.0
119.90	758.0
120.90	287.0
122.00	4519.0
123.00	6800.0
124.00	2991.0
125.00	3022.0
127.00	227200.0
128.00	17808.0
129.00	86472.0
130.00	7993.0
131.00	1686.0
132.00	916.0
132.90	622.0
133.90	2613.0
135.00	7076.0
136.00	3010.0
137.00	4207.0
138.00	715.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

138.90	378.0
140.00	998.0
141.00	10841.0
142.00	4028.0
142.90	2736.0
144.00	602.0
144.90	590.0
145.90	2032.0
147.00	6097.0
147.90	14563.0
149.00	3105.0
149.90	883.0
151.00	1470.0
152.00	917.0
153.00	3628.0
154.00	2935.0
155.00	6768.0
156.00	9077.0
157.00	1989.0
157.90	2132.0
159.00	1544.0
160.00	3422.0
161.00	5119.0
162.00	1686.0
162.90	490.0
163.90	478.0
164.90	4112.0
166.00	2813.0
167.00	24080.0
168.00	14714.0
169.00	2177.0
170.00	872.0
170.80	844.0
172.00	2211.0
173.00	2461.0
174.00	4674.0
175.00	9355.0
176.00	2647.0
177.00	4015.0
177.90	1170.0
179.00	16568.0
180.00	12458.0
181.00	5556.0
182.10	798.0
183.00	358.0
184.00	1315.0
185.00	7765.0
186.00	64248.0
187.00	18904.0
187.90	1866.0
188.90	3992.0
189.90	867.0
191.00	2183.0
192.00	5046.0
193.00	6398.0
194.00	1289.0
195.10	637.0
196.00	13279.0
197.90	442944.0
198.90	28904.0
200.00	2701.0
201.50	2039.0
202.00	554.0
202.90	3178.0
204.00	15909.0
205.00	26840.0
206.00	119824.0
207.00	16560.0
208.00	4837.0
209.00	1572.0
210.10	1605.0
211.00	4584.0
212.90	257.0
214.90	1410.0
216.00	3029.0
216.90	29472.0
218.00	4112.0
219.00	367.0
221.00	17504.0
221.80	3981.0
223.00	6441.0
224.00	63536.0
225.00	16928.0
226.00	2019.0
227.00	25960.0
227.90	3918.0
229.00	5556.0
229.90	692.0
231.00	2114.0
232.00	368.0
232.90	360.0
234.00	1821.0
235.00	1908.0
236.00	1458.0
237.00	2920.0
238.00	454.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

239.00	1216.0
240.00	788.0
241.00	1463.0
242.00	3389.0
243.00	3643.0
244.00	49792.0
245.00	6841.0
245.90	11140.0
247.00	1993.0
248.00	420.0
248.90	1692.0
250.00	446.0
251.10	543.0
252.00	537.0
253.00	888.0
254.10	2281.0
255.00	257216.0
256.00	35440.0
257.00	2943.0
258.00	14780.0
259.00	2414.0
259.90	344.0
260.90	457.0
263.00	223.0
263.80	368.0
265.00	6272.0
265.90	862.0
266.90	233.0
269.90	288.0
271.00	737.0
272.00	682.0
272.90	7206.0
273.90	20168.0
275.00	106832.0
276.00	14980.0
276.90	10182.0
278.00	1927.0
279.00	459.0
281.00	578.0
281.80	374.0
283.00	938.0
284.00	981.0
285.00	1564.0
285.90	378.0
288.80	366.0
289.80	280.0
290.90	327.0
292.00	398.0
293.00	2093.0
294.00	711.0
295.00	634.0
296.00	34320.0
297.00	4961.0
297.90	319.0
300.90	341.0
301.90	613.0
303.00	3921.0
304.00	1116.0
304.90	158.0
307.90	348.0
308.90	449.0
309.90	411.0
312.80	271.0
313.90	1712.0
315.00	3032.0
316.00	2075.0
316.90	450.0
320.90	941.0
322.00	588.0
323.00	9218.0
323.90	2054.0
324.90	241.0
325.90	283.0
326.90	2171.0
328.00	1119.0
328.80	264.0
332.00	770.0
332.90	1061.0
334.00	6280.0
335.00	2056.0
336.00	164.0
339.00	215.0
340.90	1317.0
342.00	343.0
345.90	2087.0
346.90	442.0
350.90	177.0
352.00	2680.0
352.90	1692.0
354.00	3363.0
355.00	735.0
358.80	248.0
364.00	254.0
364.90	14211.0
365.90	2280.0
366.90	153.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

369.90	243.0
371.00	867.0
372.00	4493.0
373.10	1238.0
383.00	1264.0
383.90	447.0
390.00	853.0
391.00	492.0
391.80	315.0
401.00	324.0
401.90	2218.0
402.90	3122.0
403.90	1114.0
405.10	244.0
415.00	153.0
421.00	2377.0
422.00	2548.0
423.00	18128.0
423.90	4258.0
424.90	862.0
441.00	43944.0
442.00	278720.0
443.00	53840.0
444.00	5486.0
445.00	379.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

m/z	Abundance
36.00	206.0
37.10	1011.0
38.10	2687.0
39.00	14595.0
40.00	2650.0
41.00	586.0
42.10	185.0
43.00	426.0
44.00	4323.0
45.00	559.0
49.00	1576.0
50.00	49480.0
51.00	184896.0
52.00	9295.0
53.00	549.0
54.90	1227.0
55.90	5507.0
57.00	12279.0
57.90	641.0
59.00	225.0
60.00	210.0
61.00	2779.0
62.00	2747.0
63.00	7914.0
64.00	1007.0
65.00	3461.0
65.90	523.0
67.00	336.0
68.00	3463.0
69.00	194048.0
69.90	1325.0
70.90	178.0
73.00	1816.0
74.00	21144.0
75.00	31968.0
76.00	10836.0
77.00	232064.0
78.00	15439.0
79.00	15524.0
80.00	11639.0
81.00	16456.0
82.00	4311.0
83.00	3990.0
84.00	726.0
85.00	2860.0
86.00	4948.0
87.00	1992.0
88.00	887.0
89.00	368.0
91.00	3867.0
91.90	4468.0
93.00	27592.0
94.00	2062.0
95.00	498.0
96.00	1447.0
97.00	544.0
98.00	21440.0
99.00	17840.0
99.90	1542.0
101.00	9315.0
101.90	436.0
103.00	3114.0
104.00	6529.0
105.00	5600.0
106.00	1867.0
107.00	72816.0
108.00	10956.0
109.00	1807.0
110.00	131904.0
111.00	20328.0
112.00	2628.0
112.90	921.0
114.90	348.0
116.00	4335.0
117.00	67536.0
118.00	5208.0
118.90	710.0
119.90	919.0
121.10	271.0
122.00	4740.0
123.00	7121.0
124.00	3716.0
125.00	3222.0
127.00	253184.0
128.00	19072.0
129.00	100920.0
130.00	8355.0
131.00	1773.0
132.00	1040.0
133.00	458.0
133.90	3157.0
135.00	8128.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

136.00	3652.0
137.00	4394.0
138.00	991.0
139.00	625.0
139.80	1356.0
141.00	13372.0
142.00	4324.0
143.00	3311.0
144.00	785.0
144.80	620.0
146.00	2449.0
147.00	7250.0
147.90	17208.0
149.00	3396.0
150.00	1064.0
151.00	1638.0
151.60	1101.0
152.90	4033.0
154.00	2898.0
155.00	7358.0
156.00	11953.0
157.00	1898.0
157.90	2588.0
158.90	2134.0
160.00	4295.0
161.00	6136.0
162.00	2050.0
163.00	717.0
163.90	580.0
165.00	5158.0
166.00	4329.0
167.00	29664.0
168.00	15899.0
168.90	2394.0
170.00	1000.0
170.90	1342.0
172.00	2128.0
173.00	3512.0
174.00	6190.0
175.00	10642.0
176.00	3093.0
176.90	5146.0
177.90	1925.0
178.90	19552.0
180.00	13819.0
181.00	6667.0
181.90	1065.0
182.80	625.0
184.00	1723.0
185.00	9492.0
186.00	77280.0
187.00	23040.0
188.00	2792.0
189.00	4940.0
189.90	736.0
191.00	2524.0
191.90	6632.0
193.00	7266.0
193.90	1521.0
195.00	968.0
196.00	15067.0
197.90	556864.0
198.90	37496.0
200.00	2652.0
201.50	2659.0
202.90	4077.0
204.00	19848.0
205.00	33704.0
206.00	147584.0
207.00	21408.0
208.00	5469.0
209.00	1605.0
210.20	2755.0
211.00	5112.0
212.80	302.0
213.90	248.0
214.90	1832.0
216.00	3790.0
217.00	37840.0
218.00	4457.0
219.10	389.0
221.00	21200.0
221.70	5457.0
223.00	8181.0
224.00	81528.0
225.10	20000.0
226.10	2348.0
227.00	34336.0
227.90	4995.0
229.00	7026.0
230.00	960.0
231.00	3072.0
232.00	493.0
232.90	560.0
234.00	2106.0
234.90	2633.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

235.90	1582.0
237.00	2761.0
237.90	371.0
238.90	1467.0
240.00	1050.0
240.90	1710.0
242.00	4210.0
243.00	4637.0
244.00	63320.0
245.00	7905.0
246.00	13400.0
247.00	2732.0
247.90	660.0
248.90	2275.0
250.00	433.0
251.00	560.0
252.00	535.0
253.00	1495.0
255.00	322368.0
256.00	47400.0
257.00	3634.0
258.00	18896.0
259.00	3467.0
260.00	581.0
260.90	628.0
263.00	257.0
264.00	512.0
265.00	7589.0
266.00	1010.0
266.80	210.0
269.90	385.0
271.00	908.0
271.90	1017.0
273.00	10596.0
274.00	27720.0
275.00	135808.0
276.00	19368.0
277.00	12916.0
278.00	2253.0
278.90	381.0
281.00	547.0
281.90	345.0
283.00	1224.0
284.00	858.0
285.00	1943.0
285.90	428.0
288.00	206.0
288.90	570.0
289.90	478.0
291.00	487.0
292.00	440.0
293.00	2848.0
293.90	608.0
295.10	1155.0
296.00	43848.0
297.00	5458.0
298.00	297.0
301.00	475.0
301.90	719.0
303.00	4747.0
304.00	1268.0
305.10	230.0
307.90	724.0
308.90	303.0
309.90	599.0
312.00	161.0
313.00	365.0
314.00	2051.0
314.90	3868.0
315.90	2866.0
317.00	568.0
320.90	1292.0
321.80	705.0
323.00	13038.0
324.00	2189.0
324.90	220.0
325.90	298.0
326.90	2378.0
328.00	1120.0
328.80	231.0
332.00	841.0
332.90	1217.0
334.00	9565.0
335.00	2020.0
336.00	216.0
339.00	268.0
341.00	1588.0
342.00	392.0
346.00	2962.0
347.00	671.0
351.10	385.0
352.00	4077.0
353.00	3214.0
354.00	4435.0
355.00	880.0
355.90	167.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

358.90	282.0
364.90	18992.0
365.90	3206.0
366.80	288.0
369.80	397.0
371.00	1180.0
372.00	6032.0
373.00	1715.0
377.10	233.0
382.90	2050.0
384.00	506.0
389.90	893.0
390.90	707.0
392.00	529.0
400.90	580.0
402.00	2992.0
403.00	4496.0
404.00	1633.0
405.10	251.0
409.90	180.0
414.90	151.0
421.00	3506.0
422.00	4256.0
423.00	27104.0
424.00	5393.0
425.00	615.0
441.00	64656.0
442.00	369408.0
443.00	89064.0
444.00	7569.0
445.00	511.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

m/z	Abundance
37.00	673.0
38.10	1677.0
39.00	9387.0
39.90	2437.0
41.00	590.0
42.00	208.0
43.00	327.0
44.00	4322.0
45.00	443.0
47.80	161.0
49.00	1047.0
50.00	32264.0
51.00	120744.0
52.00	6380.0
53.10	321.0
54.90	888.0
56.00	3514.0
57.00	7876.0
57.90	387.0
59.00	249.0
59.90	192.0
61.00	1905.0
62.00	1860.0
63.00	4706.0
63.90	994.0
64.90	2661.0
65.90	463.0
66.90	367.0
68.00	2728.0
69.00	130328.0
70.00	891.0
71.00	157.0
73.10	1536.0
74.00	13854.0
75.00	21856.0
76.00	7565.0
77.00	156928.0
78.00	10717.0
79.00	11200.0
79.90	7433.0
81.00	11507.0
82.00	3069.0
83.00	2596.0
83.90	477.0
85.00	2142.0
86.00	3360.0
86.90	1748.0
87.90	550.0
89.00	240.0
91.00	2649.0
92.00	3063.0
93.00	19440.0
93.90	1218.0
94.90	435.0
96.00	1112.0
97.00	512.0
98.00	14908.0
99.00	11683.0
100.00	905.0
101.00	6665.0
101.90	367.0
103.00	2490.0
103.90	3881.0
105.00	3631.0
106.00	1484.0
107.00	50976.0
108.00	7503.0
109.10	1720.0
110.00	92256.0
111.00	13663.0
112.00	1752.0
113.00	616.0
113.90	179.0
114.80	338.0
116.00	3270.0
117.00	48184.0
118.00	3081.0
119.00	625.0
120.00	664.0
120.90	228.0
122.00	3335.0
123.00	5993.0
124.00	2152.0
124.90	2312.0
126.10	1125.0
127.00	179328.0
128.00	12869.0
129.00	71856.0
130.00	5693.0
131.00	1111.0
131.90	635.0
132.90	597.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

134.00	2449.0
135.00	5379.0
136.00	2507.0
137.00	3530.0
137.90	684.0
138.90	541.0
139.80	897.0
140.90	9779.0
142.00	3334.0
142.90	2166.0
143.90	522.0
144.10	529.0
144.90	371.0
146.00	1870.0
147.00	5409.0
148.00	12062.0
149.00	2850.0
150.00	673.0
151.10	1477.0
151.90	849.0
152.90	3689.0
154.00	2241.0
155.00	5056.0
156.00	7904.0
157.00	1791.0
157.80	1824.0
159.00	1380.0
160.00	2967.0
161.00	4780.0
162.00	1212.0
163.00	508.0
163.90	378.0
165.00	4119.0
166.00	3206.0
167.00	20992.0
168.00	12263.0
168.90	1723.0
169.90	603.0
171.00	850.0
171.90	1493.0
172.90	2515.0
174.00	4729.0
175.00	7802.0
176.00	2401.0
176.90	3718.0
178.00	1444.0
178.90	15430.0
180.00	10122.0
181.00	4557.0
182.00	766.0
182.90	460.0
183.90	1187.0
185.00	7432.0
186.00	55960.0
187.00	17032.0
188.00	1693.0
188.90	4114.0
190.00	621.0
191.00	2105.0
192.00	4643.0
193.00	5851.0
193.90	1347.0
194.90	508.0
196.00	11257.0
197.90	425856.0
199.00	25944.0
200.00	2250.0
201.50	1812.0
202.10	514.0
203.00	3341.0
204.00	14956.0
205.00	26936.0
206.00	113392.0
207.00	15807.0
208.00	3847.0
208.90	1357.0
210.10	1765.0
210.90	3970.0
213.00	297.0
213.90	186.0
214.90	1580.0
215.90	2115.0
217.00	28744.0
218.00	3758.0
219.10	475.0
221.00	18328.0
221.80	4725.0
223.00	6626.0
224.00	63072.0
225.00	16340.0
226.00	1656.0
227.00	26424.0
228.00	3662.0
229.00	6063.0
230.00	721.0
231.00	2294.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

231.90	427.0
232.90	572.0
234.00	1790.0
235.00	1798.0
235.90	1179.0
237.00	2291.0
237.90	360.0
238.90	882.0
239.90	635.0
241.00	1455.0
242.00	3387.0
243.00	3537.0
244.00	51392.0
245.00	6675.0
246.00	10730.0
247.00	2001.0
248.00	391.0
248.90	1637.0
249.90	370.0
251.00	574.0
252.10	505.0
252.90	1136.0
255.00	254848.0
256.00	37032.0
256.90	2563.0
258.00	14956.0
258.90	2358.0
259.90	398.0
260.90	505.0
263.00	162.0
264.00	367.0
265.00	6368.0
265.90	1182.0
266.90	168.0
270.00	206.0
271.00	722.0
272.10	781.0
273.00	7130.0
274.00	20896.0
275.00	124280.0
276.00	15485.0
277.00	11267.0
277.90	1498.0
278.90	306.0
281.00	664.0
282.00	484.0
282.90	1206.0
284.00	861.0
285.00	1622.0
286.10	300.0
288.90	518.0
289.90	353.0
291.00	288.0
292.00	495.0
293.00	2303.0
294.00	721.0
296.00	35256.0
296.90	4769.0
297.90	350.0
299.00	161.0
300.90	424.0
301.90	626.0
303.00	3811.0
303.90	1071.0
305.20	184.0
308.00	526.0
308.90	446.0
309.90	736.0
313.10	302.0
314.00	1669.0
314.90	3994.0
316.00	1929.0
316.90	401.0
320.90	1215.0
322.00	922.0
323.00	12691.0
324.00	2298.0
325.70	155.0
326.90	2391.0
327.90	1027.0
332.10	742.0
333.00	958.0
334.00	7892.0
335.00	1972.0
336.10	180.0
339.10	180.0
339.80	187.0
341.00	1434.0
342.10	333.0
345.90	2954.0
347.00	572.0
350.80	209.0
352.00	3413.0
352.90	2637.0
354.00	3942.0
355.00	911.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313

359.00	276.0
364.90	17960.0
365.90	2576.0
370.10	321.0
371.00	927.0
372.00	6136.0
373.00	1697.0
373.90	229.0
376.90	229.0
382.90	2018.0
383.90	473.0
385.00	233.0
390.00	845.0
390.90	520.0
392.00	589.0
396.90	157.0
400.90	475.0
401.90	2933.0
402.90	3777.0
404.00	1343.0
405.00	336.0
414.90	187.0
421.00	3840.0
422.00	4035.0
423.00	23544.0
424.00	6048.0
425.00	610.0
441.00	63184.0
442.00	392704.0
443.00	77296.0
444.00	6393.0
445.00	372.0

Instrument :
BNA_N
ClientSampleId :
DFTPP313