

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040721\
 Data File : BN014217.D
 Acq On : 06 Apr 2021 10:07
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 4/8/2021 8:07:49 AM

Quant Time: Apr 06 12:15:48 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 12:12:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.644	152	7339	0.40	ng	0.00
7) Naphthalene-d8	10.416	136	28215	0.40	ng	0.00
13) Acenaphthene-d10	14.283	164	16691	0.40	ng	0.00
19) Phenanthrene-d10	17.031	188	38305	0.40	ng	0.00
29) Chrysene-d12	21.218	240	37981	0.40	ng	0.00
36) Perylene-d12	23.427	264	38484	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.236	112	1753	0.10	ng	0.00
5) Phenol-d6	6.814	99	2197	0.09	ng	0.00
8) Nitrobenzene-d5	8.781	82	1761	0.09	ng	-0.01
11) 2-Methylnaphthalene-d10	12.017	152	4444	0.07	ng	0.00
14) 2,4,6-Tribromophenol	15.780	330	470	0.07	ng	0.00
15) 2-Fluorobiphenyl	12.900	172	6016	0.10	ng	0.00
27) Fluoranthene-d10	19.064	212	10347	0.07	ng	0.00
31) Terphenyl-d14	19.670	244	8264	0.09	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.174	88	969	0.11	ng	88
6) bis(2-Chloroethyl)ether	7.080	93	1961	0.09	ng	99
9) Naphthalene	10.467	128	6986	0.09	ng	98
10) Hexachlorobutadiene	10.758	225	1321	0.10	ng	# 100
12) 2-Methylnaphthalene	12.088	142	4757	0.09	ng	99
16) Acenaphthylene	14.004	152	5824	0.09	ng	100
17) Acenaphthene	14.346	154	4543	0.09	ng	100
18) Fluorene	15.336	166	5688	0.09	ng	100
20) 4,6-Dinitro-2-methylph...	15.412	198	336	0.06	ng	# 15
21) 4-Bromophenyl-phenylether	16.228	248	1899	0.09	ng	98
22) Hexachlorobenzene	16.337	284	2135	0.09	ng	99
23) Atrazine	16.495	200	1222	0.08	ng	95
25) Phenanthrene	17.067	178	10038	0.09	ng	100
26) Anthracene	17.165	178	7845	0.08	ng	100
28) Fluoranthene	19.094	202	10947	0.08	ng	100
30) Pyrene	19.456	202	11200	0.09	ng	100
32) Benzo(a)anthracene	21.207	228	9747	0.08	ng	99
33) Chrysene	21.261	228	11800	0.09	ng	99
34) Bis(2-ethylhexyl)phtha...	21.144	149	3500	0.09	ng	99
35) Indeno(1,2,3-cd)pyrene	25.632	276	13372	0.09	ng	100
37) Benzo(b)fluoranthene	22.767	252	12723m	0.09	ng	
38) Benzo(k)fluoranthene	22.811	252	12639	0.09	ng	# 92
39) Benzo(a)pyrene	23.332	252	11262	0.09	ng	# 79
40) Dibenzo(a,h)anthracene	25.646	278	11077	0.08	ng	95
41) Benzo(g,h,i)perylene	26.303	276	11583	0.08	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040721\
Data File : BN014217.D
Acq On : 06 Apr 2021 10:07
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
4/8/2021 8:07:49 AM


Quant Time: Apr 06 12:15:48 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN040721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 06 12:12:15 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.644 min Scan# 604
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion: 152 Resp: 7339
Ion Ratio Lower Upper
152 100
150 155.3 126.2 189.2
115 59.3 49.4 74.0

Manual Integrations
APPROVED



mohammad
4/8/2021 8:07:49 AM




#2
1,4-Dioxane
Concen: 0.11 ng
RT: 3.174 min Scan# 49
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07



Tgt Ion: 88 Resp: 969
Ion Ratio Lower Upper
88 100
58 73.7 70.0 105.0
43 31.7 28.3 42.5









#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.236 min Scan# 305
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion:	112	Resp:	1753
Ion Ratio	Lower	Upper	
112	100		
64	59.9	48.4	72.6
63	33.1	26.1	39.1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM



#5
Phenol-d6
Concen: 0.09 ng
RT: 6.814 min Scan# 501
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion:	99	Resp:	2197
Ion Ratio	Lower	Upper	
99	100		
42	20.5	17.9	26.9
71	33.9	27.0	40.6









#6
bis(2-Chloroethyl)ether
Concen: 0.09 ng
RT: 7.080 min Scan# 534
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion:	93	Resp:	1961
Ion	Ratio	Lower	Upper
93	100		
63	81.2	65.3	97.9
95	33.9	26.3	39.5

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.416 min Scan# 853
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07


Tgt Ion:	136	Resp:	28215
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.9	13.3
54	9.1	8.4	12.6
68	6.7	5.8	8.8





#8
Nitrobenzene-d5
Concen: 0.09 ng
RT: 8.781 min Scan# 724
Delta R.T. -0.013 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
82	100		
128	33.3	36.1	54.1#
54	64.1	39.1	58.7#

Manual Integrations
APPROVED



mohammad
4/8/2021 8:07:49 AM




#9
Naphthalene
Concen: 0.09 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Tgt Ion	Ratio	Lower	Upper
128	100		
129	12.1	8.9	13.3
127	14.0	10.6	16.0









#10
Hexachlorobutadiene
Concen: 0.10 ng
RT: 10.758 min Scan# 880
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07





Tgt Ion:225 Resp: 1321
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.2 50.6 75.8



#11
2-Methylnaphthalene-d10
Concen: 0.07 ng
RT: 12.017 min Scan# 1054
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Tgt Ion:152 Resp: 4444
Ion Ratio Lower Upper
152 100
151 20.0 16.2 24.2



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
4/8/2021 8:07:49 AM



#12
2-Methylnaphthalene
Concen: 0.09 ng
RT: 12.088 min Scan# 1070
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.0	71.4	107.2
115	37.0	29.1	43.7

Manual Integrations
APPROVED



mohammad
4/8/2021 8:07:49 AM




#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.283 min Scan# 1304
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07



Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.7	80.4	120.6
160	45.5	36.0	54.0









#14
2,4,6-Tribromophenol
Concen: 0.07 ng
RT: 15.780 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion: 330 Resp: 470
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.0 30.6 45.8

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM




#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.900 min Scan# 1195
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion: 172 Resp: 6016
Ion Ratio Lower Upper
172 100
171 35.7 27.8 41.6
170 24.3 18.7 28.1









#16
Acenaphthylene
Concen: 0.09 ng
RT: 14.004 min Scan# 1282
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07





Tgt Ion:152 Resp: 5824
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.1 10.5 15.7



#17
Acenaphthene
Concen: 0.09 ng
RT: 14.346 min Scan# 1309
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Tgt Ion:154 Resp: 4543
Ion Ratio Lower Upper
154 100
153 110.7 88.8 133.2
152 54.6 43.8 65.6



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
4/8/2021 8:07:49 AM






#18
Fluorene
Concen: 0.09 ng
RT: 15.336 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1





Tgt Ion:166 Resp: 5688
Ion Ratio Lower Upper
166 100
165 97.8 78.6 118.0
167 13.4 10.7 16.1




#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.031 min Scan# 1524
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Tgt Ion:188 Resp: 38305
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 6.2 9.2





#20
4,6-Dinitro-2-methylphenol
Concen: 0.06 ng
RT: 15.412 min Scan# 1393
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Instrument :
BNA_N
ClientSampleID :
SSTDIC0.1




Tgt Ion:198 Resp: 336
Ion Ratio Lower Upper
198 100
51 144.8 46.3 69.5#
105 72.2 35.5 53.3#

Manual Integrations
APPROVED



mohammad
4/8/2021 8:07:49 AM




#21
4-Bromophenyl-phenylether
Concen: 0.09 ng
RT: 16.228 min Scan# 1458
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Tgt Ion:248 Resp: 1899
Ion Ratio Lower Upper
248 100
250 97.3 76.1 114.1
141 56.3 45.5 68.3





#22
Hexachlorobenzene
Concen: 0.09 ng
RT: 16.337 min Scan# 1467
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:284 Resp: 2135
Ion Ratio Lower Upper
284 100
142 35.6 28.9 43.3
249 31.9 25.0 37.4

Manual Integrations
APPROVED



mohammad
4/8/2021 8:07:49 AM




#23
Atrazine
Concen: 0.08 ng
RT: 16.495 min Scan# 1480
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07



Tgt Ion:200 Resp: 1222
Ion Ratio Lower Upper
200 100
173 30.0 21.9 32.9
215 50.7 38.1 57.1









#25
Phenanthrene
Concen: 0.09 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.0	22.6
179	15.6	12.2	18.4

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM



#26
Anthracene
Concen: 0.08 ng
RT: 17.165 min Scan# 1535
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.6	21.8
179	15.4	12.3	18.5









#27
Fluoranthene-d10
Concen: 0.07 ng
RT: 19.064 min Scan# 1816
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion: 212 Resp: 10347
Ion Ratio Lower Upper
212 100
106 12.6 10.1 15.1
104 7.0 5.6 8.4

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM



#28
Fluoranthene
Concen: 0.08 ng
RT: 19.094 min Scan# 1822
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion: 202 Resp: 10947
Ion Ratio Lower Upper
202 100
101 10.6 8.6 13.0
203 17.1 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.218 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion:240 Resp: 37981
Ion Ratio Lower Upper
240 100
120 13.1 10.8 16.2
236 26.9 21.4 32.2

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM



#30
Pyrene
Concen: 0.09 ng
RT: 19.456 min Scan# 1895
Delta R.T. -0.005 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07


Tgt Ion:202 Resp: 11200
Ion Ratio Lower Upper
202 100
200 20.2 16.3 24.5
203 17.6 13.9 20.9





#31
Terphenyl-d14
Concen: 0.09 ng
RT: 19.670 min Scan# 1938
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:244 Resp: 8264
Ion Ratio Lower Upper
244 100
212 7.9 5.9 8.9
122 12.1 9.2 13.8

Manual Integrations
APPROVED



mohammad
4/8/2021 8:07:49 AM




#32
Benzo(a)anthracene
Concen: 0.08 ng
RT: 21.207 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07



Tgt Ion:228 Resp: 9747
Ion Ratio Lower Upper
228 100
226 27.3 21.1 31.7
229 19.9 15.8 23.8









#33
Chrysene
Concen: 0.09 ng
RT: 21.261 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.2	23.0	34.6
229	20.1	15.8	23.8




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.09 ng
RT: 21.144 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.9	21.4	32.0
279	4.6	3.1	4.7









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.09 ng
RT: 25.632 min Scan# 3174
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07





Tgt Ion: 276 Resp: 13372
Ion Ratio Lower Upper
276 100
138 22.9 18.6 27.8
277 24.8 19.8 29.8



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.427 min Scan# 2530
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07




Tgt Ion: 264 Resp: 38484
Ion Ratio Lower Upper
264 100
260 24.9 19.7 29.5
265 24.8 19.2 28.8



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/8/2021 8:07:49 AM







#37
Benzo(b)fluoranthene
Concen: 0.09 ng m
RT: 22.767 min Scan# 2337
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion:252 Resp: 12723
Ion Ratio Lower Upper
252 100
253 26.7 18.5 27.7
125 18.4 10.3 15.5#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
4/8/2021 8:07:49 AM



#38
Benzo(k)fluoranthene
Concen: 0.09 ng
RT: 22.811 min Scan# 2350
Delta R.T. -0.000 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07

Tgt Ion:252 Resp: 12639
Ion Ratio Lower Upper
252 100
253 26.9 18.1 27.1
125 11.7 7.8 11.6#





#39

Benzo(a)pyrene

Concen: 0.09 ng

RT: 23.332 min Scan# 2502

Delta R.T. -0.000 min

Lab File: BN014217.D

Acq: 06 Apr 2021 10:07

Tgt Ion:	252	Resp:	11262
Ion Ratio	Lower	Upper	
252	100		
253	31.1	19.4	29.2#
125	34.2	16.6	24.8#

Instrument :





BNA_N

ClientSampleId :

SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/8/2021 8:07:49 AM



#40

Dibenzo(a,h)anthracene

Concen: 0.08 ng




RT: 25.646 min Scan# 3178


Delta R.T. -0.000 min

Lab File: BN014217.D


Acq: 06 Apr 2021 10:07

Tgt Ion:	278	Resp:	11077
Ion Ratio	Lower	Upper	
278	100		
139	16.8	12.6	19.0
279	27.2	19.4	29.0







#41
Benzo(g,h,i)perylene
Concen: 0.08 ng
RT: 26.303 min Scan# 3370
Delta R.T. -0.004 min
Lab File: BN014217.D
Acq: 06 Apr 2021 10:07



Tgt Ion:	276	Resp:	11583
Ion Ratio	Lower	Upper	
276	100		
277	25.3	18.9	28.3
138	21.8	16.4	24.6



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

mohammad
4/8/2021 8:07:49 AM