

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040724\
 Data File : BN030747.D
 Acq On : 07 Apr 2024 18:41
 Operator : MA/JU
 Sample : P2017-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH5E8

Quant Time: Apr 08 02:18:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Mar 31 12:24:07 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.687	152	174392	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.469	136	812545	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.328	164	575639	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.075	188	1315878	20.000	ng/u1	-0.01
79) Chrysene-d12	21.310	240	1279793	20.000	ng/u1	-0.02
88) Perylene-d12	24.251	264	1465599	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.193	96	16369	4.401	ng/uL	0.00
4) Pyridine-d5	3.599	84	63098	5.658	ng/u1	0.00
7) Phenol-d5	6.858	99	112300	7.960	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.022	67	276738	31.760	ng/u1	-0.02
11) 2-Chlorophenol-d4	7.216	132	291803	27.266	ng/u1	-0.01
15) 4-Methylphenol-d8	8.387	113	208003	17.592	ng/u1	-0.01
21) Nitrobenzene-d5	8.840	128	193319	34.297	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.557	143	192003	37.236	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.087	165	359940	31.885	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.610	131	148750	8.004	ng/u1	-0.01
46) Dimethylphthalate-d6	13.745	166	1476027	38.405	ng/u1	-0.01
49) Acenaphthylene-d8	14.022	160	1541255	34.879	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.534	143	54268	7.382	ng/u1	0.00
60) Fluorene-d10	15.322	176	1302828	38.738	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.445	200	184844	31.825	ng/u1	0.00
73) Anthracene-d10	17.175	188	2206062	38.910	ng/u1	-0.01
81) Pyrene-d10	19.475	212	2684453	39.014	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.051	264	2782297	41.200	ng/u1	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040724\
Data File : BN030747.D
Acq On : 07 Apr 2024 18:41
Operator : MA/JU
Sample : P2017-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH5E8

Quant Time: Apr 08 02:18:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN032824.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Mar 31 12:24:07 2024
Response via : Initial Calibration

