


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN040920\
 Data File : BN010446.D
 Acq On : 09 Apr 2020 09:49
 Operator : CG/JU
 Sample : DFTPP65
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP65

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040920MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Apr 09 14:06:32 2020



AutoFind: Scans 2493, 2494, 2495; Background Corrected with Scan 2487

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	18.6	213525	PASS
68	69	0.00	2	1.1	4312	PASS
69	198	0.00	100	33.4	382467	PASS
70	69	0.00	2	0.6	2267	PASS
127	198	10	80	42.8	491136	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1146304	PASS
199	198	5	9	6.4	73114	PASS
275	198	10	60	28.2	323242	PASS
365	198	1	100	4.1	46650	PASS
441	198	0.01	100	14.1	161546	PASS
442	198	50	100	85.1	975189	PASS
443	442	15	24	19.2	187029	PASS

m/z	Abundance
37.00	833.0
37.90	3192.0
39.00	14797.0
39.90	915.0
40.90	834.0
42.90	408.0
44.00	1007.0
48.90	1932.0
50.00	42040.0
51.00	157120.0
52.00	9909.0
53.00	755.0
55.00	1436.0
56.00	6333.0
57.00	12220.0
57.90	516.0
58.90	324.0
61.00	2730.0
62.00	4033.0
63.00	8782.0
63.90	1595.0
65.00	5137.0
66.10	801.0
67.10	1000.0
68.00	5592.0
68.90	297088.0
69.90	1800.0
71.00	402.0
73.00	5665.0
74.00	33984.0
75.00	47552.0
76.00	17000.0
77.00	343680.0
78.00	22760.0
79.00	24544.0
80.00	18008.0
81.00	27144.0
82.00	5767.0
83.00	6350.0
83.90	1026.0
85.00	4386.0
86.00	7991.0
87.00	4242.0
87.90	1060.0
89.00	635.0
91.00	5953.0
92.10	7370.0
93.00	41760.0
94.00	2808.0
95.00	983.0
96.00	1934.0
98.00	34256.0
99.00	25248.0
99.90	1663.0
101.00	15035.0
102.10	671.0
103.00	4659.0
104.00	8722.0
105.00	8438.0
106.00	2499.0
107.00	113248.0
108.00	15690.0
109.00	3536.0
110.00	207488.0
111.00	28416.0
112.00	3666.0
112.90	1413.0
115.00	1115.0
116.00	6918.0
117.00	101768.0
118.00	7574.0
119.10	808.0
119.90	1615.0
121.20	449.0
121.90	7189.0
123.00	12273.0
124.00	6658.0
125.00	4569.0
127.00	401216.0
128.00	27968.0
129.00	153856.0
130.00	13411.0
131.00	2656.0
132.00	1619.0
132.90	564.0
134.00	5198.0
135.00	13383.0
136.00	4796.0
137.00	5842.0
138.00	1492.0
138.90	1306.0
140.10	2052.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

141.00	21816.0
142.00	7448.0
142.90	3402.0
144.10	1732.0
145.10	1592.0
146.10	4254.0
147.00	10891.0
148.00	24352.0
149.00	6342.0
150.00	2289.0
151.10	2766.0
151.80	2148.0
153.00	8223.0
153.90	6147.0
155.00	12776.0
156.00	19192.0
157.10	3867.0
158.00	4379.0
158.90	4027.0
160.00	6387.0
161.00	11843.0
162.00	3220.0
163.10	1223.0
163.90	1063.0
165.00	9100.0
165.90	6982.0
167.00	47600.0
168.00	22384.0
169.00	3955.0
170.10	1832.0
170.80	1607.0
172.00	4946.0
172.90	5157.0
174.00	10535.0
175.00	17176.0
176.10	4655.0
177.00	9242.0
177.90	3561.0
179.00	37376.0
180.00	27072.0
181.00	11198.0
182.00	1946.0
183.00	845.0
184.00	4233.0
185.00	18456.0
186.00	144064.0
187.00	39424.0
188.00	5409.0
189.00	9118.0
190.10	2016.0
191.00	4926.0
192.00	13084.0
193.00	14147.0
194.00	3478.0
195.00	1642.0
196.00	29728.0
198.00	1046336.0
199.00	63440.0
200.00	4449.0
201.60	3374.0
203.00	9559.0
204.00	39168.0
205.00	61392.0
206.00	270976.0
207.00	40768.0
208.00	8664.0
209.00	3882.0
210.20	4334.0
211.00	9865.0
212.10	1118.0
212.90	746.0
214.00	505.0
215.00	3207.0
216.10	7307.0
217.00	72464.0
218.00	10313.0
219.00	1562.0
221.00	43712.0
221.80	9220.0
223.00	16984.0
224.10	158592.0
225.00	39728.0
226.10	4619.0
227.00	71248.0
228.00	8491.0
229.00	13909.0
229.90	1939.0
231.00	6308.0
232.00	1772.0
232.90	1594.0
234.00	4821.0
235.00	5194.0
235.90	3682.0
237.10	5427.0
237.90	912.0
239.00	2240.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

239.90	1940.0
240.90	3519.0
242.00	8995.0
243.00	9466.0
244.10	124464.0
245.00	14730.0
246.00	26160.0
247.00	5521.0
248.00	973.0
249.00	4764.0
250.00	812.0
251.00	1359.0
252.00	1513.0
252.90	2991.0
255.00	701312.0
256.00	97752.0
257.00	8108.0
258.00	44624.0
259.00	6445.0
260.00	967.0
260.90	918.0
263.20	506.0
264.20	978.0
265.00	15560.0
266.00	2639.0
266.80	603.0
270.00	615.0
271.10	2108.0
272.10	1804.0
273.00	19776.0
274.00	52576.0
275.00	304128.0
276.00	38088.0
277.00	26416.0
278.00	3770.0
278.90	1231.0
281.00	1875.0
282.10	1064.0
283.10	3261.0
284.10	1893.0
285.00	4520.0
285.90	896.0
289.00	1286.0
290.00	1021.0
291.00	820.0
291.80	1508.0
293.00	5549.0
294.00	1181.0
296.00	88544.0
297.00	12137.0
298.00	881.0
299.10	371.0
300.90	1230.0
301.80	1655.0
303.00	9638.0
304.00	2632.0
304.90	622.0
307.90	1265.0
309.10	862.0
310.10	1493.0
311.00	314.0
312.90	1036.0
314.00	5619.0
315.00	11285.0
315.90	5325.0
317.10	962.0
319.90	652.0
321.00	2398.0
321.90	1666.0
323.00	32272.0
324.00	5702.0
324.90	752.0
326.20	764.0
327.00	5428.0
328.00	2468.0
328.90	629.0
332.00	1411.0
333.00	3090.0
334.00	20616.0
335.00	5105.0
336.10	611.0
338.90	850.0
340.00	467.0
341.00	3644.0
342.20	596.0
346.00	6415.0
347.00	1102.0
350.90	1161.0
352.00	10256.0
353.00	6868.0
354.00	10738.0
355.00	2534.0
358.80	358.0
359.20	462.0
361.00	370.0
363.10	350.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

363.90	403.0
365.00	49384.0
366.00	6335.0
367.10	496.0
370.10	1588.0
370.90	2868.0
372.00	18176.0
373.00	4790.0
374.10	640.0
376.90	349.0
383.00	4720.0
383.90	1746.0
385.00	339.0
390.00	2410.0
391.00	2220.0
392.00	1584.0
401.00	1541.0
402.00	6607.0
403.00	8608.0
403.90	4209.0
405.10	509.0
410.10	645.0
415.00	837.0
416.00	366.0
419.90	583.0
421.00	9176.0
422.00	10800.0
423.00	62704.0
424.00	13488.0
425.00	1883.0
439.30	372.0
441.00	189824.0
442.10	1071616.0
443.00	199680.0
444.00	19648.0
445.10	1237.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

m/z	Abundance
37.00	1534.0
38.00	3449.0
39.00	23824.0
40.00	1641.0
41.10	815.0
43.90	1047.0
44.90	547.0
47.90	377.0
49.10	2442.0
50.00	65632.0
51.00	250048.0
52.00	12948.0
53.10	724.0
55.00	1703.0
56.00	8284.0
57.00	18032.0
57.80	1045.0
60.00	390.0
60.90	4889.0
62.00	5298.0
63.00	15483.0
64.00	1656.0
65.00	7265.0
65.90	836.0
67.20	924.0
69.00	445056.0
70.00	3442.0
71.10	572.0
73.10	5832.0
74.00	49560.0
75.00	71760.0
76.00	25728.0
77.00	522304.0
78.00	34624.0
79.00	38840.0
80.00	28000.0
81.00	39688.0
81.90	8242.0
83.00	8688.0
83.90	1258.0
85.00	7636.0
86.00	10763.0
87.00	5594.0
87.90	1881.0
88.90	1000.0
91.00	11868.0
92.00	9600.0
93.00	64288.0
94.00	4266.0
94.90	1764.0
96.00	3103.0
96.90	1661.0
98.00	48896.0
98.90	41208.0
100.10	3407.0
101.00	19584.0
102.10	1480.0
103.00	7358.0
103.90	14393.0
105.00	13587.0
106.10	4973.0
107.00	166144.0
108.00	25392.0
109.10	5864.0
110.00	311808.0
111.00	45672.0
112.00	5699.0
113.00	2621.0
114.20	354.0
114.90	1003.0
115.00	985.0
116.00	9801.0
117.00	148608.0
118.00	10748.0
118.90	1531.0
119.90	1546.0
121.00	694.0
122.00	11668.0
123.00	17632.0
124.00	8343.0
125.00	7028.0
127.00	577536.0
128.00	45296.0
129.00	228160.0
130.00	20464.0
131.00	4261.0
132.00	2288.0
132.80	1295.0
134.00	7003.0
135.00	18720.0
136.00	7807.0
137.00	10875.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

138.00	1733.0
139.00	999.0
140.00	3373.0
141.00	33608.0
142.00	11540.0
143.00	8110.0
144.10	2181.0
145.00	2334.0
146.00	6298.0
147.00	18912.0
148.00	39352.0
149.00	9170.0
150.00	2954.0
151.00	3299.0
151.80	3618.0
153.00	10068.0
154.00	8445.0
155.00	18600.0
156.00	26888.0
157.10	5981.0
158.00	6567.0
159.00	4705.0
159.90	10493.0
161.00	15178.0
162.00	4644.0
163.00	1683.0
164.00	1732.0
165.00	12102.0
166.10	10492.0
167.00	65584.0
168.00	33744.0
169.00	5797.0
170.00	3064.0
171.00	3276.0
172.00	5603.0
173.00	9116.0
174.00	14103.0
175.10	26456.0
176.00	6562.0
177.00	12884.0
178.00	4890.0
179.00	51544.0
180.00	32408.0
181.00	16784.0
182.00	2971.0
183.00	1352.0
183.90	3554.0
185.00	24528.0
186.00	181504.0
187.10	55392.0
188.10	5800.0
189.00	10695.0
189.90	2232.0
191.00	7015.0
192.00	16832.0
193.00	17144.0
194.00	5264.0
194.90	2649.0
196.00	42912.0
198.00	1321984.0
199.00	87824.0
200.00	8765.0
201.50	6909.0
203.00	11127.0
204.00	53480.0
205.00	84168.0
206.00	381312.0
207.10	57616.0
208.00	11835.0
209.00	4260.0
210.10	5707.0
211.00	14312.0
211.80	2396.0
213.00	1570.0
213.90	536.0
214.90	5497.0
216.10	8974.0
217.00	107264.0
218.00	12411.0
219.00	1070.0
221.00	54680.0
222.00	12946.0
223.00	19904.0
224.00	217152.0
225.00	54144.0
226.10	5237.0
227.00	87680.0
228.00	13732.0
229.00	19464.0
230.10	2665.0
231.00	7691.0
232.00	1111.0
233.00	1390.0
234.00	7037.0
235.00	6613.0
235.90	4157.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

237.00	7577.0
237.90	596.0
239.00	3057.0
240.00	2209.0
241.00	5622.0
242.00	10285.0
243.10	10512.0
244.00	153024.0
245.00	20896.0
246.00	32792.0
247.00	6578.0
248.00	1331.0
249.00	5879.0
249.90	1328.0
250.90	2033.0
252.00	2143.0
253.10	4428.0
255.00	872256.0
256.00	125448.0
257.10	10269.0
258.00	50008.0
259.00	10273.0
259.90	1423.0
261.00	1921.0
261.80	343.0
262.90	602.0
264.10	1271.0
265.00	21624.0
266.00	4047.0
266.80	811.0
267.00	754.0
270.00	962.0
271.00	2550.0
272.00	2751.0
273.00	26200.0
274.00	70120.0
275.00	383936.0
276.00	53240.0
277.00	35304.0
278.10	6135.0
279.00	1460.0
281.00	1610.0
281.80	1172.0
283.00	3762.0
284.00	2581.0
285.10	6333.0
285.90	1230.0
288.90	1733.0
289.90	650.0
290.10	657.0
290.90	657.0
292.00	1931.0
293.00	7946.0
294.00	2048.0
295.00	2906.0
296.00	115280.0
297.00	14589.0
298.00	1490.0
301.00	1399.0
302.00	2359.0
303.00	13279.0
304.00	3503.0
304.80	369.0
308.10	1791.0
309.00	531.0
310.00	1990.0
311.00	332.0
312.30	341.0
313.00	1441.0
314.10	5954.0
315.00	12221.0
316.00	5956.0
316.90	1516.0
318.80	383.0
320.00	455.0
321.00	3584.0
322.00	2466.0
323.10	33448.0
324.00	8067.0
325.00	1227.0
326.20	917.0
327.00	7488.0
328.00	3771.0
329.00	477.0
332.00	2250.0
333.00	3353.0
334.00	24936.0
335.00	6207.0
336.10	752.0
339.00	646.0
340.00	488.0
341.00	5081.0
342.00	1266.0
342.90	447.0
346.00	7566.0
347.00	1606.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

351.00	739.0
352.00	10811.0
353.00	7902.0
354.00	10772.0
355.10	3293.0
356.00	332.0
356.90	302.0
358.90	1093.0
359.90	330.0
365.00	56304.0
366.00	8160.0
367.10	440.0
369.90	1210.0
371.10	2689.0
372.00	19544.0
373.10	5138.0
374.10	563.0
376.60	470.0
383.00	5332.0
383.90	1823.0
384.90	693.0
390.10	2656.0
391.10	1799.0
392.10	982.0
393.10	392.0
394.80	330.0
396.10	315.0
400.80	1267.0
402.00	7955.0
403.00	12918.0
404.00	5261.0
405.00	1144.0
409.80	661.0
415.00	570.0
421.00	8519.0
422.10	10010.0
423.00	73248.0
424.10	15422.0
425.00	1960.0
439.00	367.0
441.00	191360.0
442.00	1113600.0
443.00	208128.0
444.00	21040.0
445.00	1022.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

m/z	Abundance
36.00	380.0
36.90	1269.0
38.00	3923.0
39.00	18536.0
39.90	1116.0
41.10	638.0
42.90	421.0
43.90	1256.0
45.00	372.0
47.10	301.0
49.10	2364.0
50.00	60688.0
51.00	233408.0
52.00	12643.0
53.00	666.0
55.00	1466.0
56.00	7396.0
57.00	16480.0
57.90	793.0
58.70	323.0
61.00	3906.0
62.00	5948.0
63.00	13259.0
64.00	1965.0
65.00	7390.0
65.90	451.0
67.10	714.0
68.10	7344.0
69.00	406656.0
70.00	1559.0
72.00	434.0
73.00	4800.0
74.00	44336.0
75.00	68504.0
76.10	22336.0
77.00	454208.0
78.00	31008.0
79.00	34776.0
80.00	23736.0
81.00	33824.0
82.00	8417.0
83.00	9684.0
83.90	1018.0
84.90	5713.0
86.00	8594.0
87.00	4711.0
87.90	1182.0
89.00	1213.0
91.00	8656.0
92.00	10117.0
93.00	54360.0
94.00	3245.0
94.90	1293.0
96.10	2164.0
97.00	1333.0
98.00	41840.0
99.00	32144.0
99.90	2295.0
101.00	18400.0
102.00	1333.0
102.90	6292.0
104.00	11812.0
105.00	12960.0
106.00	4416.0
107.00	142656.0
108.00	22440.0
109.10	5293.0
110.00	253568.0
111.00	39624.0
112.00	4722.0
113.00	2191.0
113.90	358.0
115.00	1361.0
116.00	8877.0
117.00	127304.0
118.00	9853.0
119.00	1707.0
119.90	1825.0
121.00	799.0
122.00	10083.0
123.00	15769.0
124.00	6328.0
125.00	5887.0
127.00	494656.0
128.00	41376.0
129.00	197056.0
130.00	16196.0
130.90	2918.0
132.10	1836.0
133.00	948.0
133.90	5438.0
135.00	13786.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

136.10	5830.0
137.00	8099.0
137.80	1352.0
138.10	1378.0
138.80	710.0
139.20	1020.0
139.90	2635.0
141.00	25536.0
142.00	8090.0
142.90	6598.0
144.00	2009.0
144.90	1926.0
146.00	5296.0
147.00	13663.0
147.90	30216.0
149.00	7273.0
149.90	1997.0
151.10	4053.0
151.80	2280.0
152.90	8644.0
154.00	7021.0
155.00	15681.0
156.00	23112.0
157.00	4301.0
158.00	6178.0
158.90	3511.0
160.00	8814.0
161.00	11462.0
162.00	3900.0
163.00	1419.0
164.10	1517.0
164.90	10159.0
166.00	8025.0
167.00	56352.0
168.00	28248.0
169.00	4760.0
170.00	2089.0
170.90	2664.0
172.00	5478.0
173.00	6807.0
174.00	10264.0
175.00	22048.0
176.00	5442.0
176.90	10882.0
177.90	2803.0
178.20	3002.0
179.00	42344.0
180.00	27032.0
181.00	13995.0
182.00	2771.0
183.00	1413.0
184.00	3079.0
185.00	20192.0
186.00	152512.0
187.00	46712.0
188.00	5768.0
189.00	9748.0
190.10	2037.0
191.00	6626.0
192.00	12415.0
193.00	15075.0
194.00	3891.0
195.00	1987.0
196.00	27304.0
198.00	1070592.0
199.00	68080.0
200.00	6122.0
201.50	3929.0
203.00	8277.0
204.00	39088.0
205.00	69176.0
206.00	288384.0
207.10	45496.0
208.10	11221.0
209.00	3724.0
210.10	5069.0
211.10	10150.0
211.80	2307.0
213.00	853.0
213.90	392.0
215.00	3566.0
216.10	6588.0
217.00	76944.0
218.00	10385.0
219.10	1613.0
221.00	39224.0
221.90	10316.0
223.00	18168.0
224.00	145920.0
225.10	40488.0
226.00	4170.0
227.00	64168.0
228.00	9407.0
229.00	15173.0
229.90	2250.0
231.00	5582.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

232.10	826.0
233.00	1203.0
233.90	4670.0
235.00	5632.0
235.90	2547.0
237.00	5890.0
238.00	1081.0
239.00	2455.0
240.00	2204.0
241.00	4854.0
242.10	6224.0
243.00	9677.0
244.10	113304.0
245.00	16536.0
246.00	23832.0
247.00	3614.0
248.00	1225.0
249.00	4744.0
249.90	1114.0
251.00	1726.0
251.90	1192.0
252.90	3094.0
255.00	614400.0
256.00	89912.0
257.00	7136.0
258.00	35880.0
259.10	5483.0
259.90	1200.0
261.10	774.0
263.00	642.0
264.90	14171.0
265.90	2507.0
267.00	868.0
270.20	768.0
271.00	1577.0
271.90	1727.0
273.00	19968.0
274.00	52584.0
275.00	281664.0
276.00	37800.0
277.00	25176.0
278.00	3613.0
279.00	1457.0
281.00	1126.0
282.00	914.0
283.00	2341.0
284.00	1430.0
285.00	3789.0
286.00	1293.0
287.90	456.0
288.90	1543.0
290.00	663.0
290.90	507.0
292.10	1207.0
293.10	4906.0
294.00	1242.0
295.10	1822.0
296.00	75504.0
297.00	11688.0
298.00	887.0
301.10	1098.0
301.90	1638.0
303.00	8231.0
304.10	2364.0
305.00	402.0
307.90	1025.0
309.00	818.0
309.90	1029.0
310.90	326.0
312.90	829.0
314.10	2746.0
315.00	8723.0
316.10	4537.0
317.00	1040.0
321.00	2756.0
322.10	2026.0
323.10	23992.0
324.00	4139.0
324.90	831.0
326.10	801.0
327.00	5110.0
328.10	2509.0
329.00	449.0
331.90	1624.0
333.00	2236.0
334.00	16584.0
335.10	4244.0
336.00	307.0
338.90	648.0
339.90	484.0
341.10	3483.0
342.10	1116.0
346.00	5642.0
346.90	992.0
352.10	7422.0
353.00	5810.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

354.00	6390.0
354.90	1557.0
359.10	635.0
365.00	34264.0
365.90	5130.0
370.00	844.0
371.00	2154.0
372.00	15593.0
373.00	3195.0
373.90	494.0
377.00	388.0
383.00	4055.0
384.00	1213.0
384.90	413.0
389.90	1855.0
390.90	1404.0
391.80	728.0
401.00	896.0
402.10	5034.0
403.00	7702.0
404.10	3581.0
405.10	938.0
409.90	361.0
421.00	4766.0
422.00	6037.0
423.10	47272.0
424.10	11219.0
425.00	1218.0
439.50	472.0
441.10	103456.0
442.00	740352.0
443.00	153280.0
444.00	14417.0
444.80	757.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65

m/z	Abundance
39.90	788.0
41.00	539.0
43.10	429.0
43.90	1348.0
45.20	305.0
55.00	361.0
56.20	359.0
67.10	408.0
67.80	350.0
69.00	466.0
73.00	1199.0
79.10	418.0
81.10	625.0
82.10	379.0
83.70	394.0
85.90	464.0
91.10	563.0
93.10	386.0
93.90	319.0
95.20	323.0
96.00	471.0
97.20	415.0
102.90	475.0
116.80	411.0
133.00	445.0
135.10	496.0
147.00	674.0
149.30	373.0
162.00	320.0
165.00	569.0
177.00	517.0
179.10	715.0
190.90	610.0
193.10	984.0
194.10	356.0
207.00	4039.0
207.90	1545.0
209.10	933.0
225.10	321.0
267.10	346.0
281.10	1301.0
282.00	337.0
340.80	307.0

Instrument :
BNA_N
ClientSampleId :
DFTPP65