

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041019\
 Data File : BN005135.D
 Acq On : 10 Apr 2019 19:29
 Operator : JU/SJ
 Sample : K2311-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 0168-SITE-WATER

Quant Time: Apr 11 08:40:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 27 05:27:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	7365	20.00	ng	-0.02
21) Naphthalene-d8	10.43	136	29812	20.00	ng	-0.03
39) Acenaphthene-d10	14.30	164	16986	20.00	ng	-0.03
64) Phenanthrene-d10	17.05	188	35865	20.00	ng	-0.02
76) Chrysene-d12	21.25	240	30806	20.00	ng	-0.02
87) Perylene-d12	23.48	264	31069	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	34304	80.53	ng	-0.03
7) Phenol-d6	6.83	99	33982	60.14	ng	-0.03
23) Nitrobenzene-d5	8.82	82	42438	86.11	ng	-0.03
42) 2,4,6-Tribromophenol	15.80	330	39490	144.15	ng	-0.02
45) 2-Fluorobiphenyl	12.92	172	110680	86.34	ng	-0.02
79) Terphenyl-d14	19.69	244	172222	107.63	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041019\
Data File : BN005135.D
Acq On : 10 Apr 2019 19:29
Operator : JU/SJ
Sample : K2311-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
0168-SITE-WATER

Quant Time: Apr 11 08:40:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN032619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 27 05:27:09 2019
Response via : Initial Calibration

