

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041124\
 Data File : BN030853.D
 Acq On : 11 Apr 2024 21:55
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN041124

Quant Time: Apr 12 01:12:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN041124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 12 01:10:04 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	94	0.00
2	1,4-Dioxane	0.400	0.424	-6.0	101	0.00
3	n-Nitrosodimethylamine	0.400	0.396	1.0	96	0.00
4 S	2-Fluorophenol	0.400	0.384	4.0	94	0.00
5 S	Phenol-d6	0.400	0.362	9.5	90	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	97	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	93	0.00
8 S	Nitrobenzene-d5	0.400	0.406	-1.5	100	0.00
9	Naphthalene	0.400	0.400	0.0	96	0.00
10	Hexachlorobutadiene	0.400	0.388	3.0	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	98	0.00
12	2-Methylnaphthalene	0.400	0.398	0.5	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.400	0.330	17.5	87	0.00
15 S	2-Fluorobiphenyl	0.400	0.428	-7.0	101	0.00
16	Acenaphthylene	0.400	0.416	-4.0	103	0.00
17	Acenaphthene	0.400	0.397	0.8	94	0.00
18	Fluorene	0.400	0.396	1.0	94	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.396	1.0	82	0.00
21	4-Bromophenyl-phenylether	0.400	0.386	3.5	94	0.00
22	Hexachlorobenzene	0.400	0.393	1.8	94	0.00
23	Atrazine	0.400	0.358	10.5	94	0.00
24	Pentachlorophenol	0.400	0.394	1.5	84	0.00
25	Phenanthrene	0.400	0.408	-2.0	98	0.00
26	Anthracene	0.400	0.391	2.3	99	0.00
27 SURR	Fluoranthene-d10	0.400	0.399	0.3	98	0.00
28	Fluoranthene	0.400	0.384	4.0	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.396	1.0	93	0.00
31 S	Terphenyl-d14	0.400	0.449	-12.2	107	0.00
32	Benzo(a)anthracene	0.400	0.393	1.8	96	0.00
33	Chrysene	0.400	0.412	-3.0	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.325	18.8	86	0.00
35 I	Perylene-d12	0.400	0.400	0.0	85	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.409	-2.2	93	0.00
37	Benzo(b)fluoranthene	0.400	0.401	-0.3	90	0.00
38	Benzo(k)fluoranthene	0.400	0.416	-4.0	93	0.00
39 C	Benzo(a)pyrene	0.400	0.406	-1.5	92	0.00
40	Dibenzo(a,h)anthracene	0.400	0.405	-1.3	92	0.00
41	Benzo(g,h,i)perylene	0.400	0.390	2.5	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
ClientSampleId :
ICVBN041124