

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\  
 Data File : BN010473.D  
 Acq On : 14 Apr 2020 08:38  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Apr 14 11:15:57 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Apr 13 18:04:22 2020  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.294	0.312	-6.1	108	0.00
3	n-Nitrosodimethylamine	0.191	0.194	-1.6	122	0.00
4 S	2-Fluorophenol	0.876	0.917	-4.7	105	0.00
5 S	Phenol-d6	1.145	1.199	-4.7	106	0.00
6	bis(2-Chloroethyl)ether	0.992	1.101	-11.0	110	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.280	0.300	-7.1	112	-0.01
9	Naphthalene	0.993	1.080	-8.8	110	-0.01
10	Hexachlorobutadiene	0.239	0.251	-5.0	105	-0.01
11 SURR	2-Methylnaphthalene-d10	0.658	0.672	-2.1	104	0.00
12	2-Methylnaphthalene	0.714	0.792	-10.9	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.189	0.175	7.4	99	0.00
15 S	2-Fluorobiphenyl	1.456	1.638	-12.5	114	0.00
16	Acenaphthylene	1.728	1.826	-5.7	111	0.00
17	Acenaphthene	1.110	1.190	-7.2	111	0.00
18	Fluorene	1.454	1.532	-5.4	109	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.222	0.236	-6.3	106	0.00
21	Hexachlorobenzene	0.254	0.282	-11.0	106	0.00
22	Pentachlorophenol	0.100	0.090	10.0	103	0.00
23	Phenanthrene	1.080	1.209	-11.9	109	0.00
24	Anthracene	0.988	1.093	-10.6	110	0.00
25 SURR	Fluoranthene-d10	1.213	1.238	-2.1	99	0.00
26	Fluoranthene	1.345	1.490	-10.8	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	1.346	1.434	-6.5	106	0.00
29 S	Terphenyl-d14	0.915	1.073	-17.3	119	0.00
30	Benzo(a)anthracene	1.305	1.350	-3.4	104	0.00
31	Chrysene	1.300	1.404	-8.0	109	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.619	7.6	97	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.273	-10.8	109	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.317	1.397	-6.1	106	0.00
36	Benzo(k)fluoranthene	1.320	1.438	-8.9	109	0.00
37 C	Benzo(a)pyrene	1.166	1.286	-10.3	111	0.00
38	Dibenzo(a,h)anthracene	1.066	1.169	-9.7	109	0.00
39	Benzo(a,h,i)perylene	1.073	1.226	-14.3	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\  
Data File : BN010473.D  
Acq On : 14 Apr 2020 08:38  
Operator : CG/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_N  
LabSampleId :  
SSTDCCC0.4

Quant Time: Apr 14 11:15:57 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Apr 13 18:04:22 2020  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
-----				
(#) = Out of Range	SPCC's out = 0 CCC's out = 0			