

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\
 Data File : BN010485.D
 Acq On : 14 Apr 2020 19:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 14 20:12:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.294	0.341	-16.0	112	0.00
3	n-Nitrosodimethylamine	0.191	0.245	-28.3#	146	-0.02
4 S	2-Fluorophenol	0.876	0.908	-3.7	99	0.00
5 S	Phenol-d6	1.145	1.221	-6.6	102	0.00
6	bis(2-Chloroethyl)ether	0.992	1.129	-13.8	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.280	0.304	-8.6	108	-0.01
9	Naphthalene	0.993	1.073	-8.1	103	-0.01
10	Hexachlorobutadiene	0.239	0.245	-2.5	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.658	0.666	-1.2	97	0.00
12	2-Methylnaphthalene	0.714	0.782	-9.5	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.189	0.183	3.2	97	0.00
15 S	2-Fluorobiphenyl	1.456	1.660	-14.0	109	0.00
16	Acenaphthylene	1.728	1.789	-3.5	102	0.00
17	Acenaphthene	1.110	1.199	-8.0	105	0.00
18	Fluorene	1.454	1.570	-8.0	105	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.222	0.236	-6.3	103	0.00
21	Hexachlorobenzene	0.254	0.272	-7.1	99	0.00
22	Pentachlorophenol	0.100	0.103	-3.0	114	0.00
23	Phenanthrene	1.080	1.197	-10.8	104	0.00
24	Anthracene	0.988	1.112	-12.6	108	0.00
25 SURR	Fluoranthene-d10	1.213	1.235	-1.8	96	0.00
26	Fluoranthene	1.345	1.473	-9.5	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	-0.01
28	Pvrene	1.346	1.576	-17.1	103	0.00
29 S	Terphenyl-d14	0.915	1.172	-28.1#	115	0.00
30	Benzo(a)anthracene	1.305	1.357	-4.0	92	0.00
31	Chrysene	1.300	1.401	-7.8	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.695	-3.7	96	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.236	-7.6	94	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	84	0.00
35	Benzo(b)fluoranthene	1.317	1.441	-9.4	92	0.00
36	Benzo(k)fluoranthene	1.320	1.470	-11.4	94	0.00
37 C	Benzo(a)pyrene	1.166	1.308	-12.2	95	0.00
38	Dibenzo(a,h)anthracene	1.066	1.203	-12.9	95	-0.02
39	Benzo(a,h,i)perylene	1.073	1.242	-15.8	97	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\
Data File : BN010485.D
Acq On : 14 Apr 2020 19:41
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 14 20:12:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			