

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\
 Data File : BN010485.D
 Acq On : 14 Apr 2020 19:41
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 14 20:12:02 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.463	-15.8	112	0.00
3	n-Nitrosodimethylamine	0.400	0.513	-28.2#	146	-0.02
4 S	2-Fluorophenol	0.400	0.415	-3.7	99	0.00
5 S	Phenol-d6	0.400	0.427	-6.7	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.455	-13.7	107	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.435	-8.7	108	-0.01
9	Naphthalene	0.400	0.432	-8.0	103	-0.01
10	Hexachlorobutadiene	0.400	0.409	-2.2	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	97	0.00
12	2-Methylnaphthalene	0.400	0.438	-9.5	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.400	0.387	3.3	97	0.00
15 S	2-Fluorobiphenyl	0.400	0.456	-14.0	109	0.00
16	Acenaphthylene	0.400	0.414	-3.5	102	0.00
17	Acenaphthene	0.400	0.432	-8.0	105	0.00
18	Fluorene	0.400	0.432	-8.0	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	0.00
20	4-Bromophenyl-phenylether	0.400	0.425	-6.2	103	0.00
21	Hexachlorobenzene	0.400	0.429	-7.2	99	0.00
22	Pentachlorophenol	0.400	0.412	-3.0	114	0.00
23	Phenanthrene	0.400	0.443	-10.7	104	0.00
24	Anthracene	0.400	0.451	-12.7	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.407	-1.7	96	0.00
26	Fluoranthene	0.400	0.438	-9.5	103	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	87	-0.01
28	Pvrene	0.400	0.468	-17.0	103	0.00
29 S	Terphenyl-d14	0.400	0.512	-28.0#	115	0.00
30	Benzo(a)anthracene	0.400	0.416	-4.0	92	0.00
31	Chrysene	0.400	0.431	-7.7	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.415	-3.7	96	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.430	-7.5	94	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	84	0.00
35	Benzo(b)fluoranthene	0.400	0.438	-9.5	92	0.00
36	Benzo(k)fluoranthene	0.400	0.446	-11.5	94	0.00
37 C	Benzo(a)pyrene	0.400	0.449	-12.2	95	0.00
38	Dibenzo(a,h)anthracene	0.400	0.452	-13.0	95	-0.02
39	Benzo(a,h,i)perylene	0.400	0.463	-15.8	97	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\
Data File : BN010485.D
Acq On : 14 Apr 2020 19:41
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 14 20:12:02 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				