

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\
 Data File : BN010487.D
 Acq On : 14 Apr 2020 20:53
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 14 23:06:17 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.294	0.329	-11.9	97	0.00
3	n-Nitrosodimethylamine	0.191	0.229	-19.9	122	-0.02
4 S	2-Fluorophenol	0.876	0.903	-3.1	88	0.00
5 S	Phenol-d6	1.145	1.230	-7.4	92	0.00
6	bis(2-Chloroethyl)ether	0.992	1.116	-12.5	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.280	0.303	-8.2	98	-0.01
9	Naphthalene	0.993	1.090	-9.8	96	-0.01
10	Hexachlorobutadiene	0.239	0.245	-2.5	88	-0.01
11 SURR	2-Methylnaphthalene-d10	0.658	0.669	-1.7	89	0.00
12	2-Methylnaphthalene	0.714	0.790	-10.6	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.189	0.177	6.3	88	0.00
15 S	2-Fluorobiphenyl	1.456	1.638	-12.5	100	0.00
16	Acenaphthylene	1.728	1.813	-4.9	96	0.00
17	Acenaphthene	1.110	1.194	-7.6	97	0.00
18	Fluorene	1.454	1.539	-5.8	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.222	0.235	-5.9	95	0.00
21	Hexachlorobenzene	0.254	0.273	-7.5	92	0.00
22	Pentachlorophenol	0.100	0.107	-7.0	110	0.00
23	Phenanthrene	1.080	1.200	-11.1	97	0.00
24	Anthracene	0.988	1.108	-12.1	100	0.00
25 SURR	Fluoranthene-d10	1.213	1.263	-4.1	91	0.00
26	Fluoranthene	1.345	1.502	-11.7	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	87	-0.01
28	Pvrene	1.346	1.505	-11.8	98	0.00
29 S	Terphenyl-d14	0.915	1.118	-22.2	108	0.00
30	Benzo(a)anthracene	1.305	1.364	-4.5	92	0.00
31	Chrysene	1.300	1.404	-8.0	95	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.653	2.5	89	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.255	-9.2	94	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.317	1.444	-9.6	94	-0.01
36	Benzo(k)fluoranthene	1.320	1.444	-9.4	94	0.00
37 C	Benzo(a)pyrene	1.166	1.311	-12.4	97	0.00
38	Dibenzo(a,h)anthracene	1.066	1.197	-12.3	96	-0.02
39	Benzo(a,h,i)perylene	1.073	1.232	-14.8	97	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN041420\
Data File : BN010487.D
Acq On : 14 Apr 2020 20:53
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 14 23:06:17 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			