

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024945.D
 Acq On : 15 Apr 2023 21:56
 Operator : CG/JU
 Sample : PB152129BS
 Misc :
 ALS Vial : 94 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Quant Time: Apr 17 04:05:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 02:58:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	309575	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	1360659	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	844560	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	1854325	20.000	ng/u1	0.00
79) Chrysene-d12	21.521	240	1774022	20.000	ng/u1	0.00
88) Perylene-d12	23.974	264	1857918	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	51798	6.368	ng/uL	0.00
4) Pyridine-d5	3.728	84	959568	39.398	ng/u1	0.00
7) Phenol-d5	7.093	99	1086642	36.003	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	698800	34.616	ng/u1	0.00
11) 2-Chlorophenol-d4	7.463	132	814495	36.722	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	857024	35.769	ng/u1	0.00
21) Nitrobenzene-d5	9.104	128	407894	37.024	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	452890	37.952	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	807323	37.850	ng/u1	0.00
31) 4-Chloroaniline-d4	10.887	131	535318	16.538	ng/u1	0.00
46) Dimethylphthalate-d6	13.992	166	2384836	36.425	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	2803923	37.372	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	513958	38.295	ng/u1	0.00
60) Fluorene-d10	15.575	176	2029697	36.718	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	460085	38.125	ng/u1	0.00
73) Anthracene-d10	17.428	188	3196587	36.704	ng/u1	0.00
81) Pyrene-d10	19.721	212	3744833	36.352	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	3742010	38.520	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.340	88	120516	13.730	ng/uL	94
5) Pyridine	3.746	79	1123983	44.666	ng/u1	96
6) Benzaldehyde	7.069	77	333861	22.640	ng/u1	98
8) Phenol	7.122	94	1189439	37.841	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.357	93	903876	36.628	ng/u1	99
12) 2-Chlorophenol	7.493	128	883020	38.117	ng/u1	99
13) 2-Methylphenol	8.381	108	876958	38.000	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.469	45	1730693	35.717	ng/u1	100
16) Acetophenone	8.763	105	1374944	36.291	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.752	70	745730	36.492	ng/u1	99
18) 4-Methylphenol	8.716	108	960781	37.439	ng/u1	97
19) Hexachloroethane	9.016	117	345887	36.996	ng/u1	96
22) Nitrobenzene	9.146	77	1119462	37.815	ng/u1	100
23) Isophorone	9.675	82	2048698	37.701	ng/u1	100
25) 2-Nitrophenol	9.857	139	501610	38.968	ng/u1	99
26) 2,4-Dimethylphenol	9.922	107	1021597	37.694	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.157	93	1228608	36.330	ng/u1	99
29) 2,4-Dichlorophenol	10.393	162	847670	39.598	ng/u1	97
30) Naphthalene	10.798	128	2838372	37.351	ng/u1	99
32) 4-Chloroaniline	10.910	127	651771	19.891	ng/u1	99
33) Hexachlorobutadiene	11.081	225	480504	38.511	ng/u1	95
34) Caprolactam	11.693	113	162583	21.669	ng/u1	92
35) 4-Chloro-3-methylphenol	12.034	107	997497	39.742	ng/u1	99
36) 2-Methylnaphthalene	12.404	142	1900022	37.832	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024945.D
 Acq On : 15 Apr 2023 21:56
 Operator : CG/JU
 Sample : PB152129BS
 Misc :
 ALS Vial : 94 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Quant Time: Apr 17 04:05:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 02:58:51 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	1937354	38.643	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.775	216	961102	39.014	ng/ul	99
40) Hexachlorocyclopentadiene	12.751	237	546883	36.405	ng/ul	94
41) 2,4,6-Trichlorophenol	13.016	196	631797	37.535	ng/ul	97
42) 2,4,5-Trichlorophenol	13.087	196	723197	39.262	ng/ul	98
43) 1,1'-Biphenyl	13.416	154	2525030	37.524	ng/ul	97
44) 2-Chloronaphthalene	13.457	162	1982699	37.947	ng/ul	99
45) 2-Nitroaniline	13.669	65	747848	40.989	ng/ul	97
47) Dimethylphthalate	14.039	163	2549360	38.700	ng/ul	100
48) 2,6-Dinitrotoluene	14.163	165	535944	41.681	ng/ul	99
50) Acenaphthylene	14.304	152	3160552	38.353	ng/ul	99
51) 3-Nitroaniline	14.492	138	585889	43.360	ng/ul	99
52) Acenaphthene	14.645	153	2151413	37.764	ng/ul	98
53) 2,4-Dinitrophenol	14.692	184	343836	41.590	ng/ul	95
55) 4-Nitrophenol	14.798	109	451143	41.302	ng/ul	93
56) Dibenzofuran	14.981	168	2970687	37.831	ng/ul	97
57) 2,4-Dinitrotoluene	14.945	165	779575	41.884	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.204	232	577888	38.100	ng/ul	98
59) Diethylphthalate	15.398	149	2579563	38.969	ng/ul	100
61) Fluorene	15.628	166	2423800	38.375	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.622	204	1170705	39.822	ng/ul	98
63) 4-Nitroaniline	15.657	138	621477	46.671	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.710	198	482818	40.843	ng/ul	98
67) N-Nitrosodiphenylamine	15.833	169	1960587	34.850	ng/ul	98
68) 4-Bromophenyl-phenylether	16.516	248	712935	38.250	ng/ul	97
69) Hexachlorobenzene	16.633	284	816365	38.685	ng/ul	99
71) Pentachlorophenol	16.975	266	485399	37.006	ng/ul	100
72) Phenanthrene	17.369	178	3969858	37.937	ng/ul	99
74) Anthracene	17.463	178	4005720	37.989	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.381	216	948286	35.954	ng/uL	92
76) Pentachlorobenzene	14.898	250	942153	37.337	ng/uL	97
77) Carbazole	17.733	167	3604550	37.859	ng/ul	98
78) Di-n-butylphthalate	18.292	149	4555296	40.310	ng/ul	100
80) Fluoranthene	19.386	202	4745423	38.436	ng/ul	100
82) Pyrene	19.751	202	4807535	36.994	ng/ul	99
83) Butylbenzylphthalate	20.645	149	2194346	40.671	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.439	252	846354	20.793	ng/ul	98
85) Benzo(a)anthracene	21.504	228	4903089	38.482	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.427	149	3223787	40.026	ng/ul	100
87) Chrysene	21.557	228	4703723	38.759	ng/ul	99
89) Di-n-octyl phthalate	22.368	149	5626680	44.069	ng/ul	100
90) Benzo(b)fluoranthene	23.221	252	5040198	41.366	ng/ul	97
91) Benzo(k)fluoranthene	23.274	252	5047263	42.098	ng/ul	97
93) Benzo(a)pyrene	23.868	252	4315375	40.453	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.527	276	5003332	38.710	ng/ul	98
95) Dibenzo(a,h)anthracene	26.539	278	4164590	39.028	ng/ul	99
96) Benzo(g,h,i)perylene	27.321	276	3900967	37.194	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024945.D
 Acq On : 15 Apr 2023 21:56
 Operator : CG/JU
 Sample : PB152129BS
 Misc :
 ALS Vial : 94 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Quant Time: Apr 17 04:05:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 02:58:51 2023
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 13.730 ng/uL
 RT: 3.340 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56




Instrument :
 BNA_N
 ClientSampleId :
 SLCS129


Tgt Ion: 88 Resp: 120516
 Ion Ratio Lower Upper
 88 100
 43 39.0 33.0 49.4
 58 83.4 72.3 108.5



#5
 Pyridine
 Concen: 44.666 ng/uL
 RT: 3.746 min Scan# 95
 Delta R.T. -0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Tgt Ion: 79 Resp: 1123983
 Ion Ratio Lower Upper
 79 100
 52 100.1 76.1 114.1
 51 45.7 35.8 53.6









#6
Benzaldehyde
Concen: 22.640 ng/ul
RT: 7.069 min Scan# 661
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion: 77 Resp: 333861
Ion Ratio Lower Upper
77 100
105 91.8 74.3 111.5
106 88.4 73.3 109.9



#8
Phenol
Concen: 37.841 ng/ul
RT: 7.122 min Scan# 669
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion: 94 Resp: 1189439
Ion Ratio Lower Upper
94 100
65 31.9 25.1 37.7
66 41.4 31.9 47.9





#10

Bis(2-Chloroethyl)ether

Concen: 36.628 ng/ul

RT: 7.357 min Scan# 709

Delta R.T. -0.000 min

Lab File: BN024945.D


Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129






Tgt Ion: 93 Resp: 903876

Ion Ratio Lower Upper

93 100

63 87.4 70.9 106.3

95 32.7 26.0 39.0



#12

2-Chlorophenol

Concen: 38.117 ng/ul

RT: 7.493 min Scan# 732

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:128 Resp: 883020


Ion Ratio Lower Upper

128 100

64 55.6 44.8 67.2

130 32.7 26.6 39.8





#13

2-Methylphenol

Concen: 38.000 ng/ul

RT: 8.381 min Scan# 883

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :





SLCS129

Tgt Ion:108 Resp: 876958

Ion Ratio Lower Upper

108 100

107 89.1 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 35.717 ng/ul

RT: 8.469 min Scan# 898

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion: 45 Resp: 1730693


Ion Ratio Lower Upper

45 100

77 10.6 8.4 12.6

79 8.5 6.6 9.8









#16
Acetophenone
Concen: 36.291 ng/ul
RT: 8.763 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:105 Resp: 1374944
Ion Ratio Lower Upper
105 100
77 83.5 64.8 97.2
51 37.1 28.7 43.1



#17
N-Nitroso-di-n-propylamine
Concen: 36.492 ng/ul
RT: 8.752 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion: 70 Resp: 745730
Ion Ratio Lower Upper
70 100
42 79.0 64.2 96.2
101 9.3 7.3 10.9
130 18.9 15.0 22.6









#18
4-Methylphenol
Concen: 37.439 ng/ul
RT: 8.716 min Scan# 940
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:108 Resp: 960781
Ion Ratio Lower Upper
108 100
107 110.0 85.8 128.6



#19
Hexachloroethane
Concen: 36.996 ng/ul
RT: 9.016 min Scan# 991
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:117 Resp: 345887
Ion Ratio Lower Upper
117 100
201 95.8 72.2 108.4
199 57.6 46.2 69.4









#22
Nitrobenzene
Concen: 37.815 ng/ul
RT: 9.146 min Scan# 1103
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion: 77 Resp: 1119462
Ion Ratio Lower Upper
77 100
123 42.0 33.5 50.3
65 14.3 11.8 17.8



#23
Isophorone
Concen: 37.701 ng/ul
RT: 9.675 min Scan# 1103
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion: 82 Resp: 2048698
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.0
138 16.1 12.8 19.2









#25
2-Nitrophenol
Concen: 38.968 ng/ul
RT: 9.857 min Scan# 1134
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:139 Resp: 501610
Ion Ratio Lower Upper
139 100
65 62.2 49.8 74.8
109 31.4 26.2 39.2



#26
2,4-Dimethylphenol
Concen: 37.694 ng/ul
RT: 9.922 min Scan# 1145
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:107 Resp: 1021597
Ion Ratio Lower Upper
107 100
121 49.1 39.7 59.5
122 83.1 66.1 99.1





#27

Bis(2-Chloroethoxy)methane

Concen: 36.330 ng/ul

RT: 10.157 min Scan# 1185

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :





BNA_N

ClientSampleId :

SLCS129

Tgt Ion: 93 Resp: 1228608

Ion	Ratio	Lower	Upper
93	100		
95	32.3	26.1	39.1
123	10.8	8.5	12.7



#29

2,4-Dichlorophenol

Concen: 39.598 ng/ul

RT: 10.393 min Scan# 1225




Delta R.T. -0.000 min


Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Tgt Ion:162 Resp: 847670

Ion	Ratio	Lower	Upper
162	100		
164	66.7	50.9	76.3
98	39.1	31.0	46.4









#30
Naphthalene
Concen: 37.351 ng/ul
RT: 10.798 min Scan# 1294
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:128 Resp: 2838372
Ion Ratio Lower Upper
128 100
129 10.9 8.4 12.6
127 13.3 10.7 16.1



#32
4-Chloroaniline
Concen: 19.891 ng/ul
RT: 10.910 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:127 Resp: 651771
Ion Ratio Lower Upper
127 100
129 32.4 25.5 38.3





#33

Hexachlorobutadiene

Concen: 38.511 ng/ul

RT: 11.081 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:225 Resp: 480504

Ion Ratio Lower Upper

225 100

223 65.2 45.9 68.9

227 63.5 51.2 76.8



#34

Caprolactam

Concen: 21.669 ng/ul

RT: 11.693 min Scan# 1446

Delta R.T. 0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:113 Resp: 162583


Ion Ratio Lower Upper

113 100

55 220.2 188.3 282.5

56 157.7 133.5 200.3









#35
4-Chloro-3-methylphenol
Concen: 39.742 ng/ul
RT: 12.034 min Scan# 1504
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:107 Resp: 997497
Ion Ratio Lower Upper
107 100
144 25.0 19.6 29.4
142 75.3 60.7 91.1



#36
2-Methylnaphthalene
Concen: 37.832 ng/ul
RT: 12.404 min Scan# 1567
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:142 Resp: 1900022
Ion Ratio Lower Upper
142 100
141 89.0 71.4 107.2





#37

1-Methylnaphthalene

Concen: 38.643 ng/ul

RT: 12.628 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:142 Resp: 1937354

Ion Ratio Lower Upper

142 100

141 94.5 73.5 110.3

116 3.9 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 39.014 ng/ul

RT: 12.775 min Scan# 1630

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Tgt Ion:216 Resp: 961102




Ion Ratio Lower Upper


216 100

214 77.6 62.2 93.4

179 21.9 16.8 25.2

108 19.0 15.0 22.6









#40
Hexachlorocyclopentadiene
Concen: 36.405 ng/ul
RT: 12.751 min Scan# 1670
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:237 Resp: 546883
Ion Ratio Lower Upper
237 100
235 68.1 49.9 74.9
272 13.3 10.2 15.4



#41
2,4,6-Trichlorophenol
Concen: 37.535 ng/ul
RT: 13.016 min Scan# 1671
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:196 Resp: 631797
Ion Ratio Lower Upper
196 100
198 95.9 74.5 111.7
200 31.8 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 39.262 ng/ul

RT: 13.087 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:196 Resp: 723197

Ion Ratio Lower Upper

196 100

198 99.5 78.2 117.2

200 31.9 24.0 36.0



#43

1,1'-Biphenyl

Concen: 37.524 ng/ul

RT: 13.416 min Scan# 1739

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:154 Resp: 2525030


Ion Ratio Lower Upper

154 100

153 38.5 32.2 48.2

76 14.7 12.4 18.6





#44

2-Chloronaphthalene

Concen: 37.947 ng/ul

RT: 13.457 min Scan# 1746

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:162 Resp: 1982699

Ion Ratio Lower Upper

162 100

164 32.2 25.9 38.9

127 40.4 32.9 49.3



#45

2-Nitroaniline

Concen: 40.989 ng/ul

RT: 13.669 min Scan# 1782

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion: 65 Resp: 747848


Ion Ratio Lower Upper

65 100

92 63.8 50.7 76.1

138 93.3 71.4 107.2





#47

Dimethylphthalate

Concen: 38.700 ng/ul

RT: 14.039 min Scan# 1845

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:163 Resp: 2549360

Ion Ratio Lower Upper

163 100

194 4.1 3.3 4.9

164 10.1 8.2 12.2



#48

2,6-Dinitrotoluene

Concen: 41.681 ng/ul

RT: 14.163 min Scan# 1866

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:165 Resp: 535944


Ion Ratio Lower Upper

165 100

89 63.5 49.8 74.8

121 21.2 17.2 25.8









#50
Acenaphthylene
Concen: 38.353 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:152 Resp: 3160552
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.4
153 12.9 10.6 16.0



#51
3-Nitroaniline
Concen: 43.360 ng/ul
RT: 14.492 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:138 Resp: 585889
Ion Ratio Lower Upper
138 100
108 10.7 8.6 12.8
92 122.8 99.5 149.3









#52
Acenaphthene
Concen: 37.764 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:153 Resp: 2151413
Ion Ratio Lower Upper
153 100
152 48.7 38.8 58.2
154 90.7 70.3 105.5



#53
2,4-Dinitrophenol
Concen: 41.590 ng/ul
RT: 14.692 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:184 Resp: 343836
Ion Ratio Lower Upper
184 100
63 84.7 62.6 93.8
154 64.6 50.9 76.3









#55
4-Nitrophenol
Concen: 41.302 ng/ul
RT: 14.798 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:109 Resp: 451143
Ion Ratio Lower Upper
109 100
139 122.9 104.3 156.5
65 145.1 124.3 186.5



#56
Dibenzofuran
Concen: 37.831 ng/ul
RT: 14.981 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:168 Resp: 2970687
Ion Ratio Lower Upper
168 100
139 38.6 32.1 48.1









#57
2,4-Dinitrotoluene
Concen: 41.884 ng/ul
RT: 14.945 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:165 Resp: 779575
Ion Ratio Lower Upper
165 100
63 56.5 45.4 68.0
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.100 ng/ul
RT: 15.204 min Scan# 2043
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:232 Resp: 577888
Ion Ratio Lower Upper
232 100
131 49.3 40.6 61.0
130 2.7 2.2 3.2
166 32.3 27.0 40.4









#59
Diethylphthalate
Concen: 38.969 ng/ul
RT: 15.398 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:149 Resp: 2579563
Ion Ratio Lower Upper
149 100
177 20.9 16.6 24.8
150 12.9 10.1 15.1



#61
Fluorene
Concen: 38.375 ng/ul
RT: 15.628 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:166 Resp: 2423800
Ion Ratio Lower Upper
166 100
165 99.3 79.1 118.7
167 13.5 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 39.822 ng/ul
RT: 15.622 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:204 Resp: 1170705
Ion Ratio Lower Upper
204 100
206 33.1 26.0 39.0
141 63.3 51.8 77.8



#63
4-Nitroaniline
Concen: 46.671 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:138 Resp: 621477
Ion Ratio Lower Upper
138 100
92 52.8 43.3 64.9
108 80.0 64.2 96.4









#66
 4,6-Dinitro-2-methylphenol
 Concen: 40.843 ng/ul
 RT: 15.710 min Scan# 2129
 Delta R.T. -0.000 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56




Instrument :
 BNA_N
 ClientSampleId :
 SLCS129


Tgt Ion:198 Resp: 482818
 Ion Ratio Lower Upper
 198 100
 182 3.9 3.4 5.0
 77 28.5 23.7 35.5



#67
 N-Nitrosodiphenylamine
 Concen: 34.850 ng/ul
 RT: 15.833 min Scan# 2150
 Delta R.T. -0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Tgt Ion:169 Resp: 1960587
 Ion Ratio Lower Upper
 169 100
 168 69.1 54.3 81.5
 167 36.7 27.6 41.4









#68
4-Bromophenyl-phenylether
Concen: 38.250 ng/ul
RT: 16.516 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:248 Resp: 712935
Ion Ratio Lower Upper
248 100
250 98.5 77.1 115.7
141 81.0 62.3 93.5



#69
Hexachlorobenzene
Concen: 38.685 ng/ul
RT: 16.633 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:284 Resp: 816365
Ion Ratio Lower Upper
284 100
142 39.5 31.0 46.6
249 30.8 25.1 37.7





#71

Pentachlorophenol

Concen: 37.006 ng/ul

RT: 16.975 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:266 Resp: 485399

Ion Ratio Lower Upper

266 100

264 65.6 52.7 79.1

268 64.0 51.4 77.0



#72

Phenanthrene

Concen: 37.937 ng/ul

RT: 17.369 min Scan# 2411

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:178 Resp: 3969858


Ion Ratio Lower Upper

178 100

179 15.2 12.6 19.0

176 19.5 15.2 22.8





#74

Anthracene

Concen: 37.989 ng/uI

RT: 17.463 min Scan# 2428

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:178 Resp: 4005720

Ion Ratio Lower Upper

178 100

179 15.6 12.2 18.4

176 18.5 14.6 21.8



#75

1,2,3,4-Tetrachlorobenzene

Concen: 35.954 ng/uI

RT: 13.381 min Scan# 1733

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:216 Resp: 948286


Ion Ratio Lower Upper

216 100

214 85.9 62.5 93.7

218 50.9 37.7 56.5





#76

Pentachlorobenzene

Concen: 37.337 ng/uL

RT: 14.898 min Scan# 1991

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:250 Resp: 942153

Ion Ratio Lower Upper

250 100

248 58.6 50.8 76.2

252 64.7 51.8 77.6



#77

Carbazole

Concen: 37.859 ng/uL

RT: 17.733 min Scan# 2473

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:167 Resp: 3604550


Ion Ratio Lower Upper

167 100

166 21.6 16.6 24.8

139 13.5 10.2 15.4





#78

Di-n-butylphthalate

Concen: 40.310 ng/ul

RT: 18.292 min Scan# 2569

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:149 Resp: 4555296

Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 5.4 4.2 6.4



#80

Fluoranthene

Concen: 38.436 ng/ul

RT: 19.386 min Scan# 2754

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:202 Resp: 4745423


Ion Ratio Lower Upper

202 100

101 13.3 10.7 16.1

100 10.1 7.9 11.9





#82

Pyrene

Concen: 36.994 ng/ul

RT: 19.751 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:202 Resp: 4807535

Ion Ratio Lower Upper

202 100

101 15.8 13.0 19.6

100 13.1 11.0 16.6



#83

Butylbenzylphthalate

Concen: 40.671 ng/ul

RT: 20.645 min Scan# 2968

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:149 Resp: 2194346


Ion Ratio Lower Upper

149 100

91 79.4 62.6 93.8

206 20.2 16.2 24.2









#84
3,3'-Dichlorobenzidine
Concen: 20.793 ng/ul
RT: 21.439 min Scan# 3103
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Instrument :
BNA_N
ClientSampleId :
SLCS129


Tgt Ion:252 Resp: 846354
Ion Ratio Lower Upper
252 100
254 64.5 52.2 78.2
126 12.1 11.4 17.2



#85
Benzo(a)anthracene
Concen: 38.482 ng/ul
RT: 21.504 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:228 Resp: 4903089
Ion Ratio Lower Upper
228 100
229 20.7 16.2 24.2
226 26.9 20.7 31.1









#86
 Bis(2-ethylhexyl)phthalate
 Concen: 40.026 ng/ul
 RT: 21.427 min Scan# 3101
 Delta R.T. -0.000 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56




Instrument :
 BNA_N
 ClientSampleId :
 SLCS129


Tgt Ion:149 Resp: 3223787
 Ion Ratio Lower Upper
 149 100
 167 27.4 21.8 32.8
 279 4.1 2.9 4.3



#87
 Chrysene
 Concen: 38.759 ng/ul
 RT: 21.557 min Scan# 3123
 Delta R.T. -0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Tgt Ion:228 Resp: 4703723
 Ion Ratio Lower Upper
 228 100
 226 29.3 22.4 33.6
 229 19.5 15.6 23.4





#89

Di-n-octyl phthalate

Concen: 44.069 ng/ul

RT: 22.368 min Scan# 31

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56





Instrument :

BNA_N

ClientSampleId :

SLCS129

Tgt Ion:149 Resp: 5626680



#90

Benzo(b)fluoranthene

Concen: 41.366 ng/ul

RT: 23.221 min Scan# 3406

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:252 Resp: 5040198


Ion Ratio Lower Upper

252 100

253 22.0 19.2 28.8

125 12.1 10.4 15.6





#91

Benzo(k)fluoranthene

Concen: 42.098 ng/ul

RT: 23.274 min Scan# 3416

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:252 Resp: 5047263

Ion Ratio Lower Upper

252 100

253 22.8 16.7 25.1

125 11.7 9.3 13.9



#93

Benzo(a)pyrene

Concen: 40.453 ng/ul

RT: 23.868 min Scan# 3516

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:252 Resp: 4315375


Ion Ratio Lower Upper

252 100

253 20.9 17.4 26.0

125 12.6 10.8 16.2





#94

Indeno(1,2,3-cd)pyrene

Concen: 38.710 ng/ul

RT: 26.527 min Scan# 3968

Delta R.T. -0.012 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129





Tgt Ion:276 Resp: 5003332

Ion Ratio Lower Upper

276 100

138 27.9 21.9 32.9

277 26.5 19.8 29.6



#95

Dibenzo(a,h)anthracene

Concen: 39.028 ng/ul

RT: 26.539 min Scan# 3970

Delta R.T. -0.018 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:278 Resp: 4164590


Ion Ratio Lower Upper

278 100

139 18.9 14.6 21.8

279 23.4 18.2 27.4





#96

Benzo(g,h,i)perylene

Concen: 37.194 ng/ul

RT: 27.321 min Scan# 41

Delta R.T. -0.006 min

Lab File: BN024945.D


Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129



Tgt Ion:276 Resp: 3900967

Ion Ratio Lower Upper

276 100

138 23.0 19.8 29.8

277 21.6 17.6 26.4

