

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024918.D
 Acq On : 15 Apr 2023 04:18
 Operator : CG/JU
 Sample : PB152087BS
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS087

Quant Time: Apr 15 05:22:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 15 00:01:14 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	296308	20.000	ng/u1	0.00
20) Naphthalene-d8	10.752	136	1307558	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	817825	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	1752369	20.000	ng/u1	0.00
79) Chrysene-d12	21.516	240	1700919	20.000	ng/u1	0.00
88) Perylene-d12	23.963	264	1836959	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	41377	5.315	ng/uL	0.00
4) Pyridine-d5	3.734	84	567585	24.347	ng/u1	0.00
7) Phenol-d5	7.099	99	801807	27.755	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.264	67	534667	27.671	ng/u1	0.00
11) 2-Chlorophenol-d4	7.464	132	601051	28.312	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	638881	27.858	ng/u1	0.00
21) Nitrobenzene-d5	9.105	128	298807	28.224	ng/u1	0.00
24) 2-Nitrophenol-d4	9.834	143	326715	28.491	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	603844	29.460	ng/u1	0.00
31) 4-Chloroaniline-d4	10.893	131	676708	21.755	ng/u1	0.00
46) Dimethylphthalate-d6	13.993	166	1827669	28.828	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	2083686	28.681	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	384214	29.564	ng/u1	0.00
60) Fluorene-d10	15.575	176	1512175	28.250	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	322123	28.246	ng/u1	0.00
73) Anthracene-d10	17.428	188	2458957	29.877	ng/u1	0.00
81) Pyrene-d10	19.722	212	2897626	29.337	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.816	264	2958250	30.800	ng/u1	0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.340	88	91627	10.906	ng/uL	98
5) Pyridine	3.752	79	743793	30.881	ng/u1	97
6) Benzaldehyde	7.069	77	256371	18.163	ng/u1	95
8) Phenol	7.128	94	868822	28.878	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.358	93	656780	27.807	ng/u1	99
12) 2-Chlorophenol	7.499	128	629800	28.404	ng/u1	97
13) 2-Methylphenol	8.387	108	621872	28.154	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.475	45	1291075	27.838	ng/u1	100
16) Acetophenone	8.769	105	1015635	28.008	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.752	70	547089	27.970	ng/u1	100
18) 4-Methylphenol	8.716	108	696656	28.362	ng/u1	95
19) Hexachloroethane	9.022	117	251129	28.063	ng/u1	96
22) Nitrobenzene	9.152	77	811915	28.540	ng/u1	100
23) Isophorone	9.675	82	1514063	28.994	ng/u1	100
25) 2-Nitrophenol	9.863	139	365112	29.516	ng/u1	97
26) 2,4-Dimethylphenol	9.922	107	725789	27.867	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.163	93	922695	28.392	ng/u1	99
29) 2,4-Dichlorophenol	10.399	162	612454	29.772	ng/u1	98
30) Naphthalene	10.799	128	2065817	28.289	ng/u1	99
32) 4-Chloroaniline	10.916	127	557001	17.689	ng/u1	99
33) Hexachlorobutadiene	11.087	225	340671	28.412	ng/u1	95
34) Caprolactam	11.693	113	208167m	28.871	ng/u1	
35) 4-Chloro-3-methylphenol	12.040	107	733766	30.422	ng/u1	99
36) 2-Methylnaphthalene	12.410	142	1369382	28.374	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024918.D
 Acq On : 15 Apr 2023 04:18
 Operator : CG/JU
 Sample : PB152087BS
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS087


Quant Time: Apr 15 05:22:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 15 00:01:14 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian
 Giraldo


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
37) 1-Methylnaphthalene	12.628	142	1434728	29.780	ng/ul	98	04/17/2023
39) 1,2,4,5-Tetrachloroben...	12.775	216	687189	28.807	ng/ul	99	Supervised By :Jagrut Upadhyay
40) Hexachlorocyclopentadiene	12.752	237	366964	25.227	ng/ul	96	
41) 2,4,6-Trichlorophenol	13.016	196	478061	29.330	ng/ul	98	
42) 2,4,5-Trichlorophenol	13.087	196	538003	30.163	ng/ul	98	
43) 1,1'-Biphenyl	13.416	154	1864542	28.615	ng/ul	98	
44) 2-Chloronaphthalene	13.463	162	1443855	28.537	ng/ul	98	04/17/2023
45) 2-Nitroaniline	13.669	65	538799	30.497	ng/ul	99	
47) Dimethylphthalate	14.040	163	1871589	29.340	ng/ul	100	
48) 2,6-Dinitrotoluene	14.163	165	379249	30.459	ng/ul	95	
50) Acenaphthylene	14.304	152	2305347	28.890	ng/ul	100	
51) 3-Nitroaniline	14.493	138	374160	28.595	ng/ul	97	
52) Acenaphthene	14.646	153	1620378	29.372	ng/ul	99	
53) 2,4-Dinitrophenol	14.698	184	229648	28.686	ng/ul	97	
55) 4-Nitrophenol	14.799	109	323985	30.630	ng/ul	96	
56) Dibenzofuran	14.981	168	2224375	29.253	ng/ul	94	
57) 2,4-Dinitrotoluene	14.946	165	558369	30.980	ng/ul	95	
58) 2,3,4,6-Tetrachlorophenol	15.210	232	435583	29.657	ng/ul	93	
59) Diethylphthalate	15.404	149	1891928	29.516	ng/ul	99	
61) Fluorene	15.628	166	1806677	29.540	ng/ul	97	
62) 4-Chlorophenyl-phenyle...	15.622	204	837733	29.427	ng/ul	95	
63) 4-Nitroaniline	15.657	138	421001	32.650	ng/ul	97	
66) 4,6-Dinitro-2-methylph...	15.710	198	320408	28.681	ng/ul	98	
67) N-Nitrosodiphenylamine	15.840	169	1460712	27.476	ng/ul	98	
68) 4-Bromophenyl-phenylether	16.516	248	529258	30.047	ng/ul	96	
69) Hexachlorobenzene	16.634	284	592933	29.732	ng/ul	98	
70) Atrazine	16.787	200	25489	1.409	ng/ul	100	
71) Pentachlorophenol	16.981	266	351669	28.370	ng/ul	96	
72) Phenanthrene	17.375	178	2965854	29.992	ng/ul	99	
74) Anthracene	17.463	178	2955928	29.664	ng/ul	99	
75) 1,2,3,4-Tetrachloroben...	13.387	216	714921	28.683	ng/ul	99	
76) Pentachlorobenzene	14.904	250	704680	29.551	ng/ul	99	
77) Carbazole	17.734	167	2744113	30.499	ng/ul	99	
78) Di-n-butylphthalate	18.298	149	3268831	30.609	ng/ul	99	
80) Fluoranthene	19.386	202	3441487	29.073	ng/ul	99	
82) Pyrene	19.751	202	3680663	29.540	ng/ul	98	
83) Butylbenzylphthalate	20.645	149	1564488	30.243	ng/ul	100	
84) 3,3'-Dichlorobenzidine	21.433	252	812283	20.814	ng/ul	95	
85) Benzo(a)anthracene	21.498	228	3656417	29.931	ng/ul	99	
86) Bis(2-ethylhexyl)phtha...	21.422	149	2293160	29.695	ng/ul	99	
87) Chrysene	21.557	228	3458117	29.720	ng/ul	96	
89) Di-n-octyl phthalate	22.369	149	4012387	31.784	ng/ul	100	
90) Benzo(b)fluoranthene	23.216	252	3790534	31.465	ng/ul	97	
91) Benzo(k)fluoranthene	23.269	252	3799800	32.055	ng/ul	99	
93) Benzo(a)pyrene	23.863	252	3343954	31.705	ng/ul	99	
94) Indeno(1,2,3-cd)pyrene	26.539	276	3888276m	30.427	ng/ul		
95) Dibenzo(a,h)anthracene	26.545	278	3289574m	31.180	ng/ul		
96) Benzo(g,h,i)perylene	27.315	276	3141472m	30.295	ng/ul		

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.906 ng/uL
RT: 3.340 min Scan# 20
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion: 88 Resp: 9162
Ion Ratio Lower Upper
88 100
43 42.1 33.0 49.4
58 88.6 72.3 108.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#5
Pyridine
Concen: 30.881 ng/uL
RT: 3.752 min Scan# 96
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion: 79 Resp: 743793
Ion Ratio Lower Upper
79 100
52 98.6 76.1 114.1
51 45.6 35.8 53.6





#6
Benzaldehyde
Concen: 18.163 ng/ul
RT: 7.069 min Scan# 661
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion: 77 Resp: 25637
Ion Ratio Lower Upper
77 100
105 89.4 74.3 111.5
106 85.6 73.3 109.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#8
Phenol
Concen: 28.878 ng/ul
RT: 7.128 min Scan# 670
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion: 94 Resp: 868822
Ion Ratio Lower Upper
94 100
65 31.6 25.1 37.7
66 40.7 31.9 47.9





#10
Bis(2-Chloroethyl)ether
Concen: 27.807 ng/ul
RT: 7.358 min Scan# 709
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion: 93 Resp: 656780
Ion Ratio Lower Upper
93 100
63 89.9 70.9 106.3
95 32.2 26.0 39.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#12
2-Chlorophenol
Concen: 28.404 ng/ul
RT: 7.499 min Scan# 733
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Tgt Ion:128 Resp: 629800
Ion Ratio Lower Upper
128 100
64 58.3 44.8 67.2
130 32.2 26.6 39.8





#13

2-Methylphenol

Concen: 28.154 ng/ul

RT: 8.387 min Scan# 884

Delta R.T. 0.000 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18

Instrument :

BNA_N

ClientSampleId :

SLCS087

Tgt Ion:108 Resp: 621872

Ion Ratio Lower Upper



108 100

107 88.8 71.5 107.3

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.838 ng/ul

RT: 8.475 min Scan# 899

Delta R.T. 0.000 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18




Tgt Ion: 45 Resp: 1291075


Ion Ratio Lower Upper

45 100

77 10.5 8.4 12.6


79 8.0 6.6 9.8





#16
Acetophenone
Concen: 28.008 ng/ul
RT: 8.769 min Scan# 949
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:105 Resp: 101563
Ion Ratio Lower Upper
105 100
77 82.9 64.8 97.2
51 38.2 28.7 43.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#17
N-Nitroso-di-n-propylamine
Concen: 27.970 ng/ul
RT: 8.752 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion: 70 Resp: 547089
Ion Ratio Lower Upper
70 100
42 79.9 64.2 96.2
101 8.9 7.3 10.9
130 18.5 15.0 22.6





#18
4-Methylphenol
Concen: 28.362 ng/ul
RT: 8.716 min Scan# 940
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:108 Resp: 696650
Ion Ratio Lower Upper
108 100
107 112.6 85.8 128.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#19
Hexachloroethane
Concen: 28.063 ng/ul
RT: 9.022 min Scan# 992
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Tgt Ion:117 Resp: 251129
Ion Ratio Lower Upper
117 100
201 94.9 72.2 108.4
199 56.6 46.2 69.4






#22
Nitrobenzene
Concen: 28.540 ng/ul
RT: 9.152 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Instrument :

BNA_N

ClientSampleId :


SLCS087



Tgt Ion: 77 Resp: 811915
Ion Ratio Lower Upper
77 100
123 41.9 33.5 50.3
65 14.4 11.8 17.8



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#23
Isophorone
Concen: 28.994 ng/ul
RT: 9.675 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Tgt Ion: 82 Resp: 1514063
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 15.9 12.8 19.2





#25

2-Nitrophenol

Concen: 29.516 ng/ul

RT: 9.863 min Scan# 1135

Delta R.T. 0.000 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18

Instrument :

BN_A_N



ClientSampleId :

SLCS087

Tgt Ion:139 Resp: 365112
 Ion Ratio Lower Upper
 139 100
 65 60.5 49.8 74.8
 109 30.6 26.2 39.2



Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo



04/17/2023
 Supervised By :Jagrut
 Upadhyay

04/17/2023



#26

2,4-Dimethylphenol

Concen: 27.867 ng/ul




RT: 9.922 min Scan# 1145


Delta R.T. 0.000 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18


Tgt Ion:107 Resp: 725789
 Ion Ratio Lower Upper
 107 100
 121 48.7 39.7 59.5
 122 83.8 66.1 99.1





#27
 Bis(2-Chloroethoxy)methane
 Concen: 28.392 ng/ul
 RT: 10.163 min Scan# 1186
 Delta R.T. 0.000 min
 Lab File: BN024918.D
 Acq: 15 Apr 2023 04:18



Instrument :
 BNA_N
 ClientSampleId :
 SLCS087



Tgt Ion: 93 Resp: 92269
 Ion Ratio Lower Upper
 93 100
 95 32.2 26.1 39.1
 123 10.7 8.5 12.7


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




04/17/2023
 Supervised By :Jagrut
 Upadhyay



04/17/2023




#29
 2,4-Dichlorophenol
 Concen: 29.772 ng/ul
 RT: 10.399 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BN024918.D
 Acq: 15 Apr 2023 04:18




Tgt Ion:162 Resp: 612454
 Ion Ratio Lower Upper
 162 100
 164 65.6 50.9 76.3
 98 37.9 31.0 46.4





#30
Naphthalene
Concen: 28.289 ng/ul
RT: 10.799 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:128 Resp: 206581
Ion Ratio Lower Upper
128 100
129 11.2 8.4 12.6
127 13.3 10.7 16.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#32
4-Chloroaniline
Concen: 17.689 ng/ul
RT: 10.916 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:127 Resp: 557001
Ion Ratio Lower Upper
127 100
129 32.3 25.5 38.3





#33
Hexachlorobutadiene
Concen: 28.412 ng/ul
RT: 11.087 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:225 Resp: 34067
Ion Ratio Lower Upper
225 100
223 62.5 45.9 68.9
227 66.1 51.2 76.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#34
Caprolactam
Concen: 28.871 ng/ul m
RT: 11.693 min Scan# 1446
Delta R.T. 0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion:113 Resp: 208167
Ion Ratio Lower Upper
113 100
55 224.7 188.3 282.5
56 166.0 133.5 200.3





#35
4-Chloro-3-methylphenol
Concen: 30.422 ng/ul
RT: 12.040 min Scan# 1505
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:107 Resp: 733760
Ion Ratio Lower Upper
107 100
144 24.7 19.6 29.4
142 74.9 60.7 91.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#36
2-Methylnaphthalene
Concen: 28.374 ng/ul
RT: 12.410 min Scan# 1568
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:142 Resp: 1369382
Ion Ratio Lower Upper
142 100
141 91.6 71.4 107.2





#37
1-Methylnaphthalene
Concen: 29.780 ng/ul
RT: 12.628 min Scan# 1605
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:142 Resp: 1434728
Ion Ratio Lower Upper
142 100
141 94.2 73.5 110.3
116 4.2 3.1 4.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#39
1,2,4,5-Tetrachlorobenzene
Concen: 28.807 ng/ul
RT: 12.775 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:216 Resp: 687189
Ion Ratio Lower Upper
216 100
214 78.0 62.2 93.4
179 22.3 16.8 25.2
108 19.6 15.0 22.6





#40
Hexachlorocyclopentadiene
Concen: 25.227 ng/ul
RT: 12.752 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:237 Resp: 366964
Ion Ratio Lower Upper
237 100
235 66.2 49.9 74.9
272 13.5 10.2 15.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#41
2,4,6-Trichlorophenol
Concen: 29.330 ng/ul
RT: 13.016 min Scan# 1671
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Tgt Ion:196 Resp: 478061
Ion Ratio Lower Upper
196 100
198 90.7 74.5 111.7
200 30.6 24.3 36.5






#42
2,4,5-Trichlorophenol
Concen: 30.163 ng/ul
RT: 13.087 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Instrument :

BNA_N

ClientSampleId :



SLCS087



Tgt Ion:196 Resp: 538000
Ion Ratio Lower Upper
196 100
198 95.0 78.2 117.2
200 30.3 24.0 36.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#43
1,1'-Biphenyl
Concen: 28.615 ng/ul
RT: 13.416 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion:154 Resp: 1864542
Ion Ratio Lower Upper
154 100
153 41.4 32.2 48.2
76 15.9 12.4 18.6





#44
2-Chloronaphthalene
Concen: 28.537 ng/ul
RT: 13.463 min Scan# 1747
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:162 Resp: 144385
Ion Ratio Lower Upper
162 100
164 31.9 25.9 38.9
127 39.2 32.9 49.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#45
2-Nitroaniline
Concen: 30.497 ng/ul
RT: 13.669 min Scan# 1782
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion: 65 Resp: 538799
Ion Ratio Lower Upper
65 100
92 63.1 50.7 76.1
138 90.0 71.4 107.2





#47
Dimethylphthalate
Concen: 29.340 ng/ul
RT: 14.040 min Scan# 1845
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:163 Resp: 1871589
Ion Ratio Lower Upper
163 100
194 3.8 3.3 4.9
164 10.2 8.2 12.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#48
2,6-Dinitrotoluene
Concen: 30.459 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion:165 Resp: 379249
Ion Ratio Lower Upper
165 100
89 66.3 49.8 74.8
121 22.7 17.2 25.8





#50
Acenaphthylene
Concen: 28.890 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:152 Resp: 230534
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4
153 13.1 10.6 16.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#51
3-Nitroaniline
Concen: 28.595 ng/ul
RT: 14.493 min Scan# 1922
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Tgt Ion:138 Resp: 374160
Ion Ratio Lower Upper
138 100
108 10.9 8.6 12.8
92 127.7 99.5 149.3






#52
Acenaphthene
Concen: 29.372 ng/ul
RT: 14.646 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Instrument :

BNA_N

ClientSampleId :

SLCS087





Tgt Ion:153 Resp: 1620378

Ion	Ratio	Lower	Upper
153	100		
152	48.4	38.8	58.2
154	88.6	70.3	105.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay




04/17/2023




#53
2,4-Dinitrophenol
Concen: 28.686 ng/ul
RT: 14.698 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Tgt Ion:184 Resp: 229648


Ion	Ratio	Lower	Upper
184	100		
63	83.0	62.6	93.8
154	63.4	50.9	76.3





#55
4-Nitrophenol
Concen: 30.630 ng/u1
RT: 14.799 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:109 Resp: 323985
Ion Ratio Lower Upper
109 100
139 124.3 104.3 156.5
65 151.5 124.3 186.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#56
Dibenzofuran
Concen: 29.253 ng/u1
RT: 14.981 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:168 Resp: 2224375
Ion Ratio Lower Upper
168 100
139 36.2 32.1 48.1





#57
2,4-Dinitrotoluene
Concen: 30.980 ng/ul
RT: 14.946 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:165 Resp: 558369
Ion Ratio Lower Upper
165 100
63 60.6 45.4 68.0
182 2.6 2.2 3.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#58
2,3,4,6-Tetrachlorophenol
Concen: 29.657 ng/ul
RT: 15.210 min Scan# 2044
Delta R.T. 0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion:232 Resp: 435583
Ion Ratio Lower Upper
232 100
131 46.5 40.6 61.0
130 2.2 2.2 3.2
166 29.4 27.0 40.4





#59
Diethylphthalate
Concen: 29.516 ng/ul
RT: 15.404 min Scan# 2077
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:149 Resp: 1891928
Ion Ratio Lower Upper
149 100
177 21.4 16.6 24.8
150 12.2 10.1 15.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#61
Fluorene
Concen: 29.540 ng/ul
RT: 15.628 min Scan# 2115
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18




Tgt Ion:166 Resp: 1806677
Ion Ratio Lower Upper
166 100
165 96.0 79.1 118.7
167 13.7 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 29.427 ng/ul
RT: 15.622 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:204 Resp: 83773
Ion Ratio Lower Upper
204 100
206 35.2 26.0 39.0
141 68.8 51.8 77.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#63
4-Nitroaniline
Concen: 32.650 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:138 Resp: 421001
Ion Ratio Lower Upper
138 100
92 53.7 43.3 64.9
108 84.2 64.2 96.4





#66
4,6-Dinitro-2-methylphenol
Concen: 28.681 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:198 Resp: 320408
Ion Ratio Lower Upper
198 100
182 4.3 3.4 5.0
77 30.5 23.7 35.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#67
N-Nitrosodiphenylamine
Concen: 27.476 ng/ul
RT: 15.840 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:169 Resp: 1460712
Ion Ratio Lower Upper
169 100
168 67.3 54.3 81.5
167 37.1 27.6 41.4





#68
4-Bromophenyl-phenylether
Concen: 30.047 ng/ul
RT: 16.516 min Scan# 2266
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:248 Resp: 529258
Ion Ratio Lower Upper
248 100
250 90.5 77.1 115.7
141 78.5 62.3 93.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#69
Hexachlorobenzene
Concen: 29.732 ng/ul
RT: 16.634 min Scan# 2286
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:284 Resp: 592933
Ion Ratio Lower Upper
284 100
142 38.3 31.0 46.6
249 32.8 25.1 37.7





#70
 Atrazine
 Concen: 1.409 ng/ul
 RT: 16.787 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN024918.D
 Acq: 15 Apr 2023 04:18



Instrument :
 BNA_N
 ClientSampleId :
 SLCS087



Tgt Ion:200 Resp: 25489
 Ion Ratio Lower Upper
 200 100
 173 27.1 21.4 32.2
 215 47.7 38.2 57.2


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo






04/17/2023
 Supervised By :Jagrut
 Upadhyay


04/17/2023



#71
 Pentachlorophenol
 Concen: 28.370 ng/ul
 RT: 16.981 min Scan# 2345
 Delta R.T. 0.000 min
 Lab File: BN024918.D
 Acq: 15 Apr 2023 04:18


Tgt Ion:266 Resp: 351669
 Ion Ratio Lower Upper
 266 100
 264 62.4 52.7 79.1
 268 62.0 51.4 77.0





#72
 Phenanthrene
 Concen: 29.992 ng/u1
 RT: 17.375 min Scan# 2412
 Delta R.T. 0.000 min
 Lab File: BN024918.D
 Acq: 15 Apr 2023 04:18



Instrument : BNA_N
 ClientSampleId : SLCS087



Tgt Ion:178 Resp: 2965854


Ion	Ratio	Lower	Upper
178	100		
179	15.1	12.6	19.0
176	19.0	15.2	22.8

Manual Integrations
APPROVED
 Reviewed By :Christian Giraldo



04/17/2023
 Supervised By :Jagrut Upadhyay




04/17/2023




#74
 Anthracene
 Concen: 29.664 ng/u1
 RT: 17.463 min Scan# 2427
 Delta R.T. 0.000 min
 Lab File: BN024918.D
 Acq: 15 Apr 2023 04:18

Tgt Ion:178 Resp: 2955928


Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.2	18.4
176	19.2	14.6	21.8





#75
1,2,3,4-Tetrachlorobenzene
Concen: 28.683 ng/uL
RT: 13.387 min Scan# 1734
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:216 Resp: 71492
Ion Ratio Lower Upper
216 100
214 76.8 62.5 93.7
218 46.9 37.7 56.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#76
Pentachlorobenzene
Concen: 29.551 ng/uL
RT: 14.904 min Scan# 1992
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Tgt Ion:250 Resp: 704680
Ion Ratio Lower Upper
250 100
248 62.9 50.8 76.2
252 65.5 51.8 77.6





#77

Carbazole

Concen: 30.499 ng/ul

RT: 17.734 min Scan# 2474

Delta R.T. -0.006 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18

Instrument :

BNA_N

ClientSampleId :

SLCS087

Tgt Ion:167 Resp: 274411

Ion Ratio Lower Upper

167 100



166 21.4 16.6 24.8

139 12.9 10.2 15.4

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#78

Di-n-butylphthalate

Concen: 30.609 ng/ul

RT: 18.298 min Scan# 2569

Delta R.T. 0.000 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18




Tgt Ion:149 Resp: 3268831


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.2


104 5.3 4.2 6.4





#80
Fluoranthene
Concen: 29.073 ng/u1
RT: 19.386 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:202 Resp: 344148
Ion Ratio Lower Upper
202 100
101 13.1 10.7 16.1
100 10.5 7.9 11.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#82
Pyrene
Concen: 29.540 ng/u1
RT: 19.751 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:202 Resp: 3680663
Ion Ratio Lower Upper
202 100
101 15.6 13.0 19.6
100 13.0 11.0 16.6





#83
Butylbenzylphthalate
Concen: 30.243 ng/ul
RT: 20.645 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:149 Resp: 1564488
Ion Ratio Lower Upper
149 100
91 78.1 62.6 93.8
206 20.0 16.2 24.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#84
3,3'-Dichlorobenzidine
Concen: 20.814 ng/ul
RT: 21.433 min Scan# 3102
Delta R.T. -0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:252 Resp: 812283
Ion Ratio Lower Upper
252 100
254 69.7 52.2 78.2
126 13.9 11.4 17.2





#85
Benzo(a)anthracene
Concen: 29.931 ng/ul
RT: 21.498 min Scan# 3113
Delta R.T. -0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:228 Resp: 365641
Ion Ratio Lower Upper
228 100
229 19.6 16.2 24.2
226 26.0 20.7 31.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#86
Bis(2-ethylhexyl)phthalate
Concen: 29.695 ng/ul
RT: 21.422 min Scan# 3100
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:149 Resp: 2293160
Ion Ratio Lower Upper
149 100
167 27.6 21.8 32.8
279 3.7 2.9 4.3





#87
Chrysene
Concen: 29.720 ng/ul
RT: 21.557 min Scan# 3123
Delta R.T. 0.000 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:228 Resp: 345811
Ion Ratio Lower Upper
228 100
226 31.6 22.4 33.6
229 19.8 15.6 23.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#89
Di-n-octyl phthalate
Concen: 31.784 ng/ul
RT: 22.369 min Scan# 3261
Delta R.T. 0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:149 Resp: 4012387





#90
Benzo(b)fluoranthene
Concen: 31.465 ng/ul
RT: 23.216 min Scan# 3404
Delta R.T. 0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18



Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:252 Resp: 3790534
Ion Ratio Lower Upper
252 100
253 22.3 19.2 28.8
125 11.9 10.4 15.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#91
Benzo(k)fluoranthene
Concen: 32.055 ng/ul
RT: 23.269 min Scan# 3414
Delta R.T. 0.006 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Tgt Ion:252 Resp: 3799800
Ion Ratio Lower Upper
252 100
253 20.6 16.7 25.1
125 11.1 9.3 13.9





#93
Benzo(a)pyrene
Concen: 31.705 ng/ul
RT: 23.863 min Scan# 31
Delta R.T. 0.018 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18


Instrument :
BNA_N
ClientSampleId :
SLCS087



Tgt Ion:252 Resp: 3343954
Ion Ratio Lower Upper
252 100
253 21.2 17.4 26.0
125 12.7 10.8 16.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.427 ng/ul m
RT: 26.539 min Scan# 3970
Delta R.T. 0.047 min
Lab File: BN024918.D
Acq: 15 Apr 2023 04:18

Tgt Ion:276 Resp: 3888276
Ion Ratio Lower Upper
276 100
138 28.3 21.9 32.9
277 21.8 19.8 29.6





#95

Dibenzo(a,h)anthracene

Concen: 31.180 ng/ul m

RT: 26.545 min Scan# 3965

Delta R.T. 0.035 min

Lab File: BN024918.D


Acq: 15 Apr 2023 04:18

Instrument :

BNA_N

ClientSampleId :

SLCS087



Tgt Ion:278 Resp: 3289574

Ion Ratio Lower Upper

278 100



139 18.0 14.6 21.8

279 23.4 18.2 27.4

Manual Integrations


APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#96

Benzo(g,h,i)perylene

Concen: 30.295 ng/ul m

RT: 27.315 min Scan# 4102

Delta R.T. 0.041 min

Lab File: BN024918.D

Acq: 15 Apr 2023 04:18

Tgt Ion:276 Resp: 3141472

Ion Ratio Lower Upper

276 100

138 24.3 19.8 29.8

277 23.8 17.6 26.4

