

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024945.D
 Acq On : 15 Apr 2023 21:56
 Operator : CG/JU
 Sample : PB152129BS
 Misc :
 ALS Vial : 94 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Quant Time: Apr 17 04:13:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 02:58:51 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : Christian
 Giraldo

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	309575	20.000	ng/u1	0.00
20) Naphthalene-d8	10.746	136	1360659	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.581	164	844560	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.328	188	1854325	20.000	ng/u1	0.00
79) Chrysene-d12	21.521	240	1773707	20.000	ng/u1	0.00
88) Perylene-d12	23.974	264	1855370	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.305	96	51798	6.368	ng/uL	0.00
4) Pyridine-d5	3.728	84	959568	39.398	ng/u1	0.00
7) Phenol-d5	7.093	99	1086642	36.003	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.263	67	698800	34.616	ng/u1	0.00
11) 2-Chlorophenol-d4	7.463	132	814394	36.717	ng/u1	0.00
15) 4-Methylphenol-d8	8.652	113	857024	35.769	ng/u1	0.00
21) Nitrobenzene-d5	9.104	128	407894	37.024	ng/u1	0.00
24) 2-Nitrophenol-d4	9.828	143	452890	37.952	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.369	165	807323	37.850	ng/u1	0.00
31) 4-Chloroaniline-d4	10.887	131	535318	16.538	ng/u1	0.00
46) Dimethylphthalate-d6	13.992	166	2384836	36.425	ng/u1	0.00
49) Acenaphthylene-d8	14.275	160	2803923	37.372	ng/u1	0.00
54) 4-Nitrophenol-d4	14.781	143	513958	38.295	ng/u1	0.00
60) Fluorene-d10	15.575	176	2029697	36.718	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	460085	38.125	ng/u1	0.00
73) Anthracene-d10	17.428	188	3196587	36.704	ng/u1	0.00
81) Pyrene-d10	19.721	212	3744833	36.359	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	3739663	38.549	ng/u1	0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.340	88	120516	13.730	ng/uL	94
5) Pyridine	3.746	79	1123983	44.666	ng/u1	96
6) Benzaldehyde	7.069	77	333861	22.640	ng/u1	98
8) Phenol	7.122	94	1189439	37.841	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.357	93	903876	36.628	ng/u1	99
12) 2-Chlorophenol	7.493	128	883020	38.117	ng/u1	99
13) 2-Methylphenol	8.381	108	876958	38.000	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.469	45	1730693	35.717	ng/u1	100
16) Acetophenone	8.763	105	1374944	36.291	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.752	70	745730	36.492	ng/u1	99
18) 4-Methylphenol	8.716	108	960781	37.439	ng/u1	97
19) Hexachloroethane	9.016	117	345887	36.996	ng/u1	96
22) Nitrobenzene	9.146	77	1119462	37.815	ng/u1	100
23) Isophorone	9.675	82	2048698	37.701	ng/u1	100
25) 2-Nitrophenol	9.857	139	501610	38.968	ng/u1	99
26) 2,4-Dimethylphenol	9.922	107	1021597	37.694	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.157	93	1228608	36.330	ng/u1	99
29) 2,4-Dichlorophenol	10.393	162	847857	39.607	ng/u1	97
30) Naphthalene	10.798	128	2838372	37.351	ng/u1	99
32) 4-Chloroaniline	10.910	127	651771	19.891	ng/u1	99
33) Hexachlorobutadiene	11.081	225	480504	38.511	ng/u1	95
34) Caprolactam	11.693	113	293623m	39.134	ng/u1	
35) 4-Chloro-3-methylphenol	12.034	107	997893	39.758	ng/u1	99
36) 2-Methylnaphthalene	12.404	142	1900022	37.832	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024945.D
 Acq On : 15 Apr 2023 21:56
 Operator : CG/JU
 Sample : PB152129BS
 Misc :
 ALS Vial : 94 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Quant Time: Apr 17 04:13:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 02:58:51 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian
 Giraldo

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.628	142	1937354	38.643	ng/u1	97
39) 1,2,4,5-Tetrachloroben...	12.775	216	961102	39.014	ng/u1	99
40) Hexachlorocyclopentadiene	12.751	237	546883	36.405	ng/u1	94
41) 2,4,6-Trichlorophenol	13.016	196	631797	37.535	ng/u1	97
42) 2,4,5-Trichlorophenol	13.087	196	723197	39.262	ng/u1	98
43) 1,1'-Biphenyl	13.416	154	2525030	37.524	ng/u1	97
44) 2-Chloronaphthalene	13.457	162	1982699	37.947	ng/u1	99
45) 2-Nitroaniline	13.669	65	747477	40.969	ng/u1	97
47) Dimethylphthalate	14.039	163	2549360	38.700	ng/u1	100
48) 2,6-Dinitrotoluene	14.163	165	535944	41.681	ng/u1	99
50) Acenaphthylene	14.304	152	3160552	38.353	ng/u1	99
51) 3-Nitroaniline	14.492	138	585889	43.360	ng/u1	99
52) Acenaphthene	14.645	153	2151413	37.764	ng/u1	98
53) 2,4-Dinitrophenol	14.692	184	343836	41.590	ng/u1	95
55) 4-Nitrophenol	14.798	109	451143	41.302	ng/u1	93
56) Dibenzofuran	14.981	168	2970687	37.831	ng/u1	97
57) 2,4-Dinitrotoluene	14.945	165	779844	41.898	ng/u1	100
58) 2,3,4,6-Tetrachlorophenol	15.204	232	577888	38.100	ng/u1	98
59) Diethylphthalate	15.398	149	2579563	38.969	ng/u1	100
61) Fluorene	15.628	166	2423800	38.375	ng/u1	100
62) 4-Chlorophenyl-phenyle...	15.622	204	1170705	39.822	ng/u1	98
63) 4-Nitroaniline	15.657	138	620970	46.633	ng/u1	99
66) 4,6-Dinitro-2-methylph...	15.710	198	482818	40.843	ng/u1	98
67) N-Nitrosodiphenylamine	15.833	169	1960587	34.850	ng/u1	98
68) 4-Bromophenyl-phenylether	16.516	248	712935	38.250	ng/u1	97
69) Hexachlorobenzene	16.633	284	816365	38.685	ng/u1	99
71) Pentachlorophenol	16.975	266	485185	36.989	ng/u1	100
72) Phenanthrene	17.369	178	3969858	37.937	ng/u1	99
74) Anthracene	17.463	178	4005351	37.986	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.381	216	948286	35.954	ng/uL	92
76) Pentachlorobenzene	14.898	250	942153	37.337	ng/uL	97
77) Carbazole	17.733	167	3604550	37.859	ng/u1	98
78) Di-n-butylphthalate	18.292	149	4555296	40.310	ng/u1	100
80) Fluoranthene	19.386	202	4745423	38.443	ng/u1	100
82) Pyrene	19.751	202	4807535	37.000	ng/u1	99
83) Butylbenzylphthalate	20.645	149	2194346	40.678	ng/u1	99
84) 3,3'-Dichlorobenzidine	21.439	252	846354	20.797	ng/u1	98
85) Benzo(a)anthracene	21.504	228	4903089	38.489	ng/u1	98
86) Bis(2-ethylhexyl)phtha...	21.427	149	3223787	40.033	ng/u1	100
87) Chrysene	21.557	228	4707028	38.793	ng/u1	99
89) Di-n-octyl phthalate	22.368	149	5626680	44.129	ng/u1	100
90) Benzo(b)fluoranthene	23.221	252	5040198	41.423	ng/u1	97
91) Benzo(k)fluoranthene	23.274	252	5047263	42.156	ng/u1	97
93) Benzo(a)pyrene	23.868	252	4323891	40.589	ng/u1	98
94) Indeno(1,2,3-cd)pyrene	26.527	276	5006273	38.786	ng/u1	98
95) Dibenzo(a,h)anthracene	26.539	278	4168365	39.117	ng/u1	99
96) Benzo(g,h,i)perylene	27.321	276	3920326	37.430	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041423\
 Data File : BN024945.D
 Acq On : 15 Apr 2023 21:56
 Operator : CG/JU
 Sample : PB152129BS
 Misc :
 ALS Vial : 94 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Quant Time: Apr 17 04:13:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Apr 17 02:58:51 2023
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Christian
 Giraldo





#2
1,4-Dioxane
Concen: 13.730 ng/uL
RT: 3.340 min Scan# 20
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion: 88 Resp: 120510
Ion Ratio Lower Upper
88 100
43 39.0 33.0 49.4
58 83.4 72.3 108.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#5
Pyridine
Concen: 44.666 ng/uL
RT: 3.746 min Scan# 95
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion: 79 Resp: 1123983
Ion Ratio Lower Upper
79 100
52 100.1 76.1 114.1
51 45.7 35.8 53.6





#6
Benzaldehyde
Concen: 22.640 ng/ul
RT: 7.069 min Scan# 660
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion: 77 Resp: 33386
Ion Ratio Lower Upper
77 100
105 91.8 74.3 111.5
106 88.4 73.3 109.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#8
Phenol
Concen: 37.841 ng/ul
RT: 7.122 min Scan# 669
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion: 94 Resp: 1189439
Ion Ratio Lower Upper
94 100
65 31.9 25.1 37.7
66 41.4 31.9 47.9





#10
Bis(2-Chloroethyl)ether
Concen: 36.628 ng/ul
RT: 7.357 min Scan# 709
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion: 93 Resp: 903870
Ion Ratio Lower Upper
93 100
63 87.4 70.9 106.3
95 32.7 26.0 39.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#12
2-Chlorophenol
Concen: 38.117 ng/ul
RT: 7.493 min Scan# 732
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Tgt Ion:128 Resp: 883020
Ion Ratio Lower Upper
128 100
64 55.6 44.8 67.2
130 32.7 26.6 39.8





#13

2-Methylphenol

Concen: 38.000 ng/ul

RT: 8.381 min Scan# 883

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129

Tgt Ion:108 Resp: 876958

Ion Ratio Lower Upper



108 100

107 89.1 71.5 107.3

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#14

2,2'-oxybis(1-Chloropropane)

Concen: 35.717 ng/ul

RT: 8.469 min Scan# 898

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion: 45 Resp: 1730693


Ion Ratio Lower Upper

45 100

77 10.6 8.4 12.6


79 8.5 6.6 9.8





#16
Acetophenone
Concen: 36.291 ng/ul
RT: 8.763 min Scan# 948
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:105 Resp: 1374944
Ion Ratio Lower Upper
105 100
77 83.5 64.8 97.2
51 37.1 28.7 43.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#17
N-Nitroso-di-n-propylamine
Concen: 36.492 ng/ul
RT: 8.752 min Scan# 946
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion: 70 Resp: 745730
Ion Ratio Lower Upper
70 100
42 79.0 64.2 96.2
101 9.3 7.3 10.9
130 18.9 15.0 22.6





#18
4-Methylphenol
Concen: 37.439 ng/ul
RT: 8.716 min Scan# 940
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:108 Resp: 96078
Ion Ratio Lower Upper
108 100
107 110.0 85.8 128.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#19
Hexachloroethane
Concen: 36.996 ng/ul
RT: 9.016 min Scan# 991
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:117 Resp: 345887
Ion Ratio Lower Upper
117 100
201 95.8 72.2 108.4
199 57.6 46.2 69.4





#22
Nitrobenzene
Concen: 37.815 ng/ul
RT: 9.146 min Scan# 1103
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion: 77 Resp: 111946
Ion Ratio Lower Upper
77 100
123 42.0 33.5 50.3
65 14.3 11.8 17.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#23
Isophorone
Concen: 37.701 ng/ul
RT: 9.675 min Scan# 1103
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion: 82 Resp: 2048698
Ion Ratio Lower Upper
82 100
95 6.9 5.4 8.0
138 16.1 12.8 19.2





#25

2-Nitrophenol

Concen: 38.968 ng/ul

RT: 9.857 min Scan# 1134

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129

Tgt Ion:139 Resp: 501610

Ion Ratio Lower Upper

139 100



65 62.2 49.8 74.8

109 31.4 26.2 39.2

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#26

2,4-Dimethylphenol

Concen: 37.694 ng/ul

RT: 9.922 min Scan# 1145

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:107 Resp: 1021597


Ion Ratio Lower Upper

107 100

121 49.1 39.7 59.5


122 83.1 66.1 99.1





#27
 Bis(2-Chloroethoxy)methane
 Concen: 36.330 ng/ul
 RT: 10.157 min Scan# 1185
 Delta R.T. -0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56



Instrument :
 BNA_N
 ClientSampleId :
 SLCS129



Tgt Ion: 93 Resp: 1228608
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.1 39.1
 123 10.8 8.5 12.7


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo






04/17/2023
 Supervised By :Jagrut
 Upadhyay


04/17/2023



#29
 2,4-Dichlorophenol
 Concen: 39.607 ng/ul
 RT: 10.393 min Scan# 1225
 Delta R.T. -0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56


Tgt Ion:162 Resp: 847857
 Ion Ratio Lower Upper
 162 100
 164 66.7 50.9 76.3
 98 39.1 31.0 46.4





#30
Naphthalene
Concen: 37.351 ng/ul
RT: 10.798 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:128 Resp: 283837
Ion Ratio Lower Upper
128 100
129 10.9 8.4 12.6
127 13.3 10.7 16.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#32
4-Chloroaniline
Concen: 19.891 ng/ul
RT: 10.910 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:127 Resp: 651771
Ion Ratio Lower Upper
127 100
129 32.4 25.5 38.3





#33
Hexachlorobutadiene
Concen: 38.511 ng/ul
RT: 11.081 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Instrument :

BNA_N



ClientSampleId :

SLCS129

Tgt Ion:225 Resp: 480504
Ion Ratio Lower Upper
225 100
223 65.2 45.9 68.9
227 63.5 51.2 76.8



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#34
Caprolactam
Concen: 39.134 ng/ul m
RT: 11.693 min Scan# 1446
Delta R.T. 0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:113 Resp: 293623
Ion Ratio Lower Upper
113 100
55 220.2 188.3 282.5
56 157.7 133.5 200.3





#35
4-Chloro-3-methylphenol
Concen: 39.758 ng/ul
RT: 12.034 min Scan# 1504
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:107 Resp: 99789
Ion Ratio Lower Upper
107 100
144 25.0 19.6 29.4
142 75.3 60.7 91.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#36
2-Methylnaphthalene
Concen: 37.832 ng/ul
RT: 12.404 min Scan# 1567
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:142 Resp: 1900022
Ion Ratio Lower Upper
142 100
141 89.0 71.4 107.2





#37
1-Methylnaphthalene
Concen: 38.643 ng/ul
RT: 12.628 min Scan# 1605
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:142 Resp: 1937354
Ion Ratio Lower Upper
142 100
141 94.5 73.5 110.3
116 3.9 3.1 4.7


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#39
1,2,4,5-Tetrachlorobenzene
Concen: 39.014 ng/ul
RT: 12.775 min Scan# 1630
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:216 Resp: 961102
Ion Ratio Lower Upper
216 100
214 77.6 62.2 93.4
179 21.9 16.8 25.2
108 19.0 15.0 22.6





#40
Hexachlorocyclopentadiene
Concen: 36.405 ng/ul
RT: 12.751 min Scan# 1626
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:237 Resp: 54688
Ion Ratio Lower Upper
237 100
235 68.1 49.9 74.9
272 13.3 10.2 15.4


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#41
2,4,6-Trichlorophenol
Concen: 37.535 ng/ul
RT: 13.016 min Scan# 1671
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Tgt Ion:196 Resp: 631797
Ion Ratio Lower Upper
196 100
198 95.9 74.5 111.7
200 31.8 24.3 36.5






#42
2,4,5-Trichlorophenol
Concen: 39.262 ng/u1
RT: 13.087 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :



SLCS129



Tgt Ion:196 Resp: 723197
Ion Ratio Lower Upper
196 100
198 99.5 78.2 117.2
200 31.9 24.0 36.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#43
1,1'-Biphenyl
Concen: 37.524 ng/u1
RT: 13.416 min Scan# 1739
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:154 Resp: 2525030
Ion Ratio Lower Upper
154 100
153 38.5 32.2 48.2
76 14.7 12.4 18.6





#44
2-Chloronaphthalene
Concen: 37.947 ng/ul
RT: 13.457 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:162 Resp: 1982699
Ion Ratio Lower Upper
162 100
164 32.2 25.9 38.9
127 40.4 32.9 49.3


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#45
2-Nitroaniline
Concen: 40.969 ng/ul
RT: 13.669 min Scan# 1782
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion: 65 Resp: 747477
Ion Ratio Lower Upper
65 100
92 63.8 50.7 76.1
138 93.3 71.4 107.2





#47
Dimethylphthalate
Concen: 38.700 ng/ul
RT: 14.039 min Scan# 1845
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:163 Resp: 2549360
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.2 12.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#48
2,6-Dinitrotoluene
Concen: 41.681 ng/ul
RT: 14.163 min Scan# 1866
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:165 Resp: 535944
Ion Ratio Lower Upper
165 100
89 63.5 49.8 74.8
121 21.2 17.2 25.8





#50
Acenaphthylene
Concen: 38.353 ng/ul
RT: 14.304 min Scan# 1890
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:152 Resp: 3160552
Ion Ratio Lower Upper
152 100
151 19.9 16.2 24.4
153 12.9 10.6 16.0


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#51
3-Nitroaniline
Concen: 43.360 ng/ul
RT: 14.492 min Scan# 1922
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:138 Resp: 585889
Ion Ratio Lower Upper
138 100
108 10.7 8.6 12.8
92 122.8 99.5 149.3





#52
Acenaphthene
Concen: 37.764 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:153 Resp: 215141
Ion Ratio Lower Upper
153 100
152 48.7 38.8 58.2
154 90.7 70.3 105.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#53
2,4-Dinitrophenol
Concen: 41.590 ng/ul
RT: 14.692 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:184 Resp: 343836
Ion Ratio Lower Upper
184 100
63 84.7 62.6 93.8
154 64.6 50.9 76.3





#55
4-Nitrophenol
Concen: 41.302 ng/ul
RT: 14.798 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:109 Resp: 45114
Ion Ratio Lower Upper
109 100
139 122.9 104.3 156.5
65 145.1 124.3 186.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#56
Dibenzofuran
Concen: 37.831 ng/ul
RT: 14.981 min Scan# 2005
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:168 Resp: 2970687
Ion Ratio Lower Upper
168 100
139 38.6 32.1 48.1





#57
2,4-Dinitrotoluene
Concen: 41.898 ng/ul
RT: 14.945 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:165 Resp: 779844
Ion Ratio Lower Upper
165 100
63 56.5 45.4 68.0
182 2.9 2.2 3.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#58
2,3,4,6-Tetrachlorophenol
Concen: 38.100 ng/ul
RT: 15.204 min Scan# 2043
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Tgt Ion:232 Resp: 577888
Ion Ratio Lower Upper
232 100
131 49.3 40.6 61.0
130 2.7 2.2 3.2
166 32.3 27.0 40.4






#59
Diethylphthalate
Concen: 38.969 ng/ul
RT: 15.398 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :



SLCS129



Tgt Ion:149 Resp: 257956
Ion Ratio Lower Upper
149 100
177 20.9 16.6 24.8
150 12.9 10.1 15.1


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#61
Fluorene
Concen: 38.375 ng/ul
RT: 15.628 min Scan# 2115
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:166 Resp: 2423800
Ion Ratio Lower Upper
166 100
165 99.3 79.1 118.7
167 13.5 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 39.822 ng/ul
RT: 15.622 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:204 Resp: 117070
Ion Ratio Lower Upper
204 100
206 33.1 26.0 39.0
141 63.3 51.8 77.8


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#63
4-Nitroaniline
Concen: 46.633 ng/ul
RT: 15.657 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56




Tgt Ion:138 Resp: 620970
Ion Ratio Lower Upper
138 100
92 52.8 43.3 64.9
108 80.0 64.2 96.4





#66
4,6-Dinitro-2-methylphenol
Concen: 40.843 ng/ul
RT: 15.710 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:198 Resp: 482818
Ion Ratio Lower Upper
198 100
182 3.9 3.4 5.0
77 28.5 23.7 35.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#67
N-Nitrosodiphenylamine
Concen: 34.850 ng/ul
RT: 15.833 min Scan# 2150
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:169 Resp: 1960587
Ion Ratio Lower Upper
169 100
168 69.1 54.3 81.5
167 36.7 27.6 41.4





#68
4-Bromophenyl-phenylether
Concen: 38.250 ng/ul
RT: 16.516 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:248 Resp: 712935
Ion Ratio Lower Upper
248 100
250 98.5 77.1 115.7
141 81.0 62.3 93.5


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#69
Hexachlorobenzene
Concen: 38.685 ng/ul
RT: 16.633 min Scan# 2286
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Tgt Ion:284 Resp: 816365
Ion Ratio Lower Upper
284 100
142 39.5 31.0 46.6
249 30.8 25.1 37.7





#71

Pentachlorophenol

Concen: 36.989 ng/ul

RT: 16.975 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129

Tgt Ion:266 Resp: 48518

Ion Ratio Lower Upper

266 100



264 65.6 52.7 79.1

268 64.0 51.4 77.0

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#72

Phenanthrene

Concen: 37.937 ng/ul

RT: 17.369 min Scan# 2411

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:178 Resp: 3969858


Ion Ratio Lower Upper

178 100

179 15.2 12.6 19.0

176 19.5 15.2 22.8






#74
 Anthracene
 Concen: 37.986 ng/u1
 RT: 17.463 min Scan# 2427
 Delta R.T. -0.000 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :



SLCS129



Tgt Ion:178 Resp: 400535
 Ion Ratio Lower Upper
 178 100
 179 15.6 12.2 18.4
 176 18.5 14.6 21.8


Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo




04/17/2023
 Supervised By :Jagrut
 Upadhyay



04/17/2023




#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 35.954 ng/uL
 RT: 13.381 min Scan# 1733
 Delta R.T. -0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56



Tgt Ion:216 Resp: 948286
 Ion Ratio Lower Upper
 216 100
 214 85.9 62.5 93.7
 218 50.9 37.7 56.5






#76
Pentachlorobenzene
Concen: 37.337 ng/uL
RT: 14.898 min Scan# 1991
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :



SLCS129



Tgt Ion:250 Resp: 94215
Ion Ratio Lower Upper
250 100
248 58.6 50.8 76.2
252 64.7 51.8 77.6


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo




04/17/2023
Supervised By :Jagrut
Upadhyay



04/17/2023




#77
Carbazole
Concen: 37.859 ng/uL
RT: 17.733 min Scan# 2473
Delta R.T. -0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Tgt Ion:167 Resp: 3604550
Ion Ratio Lower Upper
167 100
166 21.6 16.6 24.8
139 13.5 10.2 15.4





#78

Di-n-butylphthalate

Concen: 40.310 ng/ul

RT: 18.292 min Scan# 2568

Delta R.T. -0.006 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129

Tgt Ion:149 Resp: 4555290

Ion Ratio Lower Upper

149 100



150 9.1 7.4 11.2

104 5.4 4.2 6.4

Manual Integrations



APPROVED

Reviewed By :Christian
Giraldo



04/17/2023
Supervised By :Jagrut
Upadhyay

04/17/2023



#80

Fluoranthene

Concen: 38.443 ng/ul

RT: 19.386 min Scan# 2754

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:202 Resp: 4745423


Ion Ratio Lower Upper

202 100

101 13.3 10.7 16.1

100 10.1 7.9 11.9





#82

Pyrene

Concen: 37.000 ng/ul

RT: 19.751 min Scan# 2816

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56

Instrument :

BNA_N

ClientSampleId :

SLCS129

Tgt Ion:202 Resp: 480753

Ion Ratio Lower Upper

202 100



101 15.8 13.0 19.6

100 13.1 11.0 16.6

Manual Integrations

APPROVED



Reviewed By :Christian
Giraldo



04/17/2023

Supervised By :Jagrut
Upadhyay

04/17/2023



#83

Butylbenzylphthalate

Concen: 40.678 ng/ul

RT: 20.645 min Scan# 2968

Delta R.T. -0.000 min

Lab File: BN024945.D

Acq: 15 Apr 2023 21:56




Tgt Ion:149 Resp: 2194346


Ion Ratio Lower Upper

149 100

91 79.4 62.6 93.8


206 20.2 16.2 24.2





#84
3,3'-Dichlorobenzidine
Concen: 20.797 ng/ul
RT: 21.439 min Scan# 3103
Delta R.T. 0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:252 Resp: 846354
Ion Ratio Lower Upper
252 100
254 64.5 52.2 78.2
126 12.1 11.4 17.2


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#85
Benzo(a)anthracene
Concen: 38.489 ng/ul
RT: 21.504 min Scan# 3114
Delta R.T. -0.000 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Tgt Ion:228 Resp: 4903089
Ion Ratio Lower Upper
228 100
229 20.7 16.2 24.2
226 26.9 20.7 31.1







#86
 Bis(2-ethylhexyl)phthalate
 Concen: 40.033 ng/ul
 RT: 21.427 min Scan# 3101
 Delta R.T. 0.006 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129

Tgt Ion:149 Resp: 322378
 Ion Ratio Lower Upper
 149 100
 167 27.4 21.8 32.8
 279 4.1 2.9 4.3



Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo






04/17/2023
 Supervised By :Jagrut
 Upadhyay


04/17/2023



#87
 Chrysene
 Concen: 38.793 ng/ul
 RT: 21.557 min Scan# 3123
 Delta R.T. -0.000 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Tgt Ion:228 Resp: 4707028
 Ion Ratio Lower Upper
 228 100
 226 29.3 22.4 33.6
 229 19.5 15.6 23.4







#89
Di-n-octyl phthalate
Concen: 44.129 ng/ul
RT: 22.368 min Scan# 31
Delta R.T. 0.006 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56

Instrument :
BNA_N
ClientSampleId :
SLCS129

Tgt Ion:149 Resp: 5626680



Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#90
Benzo(b)fluoranthene
Concen: 41.423 ng/ul
RT: 23.221 min Scan# 3406
Delta R.T. 0.012 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:252 Resp: 5040198
Ion Ratio Lower Upper
252 100
253 22.0 19.2 28.8
125 12.1 10.4 15.6





#91
Benzo(k)fluoranthene
Concen: 42.156 ng/u1
RT: 23.274 min Scan# 3415
Delta R.T. 0.012 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56



Instrument :
BNA_N
ClientSampleId :
SLCS129



Tgt Ion:252 Resp: 504726
Ion Ratio Lower Upper
252 100
253 22.8 16.7 25.1
125 11.7 9.3 13.9


Manual Integrations
APPROVED

Reviewed By :Christian
Giraldo






04/17/2023
Supervised By :Jagrut
Upadhyay


04/17/2023



#93
Benzo(a)pyrene
Concen: 40.589 ng/u1
RT: 23.868 min Scan# 3516
Delta R.T. 0.023 min
Lab File: BN024945.D
Acq: 15 Apr 2023 21:56


Tgt Ion:252 Resp: 4323891
Ion Ratio Lower Upper
252 100
253 20.9 17.4 26.0
125 12.6 10.8 16.2





#94
 Indeno(1,2,3-cd)pyrene
 Concen: 38.786 ng/ul
 RT: 26.527 min Scan# 3968
 Delta R.T. 0.035 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Instrument :
 BNA_N
 ClientSampleId :
 SLCS129





Tgt Ion: 276 Resp: 500627

Ion	Ratio	Lower	Upper
276	100		
138	27.9	21.9	32.9
277	26.5	19.8	29.6


Manual Integrations
 APPROVED

Reviewed By : Christian
 Giraldo



04/17/2023
 Supervised By : Jagrut
 Upadhyay




04/17/2023




#95
 Dibenzo(a,h)anthracene
 Concen: 39.117 ng/ul
 RT: 26.539 min Scan# 3970
 Delta R.T. 0.029 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56

Tgt Ion: 278 Resp: 4168365


Ion	Ratio	Lower	Upper
278	100		
139	18.9	14.6	21.8
279	23.4	18.2	27.4





#96
 Benzo(g,h,i)perylene
 Concen: 37.430 ng/ul
 RT: 27.321 min Scan# 4103
 Delta R.T. 0.047 min
 Lab File: BN024945.D
 Acq: 15 Apr 2023 21:56



Instrument :
 BNA_N
 ClientSampleId :
 SLCS129



Tgt Ion:276 Resp: 3920320
 Ion Ratio Lower Upper
 276 100
 138 23.0 19.8 29.8
 277 21.6 17.6 26.4

Manual Integrations
 APPROVED

Reviewed By :Christian
 Giraldo



04/17/2023
 Supervised By :Jagrut
 Upadhyay

04/17/2023