

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041520\
 Data File : BN010503.D
 Acq On : 15 Apr 2020 10:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 15 11:31:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	-0.02
2	1,4-Dioxane	0.294	0.317	-7.8	95	0.00
3	n-Nitrosodimethylamine	0.191	0.241	-26.2#	131	-0.02
4 S	2-Fluorophenol	0.876	0.893	-1.9	89	0.00
5 S	Phenol-d6	1.145	1.244	-8.6	95	-0.02
6	bis(2-Chloroethyl)ether	0.992	1.093	-10.2	95	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	90	-0.01
8 S	Nitrobenzene-d5	0.280	0.297	-6.1	100	-0.01
9	Naphthalene	0.993	1.081	-8.9	99	-0.01
10	Hexachlorobutadiene	0.239	0.243	-1.7	91	-0.01
11 SURR	2-Methylnaphthalene-d10	0.658	0.671	-2.0	93	0.00
12	2-Methylnaphthalene	0.714	0.791	-10.8	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.189	0.184	2.6	97	0.00
15 S	2-Fluorobiphenyl	1.456	1.618	-11.1	105	-0.01
16	Acenaphthylene	1.728	1.811	-4.8	102	0.00
17	Acenaphthene	1.110	1.182	-6.5	102	0.00
18	Fluorene	1.454	1.559	-7.2	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	-0.01
20	4-Bromophenyl-phenylether	0.222	0.227	-2.3	102	-0.01
21	Hexachlorobenzene	0.254	0.269	-5.9	101	-0.01
22	Pentachlorophenol	0.100	0.103	-3.0	118	-0.01
23	Phenanthrene	1.080	1.206	-11.7	108	-0.01
24	Anthracene	0.988	1.102	-11.5	111	-0.01
25 SURR	Fluoranthene-d10	1.213	1.314	-8.3	105	0.00
26	Fluoranthene	1.345	1.571	-16.8	113	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	-0.01
28	Pvrene	1.346	1.416	-5.2	115	0.00
29 S	Terphenyl-d14	0.915	1.042	-13.9	127	0.00
30	Benzo(a)anthracene	1.305	1.366	-4.7	115	0.00
31	Chrysene	1.300	1.402	-7.8	119	-0.01
32	Bis(2-ethylhexyl)phthalate	0.670	0.573	14.5	98	-0.01
33	Indeno(1,2,3-cd)pyrene	1.149	1.308	-13.8	123	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	111	-0.01
35	Benzo(b)fluoranthene	1.317	1.407	-6.8	119	-0.01
36	Benzo(k)fluoranthene	1.320	1.466	-11.1	124	-0.01
37 C	Benzo(a)pyrene	1.166	1.272	-9.1	123	-0.01
38	Dibenzo(a,h)anthracene	1.066	1.196	-12.2	125	-0.01
39	Benzo(a,h,i)perylene	1.073	1.220	-13.7	126	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041520\
Data File : BN010503.D
Acq On : 15 Apr 2020 10:55
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 15 11:31:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			