

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041520\
 Data File : BN010510.D
 Acq On : 15 Apr 2020 16:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 15 16:56:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.294	0.316	-7.5	103	0.00
3	n-Nitrosodimethylamine	0.191	0.255	-33.5#	151#	-0.02
4 S	2-Fluorophenol	0.876	0.905	-3.3	98	0.00
5 S	Phenol-d6	1.145	1.292	-12.8	107	0.00
6	bis(2-Chloroethyl)ether	0.992	1.144	-15.3	108	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	-0.01
8 S	Nitrobenzene-d5	0.280	0.304	-8.6	113	-0.01
9	Naphthalene	0.993	1.078	-8.6	109	-0.01
10	Hexachlorobutadiene	0.239	0.247	-3.3	102	-0.01
11 SURR	2-Methylnaphthalene-d10	0.658	0.673	-2.3	103	0.00
12	2-Methylnaphthalene	0.714	0.797	-11.6	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.189	0.196	-3.7	114	0.00
15 S	2-Fluorobiphenyl	1.456	1.622	-11.4	116	0.00
16	Acenaphthylene	1.728	1.848	-6.9	115	0.00
17	Acenaphthene	1.110	1.207	-8.7	115	0.00
18	Fluorene	1.454	1.583	-8.9	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.222	0.226	-1.8	112	0.00
21	Hexachlorobenzene	0.254	0.269	-5.9	112	0.00
22	Pentachlorophenol	0.100	0.110	-10.0	139	0.00
23	Phenanthrene	1.080	1.211	-12.1	120	0.00
24	Anthracene	0.988	1.112	-12.6	124	0.00
25 SURR	Fluoranthene-d10	1.213	1.321	-8.9	117	0.00
26	Fluoranthene	1.345	1.579	-17.4	126	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	120	-0.01
28	Pvrene	1.346	1.419	-5.4	128	0.00
29 S	Terphenyl-d14	0.915	1.059	-15.7	142	0.00
30	Benzo(a)anthracene	1.305	1.381	-5.8	128	0.00
31	Chrysene	1.300	1.399	-7.6	131	-0.01
32	Bis(2-ethylhexyl)phthalate	0.670	0.632	5.7	120	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.303	-13.4	136	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	123	-0.01
35	Benzo(b)fluoranthene	1.317	1.402	-6.5	131	-0.01
36	Benzo(k)fluoranthene	1.320	1.452	-10.0	136	0.00
37 C	Benzo(a)pyrene	1.166	1.293	-10.9	137	-0.01
38	Dibenzo(a,h)anthracene	1.066	1.194	-12.0	137	-0.02
39	Benzo(a,h,i)perylene	1.073	1.209	-12.7	138	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041520\
Data File : BN010510.D
Acq On : 15 Apr 2020 16:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 15 16:56:58 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			