

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041520\
 Data File : BN010510.D
 Acq On : 15 Apr 2020 16:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 15 16:56:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.430	-7.5	103	0.00
3	n-Nitrosodimethylamine	0.400	0.533	-33.3#	151	-0.02
4 S	2-Fluorophenol	0.400	0.413	-3.2	98	0.00
5 S	Phenol-d6	0.400	0.451	-12.7	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.461	-15.3	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	-0.01
8 S	Nitrobenzene-d5	0.400	0.435	-8.7	113	-0.01
9	Naphthalene	0.400	0.434	-8.5	109	-0.01
10	Hexachlorobutadiene	0.400	0.413	-3.2	102	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.410	-2.5	103	0.00
12	2-Methylnaphthalene	0.400	0.447	-11.7	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	0.00
14 S	2,4,6-Tribromophenol	0.400	0.415	-3.7	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.446	-11.5	116	0.00
16	Acenaphthylene	0.400	0.428	-7.0	115	0.00
17	Acenaphthene	0.400	0.435	-8.7	115	0.00
18	Fluorene	0.400	0.435	-8.7	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	107	0.00
20	4-Bromophenyl-phenylether	0.400	0.407	-1.7	112	0.00
21	Hexachlorobenzene	0.400	0.425	-6.2	112	0.00
22	Pentachlorophenol	0.400	0.440	-10.0	139	0.00
23	Phenanthrene	0.400	0.448	-12.0	120	0.00
24	Anthracene	0.400	0.450	-12.5	124	0.00
25 SURR	Fluoranthene-d10	0.400	0.436	-9.0	117	0.00
26	Fluoranthene	0.400	0.470	-17.5	126	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	120	-0.01
28	Pyrene	0.400	0.422	-5.5	128	0.00
29 S	Terphenyl-d14	0.400	0.463	-15.8	142	0.00
30	Benzo(a)anthracene	0.400	0.423	-5.7	128	0.00
31	Chrysene	0.400	0.430	-7.5	131	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.378	5.5	120	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.453	-13.2	136	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	123	-0.01
35	Benzo(b)fluoranthene	0.400	0.426	-6.5	131	-0.01
36	Benzo(k)fluoranthene	0.400	0.440	-10.0	136	0.00
37 C	Benzo(a)pyrene	0.400	0.444	-11.0	137	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.448	-12.0	137	-0.02
39	Benzo(a,h,i)perylene	0.400	0.451	-12.7	138	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041520\
Data File : BN010510.D
Acq On : 15 Apr 2020 16:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 15 16:56:58 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				