


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041525\
Data File : BN036910.D
Acq On : 15 Apr 2025 10:20
Operator : RC/JU
Sample : DFTPP373
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP373

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041525.MA.M
Title : SVOA CALIBRATION
Last Update : Tue Apr 15 15:13:59 2025



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	70.0	112000	PASS
68	69	0.00	2	0.8	746	PASS
69	198	0.00	100	58.1	92944	PASS
70	69	0.00	2	0.6	521	PASS
127	198	10	80	52.9	84680	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	160085	PASS
199	198	5	9	6.8	10825	PASS
275	198	10	60	25.6	40944	PASS
365	198	1	100	4.2	6669	PASS
441	198	0.01	100	8.3	13347	PASS
442	198	50	100	53.9	86208	PASS
443	442	15	24	19.5	16853	PASS

m/z	Abundance
39.90	651.0
44.00	1298.0
48.90	203.0
73.10	578.0
147.00	238.0
206.90	603.0
221.00	162.0
281.10	246.0

Instrument :

BNA_N

ClientSampleId :

DFTPP373

m/z	Abundance
37.00	613.0
38.00	2156.0
39.00	10029.0
40.00	1222.0
41.00	308.0
44.00	1781.0
45.00	308.0
49.00	1249.0
50.00	35376.0
51.00	121072.0
52.00	6282.0
52.90	220.0
55.00	697.0
56.00	3503.0
57.00	6445.0
57.90	197.0
61.00	1566.0
62.00	1724.0
63.00	4405.0
63.90	468.0
65.00	1777.0
67.10	203.0
68.10	2237.0
69.00	98864.0
70.00	595.0
73.00	1359.0
74.00	12017.0
75.00	16295.0
76.00	6132.0
77.00	101168.0
78.00	7086.0
79.00	8787.0
80.00	5907.0
81.00	8831.0
82.00	1874.0
82.90	1652.0
83.80	227.0
84.90	1333.0
86.00	2919.0
87.00	1051.0
87.90	347.0
88.90	184.0
90.90	1669.0
92.00	2349.0
93.00	15753.0
93.90	1100.0
94.90	180.0
96.00	592.0
97.90	12713.0
99.00	8731.0
99.90	774.0
101.00	3548.0
101.90	244.0
103.00	1200.0
104.00	2675.0
104.90	2524.0
106.10	1012.0
107.00	31840.0
108.00	4564.0
110.00	47720.0
111.00	8513.0
111.90	1199.0
112.90	369.0
114.80	185.0
116.90	40008.0
117.90	2943.0
119.10	270.0
120.00	291.0
122.00	2221.0
123.00	3153.0
123.90	1594.0
124.90	1229.0
127.00	85368.0
128.00	6816.0
129.00	36768.0
129.90	3374.0
130.90	598.0
132.00	364.0
133.10	241.0
134.00	1016.0
134.90	3055.0
135.90	1325.0
137.00	2146.0
137.70	314.0
138.90	233.0
139.90	524.0
141.00	5537.0
142.00	1747.0
142.90	1057.0
143.80	361.0
144.90	231.0
146.00	1045.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

147.00	2685.0
148.00	8145.0
148.90	1437.0
150.00	291.0
151.00	596.0
152.00	558.0
152.90	1888.0
154.00	1070.0
155.00	2936.0
156.00	4353.0
157.00	564.0
157.90	876.0
158.90	874.0
159.90	1723.0
161.00	2211.0
161.90	714.0
162.90	220.0
163.80	211.0
165.00	2348.0
166.00	1493.0
167.00	12750.0
167.90	7604.0
169.00	893.0
169.90	338.0
170.90	568.0
171.90	975.0
172.90	1226.0
174.00	2517.0
175.00	4350.0
176.00	760.0
176.90	1632.0
178.00	646.0
178.90	8100.0
180.00	4637.0
181.00	2057.0
181.90	490.0
182.90	211.0
184.00	693.0
185.00	3632.0
186.00	23664.0
186.90	6611.0
187.90	900.0
188.90	1980.0
189.80	469.0
190.90	812.0
191.90	2536.0
193.00	2886.0
194.00	666.0
195.00	267.0
196.00	4315.0
197.90	151872.0
198.90	10412.0
199.90	823.0
201.20	385.0
202.10	163.0
203.00	1643.0
204.00	7473.0
205.00	12085.0
206.00	43976.0
207.00	6295.0
208.00	2347.0
208.80	624.0
209.90	697.0
211.00	2074.0
212.10	275.0
212.90	239.0
214.80	609.0
215.90	1234.0
217.00	13620.0
217.90	2138.0
218.90	178.0
221.00	4598.0
222.00	1695.0
222.90	2449.0
224.00	24936.0
225.00	6241.0
226.10	870.0
226.90	13404.0
227.90	2064.0
229.00	2097.0
230.00	403.0
231.00	806.0
231.90	232.0
232.90	260.0
233.90	829.0
235.00	881.0
235.90	681.0
236.90	962.0
237.80	178.0
239.00	625.0
239.90	459.0
240.90	642.0
241.90	1166.0
243.00	1502.0
244.00	17088.0
245.00	2285.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

246.00	4711.0
246.90	1021.0
247.90	188.0
249.00	537.0
250.90	253.0
252.90	548.0
255.00	84424.0
255.90	13198.0
256.90	1092.0
258.00	6583.0
259.00	1203.0
264.90	2874.0
265.90	310.0
270.90	254.0
271.80	236.0
272.90	2397.0
274.00	7122.0
275.00	34320.0
276.00	4806.0
276.90	3759.0
277.90	716.0
283.00	398.0
283.90	189.0
285.00	661.0
289.00	172.0
292.00	178.0
292.90	662.0
295.00	424.0
296.00	12173.0
296.90	2090.0
301.90	160.0
303.00	1456.0
304.00	323.0
307.90	226.0
313.90	669.0
314.90	1469.0
315.90	553.0
317.10	165.0
321.00	323.0
321.90	192.0
323.00	2884.0
324.00	539.0
326.80	702.0
328.10	275.0
331.90	234.0
332.90	272.0
334.00	1816.0
335.00	406.0
340.90	365.0
345.90	620.0
352.00	1174.0
353.00	634.0
354.00	934.0
355.00	163.0
364.90	4913.0
365.80	812.0
370.80	179.0
371.90	1322.0
372.90	319.0
382.80	349.0
390.00	199.0
402.00	418.0
402.90	885.0
403.90	291.0
420.90	615.0
422.00	677.0
422.90	3715.0
423.90	833.0
440.90	8897.0
442.00	60336.0
442.90	12559.0
444.00	1243.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

m/z	Abundance
37.00	756.0
38.00	2060.0
39.00	10988.0
40.00	1252.0
41.00	205.0
44.00	1761.0
44.90	360.0
49.00	1386.0
50.00	37936.0
51.00	134208.0
52.00	6890.0
53.00	269.0
54.90	738.0
55.90	3905.0
57.00	7929.0
58.00	318.0
58.90	173.0
61.00	1433.0
62.00	2021.0
63.00	5026.0
63.90	878.0
65.00	2205.0
66.90	245.0
68.90	110088.0
69.90	577.0
73.00	1575.0
74.00	14120.0
75.00	19392.0
76.10	6638.0
77.00	119656.0
78.00	7872.0
79.00	9369.0
79.90	6828.0
81.00	9214.0
81.90	2318.0
82.90	1557.0
84.00	279.0
85.00	1354.0
86.00	3583.0
87.00	1240.0
87.90	480.0
88.80	211.0
90.90	2441.0
91.90	2279.0
93.00	18856.0
93.90	1219.0
94.90	217.0
95.90	434.0
97.90	14206.0
99.00	10316.0
100.00	784.0
101.00	4296.0
101.90	327.0
102.90	1787.0
103.90	3110.0
105.00	3488.0
107.00	37112.0
108.00	5618.0
110.00	55640.0
111.00	8754.0
111.90	1505.0
113.00	403.0
115.00	261.0
116.00	2011.0
117.00	47696.0
117.90	2924.0
118.90	389.0
120.00	591.0
122.00	2533.0
123.00	4499.0
123.90	1798.0
125.00	1516.0
127.00	101128.0
128.00	8620.0
129.00	45192.0
130.00	4466.0
130.90	776.0
132.00	458.0
132.90	307.0
134.00	1584.0
135.00	3662.0
135.90	1699.0
137.00	2306.0
137.80	406.0
138.90	290.0
139.90	776.0
141.00	7554.0
141.90	2402.0
142.90	1548.0
143.90	466.0
145.00	466.0
145.90	1187.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

147.00	4108.0
147.90	9932.0
148.90	1646.0
150.00	572.0
151.00	849.0
151.90	513.0
152.90	1864.0
154.00	1532.0
155.00	3575.0
156.00	5379.0
157.00	941.0
158.00	1402.0
158.90	902.0
159.90	2165.0
161.00	2960.0
161.90	616.0
163.00	317.0
164.00	294.0
165.00	2737.0
166.00	1849.0
167.00	15627.0
167.90	8376.0
168.90	1335.0
169.90	425.0
170.90	586.0
171.90	1192.0
173.00	1467.0
174.00	2868.0
175.00	5300.0
176.00	1172.0
176.90	2603.0
178.00	825.0
178.90	10523.0
180.00	5816.0
180.90	2900.0
181.90	604.0
182.70	198.0
183.90	822.0
185.00	4750.0
186.00	34072.0
187.00	9715.0
187.90	900.0
189.00	2471.0
189.90	496.0
191.00	1182.0
192.00	3217.0
193.00	3797.0
193.90	852.0
194.90	441.0
195.90	5084.0
197.90	193088.0
198.90	13090.0
199.90	1111.0
201.20	791.0
203.00	1974.0
204.00	8874.0
205.00	17048.0
206.00	60544.0
207.00	8569.0
207.90	2795.0
208.90	1373.0
210.00	1058.0
211.00	2612.0
211.70	413.0
212.90	251.0
214.10	160.0
214.90	1066.0
216.00	1705.0
216.90	19568.0
217.90	2546.0
218.80	274.0
220.10	524.0
221.00	6270.0
221.90	2111.0
223.00	4191.0
224.00	33528.0
225.00	9568.0
226.00	1145.0
226.90	18232.0
227.90	2469.0
228.90	3150.0
229.90	582.0
231.00	1067.0
231.80	274.0
232.90	357.0
234.00	1317.0
235.00	1288.0
236.00	1110.0
237.00	1480.0
238.00	228.0
238.90	675.0
240.00	565.0
240.90	911.0
242.00	2244.0
243.10	2032.0
244.00	24552.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

245.00	3332.0
246.00	6188.0
246.80	1045.0
247.80	246.0
248.90	866.0
250.90	181.0
251.90	282.0
253.00	750.0
255.00	123224.0
256.00	17728.0
256.90	1504.0
258.00	9073.0
258.90	1410.0
259.90	232.0
260.90	209.0
264.90	3910.0
265.90	658.0
270.00	274.0
270.90	331.0
272.10	362.0
273.00	3566.0
274.00	9023.0
275.00	49824.0
275.90	6019.0
276.90	5969.0
277.90	1101.0
281.00	304.0
281.90	259.0
282.90	555.0
283.90	311.0
285.00	847.0
285.80	186.0
288.90	217.0
290.00	166.0
291.90	172.0
292.90	1137.0
293.90	233.0
296.00	17928.0
296.90	3101.0
297.80	299.0
300.90	227.0
302.00	229.0
303.00	1666.0
304.00	453.0
307.90	195.0
309.00	161.0
309.90	173.0
312.80	180.0
313.90	845.0
314.90	2238.0
316.10	733.0
320.90	377.0
322.00	241.0
323.00	4433.0
324.00	810.0
326.90	888.0
327.90	609.0
331.90	386.0
333.00	339.0
334.00	3653.0
335.00	823.0
336.00	162.0
340.90	457.0
342.00	157.0
345.90	1162.0
346.90	265.0
351.90	1293.0
352.90	935.0
354.00	1053.0
354.90	283.0
364.90	8062.0
365.90	1182.0
370.80	319.0
371.90	2090.0
373.00	399.0
383.00	599.0
389.90	282.0
390.90	183.0
391.90	197.0
401.10	175.0
402.00	904.0
402.90	987.0
403.90	455.0
420.90	1017.0
422.00	936.0
422.90	5901.0
423.90	1416.0
424.90	173.0
441.00	16344.0
442.00	101752.0
443.00	19896.0
443.90	2062.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

m/z	Abundance
37.00	491.0
38.00	1203.0
39.00	6414.0
40.00	1121.0
41.20	211.0
44.00	1768.0
44.90	155.0
49.10	903.0
50.00	23264.0
51.00	80720.0
52.00	3778.0
55.00	472.0
55.90	2284.0
57.00	4643.0
57.90	238.0
61.00	815.0
62.00	1169.0
63.00	3294.0
63.90	331.0
65.00	1077.0
69.00	69880.0
69.90	390.0
72.90	781.0
74.00	8433.0
75.00	11333.0
76.00	4223.0
77.00	73024.0
78.00	4873.0
79.00	5601.0
79.90	4791.0
80.90	6089.0
82.00	1354.0
82.90	997.0
83.90	294.0
85.00	1035.0
86.00	2007.0
87.00	716.0
87.90	286.0
88.90	183.0
90.90	1346.0
92.00	1669.0
93.00	11892.0
94.00	758.0
94.90	182.0
96.00	424.0
97.10	219.0
98.00	9700.0
99.00	6431.0
99.90	694.0
101.00	3039.0
101.90	165.0
102.90	696.0
103.90	2187.0
104.90	2290.0
107.00	23832.0
107.90	3237.0
109.90	34976.0
111.00	5975.0
111.90	826.0
112.90	355.0
116.00	1453.0
116.90	30544.0
117.90	2084.0
118.90	197.0
119.90	329.0
121.00	209.0
121.90	2053.0
123.00	2982.0
123.90	1304.0
124.90	822.0
127.00	67544.0
128.00	5564.0
129.00	30008.0
129.90	2272.0
130.80	441.0
131.90	207.0
133.20	201.0
133.90	828.0
134.90	2495.0
135.90	1055.0
137.00	1577.0
137.90	224.0
138.80	252.0
139.90	371.0
140.90	5036.0
142.00	1398.0
142.90	1195.0
144.10	215.0
144.90	226.0
145.90	805.0
147.00	2608.0
148.00	6465.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

149.00	1241.0
149.90	338.0
151.00	607.0
151.90	394.0
153.00	1193.0
153.90	961.0
155.00	2431.0
156.00	3413.0
157.00	613.0
157.90	910.0
158.90	551.0
160.00	1382.0
160.90	1785.0
161.90	646.0
162.90	217.0
164.90	1907.0
165.90	1316.0
166.90	10842.0
168.00	6554.0
169.00	795.0
169.80	267.0
170.80	353.0
172.00	786.0
172.90	1214.0
173.90	1981.0
174.90	3562.0
176.00	983.0
176.90	1893.0
178.90	6666.0
179.90	4154.0
181.00	2200.0
182.00	308.0
184.00	690.0
185.00	3443.0
186.00	23008.0
187.00	6468.0
188.00	679.0
189.00	2100.0
189.90	342.0
190.90	907.0
191.90	1715.0
193.00	2512.0
193.90	370.0
194.90	216.0
195.90	4077.0
197.90	135296.0
199.00	8973.0
199.90	649.0
201.50	678.0
202.00	262.0
202.90	1383.0
204.00	6595.0
205.00	10576.0
206.00	41848.0
207.00	6126.0
207.90	2044.0
208.90	613.0
210.10	768.0
210.90	1887.0
211.90	252.0
212.90	185.0
214.90	746.0
215.90	989.0
216.90	12775.0
217.90	1271.0
221.00	4279.0
221.90	1581.0
223.00	3235.0
224.00	24656.0
225.00	5661.0
227.00	13726.0
227.90	1864.0
229.00	1985.0
229.90	301.0
230.90	925.0
232.90	206.0
233.90	912.0
234.90	956.0
235.90	835.0
237.00	940.0
238.90	422.0
239.90	473.0
241.00	698.0
242.00	1463.0
242.90	1356.0
244.00	18576.0
245.00	2350.0
245.90	4481.0
246.90	873.0
248.00	239.0
248.90	701.0
250.90	199.0
251.90	250.0
253.00	603.0
254.90	90768.0
256.00	13131.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373

256.90	940.0
258.00	7072.0
258.90	1148.0
260.00	155.0
264.00	206.0
265.00	3108.0
265.90	461.0
270.80	221.0
272.90	2479.0
274.00	7045.0
275.00	38688.0
276.00	4926.0
277.00	4666.0
277.90	728.0
281.00	153.0
282.90	417.0
283.90	397.0
285.00	679.0
285.80	157.0
288.80	186.0
291.00	176.0
292.00	279.0
292.90	1053.0
293.90	209.0
296.00	14726.0
297.00	2337.0
300.90	178.0
302.00	288.0
303.00	1257.0
303.90	467.0
307.90	172.0
309.90	226.0
313.90	689.0
314.90	1582.0
316.00	702.0
316.90	175.0
321.00	437.0
321.90	274.0
323.00	3181.0
324.00	762.0
326.90	759.0
327.90	323.0
331.90	308.0
332.90	477.0
334.00	2534.0
335.00	596.0
341.00	426.0
345.90	887.0
346.90	317.0
351.90	999.0
352.90	766.0
354.00	1022.0
354.90	245.0
364.90	7033.0
366.00	979.0
371.00	285.0
372.00	1622.0
373.00	637.0
383.00	497.0
390.00	243.0
401.90	661.0
402.90	1042.0
403.90	337.0
420.90	1030.0
421.90	753.0
423.00	4991.0
423.90	1168.0
439.70	234.0
441.00	14801.0
441.90	96536.0
443.00	18104.0
444.00	1397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP373