

Quantitation Report (QT Reviewed)


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041620\
 Data File : BN010526.D
 Acq On : 16 Apr 2020 20:59
 Operator : CG/JU
 Sample : L2191-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB799

Manual Integrations
 APPROVED

mohammad
 4/17/2020 1:52:10 PM

Quant Time: Apr 17 02:33:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 16 10:21:51 2020
 Response via : Initial Calibration



Quantitation Report (Qedit)



Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041620\
 Data File : BN010526.D
 Acq On : 16 Apr 2020 20:59
 Operator : CG/JU
 Sample : L2191-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB799

Manual Integrations
 APPROVED

mohammad
 4/17/2020 1:52:10 PM

Quant Time: Apr 17 02:30:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 16 10:21:51 2020
 Response via : Initial Calibration



(89) Benzo(k)fluoranthene
 22.704min (-0.047) 3.60ng/ui
 response 653431

Ion	Exp%	Act%
252.00	100	100
253.00	21.80	23.47
125.00	8.60	10.55#
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN041620\
 Data File : BN010526.D
 Acq On : 16 Apr 2020 20:59
 Operator : CG/JU
 Sample : L2191-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampled :
 CB799

Manual Integrations
 APPROVED

mohammad
 4/17/2020 1:52:10 PM

Quant Time: Apr 17 02:30:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 16 10:21:51 2020
 Response via : Initial Calibration



(89) Benzo(k)fluoranthene

22.739min (-0.012) 1.41ng/ul m

response 255551

Ion	Exp%	Act%
252.00	100	100
253.00	21.80	24.52
125.00	8.60	18.76#
0.00	0.00	0.00

AP
 04/20/2020

Quantitation Report (QT Reviewed)

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN041620\
 Data File : BN010526.D
 Acq On : 16 Apr 2020 20:59
 Operator : CG/JU
 Sample : L2191-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 CB799

Manual Integrations
 APPROVED

mohammad
 4/17/2020 1:52:10 PM

Quant Time: Apr 17 02:33:54 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN040920MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 16 10:21:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	486152	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	2143381	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.22	164	1383739	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.97	188	3193238	20.00	ng/ul	0.00
78) Chrysene-d12	21.17	240	3059631	20.00	ng/ul	0.00
86) Perylene-d12	23.35	264	2805565	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.20	96	58356	6.28	ng/uL	0.00
5) Phenol-d5	6.78	99	1275603	33.23	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	680686	35.06	ng/ul	0.00
9) 2-Chlorophenol-d4	7.14	132	1117808	34.16	ng/ul	0.00
13) 4-Methylphenol-d8	8.30	113	1016901	31.42	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	564351	34.17	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	641184	34.83	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.98	165	1162821	32.23	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	1099273	26.57	ng/ul	0.00
44) Dimethylphthalate-d6	13.64	166	3734636	36.38	ng/ul	0.00
47) Acenaphthylene-d8	13.91	160	4563642	36.18	ng/ul	0.00
52) 4-Nitrophenol-d4	14.43	143	634371	31.73	ng/ul	0.00
58) Fluorene-d10	15.22	176	3293208	35.74	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.35	200	526702	31.58	ng/ul	0.00
71) Anthracene-d10	17.07	188	5108324	34.54	ng/ul	0.00
79) Pyrene-d10	19.37	212	6003707	34.76	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.22	264	5166735	35.50	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)	Ovalue
45) Dimethylphthalate	13.69	163	911478	8.551	ng/ul	99	
70) Phenanthrene	17.01	178	413548	2.332	ng/ul	98	
77) Fluoranthene	19.03	202	1050304	5.451	ng/ul	98	
80) Pyrene	19.40	202	951487	4.304	ng/ul	98	
83) Benzo(a)anthracene	21.16	228	584447	2.754	ng/ul	99	
85) Chrysene	21.21	228	568548	2.849	ng/ul	97	
88) Benzo(b)fluoranthene	22.70	252	653431	3.569	ng/ul	95	
89) Benzo(k)fluoranthene	22.74	252	255551m	1.407	ng/ul	98	
91) Benzo(a)pyrene	23.25	252	410562	2.524	ng/ul	98	
92) Indeno(1,2,3-cd)pyrene	25.48	276	267313	1.460	ng/ul	95	
94) Benzo(g,h,i)perylene	26.13	276	194654	1.285	ng/ul#	89	

AP
 04/20/2020

(#) = qualifier out of range (m) = manual integration (+) = signals summed