

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN041819\
 Data File : BN005150.D
 Acq On : 18 Apr 2019 12:00
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV58

Quant Time: Apr 18 12:34:42 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 18 11:57:47 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
3	Naphthalene	1.105	1.098	0.6	89	0.00
4 SURR	2-Methylnaphthalene-d10	0.644	0.632	1.9	87	0.00
5	2-Methylnaphthalene	0.784	0.767	2.2	87	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
7	Acenaphthylene	2.006	1.941	3.2	84	0.00
8 C	Acenaphthene	1.501	1.491	0.7	85	0.00
9	Fluorene	1.738	1.711	1.6	84	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
11	Pentachlorophenol	0.097	0.088	9.3	87	0.00
12	Phenanthrene	1.243	1.228	1.2	84	0.00
13	Anthracene	1.136	1.108	2.5	84	0.00
14 SURR	Fluoranthene-d10	1.121	1.132	-1.0	87	0.00
15 C	Fluoranthene	1.406	1.404	0.1	86	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
17	Pyrene	1.723	1.573	8.7	86	0.00
18	Benzo(a)anthracene	1.526	1.467	3.9	92	0.00
19	Chrysene	1.568	1.552	1.0	94	0.00
20 I	Perylene-d12	1.000	1.000	0.0	102	0.00
21	Benzo(b)fluoranthene	1.541	1.482	3.8	100	0.00
22	Benzo(k)fluoranthene	1.515	1.507	0.5	102	0.00
23 C	Benzo(a)pyrene	1.367	1.360	0.5	102	0.00
24	Indeno(1,2,3-cd)pyrene	1.516	1.591	-4.9	109	0.00
25	Dibenzo(a,h)anthracene	1.241	1.321	-6.4	110	0.00
26	Benzo(g,h,i)perylene	1.293	1.380	-6.7	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0