

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041923\
 Data File : BN025017.D
 Acq On : 20 Apr 2023 00:37
 Operator : CG/JU
 Sample : 02296-08DL 5X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DCBL0DL

Quant Time: Apr 20 05:11:19 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 05:08:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.930	152	7272	0.400	ng/u1	0.00
4) Naphthalene-d8	10.748	136	23601	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.575	164	19279	0.400	ng/u1	# 0.00
13) Phenanthrene-d10	17.322	188	32916	0.400	ng/u1	0.00
17) Chrysene-d12	21.511	240	27193	0.400	ng/u1	0.00
23) Perylene-d12	23.952	264	23597	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.306	96	5770	0.727	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.332	152	2195	0.072	ng/u1	0.00
18) Fluoranthene-d10	19.349	212	5371	0.080	ng/u1	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.798	128	27775289	461.751	ng/u1#	94
7) 2-Methylnaphthalene	12.403	142	2293362	61.315	ng/u1	100
8) 1-Methylnaphthalene	12.623	142	4148281	104.613	ng/u1	98
10) Acenaphthylene	14.293	152	42749	0.591	ng/u1#	51
11) Acenaphthene	14.640	153	4088591	65.968	ng/u1	97
12) Fluorene	15.625	166	2148328	30.949	ng/u1	97
15) Phenanthrene	17.364	178	3421099	36.882	ng/u1	99
19) Fluoranthene	19.377	202	83975	0.897	ng/u1	98
20) Pyrene	19.740	202	34264	0.357	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041923\
Data File : BN025017.D
Acq On : 20 Apr 2023 00:37
Operator : CG/JU
Sample : 02296-08DL 5X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DCBL0DL

Quant Time: Apr 20 05:11:19 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 20 05:08:16 2023
Response via : Initial Calibration

