

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041923\
 Data File : BN024994.D
 Acq On : 19 Apr 2023 10:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4270

Quant Time: Apr 20 04:19:53 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 19 18:38:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	9412	0.400	ng/u1	0.00
4) Naphthalene-d8	10.743	136	29121	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.575	164	15730	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.326	188	33023	0.400	ng/u1	0.00
17) Chrysene-d12	21.520	240	24670	0.400	ng/u1	0.00
23) Perylene-d12	23.970	264	19177	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.310	96	4343	0.423	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.332	152	16213	0.434	ng/u1	0.00
18) Fluoranthene-d10	19.354	212	31825	0.519	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.343	88	4794	0.423	ng/u1	96
5) Naphthalene	10.792	128	32058	0.432	ng/u1	100
7) 2-Methylnaphthalene	12.403	142	19888	0.431	ng/u1	100
8) 1-Methylnaphthalene	12.623	142	21026	0.430	ng/u1	99
10) Acenaphthylene	14.297	152	25268	0.428	ng/u1	100
11) Acenaphthene	14.639	153	22092	0.437	ng/u1	98
12) Fluorene	15.625	166	24347	0.430	ng/u1	98
14) Pentachlorophenol	16.975	266	2616	0.371	ng/u1	99
15) Phenanthrene	17.364	178	39906	0.429	ng/u1	99
16) Anthracene	17.456	178	32151	0.416	ng/u1	99
19) Fluoranthene	19.387	202	44083	0.519	ng/u1	97
20) Pyrene	19.749	202	44368	0.510	ng/u1	96
21) Benzo(a)anthracene	21.505	228	33845	0.434	ng/u1	100
22) Chrysene	21.558	228	37829	0.449	ng/u1	100
24) Benzo(b)fluoranthene	23.218	252	33901	0.441	ng/u1	99
25) Benzo(k)fluoranthene	23.268	252	32956	0.439	ng/u1	99
26) Benzo(a)pyrene	23.864	252	26164	0.410	ng/u1	97
27) Indeno(1,2,3-cd)pyrene	26.521	276	31160	0.414	ng/u1#	98
28) Dibenzo(a,h)anthracene	26.544	278	24218	0.410	ng/u1	99
29) Benzo(g,h,i)perylene	27.299	276	27731	0.436	ng/u1	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041923\
Data File : BN024994.D
Acq On : 19 Apr 2023 10:19
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4270

Quant Time: Apr 20 04:19:53 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN040723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 19 18:38:47 2023
Response via : Initial Calibration

