

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN041923\  
 Data File : BN024997.D  
 Acq On : 19 Apr 2023 12:06  
 Operator : CG/JU  
 Sample : 02340-01  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 BHE02

Quant Time: Apr 20 04:20:19 2023  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN040723.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Apr 19 18:38:47 2023  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.934  | 152  | 8987     | 0.400 | ng/u1 | 0.00     |
| 4) Naphthalene-d8           | 10.743 | 136  | 27809    | 0.400 | ng/u1 | 0.00     |
| 9) Acenaphthene-d10         | 14.575 | 164  | 16774    | 0.400 | ng/u1 | 0.00     |
| 13) Phenanthrene-d10        | 17.322 | 188  | 34491    | 0.400 | ng/u1 | 0.00     |
| 17) Chrysene-d12            | 21.517 | 240  | 27691    | 0.400 | ng/u1 | 0.00     |
| 23) Perylene-d12            | 23.952 | 264  | 28144    | 0.400 | ng/u1 | #-0.02   |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.306  | 96   | 52516    | 5.351 | ng/u1 | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.332 | 152  | 12033    | 0.337 | ng/u1 | 0.00     |
| 18) Fluoranthene-d10        | 19.354 | 212  | 33387    | 0.485 | ng/u1 | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN041923\  
Data File : BN024997.D  
Acq On : 19 Apr 2023 12:06  
Operator : CG/JU  
Sample : 02340-01  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
BHE02

Quant Time: Apr 20 04:20:19 2023  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SFAM-EPA-SIM-BN040723.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Apr 19 18:38:47 2023  
Response via : Initial Calibration

