

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031044.D
 Acq On : 19 Apr 2024 16:58
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020368

Quant Time: Apr 19 22:26:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:25:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	111656	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	518078	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.287	164	362273	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.034	188	803288	20.000	ng/u1	0.00
79) Chrysene-d12	21.274	240	938293	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1264550	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	19559	8.213	ng/uL	0.00
4) Pyridine-d5	3.552	84	147450	20.650	ng/u1	0.00
7) Phenol-d5	6.811	99	195077	21.597	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	114929	20.601	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	155857	22.746	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	163754	21.632	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	81098	22.566	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	83929	25.528	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	163341	22.694	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	251203	21.200	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	528866	21.865	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	613219	22.050	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	110222	23.824	ng/u1	0.00
60) Fluorene-d10	15.281	176	471214	22.263	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	82172	23.175	ng/u1	0.00
73) Anthracene-d10	17.133	188	764214	22.080	ng/u1	0.00
81) Pyrene-d10	19.439	212	959065	19.012	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	1258576	21.600	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.175	88	20553	8.057	ng/uL	98
5) Pyridine	3.570	79	145605	20.209	ng/u1	97
6) Benzaldehyde	6.787	77	117888	28.026	ng/u1	99
8) Phenol	6.834	94	201462	21.410	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.069	93	161759	21.081	ng/u1	99
12) 2-Chlorophenol	7.199	128	163333	22.221	ng/u1	98
13) 2-Methylphenol	8.075	108	156241	21.548	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.158	45	212015	19.623	ng/u1	98
16) Acetophenone	8.457	105	262472	21.960	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.440	70	130652	21.704	ng/u1	98
18) 4-Methylphenol	8.405	108	174491	21.497	ng/u1	97
19) Hexachloroethane	8.705	117	66068	21.439	ng/u1	98
22) Nitrobenzene	8.834	77	191011	21.772	ng/u1	100
23) Isophorone	9.352	82	383283	21.907	ng/u1	99
25) 2-Nitrophenol	9.540	139	93905	24.303	ng/u1	96
26) 2,4-Dimethylphenol	9.599	107	187284	21.808	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.840	93	229253	20.609	ng/u1	99
29) 2,4-Dichlorophenol	10.069	162	161243	21.948	ng/u1	97
30) Naphthalene	10.463	128	570144	21.457	ng/u1	98
32) 4-Chloroaniline	10.587	127	246675	21.601	ng/u1	100
33) Hexachlorobutadiene	10.746	225	98695	21.343	ng/u1	96
34) Caprolactam	11.357	113	60828	22.163	ng/u1	95
35) 4-Chloro-3-methylphenol	11.710	107	188504	22.781	ng/u1	97
36) 2-Methylnaphthalene	12.087	142	397036	21.396	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031044.D
 Acq On : 19 Apr 2024 16:58
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020368

Quant Time: Apr 19 22:26:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:25:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	406201	21.564	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.457	216	199902	21.382	ng/ul	99
40) Hexachlorocyclopentadiene	12.434	237	104500	19.682	ng/ul	96
41) 2,4,6-Trichlorophenol	12.704	196	132166	22.623	ng/ul	92
42) 2,4,5-Trichlorophenol	12.775	196	148990	22.582	ng/ul	98
43) 1,1'-Biphenyl	13.110	154	531361	21.519	ng/ul	99
44) 2-Chloronaphthalene	13.151	162	419270	21.569	ng/ul	98
45) 2-Nitroaniline	13.369	65	118605	22.928	ng/ul	98
47) Dimethylphthalate	13.745	163	547105	21.897	ng/ul	100
48) 2,6-Dinitrotoluene	13.869	165	102921	22.217	ng/ul	96
50) Acenaphthylene	14.004	152	710205	22.103	ng/ul	99
51) 3-Nitroaniline	14.198	138	121154	24.640	ng/ul	96
52) Acenaphthene	14.345	153	465650	21.746	ng/ul	97
53) 2,4-Dinitrophenol	14.404	184	50358	23.668	ng/ul	96
55) 4-Nitrophenol	14.510	109	89506	24.419	ng/ul	95
56) Dibenzofuran	14.687	168	643923	21.633	ng/ul	99
57) 2,4-Dinitrotoluene	14.657	165	161395	23.863	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.910	232	129833	23.694	ng/ul	98
59) Diethylphthalate	15.122	149	573178	22.581	ng/ul	97
61) Fluorene	15.334	166	533500	22.550	ng/ul	94
62) 4-Chlorophenyl-phenyle...	15.334	204	255510	22.013	ng/ul	98
63) 4-Nitroaniline	15.369	138	140826	29.460	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.416	198	87520	22.972	ng/ul	98
67) N-Nitrosodiphenylamine	15.551	169	466935	21.344	ng/ul	100
68) 4-Bromophenyl-phenylether	16.234	248	157650	20.668	ng/ul	93
69) Hexachlorobenzene	16.334	284	186995	20.819	ng/ul	96
70) Atrazine	16.510	200	168817	22.264	ng/ul	97
71) Pentachlorophenol	16.681	266	121280	22.901	ng/ul	98
72) Phenanthrene	17.075	178	914427	22.157	ng/ul	98
74) Anthracene	17.169	178	929250	22.273	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.069	216	201544	20.423	ng/uL	99
76) Pentachlorobenzene	14.598	250	208455	20.796	ng/uL	96
77) Carbazole	17.445	167	886244	24.222	ng/ul	98
78) Di-n-butylphthalate	18.010	149	1055553	23.648	ng/ul	100
80) Fluoranthene	19.104	202	1146882	19.125	ng/ul	99
82) Pyrene	19.469	202	1204760	19.023	ng/ul	98
83) Butylbenzylphthalate	20.380	149	542126	22.415	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.186	252	446992	25.926	ng/ul	94
85) Benzo(a)anthracene	21.257	228	1376838	21.987	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.186	149	832493	23.340	ng/ul	99
87) Chrysene	21.316	228	1286524	21.767	ng/ul	97
89) Di-n-octyl phthalate	22.286	149	1585325	19.797	ng/ul	100
90) Benzo(b)fluoranthene	23.257	252	1529347	20.550	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	1480524	19.589	ng/ul	99
93) Benzo(a)pyrene	24.039	252	1486956	21.151	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.433	276	1852699	24.379	ng/ul	99
95) Dibenzo(a,h)anthracene	27.492	278	1545110	24.152	ng/ul	98
96) Benzo(g,h,i)perylene	28.456	276	1464792	24.552	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 8.057 ng/uL
RT: 3.175 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:	88	Resp:	20553
Ion	Ratio	Lower	Upper
88	100		
43	27.7	22.2	33.2
58	65.2	50.2	75.2



#5
Pyridine
Concen: 20.209 ng/uL
RT: 3.570 min Scan# 99
Delta R.T. -0.006 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:	79	Resp:	145605
Ion	Ratio	Lower	Upper
79	100		
52	76.7	59.9	89.9
51	33.9	29.0	43.6









#6
Benzaldehyde
Concen: 28.026 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion: 77 Resp: 117888
Ion Ratio Lower Upper
77 100
105 98.9 80.0 120.0
106 94.2 75.0 112.6



#8
Phenol
Concen: 21.410 ng/ul
RT: 6.834 min Scan# 654
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion: 94 Resp: 201462
Ion Ratio Lower Upper
94 100
65 26.5 22.0 33.0
66 35.1 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.081 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion: 93 Resp: 161759
Ion Ratio Lower Upper
93 100
63 70.6 57.1 85.7
95 32.2 24.9 37.3



#12
2-Chlorophenol
Concen: 22.221 ng/ul
RT: 7.199 min Scan# 716
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:128 Resp: 163333
Ion Ratio Lower Upper
128 100
64 42.2 34.9 52.3
130 32.7 27.0 40.4





#13

2-Methylphenol

Concen: 21.548 ng/ul

RT: 8.075 min Scan# 80

Delta R.T. -0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :





SSTD020368

Tgt Ion:108 Resp: 156241

Ion Ratio Lower Upper

108 100

107 88.6 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.623 ng/ul

RT: 8.158 min Scan# 879

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion: 45 Resp: 212015


Ion Ratio Lower Upper

45 100

77 15.1 12.5 18.7

79 11.2 9.6 14.4









#16
Acetophenone
Concen: 21.960 ng/ul
RT: 8.457 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:105 Resp: 262472
Ion Ratio Lower Upper
105 100
77 74.9 60.2 90.4
51 25.0 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.704 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion: 70 Resp: 130652
Ion Ratio Lower Upper
70 100
42 57.0 45.0 67.6
101 10.5 8.1 12.1
130 22.4 19.2 28.8









#18
4-Methylphenol
Concen: 21.497 ng/ul
RT: 8.405 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:108 Resp: 174491
Ion Ratio Lower Upper
108 100
107 115.7 89.8 134.8



#19
Hexachloroethane
Concen: 21.439 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:117 Resp: 66068
Ion Ratio Lower Upper
117 100
201 99.3 77.4 116.0
199 62.0 49.6 74.4









#22
Nitrobenzene
Concen: 21.772 ng/ul
RT: 8.834 min Scan# 994
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion: 77 Resp: 191011
Ion Ratio Lower Upper
77 100
123 47.7 38.2 57.2
65 13.2 10.7 16.1



#23
Isophorone
Concen: 21.907 ng/ul
RT: 9.352 min Scan# 1082
Delta R.T. -0.006 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion: 82 Resp: 383283
Ion Ratio Lower Upper
82 100
95 6.9 5.9 8.9
138 17.8 14.7 22.1





#25

2-Nitrophenol

Concen: 24.303 ng/ul

RT: 9.540 min Scan# 1114

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:139 Resp: 93905

Ion Ratio Lower Upper

139 100

65 47.2 35.2 52.8

109 30.8 25.4 38.0



#26

2,4-Dimethylphenol

Concen: 21.808 ng/ul

RT: 9.599 min Scan# 1124

Delta R.T. -0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:107 Resp: 187284


Ion Ratio Lower Upper

107 100

121 51.9 40.3 60.5

122 85.6 69.8 104.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.609 ng/ul
 RT: 9.840 min Scan# 1165
 Delta R.T. -0.000 min
 Lab File: BN031044.D
 Acq: 19 Apr 2024 16:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020368


Tgt Ion: 93 Resp: 229253
 Ion Ratio Lower Upper
 93 100
 95 32.9 25.8 38.6
 123 12.1 9.8 14.6



#29
 2,4-Dichlorophenol
 Concen: 21.948 ng/ul
 RT: 10.069 min Scan# 1204
 Delta R.T. -0.000 min
 Lab File: BN031044.D
 Acq: 19 Apr 2024 16:58

Tgt Ion: 162 Resp: 161243
 Ion Ratio Lower Upper
 162 100
 164 67.0 51.0 76.6
 98 35.8 28.6 42.8





#30

Naphthalene

Concen: 21.457 ng/ul

RT: 10.463 min Scan# 1271

Delta R.T. -0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:128 Resp: 570144

Ion Ratio Lower Upper

128 100

129 10.2 9.0 13.4

127 12.8 10.5 15.7



#32

4-Chloroaniline

Concen: 21.601 ng/ul

RT: 10.587 min Scan# 1292

Delta R.T. -0.000 min

Lab File: BN031044.D




Acq: 19 Apr 2024 16:58


Tgt Ion:127 Resp: 246675

Ion Ratio Lower Upper

127 100

129 31.9 25.4 38.2









#33
Hexachlorobutadiene
Concen: 21.343 ng/ul
RT: 10.746 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:225 Resp: 98695
Ion Ratio Lower Upper
225 100
223 61.6 49.8 74.6
227 66.9 48.8 73.2



#34
Caprolactam
Concen: 22.163 ng/ul
RT: 11.357 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:113 Resp: 60828
Ion Ratio Lower Upper
113 100
55 149.1 112.4 168.6
56 111.4 88.0 132.0









#35
4-Chloro-3-methylphenol
Concen: 22.781 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:107 Resp: 188504
Ion Ratio Lower Upper
107 100
144 25.1 20.6 31.0
142 81.0 67.4 101.2



#36
2-Methylnaphthalene
Concen: 21.396 ng/ul
RT: 12.087 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:142 Resp: 397036
Ion Ratio Lower Upper
142 100
141 87.4 69.6 104.4





#37

1-Methylnaphthalene

Concen: 21.564 ng/ul

RT: 12.310 min Scan# 1585

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:142 Resp: 406201

Ion Ratio Lower Upper

142 100

141 88.5 70.2 105.4

116 3.6 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.382 ng/ul

RT: 12.457 min Scan# 1610

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Tgt Ion:216 Resp: 199902




Ion Ratio Lower Upper


216 100

214 77.3 63.1 94.7

179 21.3 17.0 25.6

108 17.1 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 19.682 ng/ul

RT: 12.434 min Scan# 1606

Delta R.T. -0.000 min

Lab File: BN031044.D


Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368






Tgt Ion:237 Resp: 104500

Ion Ratio Lower Upper

237 100

235 63.2 47.6 71.4

272 14.5 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 22.623 ng/ul

RT: 12.704 min Scan# 1652

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:196 Resp: 132166


Ion Ratio Lower Upper

196 100

198 97.2 69.9 104.9


200 29.5 23.4 35.2








#42
2,4,5-Trichlorophenol
Concen: 22.582 ng/ul
RT: 12.775 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Instrument :
BNA_N
ClientSampleId :
SSTD020368







Tgt Ion:196 Resp: 148990
Ion Ratio Lower Upper
196 100
198 96.8 75.9 113.9
200 31.2 24.7 37.1



#43
1,1'-Biphenyl
Concen: 21.519 ng/ul
RT: 13.110 min Scan# 1721
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:154 Resp: 531361
Ion Ratio Lower Upper
154 100
153 39.5 31.8 47.8
76 12.6 10.6 15.8





#44

2-Chloronaphthalene

Concen: 21.569 ng/ul

RT: 13.151 min Scan# 1728

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:162 Resp: 419270

Ion Ratio Lower Upper

162 100

164 32.3 25.8 38.8

127 40.1 33.7 50.5



#45

2-Nitroaniline

Concen: 22.928 ng/ul

RT: 13.369 min Scan# 1765

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion: 65 Resp: 118605


Ion Ratio Lower Upper

65 100

92 71.8 57.4 86.2

138 111.7 91.6 137.4









#47
Dimethylphthalate
Concen: 21.897 ng/ul
RT: 13.745 min Scan# 1829
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:163 Resp: 547105
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 22.217 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:165 Resp: 102921
Ion Ratio Lower Upper
165 100
89 59.8 44.7 67.1
121 21.9 17.0 25.4









#50
Acenaphthylene
Concen: 22.103 ng/ul
RT: 14.004 min Scan# 1873
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument : BNA_N
ClientSampleId : SST020368


Tgt Ion:	152	Resp:	710205
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.8	23.8
153	12.7	10.5	15.7



#51
3-Nitroaniline
Concen: 24.640 ng/ul
RT: 14.198 min Scan# 1906
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:	138	Resp:	121154
Ion Ratio	Lower	Upper	
138	100		
108	11.8	11.1	16.7
92	113.5	93.9	140.9









#52
Acenaphthene
Concen: 21.746 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:153 Resp: 465650
Ion Ratio Lower Upper
153 100
152 47.0 36.3 54.5
154 87.1 72.4 108.6



#53
2,4-Dinitrophenol
Concen: 23.668 ng/ul
RT: 14.404 min Scan# 1941
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:184 Resp: 50358
Ion Ratio Lower Upper
184 100
63 68.8 50.4 75.6
154 65.2 53.0 79.6









#55
4-Nitrophenol
Concen: 24.419 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:109 Resp: 89506
Ion Ratio Lower Upper
109 100
139 139.8 105.6 158.4
65 125.3 98.2 147.2



#56
Dibenzofuran
Concen: 21.633 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:168 Resp: 643923
Ion Ratio Lower Upper
168 100
139 38.1 31.0 46.6





#57

2,4-Dinitrotoluene

Concen: 23.863 ng/ul

RT: 14.657 min Scan# 1984

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:165 Resp: 161395

Ion Ratio Lower Upper

165 100

63 42.3 33.9 50.9

182 2.9 2.6 4.0



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.694 ng/ul

RT: 14.910 min Scan# 2027

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Tgt Ion:232 Resp: 129833




Ion Ratio Lower Upper


232 100

131 53.4 41.2 61.8

130 2.8 2.2 3.2

166 34.5 27.3 40.9





#59

Diethylphthalate

Concen: 22.581 ng/ul

RT: 15.122 min Scan# 2063

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:149 Resp: 573178

Ion Ratio Lower Upper

149 100

177 22.2 16.8 25.2

150 11.5 10.1 15.1



#61

Fluorene

Concen: 22.550 ng/ul

RT: 15.334 min Scan# 2099

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:166 Resp: 533500


Ion Ratio Lower Upper

166 100

165 99.9 74.6 112.0

167 13.9 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 22.013 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:204 Resp: 255510
Ion Ratio Lower Upper
204 100
206 32.6 26.4 39.6
141 64.9 50.7 76.1



#63
4-Nitroaniline
Concen: 29.460 ng/ul
RT: 15.369 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:138 Resp: 140826
Ion Ratio Lower Upper
138 100
92 47.6 38.4 57.6
108 77.6 63.8 95.8









#66
4,6-Dinitro-2-methylphenol
Concen: 22.972 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:198 Resp: 87520
Ion Ratio Lower Upper
198 100
182 4.7 4.0 6.0
77 28.8 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 21.344 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:169 Resp: 466935
Ion Ratio Lower Upper
169 100
168 65.5 52.6 79.0
167 34.4 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 20.668 ng/ul
RT: 16.234 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:248 Resp: 157650
Ion Ratio Lower Upper
248 100
250 101.5 78.2 117.2
141 79.7 71.4 107.0



#69
Hexachlorobenzene
Concen: 20.819 ng/ul
RT: 16.334 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:284 Resp: 186995
Ion Ratio Lower Upper
284 100
142 39.7 30.0 45.0
249 33.9 25.5 38.3









#70
 Atrazine
 Concen: 22.264 ng/ul
 RT: 16.510 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN031044.D
 Acq: 19 Apr 2024 16:58




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020368


Tgt Ion:	200	Resp:	168817
Ion Ratio	Lower	Upper	
200	100		
173	28.2	20.6	31.0
215	45.3	37.6	56.4



#71
 Pentachlorophenol
 Concen: 22.901 ng/ul
 RT: 16.681 min Scan# 2328
 Delta R.T. -0.000 min
 Lab File: BN031044.D
 Acq: 19 Apr 2024 16:58

Tgt Ion:	266	Resp:	121280
Ion Ratio	Lower	Upper	
266	100		
264	58.5	49.1	73.7
268	62.4	49.2	73.8









#72
 Phenanthrene
 Concen: 22.157 ng/ul
 RT: 17.075 min Scan# 2395
 Delta R.T. -0.000 min
 Lab File: BN031044.D
 Acq: 19 Apr 2024 16:58

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020368

Tgt Ion:178 Resp: 914427




Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.4	18.6
176	19.6	14.6	21.8




#74
 Anthracene
 Concen: 22.273 ng/ul
 RT: 17.169 min Scan# 2411
 Delta R.T. -0.000 min
 Lab File: BN031044.D
 Acq: 19 Apr 2024 16:58

Tgt Ion:178 Resp: 929250

Ion	Ratio	Lower	Upper
178	100		
179	15.7	12.2	18.4
176	18.3	14.4	21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.423 ng/uL

RT: 13.069 min Scan# 1714

Delta R.T. -0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:216 Resp: 201544

Ion Ratio Lower Upper

216 100

214 80.0 63.0 94.6

218 48.2 38.3 57.5



#76

Pentachlorobenzene

Concen: 20.796 ng/uL

RT: 14.598 min Scan# 1974

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:250 Resp: 208455


Ion Ratio Lower Upper

250 100

248 62.6 49.3 73.9

252 60.3 52.9 79.3





#77

Carbazole

Concen: 24.222 ng/ul

RT: 17.445 min Scan# 2458

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:167 Resp: 886244

Ion Ratio Lower Upper

167 100

166 21.2 16.2 24.2

139 12.9 10.6 15.8



#78

Di-n-butylphthalate

Concen: 23.648 ng/ul

RT: 18.010 min Scan# 2554

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:149 Resp: 1055553


Ion Ratio Lower Upper

149 100

150 8.9 7.1 10.7

104 5.3 3.9 5.9









#80
Fluoranthene
Concen: 19.125 ng/ul
RT: 19.104 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:202 Resp: 1146882
Ion Ratio Lower Upper
202 100
101 10.9 9.0 13.6
100 8.2 6.9 10.3



#82
Pyrene
Concen: 19.023 ng/ul
RT: 19.469 min Scan# 2802
Delta R.T. 0.006 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:202 Resp: 1204760
Ion Ratio Lower Upper
202 100
101 13.0 11.1 16.7
100 10.8 9.3 13.9





#83

Butylbenzylphthalate

Concen: 22.415 ng/ul

RT: 20.380 min Scan# 2956

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:149 Resp: 542126

Ion Ratio Lower Upper

149 100

91 68.0 57.3 85.9

206 20.1 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 25.926 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. -0.000 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:252 Resp: 446992


Ion Ratio Lower Upper

252 100

254 62.5 54.5 81.7

126 13.1 9.8 14.6





#85

Benzo(a)anthracene

Concen: 21.987 ng/ul

RT: 21.257 min Scan# 3105

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:228 Resp: 1376838

Ion Ratio Lower Upper

228 100

229 19.1 16.7 25.1

226 27.4 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.340 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:149 Resp: 832493


Ion Ratio Lower Upper

149 100

167 28.5 22.3 33.5

279 4.5 3.0 4.6





#87

Chrysene

Concen: 21.767 ng/ul

RT: 21.316 min Scan# 3116

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:228 Resp: 1286524

Ion Ratio Lower Upper

228 100

226 31.2 22.6 34.0

229 20.0 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 19.797 ng/ul




RT: 22.286 min Scan# 3281


Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Tgt Ion:149 Resp: 1585325









#90
Benzo(b)fluoranthene
Concen: 20.550 ng/ul
RT: 23.257 min Scan# 3446
Delta R.T. -0.000 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58




Instrument :
BNA_N
ClientSampleId :
SSTD020368


Tgt Ion:252 Resp: 1529347
Ion Ratio Lower Upper
252 100
253 21.6 17.6 26.4
125 11.0 8.6 12.8



#91
Benzo(k)fluoranthene
Concen: 19.589 ng/ul
RT: 23.321 min Scan# 3457
Delta R.T. 0.012 min
Lab File: BN031044.D
Acq: 19 Apr 2024 16:58

Tgt Ion:252 Resp: 1480524
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 10.8 9.4 14.2





#93

Benzo(a)pyrene

Concen: 21.151 ng/ul

RT: 24.039 min Scan# 3

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:252 Resp: 1486956

Ion Ratio Lower Upper

252 100

253 21.4 17.7 26.5

125 11.7 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 24.379 ng/ul

RT: 27.433 min Scan# 4156

Delta R.T. 0.012 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58




Tgt Ion:276 Resp: 1852699


Ion Ratio Lower Upper

276 100

138 22.4 17.4 26.2

277 22.7 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 24.152 ng/ul

RT: 27.492 min Scan# 4166

Delta R.T. 0.006 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Instrument :

BNA_N

ClientSampleId :

SSTD020368





Tgt Ion:278 Resp: 1545110

Ion Ratio Lower Upper

278 100

139 17.7 13.3 19.9

279 22.3 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 24.552 ng/ul

RT: 28.456 min Scan# 4330

Delta R.T. 0.018 min

Lab File: BN031044.D

Acq: 19 Apr 2024 16:58

Tgt Ion:276 Resp: 1464792

Ion Ratio Lower Upper

276 100

138 24.2 19.0 28.6

277 24.4 19.3 28.9

