

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031046.D
 Acq On : 19 Apr 2024 18:24
 Operator : MA/JU
 Sample : SSTD01039
 Misc :
 ALS Vial : 93 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218

Quant Time: Apr 19 22:27:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:25:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	99524	20.000	ng/ul	0.00
20) Naphthalene-d8	10.416	136	468206	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.281	164	336797	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.034	188	738579	20.000	ng/ul	0.00
79) Chrysene-d12	21.269	240	846152	20.000	ng/ul	0.00
88) Perylene-d12	24.169	264	1095169	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	8381	3.948	ng/uL	0.00
4) Pyridine-d5	3.558	84	53227	8.363	ng/ul	0.00
7) Phenol-d5	6.811	99	74382	9.239	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	47497	9.552	ng/ul	0.00
11) 2-Chlorophenol-d4	7.169	132	62261	10.194	ng/ul	0.00
15) 4-Methylphenol-d8	8.340	113	64311	9.531	ng/ul	0.00
21) Nitrobenzene-d5	8.793	128	31717	9.765	ng/ul	0.00
24) 2-Nitrophenol-d4	9.510	143	31770	10.693	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	65871	10.127	ng/ul	0.00
31) 4-Chloroaniline-d4	10.563	131	105106	9.815	ng/ul	0.00
46) Dimethylphthalate-d6	13.698	166	227976	10.138	ng/ul	0.00
49) Acenaphthylene-d8	13.975	160	258458	9.997	ng/ul	0.00
54) 4-Nitrophenol-d4	14.493	143	41157	9.569	ng/ul	0.00
60) Fluorene-d10	15.281	176	200445	10.187	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	28903	8.866	ng/ul	0.00
73) Anthracene-d10	17.134	188	319594	10.043	ng/ul	0.00
81) Pyrene-d10	19.433	212	414583	9.113	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.969	264	519976	10.304	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.176	88	8885	3.907	ng/uL	97
5) Pyridine	3.575	79	55268	8.606	ng/ul	96
6) Benzaldehyde	6.787	77	43164	11.513	ng/ul	96
8) Phenol	6.840	94	79845	9.520	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.069	93	66676	9.749	ng/ul	98
12) 2-Chlorophenol	7.199	128	65484	9.995	ng/ul	98
13) 2-Methylphenol	8.081	108	61131	9.459	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.163	45	88931	9.234	ng/ul	98
16) Acetophenone	8.458	105	109057	10.237	ng/ul	96
17) N-Nitroso-di-n-propyla...	8.440	70	53032	9.884	ng/ul	98
18) 4-Methylphenol	8.405	108	70493	9.743	ng/ul	98
19) Hexachloroethane	8.705	117	27744	10.100	ng/ul	97
22) Nitrobenzene	8.834	77	75317	9.499	ng/ul	98
23) Isophorone	9.358	82	156296	9.885	ng/ul	100
25) 2-Nitrophenol	9.540	139	34919	10.000	ng/ul	94
26) 2,4-Dimethylphenol	9.599	107	76238	9.823	ng/ul	97
27) Bis(2-Chloroethoxy)met...	9.840	93	95788	9.528	ng/ul	99
29) 2,4-Dichlorophenol	10.069	162	65568	9.876	ng/ul	96
30) Naphthalene	10.469	128	241241	10.046	ng/ul	99
32) 4-Chloroaniline	10.587	127	101758	9.860	ng/ul	94
33) Hexachlorobutadiene	10.746	225	41704	9.979	ng/ul	95
34) Caprolactam	11.357	113	23171	9.342	ng/ul	93
35) 4-Chloro-3-methylphenol	11.710	107	76516	10.232	ng/ul	94
36) 2-Methylnaphthalene	12.087	142	168726	10.061	ng/ul	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031046.D
 Acq On : 19 Apr 2024 18:24
 Operator : MA/JU
 Sample : SSTD01039
 Misc :
 ALS Vial : 93 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218

Quant Time: Apr 19 22:27:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:25:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	172410	10.128	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.457	216	85285	9.812	ng/ul	99
40) Hexachlorocyclopentadiene	12.434	237	41400	8.387	ng/ul	99
41) 2,4,6-Trichlorophenol	12.704	196	55090	10.143	ng/ul	96
42) 2,4,5-Trichlorophenol	12.775	196	62373	10.169	ng/ul	99
43) 1,1'-Biphenyl	13.110	154	232248	10.117	ng/ul	98
44) 2-Chloronaphthalene	13.151	162	182164	10.080	ng/ul	99
45) 2-Nitroaniline	13.369	65	46042	9.574	ng/ul	95
47) Dimethylphthalate	13.746	163	235314	10.130	ng/ul	99
48) 2,6-Dinitrotoluene	13.869	165	39366	9.141	ng/ul	96
50) Acenaphthylene	14.004	152	307932	10.308	ng/ul	99
51) 3-Nitroaniline	14.198	138	44085	9.644	ng/ul	94
52) Acenaphthene	14.345	153	200191	10.056	ng/ul	95
53) 2,4-Dinitrophenol	14.404	184	15974	8.075	ng/ul	91
55) 4-Nitrophenol	14.510	109	34973	10.263	ng/ul	93
56) Dibenzofuran	14.687	168	280914	10.152	ng/ul	99
57) 2,4-Dinitrotoluene	14.657	165	61398	9.765	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	14.910	232	52319	10.270	ng/ul	98
59) Diethylphthalate	15.116	149	242111	10.260	ng/ul	98
61) Fluorene	15.334	166	238798	10.857	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.334	204	114577	10.618	ng/ul	98
63) 4-Nitroaniline	15.363	138	52114	11.727	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.416	198	31733	9.059	ng/ul#	96
67) N-Nitrosodiphenylamine	15.551	169	202941	10.089	ng/ul	100
68) 4-Bromophenyl-phenylether	16.228	248	70298	10.023	ng/ul	96
69) Hexachlorobenzene	16.334	284	81493	9.868	ng/ul	98
70) Atrazine	16.504	200	68653	9.847	ng/ul	96
71) Pentachlorophenol	16.681	266	46969	9.646	ng/ul	96
72) Phenanthrene	17.075	178	403806	10.642	ng/ul	99
74) Anthracene	17.169	178	399810	10.422	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.069	216	87827	9.679	ng/uL	97
76) Pentachlorobenzene	14.598	250	92732	10.062	ng/uL	94
77) Carbazole	17.439	167	377815	11.231	ng/ul	98
78) Di-n-butylphthalate	18.010	149	426806	10.400	ng/ul	99
80) Fluoranthene	19.098	202	479598	8.868	ng/ul	98
82) Pyrene	19.463	202	528546	9.254	ng/ul	98
83) Butylbenzylphthalate	20.375	149	214702	9.844	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.186	252	178979	11.512	ng/ul	97
85) Benzo(a)anthracene	21.251	228	571321	10.117	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.180	149	336466	10.461	ng/ul	98
87) Chrysene	21.310	228	554015	10.394	ng/ul	99
89) Di-n-octyl phthalate	22.280	149	607569	8.760	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	622640	9.660	ng/ul	98
91) Benzo(k)fluoranthene	23.310	252	654658	10.002	ng/ul	99
93) Benzo(a)pyrene	24.027	252	621093	10.201	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.421	276	758689	11.528	ng/ul	99
95) Dibenzo(a,h)anthracene	27.480	278	622017	11.226	ng/ul	97
96) Benzo(g,h,i)perylene	28.433	276	607699	11.761	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 3.907 ng/uL
RT: 3.176 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion: 88	Resp: 8885
Ion Ratio	Lower Upper
88 100	
43 31.9	22.2 33.2
58 62.8	50.2 75.2



#5
Pyridine
Concen: 8.606 ng/uL
RT: 3.575 min Scan# 100
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion: 79	Resp: 55268
Ion Ratio	Lower Upper
79 100	
52 78.4	59.9 89.9
51 34.8	29.0 43.6





#6

Benzaldehyde

Concen: 11.513 ng/ul

RT: 6.787 min Scan# 646

Delta R.T. 0.000 min

Lab File: BN031046.D


Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218






Tgt Ion: 77 Resp: 43164

Ion Ratio Lower Upper

77 100

105 101.2 80.0 120.0

106 99.8 75.0 112.6



#8

Phenol

Concen: 9.520 ng/ul

RT: 6.840 min Scan# 655

Delta R.T. 0.006 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion: 94 Resp: 79845


Ion Ratio Lower Upper

94 100

65 26.1 22.0 33.0

66 36.3 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 9.749 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion: 93 Resp: 66676
Ion Ratio Lower Upper
93 100
63 71.9 57.1 85.7
95 33.3 24.9 37.3



#12
2-Chlorophenol
Concen: 9.995 ng/ul
RT: 7.199 min Scan# 716
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:128 Resp: 65484
Ion Ratio Lower Upper
128 100
64 44.9 34.9 52.3
130 32.8 27.0 40.4





#13

2-Methylphenol

Concen: 9.459 ng/ul

RT: 8.081 min Scan# 80

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :





SSTD010218

Tgt Ion:108 Resp: 61131

Ion Ratio Lower Upper

108 100

107 89.3 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 9.234 ng/ul

RT: 8.163 min Scan# 880

Delta R.T. 0.006 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion: 45 Resp: 88931


Ion Ratio Lower Upper

45 100

77 14.8 12.5 18.7

79 11.1 9.6 14.4









#16
Acetophenone
Concen: 10.237 ng/ul
RT: 8.458 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:105 Resp: 109057
Ion Ratio Lower Upper
105 100
77 79.2 60.2 90.4
51 25.6 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 9.884 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion: 70 Resp: 53032
Ion Ratio Lower Upper
70 100
42 55.3 45.0 67.6
101 10.5 8.1 12.1
130 22.8 19.2 28.8









#18
4-Methylphenol
Concen: 9.743 ng/ul
RT: 8.405 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:108 Resp: 70493
Ion Ratio Lower Upper
108 100
107 110.5 89.8 134.8



#19
Hexachloroethane
Concen: 10.100 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:117 Resp: 27744
Ion Ratio Lower Upper
117 100
201 95.8 77.4 116.0
199 56.8 49.6 74.4









#22
Nitrobenzene
Concen: 9.499 ng/ul
RT: 8.834 min Scan# 994
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion: 77 Resp: 75317
Ion Ratio Lower Upper
77 100
123 46.0 38.2 57.2
65 13.0 10.7 16.1



#23
Isophorone
Concen: 9.885 ng/ul
RT: 9.358 min Scan# 1083
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion: 82 Resp: 156296
Ion Ratio Lower Upper
82 100
95 7.2 5.9 8.9
138 18.4 14.7 22.1









#25
2-Nitrophenol
Concen: 10.000 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:139 Resp: 34919
Ion Ratio Lower Upper
139 100
65 50.6 35.2 52.8
109 32.3 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 9.823 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:107 Resp: 76238
Ion Ratio Lower Upper
107 100
121 52.7 40.3 60.5
122 89.5 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 9.528 ng/ul

RT: 9.840 min Scan# 1165

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion: 93 Resp: 95788

Ion Ratio Lower Upper

93 100

95 32.5 25.8 38.6

123 11.5 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 9.876 ng/ul

RT: 10.069 min Scan# 1204

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion: 162 Resp: 65568


Ion Ratio Lower Upper

162 100

164 60.6 51.0 76.6

98 34.0 28.6 42.8









#30
Naphthalene
Concen: 10.046 ng/ul
RT: 10.469 min Scan# 1272
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:128 Resp: 241241
Ion Ratio Lower Upper
128 100
129 11.8 9.0 13.4
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 9.860 ng/ul
RT: 10.587 min Scan# 1292
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:127 Resp: 101758
Ion Ratio Lower Upper
127 100
129 35.1 25.4 38.2





#33

Hexachlorobutadiene

Concen: 9.979 ng/ul

RT: 10.746 min Scan# 1319

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:225 Resp: 41704

Ion Ratio Lower Upper

225 100

223 64.4 49.8 74.6

227 66.0 48.8 73.2



#34

Caprolactam

Concen: 9.342 ng/ul

RT: 11.357 min Scan# 1423

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:113 Resp: 23171


Ion Ratio Lower Upper

113 100

55 154.1 112.4 168.6

56 112.2 88.0 132.0









#35
 4-Chloro-3-methylphenol
 Concen: 10.232 ng/ul
 RT: 11.710 min Scan# 1483
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SST010218


Tgt Ion:107 Resp: 76516
 Ion Ratio Lower Upper
 107 100
 144 26.9 20.6 31.0
 142 77.6 67.4 101.2



#36
 2-Methylnaphthalene
 Concen: 10.061 ng/ul
 RT: 12.087 min Scan# 1547
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24

Tgt Ion:142 Resp: 168726
 Ion Ratio Lower Upper
 142 100
 141 87.4 69.6 104.4









#37
1-Methylnaphthalene
Concen: 10.128 ng/ul
RT: 12.310 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:142 Resp: 172410
Ion Ratio Lower Upper
142 100
141 93.5 70.2 105.4
116 3.7 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.812 ng/ul
RT: 12.457 min Scan# 1610
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:216 Resp: 85285
Ion Ratio Lower Upper
216 100
214 77.9 63.1 94.7
179 21.9 17.0 25.6
108 18.3 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 8.387 ng/ul

RT: 12.434 min Scan# 1606

Delta R.T. 0.000 min

Lab File: BN031046.D


Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218






Tgt Ion:237 Resp: 41400

Ion Ratio Lower Upper

237 100

235 59.0 47.6 71.4

272 13.7 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 10.143 ng/ul

RT: 12.704 min Scan# 1652

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:196 Resp: 55090


Ion Ratio Lower Upper

196 100

198 91.1 69.9 104.9

200 28.2 23.4 35.2





#42

2,4,5-Trichlorophenol

Concen: 10.169 ng/ul

RT: 12.775 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:196 Resp: 62373

Ion Ratio Lower Upper

196 100

198 94.1 75.9 113.9

200 29.8 24.7 37.1



#43

1,1'-Biphenyl

Concen: 10.117 ng/ul

RT: 13.110 min Scan# 1721

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:154 Resp: 232248


Ion Ratio Lower Upper

154 100

153 38.9 31.8 47.8

76 12.2 10.6 15.8





#44

2-Chloronaphthalene

Concen: 10.080 ng/ul

RT: 13.151 min Scan# 1728

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:162 Resp: 182164

Ion Ratio Lower Upper

162 100

164 31.7 25.8 38.8

127 41.1 33.7 50.5



#45

2-Nitroaniline

Concen: 9.574 ng/ul

RT: 13.369 min Scan# 1765

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion: 65 Resp: 46042


Ion Ratio Lower Upper

65 100

92 70.8 57.4 86.2

138 107.2 91.6 137.4









#47
Dimethylphthalate
Concen: 10.130 ng/ul
RT: 13.746 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:163 Resp: 235314
Ion Ratio Lower Upper
163 100
194 4.2 3.0 4.6
164 9.8 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 9.141 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:165 Resp: 39366
Ion Ratio Lower Upper
165 100
89 58.6 44.7 67.1
121 24.2 17.0 25.4









#50
Acenaphthylene
Concen: 10.308 ng/ul
RT: 14.004 min Scan# 1873
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:152 Resp: 307932
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 12.4 10.5 15.7



#51
3-Nitroaniline
Concen: 9.644 ng/ul
RT: 14.198 min Scan# 1906
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:138 Resp: 44085
Ion Ratio Lower Upper
138 100
108 13.8 11.1 16.7
92 109.9 93.9 140.9









#52
Acenaphthene
Concen: 10.056 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:153 Resp: 200191
Ion Ratio Lower Upper
153 100
152 50.8 36.3 54.5
154 93.6 72.4 108.6



#53
2,4-Dinitrophenol
Concen: 8.075 ng/ul
RT: 14.404 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:184 Resp: 15974
Ion Ratio Lower Upper
184 100
63 69.3 50.4 75.6
154 74.5 53.0 79.6









#55
 4-Nitrophenol
 Concen: 10.263 ng/ul
 RT: 14.510 min Scan# 1959
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218


Tgt Ion:	109	Resp:	34973
Ion	Ratio	Lower	Upper
109	100		
139	121.9	105.6	158.4
65	115.9	98.2	147.2



#56
 Dibenzofuran
 Concen: 10.152 ng/ul
 RT: 14.687 min Scan# 1989
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24

Tgt Ion:	168	Resp:	280914
Ion	Ratio	Lower	Upper
168	100		
139	38.5	31.0	46.6









#57
2,4-Dinitrotoluene
Concen: 9.765 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:165 Resp: 61398
Ion Ratio Lower Upper
165 100
63 41.6 33.9 50.9
182 2.2 2.6 4.0#



#58
2,3,4,6-Tetrachlorophenol
Concen: 10.270 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:232 Resp: 52319
Ion Ratio Lower Upper
232 100
131 50.6 41.2 61.8
130 2.8 2.2 3.2
166 32.8 27.3 40.9









#59
 Diethylphthalate
 Concen: 10.260 ng/ul
 RT: 15.116 min Scan# 2062
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010218

Tgt Ion:149 Resp: 242111




Ion	Ratio	Lower	Upper
149	100		
177	22.4	16.8	25.2
150	12.4	10.1	15.1




#61
 Fluorene
 Concen: 10.857 ng/ul
 RT: 15.334 min Scan# 2099
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24

Tgt Ion:166 Resp: 238798

Ion	Ratio	Lower	Upper
166	100		
165	95.6	74.6	112.0
167	13.9	10.6	16.0









#62
4-Chlorophenyl-phenylether
Concen: 10.618 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:204 Resp: 114577
Ion Ratio Lower Upper
204 100
206 32.5 26.4 39.6
141 61.7 50.7 76.1



#63
4-Nitroaniline
Concen: 11.727 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:138 Resp: 52114
Ion Ratio Lower Upper
138 100
92 56.6 38.4 57.6
108 84.8 63.8 95.8









#66
4,6-Dinitro-2-methylphenol
Concen: 9.059 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:198 Resp: 31733
Ion Ratio Lower Upper
198 100
182 6.2 4.0 6.0#
77 30.1 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 10.089 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:169 Resp: 202941
Ion Ratio Lower Upper
169 100
168 65.9 52.6 79.0
167 34.8 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 10.023 ng/ul
RT: 16.228 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:248 Resp: 70298
Ion Ratio Lower Upper
248 100
250 97.1 78.2 117.2
141 81.7 71.4 107.0



#69
Hexachlorobenzene
Concen: 9.868 ng/ul
RT: 16.334 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:284 Resp: 81493
Ion Ratio Lower Upper
284 100
142 39.0 30.0 45.0
249 32.8 25.5 38.3









#70
 Atrazine
 Concen: 9.847 ng/ul
 RT: 16.504 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SST010218


Tgt Ion:200 Resp: 68653
 Ion Ratio Lower Upper
 200 100
 173 26.8 20.6 31.0
 215 50.0 37.6 56.4



#71
 Pentachlorophenol
 Concen: 9.646 ng/ul
 RT: 16.681 min Scan# 2328
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24

Tgt Ion:266 Resp: 46969
 Ion Ratio Lower Upper
 266 100
 264 64.5 49.1 73.7
 268 65.2 49.2 73.8









#72
 Phenanthrene
 Concen: 10.642 ng/u1
 RT: 17.075 min Scan# 2395
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SST010218


Tgt Ion:178	Resp:	403806
Ion Ratio	Lower	Upper
178	100	
179	14.5	12.4 18.6
176	18.3	14.6 21.8



#74
 Anthracene
 Concen: 10.422 ng/u1
 RT: 17.169 min Scan# 2411
 Delta R.T. 0.000 min
 Lab File: BN031046.D
 Acq: 19 Apr 2024 18:24

Tgt Ion:178	Resp:	399810
Ion Ratio	Lower	Upper
178	100	
179	15.8	12.2 18.4
176	18.1	14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 9.679 ng/uL
RT: 13.069 min Scan# 1714
Delta R.T. -0.006 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:216 Resp: 87827
Ion Ratio Lower Upper
216 100
214 75.8 63.0 94.6
218 46.1 38.3 57.5



#76
Pentachlorobenzene
Concen: 10.062 ng/uL
RT: 14.598 min Scan# 1974
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:250 Resp: 92732
Ion Ratio Lower Upper
250 100
248 58.9 49.3 73.9
252 59.4 52.9 79.3





#77

Carbazole

Concen: 11.231 ng/ul

RT: 17.439 min Scan# 2457

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:167 Resp: 377815

Ion Ratio Lower Upper

167 100

166 21.6 16.2 24.2

139 12.8 10.6 15.8



#78

Di-n-butylphthalate

Concen: 10.400 ng/ul

RT: 18.010 min Scan# 2554

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:149 Resp: 426806


Ion Ratio Lower Upper

149 100

150 8.7 7.1 10.7

104 5.1 3.9 5.9





#80

Fluoranthene

Concen: 8.868 ng/ul

RT: 19.098 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:202 Resp: 479598

Ion Ratio Lower Upper

202 100

101 12.1 9.0 13.6

100 9.0 6.9 10.3



#82

Pyrene

Concen: 9.254 ng/ul

RT: 19.463 min Scan# 2801

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:202 Resp: 528546


Ion Ratio Lower Upper

202 100

101 13.1 11.1 16.7

100 10.8 9.3 13.9





#83

Butylbenzylphthalate

Concen: 9.844 ng/ul

RT: 20.375 min Scan# 2956

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:149 Resp: 214702

Ion Ratio Lower Upper

149 100

91 70.5 57.3 85.9

206 20.9 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 11.512 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:252 Resp: 178979


Ion Ratio Lower Upper

252 100

254 65.5 54.5 81.7

126 11.7 9.8 14.6





#85

Benzo(a)anthracene

Concen: 10.117 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:228 Resp: 571321

Ion Ratio Lower Upper

228 100

229 19.8 16.7 25.1

226 27.7 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 10.461 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24




Tgt Ion:149 Resp: 336466


Ion Ratio Lower Upper

149 100

167 26.6 22.3 33.5

279 3.9 3.0 4.6





#87

Chrysene

Concen: 10.394 ng/ul

RT: 21.310 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:228 Resp: 554015

Ion Ratio Lower Upper

228 100

226 28.7 22.6 34.0

229 20.4 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 8.760 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. 0.000 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Tgt Ion:149 Resp: 607569









#90
Benzo(b)fluoranthene
Concen: 9.660 ng/ul
RT: 23.251 min Scan# 3445
Delta R.T. -0.006 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:252 Resp: 622640
Ion Ratio Lower Upper
252 100
253 23.1 17.6 26.4
125 11.8 8.6 12.8



#91
Benzo(k)fluoranthene
Concen: 10.002 ng/ul
RT: 23.310 min Scan# 3455
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:252 Resp: 654658
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.0
125 11.2 9.4 14.2









#93
Benzo(a)pyrene
Concen: 10.201 ng/ul
RT: 24.027 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD010218


Tgt Ion:252 Resp: 621093
Ion Ratio Lower Upper
252 100
253 20.9 17.7 26.5
125 11.8 9.9 14.9



#94
Indeno(1,2,3-cd)pyrene
Concen: 11.528 ng/ul
RT: 27.421 min Scan# 4154
Delta R.T. 0.000 min
Lab File: BN031046.D
Acq: 19 Apr 2024 18:24

Tgt Ion:276 Resp: 758689
Ion Ratio Lower Upper
276 100
138 21.8 17.4 26.2
277 22.1 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 11.226 ng/ul

RT: 27.480 min Scan# 4165

Delta R.T. -0.006 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD010218





Tgt Ion:278 Resp: 622017

Ion Ratio Lower Upper

278 100

139 17.5 13.3 19.9

279 23.1 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 11.761 ng/ul

RT: 28.433 min Scan# 4326

Delta R.T. -0.006 min

Lab File: BN031046.D

Acq: 19 Apr 2024 18:24

Tgt Ion:276 Resp: 607699

Ion Ratio Lower Upper

276 100

138 23.8 19.0 28.6

277 21.4 19.3 28.9

