

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031048.D
 Acq On : 19 Apr 2024 19:43
 Operator : MA/JU
 Sample : SSTD04041
 Misc :
 ALS Vial : 95 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040220

Quant Time: Apr 19 22:28:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:25:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	100933	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	489019	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.287	164	352926	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.033	188	783529	20.000	ng/u1	0.00
79) Chrysene-d12	21.274	240	889233	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1191956	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	33936	15.764	ng/uL	0.00
4) Pyridine-d5	3.552	84	252263	39.083	ng/u1	0.00
7) Phenol-d5	6.810	99	354530	43.421	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	205216	40.693	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	283730	45.807	ng/u1	0.00
15) 4-Methylphenol-d8	8.346	113	303499	44.351	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	147911	43.602	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	158668	51.129	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	300948	44.297	ng/u1	0.00
31) 4-Chloroaniline-d4	10.563	131	463325	41.424	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	995797	42.260	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	1139402	42.056	ng/u1	0.00
54) 4-Nitrophenol-d4	14.498	143	213985	47.476	ng/u1	0.00
60) Fluorene-d10	15.281	176	856012	41.514	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.410	200	169824	49.104	ng/u1	0.00
73) Anthracene-d10	17.133	188	1392782	41.256	ng/u1	0.00
81) Pyrene-d10	19.439	212	1759040	36.793	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	2254043	41.040	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	36122	15.664	ng/uL	93
5) Pyridine	3.569	79	259987	39.918	ng/u1	96
6) Benzaldehyde	6.787	77	185925	48.897	ng/u1	98
8) Phenol	6.840	94	365228	42.938	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.069	93	289541	41.743	ng/u1	100
12) 2-Chlorophenol	7.199	128	290616	43.738	ng/u1	99
13) 2-Methylphenol	8.081	108	281901	43.009	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.157	45	374783	38.373	ng/u1	99
16) Acetophenone	8.457	105	467671	43.285	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.446	70	234385	43.072	ng/u1	100
18) 4-Methylphenol	8.405	108	319196	43.502	ng/u1	98
19) Hexachloroethane	8.704	117	115851	41.587	ng/u1	98
22) Nitrobenzene	8.834	77	347165	41.923	ng/u1	99
23) Isophorone	9.357	82	705422	42.714	ng/u1	99
25) 2-Nitrophenol	9.540	139	176569	48.412	ng/u1	95
26) 2,4-Dimethylphenol	9.604	107	337370	41.618	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.846	93	418182	39.826	ng/u1	99
29) 2,4-Dichlorophenol	10.069	162	299429	43.179	ng/u1	98
30) Naphthalene	10.469	128	1023429	40.805	ng/u1	99
32) 4-Chloroaniline	10.587	127	456918	42.389	ng/u1	99
33) Hexachlorobutadiene	10.746	225	175134	40.123	ng/u1	98
34) Caprolactam	11.363	113	116174	44.844	ng/u1	96
35) 4-Chloro-3-methylphenol	11.710	107	350597	44.888	ng/u1	97
36) 2-Methylnaphthalene	12.087	142	725339	41.411	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031048.D
 Acq On : 19 Apr 2024 19:43
 Operator : MA/JU
 Sample : SSTD04041
 Misc :
 ALS Vial : 95 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD040220


Quant Time: Apr 19 22:28:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:25:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	729630	41.035	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.457	216	368437	40.452	ng/ul	99
40) Hexachlorocyclopentadiene	12.434	237	205129	39.658	ng/ul	99
41) 2,4,6-Trichlorophenol	12.704	196	256444	45.058	ng/ul	95
42) 2,4,5-Trichlorophenol	12.775	196	284135	44.205	ng/ul	99
43) 1,1'-Biphenyl	13.116	154	952982	39.616	ng/ul	99
44) 2-Chloronaphthalene	13.151	162	769547	40.638	ng/ul	98
45) 2-Nitroaniline	13.369	65	229911	45.622	ng/ul	99
47) Dimethylphthalate	13.751	163	1011970	41.575	ng/ul	100
48) 2,6-Dinitrotoluene	13.869	165	208720	46.249	ng/ul	97
50) Acenaphthylene	14.004	152	1297184	41.439	ng/ul	99
51) 3-Nitroaniline	14.204	138	239763	50.054	ng/ul	94
52) Acenaphthene	14.351	153	862456	41.344	ng/ul	98
53) 2,4-Dinitrophenol	14.404	184	109769	52.957	ng/ul	96
55) 4-Nitrophenol	14.510	109	172498	48.306	ng/ul	98
56) Dibenzofuran	14.687	168	1182084	40.765	ng/ul	99
57) 2,4-Dinitrotoluene	14.657	165	320084	48.580	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.916	232	244011	45.710	ng/ul	96
59) Diethylphthalate	15.122	149	1062858	42.982	ng/ul	98
61) Fluorene	15.339	166	976925	42.386	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.334	204	456359	40.357	ng/ul	99
63) 4-Nitroaniline	15.369	138	235826	50.640	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.422	198	181241	48.772	ng/ul	99
67) N-Nitrosodiphenylamine	15.551	169	853336	39.991	ng/ul	99
68) 4-Bromophenyl-phenylether	16.234	248	309897	41.652	ng/ul	90
69) Hexachlorobenzene	16.334	284	346548	39.556	ng/ul	95
70) Atrazine	16.510	200	315581	42.669	ng/ul	97
71) Pentachlorophenol	16.686	266	235319	45.555	ng/ul	98
72) Phenanthrene	17.081	178	1662632	41.302	ng/ul	100
74) Anthracene	17.169	178	1723566	42.353	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.075	216	371761	38.621	ng/uL	94
76) Pentachlorobenzene	14.598	250	378211	38.682	ng/uL	99
77) Carbazole	17.445	167	1643919	46.063	ng/ul	98
78) Di-n-butylphthalate	18.010	149	1965789	45.152	ng/ul	100
80) Fluoranthene	19.104	202	2100468	36.959	ng/ul	98
82) Pyrene	19.469	202	2164137	36.056	ng/ul	98
83) Butylbenzylphthalate	20.380	149	1019020	44.457	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.186	252	764200	46.771	ng/ul	89
85) Benzo(a)anthracene	21.257	228	2375324	40.025	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	1489131	44.054	ng/ul	99
87) Chrysene	21.316	228	2274193	40.600	ng/ul	96
89) Di-n-octyl phthalate	22.280	149	2879449	38.147	ng/ul	100
90) Benzo(b)fluoranthene	23.257	252	2692472	38.383	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	2593167	36.400	ng/ul	98
93) Benzo(a)pyrene	24.045	252	2629372	39.679	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.433	276	3354667	46.832	ng/ul	99
95) Dibenzo(a,h)anthracene	27.498	278	2771452	45.959	ng/ul	97
96) Benzo(g,h,i)perylene	28.456	276	2711220	48.212	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 19 22:28:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:25:10 2024
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 15.664 ng/uL

RT: 3.181 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion: 88 Resp: 36122

Ion Ratio Lower Upper

88 100

43 29.7 22.2 33.2

58 69.0 50.2 75.2



#5

Pyridine

Concen: 39.918 ng/uL

RT: 3.569 min Scan# 99

Delta R.T. -0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion: 79 Resp: 259987


Ion Ratio Lower Upper

79 100

52 71.2 59.9 89.9

51 33.7 29.0 43.6









#6
Benzaldehyde
Concen: 48.897 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion: 77 Resp: 185925
Ion Ratio Lower Upper
77 100
105 101.4 80.0 120.0
106 96.2 75.0 112.6



#8
Phenol
Concen: 42.938 ng/ul
RT: 6.840 min Scan# 655
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion: 94 Resp: 365228
Ion Ratio Lower Upper
94 100
65 27.6 22.0 33.0
66 37.1 29.8 44.6





#10

Bis(2-Chloroethyl)ether

Concen: 41.743 ng/ul

RT: 7.069 min Scan# 694

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion: 93 Resp: 289541

Ion Ratio Lower Upper

93 100

63 71.7 57.1 85.7

95 31.6 24.9 37.3



#12

2-Chlorophenol

Concen: 43.738 ng/ul

RT: 7.199 min Scan# 716

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:128 Resp: 290616


Ion Ratio Lower Upper

128 100

64 44.2 34.9 52.3

130 32.4 27.0 40.4





#13

2-Methylphenol

Concen: 43.009 ng/ul

RT: 8.081 min Scan# 8

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :





SSTD040220

Tgt Ion:108 Resp: 281901

Ion Ratio Lower Upper

108 100

107 89.3 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 38.373 ng/ul

RT: 8.157 min Scan# 879

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion: 45 Resp: 374783


Ion Ratio Lower Upper

45 100

77 16.0 12.5 18.7

79 12.3 9.6 14.4









#16
Acetophenone
Concen: 43.285 ng/ul
RT: 8.457 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:105 Resp: 467671
Ion Ratio Lower Upper
105 100
77 75.6 60.2 90.4
51 24.8 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 43.072 ng/ul
RT: 8.446 min Scan# 928
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion: 70 Resp: 234385
Ion Ratio Lower Upper
70 100
42 56.5 45.0 67.6
101 10.4 8.1 12.1
130 23.6 19.2 28.8









#18
4-Methylphenol
Concen: 43.502 ng/ul
RT: 8.405 min Scan# 918
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:108 Resp: 319196
Ion Ratio Lower Upper
108 100
107 114.3 89.8 134.8



#19
Hexachloroethane
Concen: 41.587 ng/ul
RT: 8.704 min Scan# 972
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:117 Resp: 115851
Ion Ratio Lower Upper
117 100
201 99.6 77.4 116.0
199 62.7 49.6 74.4









#22
Nitrobenzene
Concen: 41.923 ng/ul
RT: 8.834 min Scan# 994
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion: 77 Resp: 347165
Ion Ratio Lower Upper
77 100
123 46.7 38.2 57.2
65 13.3 10.7 16.1



#23
Isophorone
Concen: 42.714 ng/ul
RT: 9.357 min Scan# 1083
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion: 82 Resp: 705422
Ion Ratio Lower Upper
82 100
95 7.4 5.9 8.9
138 19.1 14.7 22.1









#25
2-Nitrophenol
Concen: 48.412 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:139 Resp: 176569
Ion Ratio Lower Upper
139 100
65 49.1 35.2 52.8
109 32.1 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 41.618 ng/ul
RT: 9.604 min Scan# 1125
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:107 Resp: 337370
Ion Ratio Lower Upper
107 100
121 52.3 40.3 60.5
122 89.0 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 39.826 ng/ul

RT: 9.846 min Scan# 1165

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion: 93 Resp: 418182

Ion Ratio Lower Upper

93 100

95 32.7 25.8 38.6

123 11.7 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 43.179 ng/ul

RT: 10.069 min Scan# 1204

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:162 Resp: 299429


Ion Ratio Lower Upper

162 100

164 66.5 51.0 76.6

98 35.4 28.6 42.8









#30
Naphthalene
Concen: 40.805 ng/ul
RT: 10.469 min Scan# 1272
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:128 Resp: 1023429
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.4
127 12.9 10.5 15.7



#32
4-Chloroaniline
Concen: 42.389 ng/ul
RT: 10.587 min Scan# 1292
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:127 Resp: 456918
Ion Ratio Lower Upper
127 100
129 32.4 25.4 38.2





#33

Hexachlorobutadiene

Concen: 40.123 ng/ul

RT: 10.746 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:225 Resp: 175134

Ion Ratio Lower Upper

225 100

223 62.9 49.8 74.6

227 58.5 48.8 73.2



#34

Caprolactam

Concen: 44.844 ng/ul

RT: 11.363 min Scan# 1424

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:113 Resp: 116174


Ion Ratio Lower Upper

113 100

55 146.7 112.4 168.6

56 113.9 88.0 132.0









#35
 4-Chloro-3-methylphenol
 Concen: 44.888 ng/ul
 RT: 11.710 min Scan# 1483
 Delta R.T. -0.000 min
 Lab File: BN031048.D
 Acq: 19 Apr 2024 19:43




Instrument :
 BNA_N
 ClientSampleId :
 SST040220


Tgt Ion:107 Resp: 350597
 Ion Ratio Lower Upper
 107 100
 144 25.8 20.6 31.0
 142 81.1 67.4 101.2



#36
 2-Methylnaphthalene
 Concen: 41.411 ng/ul
 RT: 12.087 min Scan# 1547
 Delta R.T. -0.000 min
 Lab File: BN031048.D
 Acq: 19 Apr 2024 19:43

Tgt Ion:142 Resp: 725339
 Ion Ratio Lower Upper
 142 100
 141 87.1 69.6 104.4









#37
1-Methylnaphthalene
Concen: 41.035 ng/ul
RT: 12.310 min Scan# 1585
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:142 Resp: 729630
Ion Ratio Lower Upper
142 100
141 92.2 70.2 105.4
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.452 ng/ul
RT: 12.457 min Scan# 1610
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:216 Resp: 368437
Ion Ratio Lower Upper
216 100
214 78.4 63.1 94.7
179 21.9 17.0 25.6
108 19.0 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 39.658 ng/ul
RT: 12.434 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:237 Resp: 205129
Ion Ratio Lower Upper
237 100
235 60.1 47.6 71.4
272 13.2 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 45.058 ng/ul
RT: 12.704 min Scan# 1652
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:196 Resp: 256444
Ion Ratio Lower Upper
196 100
198 93.7 69.9 104.9
200 29.7 23.4 35.2





#42

2,4,5-Trichlorophenol

Concen: 44.205 ng/ul

RT: 12.775 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:196 Resp: 284135

Ion Ratio Lower Upper

196 100

198 95.3 75.9 113.9

200 30.4 24.7 37.1



#43

1,1'-Biphenyl

Concen: 39.616 ng/ul

RT: 13.116 min Scan# 1722

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:154 Resp: 952982


Ion Ratio Lower Upper

154 100

153 39.6 31.8 47.8

76 12.3 10.6 15.8









#44
2-Chloronaphthalene
Concen: 40.638 ng/ul
RT: 13.151 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:162 Resp: 769547
Ion Ratio Lower Upper
162 100
164 31.5 25.8 38.8
127 40.9 33.7 50.5



#45
2-Nitroaniline
Concen: 45.622 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion: 65 Resp: 229911
Ion Ratio Lower Upper
65 100
92 70.3 57.4 86.2
138 115.5 91.6 137.4





#47

Dimethylphthalate

Concen: 41.575 ng/ul

RT: 13.751 min Scan# 1830

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:163 Resp: 1011970

Ion Ratio Lower Upper

163 100

194 4.1 3.0 4.6

164 10.1 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 46.249 ng/ul

RT: 13.869 min Scan# 1850

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:165 Resp: 208720


Ion Ratio Lower Upper

165 100

89 58.2 44.7 67.1

121 22.3 17.0 25.4









#50
Acenaphthylene
Concen: 41.439 ng/ul
RT: 14.004 min Scan# 1873
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:152 Resp: 1297184
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.8
153 13.5 10.5 15.7



#51
3-Nitroaniline
Concen: 50.054 ng/ul
RT: 14.204 min Scan# 1907
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:138 Resp: 239763
Ion Ratio Lower Upper
138 100
108 11.7 11.1 16.7
92 110.3 93.9 140.9









#52
Acenaphthene
Concen: 41.344 ng/ul
RT: 14.351 min Scan# 1932
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:153 Resp: 862456
Ion Ratio Lower Upper
153 100
152 47.3 36.3 54.5
154 88.4 72.4 108.6



#53
2,4-Dinitrophenol
Concen: 52.957 ng/ul
RT: 14.404 min Scan# 1941
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:184 Resp: 109769
Ion Ratio Lower Upper
184 100
63 65.7 50.4 75.6
154 70.5 53.0 79.6









#55
4-Nitrophenol
Concen: 48.306 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:109 Resp: 172498
Ion Ratio Lower Upper
109 100
139 129.3 105.6 158.4
65 124.4 98.2 147.2



#56
Dibenzofuran
Concen: 40.765 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:168 Resp: 1182084
Ion Ratio Lower Upper
168 100
139 38.0 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 48.580 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:165 Resp: 320084
Ion Ratio Lower Upper
165 100
63 40.8 33.9 50.9
182 2.9 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 45.710 ng/ul
RT: 14.916 min Scan# 2028
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:232 Resp: 244011
Ion Ratio Lower Upper
232 100
131 47.9 41.2 61.8
130 2.7 2.2 3.2
166 32.1 27.3 40.9









#59
Diethylphthalate
Concen: 42.982 ng/ul
RT: 15.122 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:149 Resp: 1062858
Ion Ratio Lower Upper
149 100
177 22.5 16.8 25.2
150 12.6 10.1 15.1



#61
Fluorene
Concen: 42.386 ng/ul
RT: 15.339 min Scan# 2100
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:166 Resp: 976925
Ion Ratio Lower Upper
166 100
165 93.4 74.6 112.0
167 13.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 40.357 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:204 Resp: 456359
Ion Ratio Lower Upper
204 100
206 32.7 26.4 39.6
141 62.7 50.7 76.1



#63
4-Nitroaniline
Concen: 50.640 ng/ul
RT: 15.369 min Scan# 2105
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:138 Resp: 235826
Ion Ratio Lower Upper
138 100
92 49.6 38.4 57.6
108 84.6 63.8 95.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 48.772 ng/ul
 RT: 15.422 min Scan# 2114
 Delta R.T. 0.006 min
 Lab File: BN031048.D
 Acq: 19 Apr 2024 19:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTD040220


Tgt Ion:198 Resp: 181241
 Ion Ratio Lower Upper
 198 100
 182 4.4 4.0 6.0
 77 27.1 22.2 33.4



#67
 N-Nitrosodiphenylamine
 Concen: 39.991 ng/ul
 RT: 15.551 min Scan# 2136
 Delta R.T. -0.000 min
 Lab File: BN031048.D
 Acq: 19 Apr 2024 19:43

Tgt Ion:169 Resp: 853336
 Ion Ratio Lower Upper
 169 100
 168 65.0 52.6 79.0
 167 35.0 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 41.652 ng/ul
RT: 16.234 min Scan# 2252
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:248 Resp: 309897
Ion Ratio Lower Upper
248 100
250 94.4 78.2 117.2
141 73.1 71.4 107.0



#69
Hexachlorobenzene
Concen: 39.556 ng/ul
RT: 16.334 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:284 Resp: 346548
Ion Ratio Lower Upper
284 100
142 40.1 30.0 45.0
249 34.6 25.5 38.3





#70

Atrazine

Concen: 42.669 ng/ul

RT: 16.510 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:200 Resp: 315581

Ion Ratio Lower Upper

200 100

173 26.8 20.6 31.0

215 49.4 37.6 56.4



#71

Pentachlorophenol

Concen: 45.555 ng/ul

RT: 16.686 min Scan# 2329

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:266 Resp: 235319


Ion Ratio Lower Upper

266 100

264 60.8 49.1 73.7

268 58.5 49.2 73.8





#72

Phenanthrene

Concen: 41.302 ng/u1

RT: 17.081 min Scan# 2395

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:178 Resp: 1662632

Ion Ratio Lower Upper

178 100

179 15.5 12.4 18.6

176 18.4 14.6 21.8



#74

Anthracene

Concen: 42.353 ng/u1

RT: 17.169 min Scan# 2411

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:178 Resp: 1723566


Ion Ratio Lower Upper

178 100

179 14.8 12.2 18.4

176 18.6 14.4 21.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 38.621 ng/uL
RT: 13.075 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:216 Resp: 371761
Ion Ratio Lower Upper
216 100
214 75.5 63.0 94.6
218 40.6 38.3 57.5



#76
Pentachlorobenzene
Concen: 38.682 ng/uL
RT: 14.598 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:250 Resp: 378211
Ion Ratio Lower Upper
250 100
248 63.1 49.3 73.9
252 65.8 52.9 79.3









#77
 Carbazole
 Concen: 46.063 ng/ul
 RT: 17.445 min Scan# 2458
 Delta R.T. 0.006 min
 Lab File: BN031048.D
 Acq: 19 Apr 2024 19:43

Instrument :
 BNA_N
 ClientSampleId :
 SSTD040220

Tgt Ion:167 Resp: 1643919




Ion	Ratio	Lower	Upper
167	100		
166	21.6	16.2	24.2
139	13.4	10.6	15.8




#78
 Di-n-butylphthalate
 Concen: 45.152 ng/ul
 RT: 18.010 min Scan# 2554
 Delta R.T. -0.000 min
 Lab File: BN031048.D
 Acq: 19 Apr 2024 19:43

Tgt Ion:149 Resp: 1965789

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.1	10.7
104	5.2	3.9	5.9









#80
Fluoranthene
Concen: 36.959 ng/ul
RT: 19.104 min Scan# 21
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:202 Resp: 2100468
Ion Ratio Lower Upper
202 100
101 10.5 9.0 13.6
100 7.9 6.9 10.3



#82
Pyrene
Concen: 36.056 ng/ul
RT: 19.469 min Scan# 2802
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:202 Resp: 2164137
Ion Ratio Lower Upper
202 100
101 13.2 11.1 16.7
100 10.7 9.3 13.9





#83

Butylbenzylphthalate

Concen: 44.457 ng/ul

RT: 20.380 min Scan# 2957

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:149 Resp: 1019020

Ion Ratio Lower Upper

149 100

91 68.8 57.3 85.9

206 21.3 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 46.771 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:252 Resp: 764200


Ion Ratio Lower Upper

252 100

254 57.5 54.5 81.7

126 12.4 9.8 14.6









#85
Benzo(a)anthracene
Concen: 40.025 ng/ul
RT: 21.257 min Scan# 3105
Delta R.T. 0.006 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43




Instrument :
BNA_N
ClientSampleId :
SSTD040220


Tgt Ion:228 Resp: 2375324
Ion Ratio Lower Upper
228 100
229 20.0 16.7 25.1
226 26.7 21.2 31.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 44.054 ng/ul
RT: 21.180 min Scan# 3093
Delta R.T. -0.000 min
Lab File: BN031048.D
Acq: 19 Apr 2024 19:43

Tgt Ion:149 Resp: 1489131
Ion Ratio Lower Upper
149 100
167 27.4 22.3 33.5
279 4.0 3.0 4.6





#87

Chrysene

Concen: 40.600 ng/ul

RT: 21.316 min Scan# 3116

Delta R.T. 0.006 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:228 Resp: 2274193

Ion Ratio Lower Upper

228 100

226 30.9 22.6 34.0

229 18.9 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 38.147 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Tgt Ion:149 Resp: 2879449





#90

Benzo(b)fluoranthene

Concen: 38.383 ng/ul

RT: 23.257 min Scan# 3446

Delta R.T. -0.000 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:252 Resp: 2692472

Ion Ratio Lower Upper

252 100

253 22.2 17.6 26.4

125 11.0 8.6 12.8



#91

Benzo(k)fluoranthene

Concen: 36.400 ng/ul

RT: 23.321 min Scan# 3457

Delta R.T. 0.012 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:252 Resp: 2593167


Ion Ratio Lower Upper

252 100

253 23.0 17.4 26.0

125 11.4 9.4 14.2





#93

Benzo(a)pyrene

Concen: 39.679 ng/u1

RT: 24.045 min Scan# 3

Delta R.T. 0.012 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:252 Resp: 2629372

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5

125 11.9 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 46.832 ng/u1

RT: 27.433 min Scan# 4156

Delta R.T. 0.012 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43




Tgt Ion:276 Resp: 3354667


Ion Ratio Lower Upper

276 100

138 21.5 17.4 26.2

277 22.3 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 45.959 ng/ul

RT: 27.498 min Scan# 4165

Delta R.T. 0.012 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Instrument :

BNA_N

ClientSampleId :

SSTD040220





Tgt Ion:278 Resp: 2771452

Ion Ratio Lower Upper

278 100

139 16.6 13.3 19.9

279 23.9 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 48.212 ng/ul

RT: 28.456 min Scan# 4330

Delta R.T. 0.018 min

Lab File: BN031048.D

Acq: 19 Apr 2024 19:43

Tgt Ion:276 Resp: 2711220

Ion Ratio Lower Upper

276 100

138 22.8 19.0 28.6

277 25.3 19.3 28.9

