

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031051.D
 Acq On : 19 Apr 2024 22:22
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV223

Quant Time: Apr 19 23:00:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	113785	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	538022	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	389553	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.034	188	867229	20.000	ng/u1	0.00
79) Chrysene-d12	21.269	240	958978	20.000	ng/u1	0.00
88) Perylene-d12	24.174	264	1258025	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	18310	7.518	ng/uL	0.00
4) Pyridine-d5	3.552	84	149280	21.283	ng/u1	0.00
7) Phenol-d5	6.811	99	198073	20.636	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	118245	21.174	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	157500	21.074	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	168329	20.853	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	83851	21.691	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	87902	21.876	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	172692	21.793	ng/u1	0.00
31) 4-Chloroaniline-d4	10.563	131	267225	21.580	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	571452	21.081	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	653947	21.288	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	119701	21.156	ng/u1	0.00
60) Fluorene-d10	15.281	176	498581	21.128	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	89099	20.364	ng/u1	0.00
73) Anthracene-d10	17.134	188	822933	21.893	ng/u1	0.00
81) Pyrene-d10	19.433	212	1042795	22.129	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	1243664	20.847	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	20771	8.048	ng/uL	91
5) Pyridine	3.570	79	147308	20.813	ng/u1	99
6) Benzaldehyde	6.787	77	119725	25.403	ng/u1	97
8) Phenol	6.834	94	205543	20.770	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.069	93	165815	21.236	ng/u1	96
12) 2-Chlorophenol	7.199	128	166567	21.442	ng/u1	99
13) 2-Methylphenol	8.081	108	162015	21.126	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.158	45	221792	21.787	ng/u1	99
16) Acetophenone	8.457	105	273879	21.898	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.440	70	136289	21.862	ng/u1	97
18) 4-Methylphenol	8.405	108	179212	20.862	ng/u1	98
19) Hexachloroethane	8.705	117	68665	21.301	ng/u1	95
22) Nitrobenzene	8.834	77	196400	21.527	ng/u1	99
23) Isophorone	9.357	82	399578	21.461	ng/u1	98
25) 2-Nitrophenol	9.540	139	97885	22.018	ng/u1	97
26) 2,4-Dimethylphenol	9.599	107	196009	21.619	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.840	93	244159	21.589	ng/u1	98
29) 2,4-Dichlorophenol	10.063	162	171918	21.722	ng/u1	99
30) Naphthalene	10.469	128	597803	21.440	ng/u1	99
32) 4-Chloroaniline	10.581	127	261218	21.710	ng/u1	99
33) Hexachlorobutadiene	10.746	225	103242	21.398	ng/u1	98
34) Caprolactam	11.363	113	63775	20.451	ng/u1	98
35) 4-Chloro-3-methylphenol	11.710	107	199516	21.742	ng/u1	97
36) 2-Methylnaphthalene	12.087	142	423973	21.562	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031051.D
 Acq On : 19 Apr 2024 22:22
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 99 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SICV223


Quant Time: Apr 19 23:00:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.310	142	425965	21.409	ng/ul	95
39) 1,2,4,5-Tetrachloroben...	12.457	216	214809	21.073	ng/ul	98
40) Hexachlorocyclopentadiene	12.434	237	110816	19.791	ng/ul	94
41) 2,4,6-Trichlorophenol	12.704	196	142497	21.002	ng/ul	89
42) 2,4,5-Trichlorophenol	12.775	196	160097	21.282	ng/ul	96
43) 1,1'-Biphenyl	13.110	154	558467	20.845	ng/ul	99
44) 2-Chloronaphthalene	13.151	162	450440	21.040	ng/ul	97
45) 2-Nitroaniline	13.363	65	126626	21.425	ng/ul	97
47) Dimethylphthalate	13.745	163	589834	21.091	ng/ul	99
48) 2,6-Dinitrotoluene	13.869	165	114545	21.820	ng/ul	99
50) Acenaphthylene	14.004	152	765416	21.331	ng/ul	99
51) 3-Nitroaniline	14.198	138	131045	21.744	ng/ul	95
52) Acenaphthene	14.345	153	504570	21.212	ng/ul	97
53) 2,4-Dinitrophenol	14.404	184	53162	17.907	ng/ul	93
55) 4-Nitrophenol	14.510	109	98527	21.458	ng/ul	97
56) Dibenzofuran	14.687	168	694301	21.111	ng/ul	98
57) 2,4-Dinitrotoluene	14.657	165	177836	22.295	ng/ul	99
58) 2,3,4,6-Tetrachlorophenol	14.910	232	137483	21.147	ng/ul	99
59) Diethylphthalate	15.116	149	607855	21.135	ng/ul	99
61) Fluorene	15.334	166	574859	21.240	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.334	204	278881	21.633	ng/ul	96
63) 4-Nitroaniline	15.363	138	150244	23.122	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.416	198	98521	20.970	ng/ul	99
67) N-Nitrosodiphenylamine	15.551	169	495992	21.338	ng/ul	99
68) 4-Bromophenyl-phenylether	16.228	248	172855	21.076	ng/ul	97
69) Hexachlorobenzene	16.334	284	205847	21.771	ng/ul	99
70) Atrazine	16.510	200	181046	22.185	ng/ul	99
71) Pentachlorophenol	16.681	266	129858	21.139	ng/ul	98
72) Phenanthrene	17.075	178	982177	21.412	ng/ul	100
74) Anthracene	17.169	178	1010204	21.762	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.069	216	215904	21.225	ng/uL	99
76) Pentachlorobenzene	14.598	250	225512	21.469	ng/uL	95
77) Carbazole	17.445	167	944595	22.277	ng/ul	99
78) Di-n-butylphthalate	18.010	149	1129656	21.857	ng/ul	99
80) Fluoranthene	19.098	202	1194194	21.412	ng/ul	100
82) Pyrene	19.463	202	1287062	21.964	ng/ul	100
83) Butylbenzylphthalate	20.375	149	572049	22.124	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.186	252	447160	22.378	ng/ul	96
85) Benzo(a)anthracene	21.251	228	1384337	21.258	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.180	149	871824	22.456	ng/ul	99
87) Chrysene	21.310	228	1348438	21.814	ng/ul	98
89) Di-n-octyl phthalate	22.280	149	1566008	21.596	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	1505476	21.269	ng/ul	97
91) Benzo(k)fluoranthene	23.316	252	1585936	22.277	ng/ul	96
93) Benzo(a)pyrene	24.033	252	1481549	20.981	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.421	276	1055758	12.049	ng/ul	98
95) Dibenzo(a,h)anthracene	27.480	278	1509177	20.984	ng/ul	97
96) Benzo(g,h,i)perylene	28.433	276	798507	11.316	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
BNA_N
ClientSampleId :
SICV223

Quant Time: Apr 19 23:00:50 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 8.048 ng/uL
 RT: 3.181 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22




Instrument :
 BNA_N
 ClientSampleId :
 SICV223


Tgt Ion:	88	Resp:	20771
Ion	Ratio	Lower	Upper
88	100		
43	29.1	22.2	33.2
58	71.3	50.2	75.2



#5
 Pyridine
 Concen: 20.813 ng/uL
 RT: 3.570 min Scan# 99
 Delta R.T. -0.006 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22

Tgt Ion:	79	Resp:	147308
Ion	Ratio	Lower	Upper
79	100		
52	74.9	59.9	89.9
51	35.3	29.0	43.6









#6
Benzaldehyde
Concen: 25.403 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion: 77 Resp: 119725
Ion Ratio Lower Upper
77 100
105 103.1 80.0 120.0
106 97.2 75.0 112.6



#8
Phenol
Concen: 20.770 ng/ul
RT: 6.834 min Scan# 654
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion: 94 Resp: 205543
Ion Ratio Lower Upper
94 100
65 26.7 22.0 33.0
66 36.6 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.236 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion: 93 Resp: 165815
Ion Ratio Lower Upper
93 100
63 75.1 57.1 85.7
95 33.6 24.9 37.3



#12
2-Chlorophenol
Concen: 21.442 ng/ul
RT: 7.199 min Scan# 716
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:128 Resp: 166567
Ion Ratio Lower Upper
128 100
64 44.3 34.9 52.3
130 33.5 27.0 40.4





#13

2-Methylphenol

Concen: 21.126 ng/ul

RT: 8.081 min Scan# 8

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :





SICV223

Tgt Ion:108 Resp: 162015

Ion Ratio Lower Upper

108 100

107 88.2 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 21.787 ng/ul

RT: 8.158 min Scan# 879

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion: 45 Resp: 221792


Ion Ratio Lower Upper

45 100

77 16.0 12.5 18.7

79 11.3 9.6 14.4









#16
Acetophenone
Concen: 21.898 ng/ul
RT: 8.457 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:105 Resp: 273879
Ion Ratio Lower Upper
105 100
77 73.8 60.2 90.4
51 24.8 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.862 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion: 70 Resp: 136289
Ion Ratio Lower Upper
70 100
42 53.9 45.0 67.6
101 10.2 8.1 12.1
130 22.3 19.2 28.8









#18
4-Methylphenol
Concen: 20.862 ng/ul
RT: 8.405 min Scan# 912
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:108 Resp: 179212
Ion Ratio Lower Upper
108 100
107 114.2 89.8 134.8



#19
Hexachloroethane
Concen: 21.301 ng/ul
RT: 8.705 min Scan# 972
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:117 Resp: 68665
Ion Ratio Lower Upper
117 100
201 102.7 77.4 116.0
199 64.0 49.6 74.4









#22
Nitrobenzene
Concen: 21.527 ng/ul
RT: 8.834 min Scan# 994
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion: 77 Resp: 196400
Ion Ratio Lower Upper
77 100
123 46.6 38.2 57.2
65 13.4 10.7 16.1



#23
Isophorone
Concen: 21.461 ng/ul
RT: 9.357 min Scan# 1083
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion: 82 Resp: 399578
Ion Ratio Lower Upper
82 100
95 7.6 5.9 8.9
138 19.3 14.7 22.1









#25
2-Nitrophenol
Concen: 22.018 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:139 Resp: 97885
Ion Ratio Lower Upper
139 100
65 47.3 35.2 52.8
109 32.3 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 21.619 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:107 Resp: 196009
Ion Ratio Lower Upper
107 100
121 51.3 40.3 60.5
122 88.5 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.589 ng/ul

RT: 9.840 min Scan# 1165

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion: 93 Resp: 244159

Ion Ratio Lower Upper

93 100

95 33.0 25.8 38.6

123 11.4 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 21.722 ng/ul

RT: 10.063 min Scan# 1203

Delta R.T. -0.006 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion: 162 Resp: 171918


Ion Ratio Lower Upper

162 100

164 65.1 51.0 76.6

98 35.6 28.6 42.8





#30

Naphthalene

Concen: 21.440 ng/ul

RT: 10.469 min Scan# 1272

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:128 Resp: 597803

Ion Ratio Lower Upper

128 100

129 10.7 9.0 13.4

127 13.4 10.5 15.7



#32

4-Chloroaniline

Concen: 21.710 ng/ul

RT: 10.581 min Scan# 1291

Delta R.T. -0.006 min

Lab File: BN031051.D




Acq: 19 Apr 2024 22:22


Tgt Ion:127 Resp: 261218

Ion Ratio Lower Upper

127 100


129 32.5 25.4 38.2








#33
Hexachlorobutadiene
Concen: 21.398 ng/ul
RT: 10.746 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Instrument :
BNA_N
ClientSampleId :
SICV223







Tgt Ion:225 Resp: 103242
Ion Ratio Lower Upper
225 100
223 59.2 49.8 74.6
227 61.7 48.8 73.2



#34
Caprolactam
Concen: 20.451 ng/ul
RT: 11.363 min Scan# 1424
Delta R.T. 0.006 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:113 Resp: 63775
Ion Ratio Lower Upper
113 100
55 136.0 112.4 168.6
56 109.2 88.0 132.0









#35
4-Chloro-3-methylphenol
Concen: 21.742 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:107 Resp: 199516
Ion Ratio Lower Upper
107 100
144 25.4 20.6 31.0
142 80.4 67.4 101.2



#36
2-Methylnaphthalene
Concen: 21.562 ng/ul
RT: 12.087 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:142 Resp: 423973
Ion Ratio Lower Upper
142 100
141 88.1 69.6 104.4





#37

1-Methylnaphthalene

Concen: 21.409 ng/ul

RT: 12.310 min Scan# 1585

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:142 Resp: 425965

Ion Ratio Lower Upper

142 100

141 92.9 70.2 105.4

116 3.6 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.073 ng/ul

RT: 12.457 min Scan# 1610

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Tgt Ion:216 Resp: 214809




Ion Ratio Lower Upper


216 100

214 77.9 63.1 94.7

179 20.0 17.0 25.6

108 16.1 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 19.791 ng/ul
RT: 12.434 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:237 Resp: 110816
Ion Ratio Lower Upper
237 100
235 64.7 47.6 71.4
272 12.5 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 21.002 ng/ul
RT: 12.704 min Scan# 1652
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22


Tgt Ion:196 Resp: 142497
Ion Ratio Lower Upper
196 100
198 99.2 69.9 104.9
200 32.3 23.4 35.2








#42
2,4,5-Trichlorophenol
Concen: 21.282 ng/ul
RT: 12.775 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Instrument :
BNA_N
ClientSampleId :
SICV223







Tgt Ion:196 Resp: 160097
Ion Ratio Lower Upper
196 100
198 99.0 75.9 113.9
200 28.5 24.7 37.1



#43
1,1'-Biphenyl
Concen: 20.845 ng/ul
RT: 13.110 min Scan# 1721
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:154 Resp: 558467
Ion Ratio Lower Upper
154 100
153 40.7 31.8 47.8
76 12.8 10.6 15.8









#44
2-Chloronaphthalene
Concen: 21.040 ng/ul
RT: 13.151 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:162 Resp: 450440
Ion Ratio Lower Upper
162 100
164 32.1 25.8 38.8
127 39.3 33.7 50.5



#45
2-Nitroaniline
Concen: 21.425 ng/ul
RT: 13.363 min Scan# 1764
Delta R.T. -0.006 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion: 65 Resp: 126626
Ion Ratio Lower Upper
65 100
92 70.1 57.4 86.2
138 110.4 91.6 137.4









#47
Dimethylphthalate
Concen: 21.091 ng/ul
RT: 13.745 min Scan# 1829
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:163 Resp: 589834
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.6
164 10.5 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.820 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:165 Resp: 114545
Ion Ratio Lower Upper
165 100
89 56.3 44.7 67.1
121 20.8 17.0 25.4









#50
 Acenaphthylene
 Concen: 21.331 ng/ul
 RT: 14.004 min Scan# 1873
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22




Instrument : BNA_N
 ClientSampleId : SICV223


Tgt Ion:	152	Resp:	765416
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.8	23.8
153	12.8	10.5	15.7



#51
 3-Nitroaniline
 Concen: 21.744 ng/ul
 RT: 14.198 min Scan# 1906
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22

Tgt Ion:	138	Resp:	131045
Ion Ratio	Lower	Upper	
138	100		
108	12.5	11.1	16.7
92	112.1	93.9	140.9









#52
Acenaphthene
Concen: 21.212 ng/ul
RT: 14.345 min Scan# 1931
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:153 Resp: 504570
Ion Ratio Lower Upper
153 100
152 46.8 36.3 54.5
154 86.8 72.4 108.6



#53
2,4-Dinitrophenol
Concen: 17.907 ng/ul
RT: 14.404 min Scan# 1941
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:184 Resp: 53162
Ion Ratio Lower Upper
184 100
63 68.2 50.4 75.6
154 71.4 53.0 79.6









#55
4-Nitrophenol
Concen: 21.458 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:109 Resp: 98527
Ion Ratio Lower Upper
109 100
139 133.9 105.6 158.4
65 118.8 98.2 147.2



#56
Dibenzofuran
Concen: 21.111 ng/ul
RT: 14.687 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:168 Resp: 694301
Ion Ratio Lower Upper
168 100
139 37.4 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 22.295 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:165 Resp: 177836
Ion Ratio Lower Upper
165 100
63 41.9 33.9 50.9
182 3.2 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.147 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:232 Resp: 137483
Ion Ratio Lower Upper
232 100
131 52.2 41.2 61.8
130 3.0 2.2 3.2
166 34.0 27.3 40.9









#59
Diethylphthalate
Concen: 21.135 ng/ul
RT: 15.116 min Scan# 2062
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:149 Resp: 607855
Ion Ratio Lower Upper
149 100
177 20.5 16.8 25.2
150 12.6 10.1 15.1



#61
Fluorene
Concen: 21.240 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:166 Resp: 574859
Ion Ratio Lower Upper
166 100
165 97.2 74.6 112.0
167 13.2 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.633 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:204 Resp: 278881
Ion Ratio Lower Upper
204 100
206 30.2 26.4 39.6
141 60.6 50.7 76.1



#63
4-Nitroaniline
Concen: 23.122 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:138 Resp: 150244
Ion Ratio Lower Upper
138 100
92 49.5 38.4 57.6
108 80.8 63.8 95.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 20.970 ng/ul
 RT: 15.416 min Scan# 2113
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22




Instrument :
 BNA_N
 ClientSampleId :
 SICV223


Tgt Ion:198 Resp: 98521
 Ion Ratio Lower Upper
 198 100
 182 4.1 4.0 6.0
 77 27.3 22.2 33.4



#67
 N-Nitrosodiphenylamine
 Concen: 21.338 ng/ul
 RT: 15.551 min Scan# 2136
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22

Tgt Ion:169 Resp: 495992
 Ion Ratio Lower Upper
 169 100
 168 65.5 52.6 79.0
 167 36.2 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 21.076 ng/ul
RT: 16.228 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:248 Resp: 172855
Ion Ratio Lower Upper
248 100
250 98.3 78.2 117.2
141 83.6 71.4 107.0



#69
Hexachlorobenzene
Concen: 21.771 ng/ul
RT: 16.334 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:284 Resp: 205847
Ion Ratio Lower Upper
284 100
142 38.4 30.0 45.0
249 31.4 25.5 38.3









#70
 Atrazine
 Concen: 22.185 ng/ul
 RT: 16.510 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22




Instrument :
 BNA_N
 ClientSampleId :
 SICV223


Tgt Ion:	200	Resp:	181046
Ion Ratio	Lower	Upper	
200	100		
173	26.8	20.6	31.0
215	47.0	37.6	56.4



#71
 Pentachlorophenol
 Concen: 21.139 ng/ul
 RT: 16.681 min Scan# 2328
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22

Tgt Ion:	266	Resp:	129858
Ion Ratio	Lower	Upper	
266	100		
264	64.1	49.1	73.7
268	61.5	49.2	73.8





#72

Phenanthrene

Concen: 21.412 ng/ul

RT: 17.075 min Scan# 2395

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:178 Resp: 982177

Ion Ratio Lower Upper

178 100

179 15.2 12.4 18.6

176 18.2 14.6 21.8



#74

Anthracene

Concen: 21.762 ng/ul

RT: 17.169 min Scan# 2411

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion:178 Resp: 1010204


Ion Ratio Lower Upper

178 100

179 14.1 12.2 18.4

176 16.9 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.225 ng/uL

RT: 13.069 min Scan# 1714

Delta R.T. -0.006 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:216 Resp: 215904

Ion Ratio Lower Upper

216 100

214 77.9 63.0 94.6

218 49.2 38.3 57.5



#76

Pentachlorobenzene

Concen: 21.469 ng/uL

RT: 14.598 min Scan# 1974

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion:250 Resp: 225512


Ion Ratio Lower Upper

250 100

248 60.1 49.3 73.9

252 60.0 52.9 79.3





#77

Carbazole

Concen: 22.277 ng/ul

RT: 17.445 min Scan# 2457

Delta R.T. 0.006 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:167 Resp: 944595

Ion Ratio Lower Upper

167 100

166 20.4 16.2 24.2

139 12.6 10.6 15.8



#78

Di-n-butylphthalate

Concen: 21.857 ng/ul

RT: 18.010 min Scan# 2554

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion:149 Resp: 1129656


Ion Ratio Lower Upper

149 100

150 9.4 7.1 10.7

104 5.2 3.9 5.9





#80

Fluoranthene

Concen: 21.412 ng/ul

RT: 19.098 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:202 Resp: 1194194

Ion Ratio Lower Upper

202 100

101 11.4 9.0 13.6

100 8.8 6.9 10.3



#82

Pyrene

Concen: 21.964 ng/ul

RT: 19.463 min Scan# 2801

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion:202 Resp: 1287062


Ion Ratio Lower Upper

202 100

101 13.7 11.1 16.7

100 11.5 9.3 13.9









#83
 Butylbenzylphthalate
 Concen: 22.124 ng/ul
 RT: 20.375 min Scan# 2956
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22




Instrument :
 BNA_N
 ClientSampleId :
 SICV223


Tgt Ion:149	Resp:	572049
Ion	Ratio	Lower Upper
149	100	
91	72.6	57.3 85.9
206	19.8	16.2 24.4



#84
 3,3'-Dichlorobenzidine
 Concen: 22.378 ng/ul
 RT: 21.186 min Scan# 3094
 Delta R.T. -0.000 min
 Lab File: BN031051.D
 Acq: 19 Apr 2024 22:22

Tgt Ion:252	Resp:	447160
Ion	Ratio	Lower Upper
252	100	
254	64.2	54.5 81.7
126	11.9	9.8 14.6





#85

Benzo(a)anthracene

Concen: 21.258 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:228 Resp: 1384337

Ion Ratio Lower Upper

228 100

229 20.4 16.7 25.1

226 29.0 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.456 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion:149 Resp: 871824


Ion Ratio Lower Upper

149 100

167 27.1 22.3 33.5

279 3.8 3.0 4.6





#87

Chrysene

Concen: 21.814 ng/ul

RT: 21.310 min Scan# 3115

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:228 Resp: 1348438

Ion Ratio Lower Upper

228 100

226 30.0 22.6 34.0

229 19.3 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 21.596 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Tgt Ion:149 Resp: 1566008









#90
Benzo(b)fluoranthene
Concen: 21.269 ng/ul
RT: 23.251 min Scan# 3445
Delta R.T. -0.006 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22




Instrument :
BNA_N
ClientSampleId :
SICV223


Tgt Ion:252 Resp: 1505476
Ion Ratio Lower Upper
252 100
253 20.3 17.6 26.4
125 10.8 8.6 12.8



#91
Benzo(k)fluoranthene
Concen: 22.277 ng/ul
RT: 23.316 min Scan# 3456
Delta R.T. 0.006 min
Lab File: BN031051.D
Acq: 19 Apr 2024 22:22

Tgt Ion:252 Resp: 1585936
Ion Ratio Lower Upper
252 100
253 20.0 17.4 26.0
125 10.1 9.4 14.2





#93

Benzo(a)pyrene

Concen: 20.981 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:252 Resp: 1481549

Ion Ratio Lower Upper

252 100

253 22.5 17.7 26.5

125 12.2 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 12.049 ng/ul

RT: 27.421 min Scan# 4154

Delta R.T. -0.000 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22




Tgt Ion:276 Resp: 1055758


Ion Ratio Lower Upper

276 100

138 23.1 17.4 26.2

277 22.6 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 20.984 ng/ul

RT: 27.480 min Scan# 4165

Delta R.T. -0.006 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Instrument :

BNA_N

ClientSampleId :

SICV223





Tgt Ion:278 Resp: 1509177

Ion Ratio Lower Upper

278 100

139 17.3 13.3 19.9

279 23.6 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 11.316 ng/ul

RT: 28.433 min Scan# 4326

Delta R.T. -0.006 min

Lab File: BN031051.D

Acq: 19 Apr 2024 22:22

Tgt Ion:276 Resp: 798507

Ion Ratio Lower Upper

276 100

138 24.6 19.0 28.6

277 24.3 19.3 28.9

