

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031113.D
 Acq On : 21 Apr 2024 18:24
 Operator : MA/JU
 Sample : SSTDCC020
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020372

Quant Time: Apr 21 22:22:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	152897	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	714102	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	477235	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.027	188	1079438	20.000	ng/u1	0.00
79) Chrysene-d12	21.268	240	1249276	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	1563288	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	27391	8.370	ng/uL	0.00
4) Pyridine-d5	3.552	84	206911	21.954	ng/u1	0.00
7) Phenol-d5	6.805	99	271463	21.047	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	162069	21.597	ng/u1	0.00
11) 2-Chlorophenol-d4	7.163	132	215638	21.473	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	222776	20.538	ng/u1	0.00
21) Nitrobenzene-d5	8.787	128	110499	21.537	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	115076	21.577	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.034	165	228039	21.682	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	342901	20.863	ng/u1	0.00
46) Dimethylphthalate-d6	13.692	166	699842	21.073	ng/u1	0.00
49) Acenaphthylene-d8	13.969	160	827665	21.993	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	150154	21.663	ng/u1	0.00
60) Fluorene-d10	15.275	176	632496	21.878	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.398	200	103616	19.026	ng/u1	0.00
73) Anthracene-d10	17.127	188	1037696	22.179	ng/u1	0.00
81) Pyrene-d10	19.433	212	1314634	21.415	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	1630534	21.995	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	30180	8.702	ng/uL	96
5) Pyridine	3.569	79	213321	22.430	ng/u1	98
6) Benzaldehyde	6.787	77	160914	25.409	ng/u1	97
8) Phenol	6.834	94	285229	21.450	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.069	93	224259	21.374	ng/u1	100
12) 2-Chlorophenol	7.199	128	228255	21.866	ng/u1	100
13) 2-Methylphenol	8.075	108	213514	20.720	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.151	45	299108	21.866	ng/u1	99
16) Acetophenone	8.451	105	356338	21.203	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.440	70	172507	20.593	ng/u1	98
18) 4-Methylphenol	8.399	108	239171	20.720	ng/u1	98
19) Hexachloroethane	8.699	117	94350	21.782	ng/u1	94
22) Nitrobenzene	8.828	77	268980	22.212	ng/u1	98
23) Isophorone	9.351	82	502947	20.352	ng/u1	100
25) 2-Nitrophenol	9.534	139	130613	22.135	ng/u1	96
26) 2,4-Dimethylphenol	9.598	107	244162	20.289	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.840	93	316704	21.099	ng/u1	99
29) 2,4-Dichlorophenol	10.063	162	226920	21.601	ng/u1	100
30) Naphthalene	10.463	128	799688	21.608	ng/u1	99
32) 4-Chloroaniline	10.581	127	337583	21.139	ng/u1	98
33) Hexachlorobutadiene	10.740	225	136710	21.348	ng/u1	99
34) Caprolactam	11.357	113	78479	18.961	ng/u1	95
35) 4-Chloro-3-methylphenol	11.704	107	251806	20.674	ng/u1	99
36) 2-Methylnaphthalene	12.081	142	542572	20.789	ng/u1	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031113.D
 Acq On : 21 Apr 2024 18:24
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 65 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTD020372


Quant Time: Apr 21 22:22:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.304	142	557917	21.127	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.451	216	278260	22.283	ng/ul	98
40) Hexachlorocyclopentadiene	12.428	237	131697	19.199	ng/ul	93
41) 2,4,6-Trichlorophenol	12.698	196	182268	21.928	ng/ul	91
42) 2,4,5-Trichlorophenol	12.769	196	202639	21.988	ng/ul	97
43) 1,1'-Biphenyl	13.104	154	729635	22.230	ng/ul	99
44) 2-Chloronaphthalene	13.145	162	580520	22.134	ng/ul	98
45) 2-Nitroaniline	13.363	65	163087	22.524	ng/ul	97
47) Dimethylphthalate	13.745	163	723034	21.104	ng/ul	100
48) 2,6-Dinitrotoluene	13.863	165	141125	21.944	ng/ul	94
50) Acenaphthylene	13.998	152	973190	22.138	ng/ul	99
51) 3-Nitroaniline	14.192	138	166521	22.554	ng/ul	95
52) Acenaphthene	14.345	153	638425	21.908	ng/ul	97
53) 2,4-Dinitrophenol	14.404	184	61948	17.033	ng/ul	97
55) 4-Nitrophenol	14.504	109	121831	21.658	ng/ul	97
56) Dibenzofuran	14.681	168	874477	21.704	ng/ul	98
57) 2,4-Dinitrotoluene	14.651	165	218612	22.371	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.910	232	169066	21.227	ng/ul	95
59) Diethylphthalate	15.116	149	752341	21.353	ng/ul	99
61) Fluorene	15.333	166	713881	21.530	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.333	204	344785	21.831	ng/ul	98
63) 4-Nitroaniline	15.363	138	186302	23.404	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.416	198	116105	19.855	ng/ul	96
67) N-Nitrosodiphenylamine	15.545	169	627333	21.683	ng/ul	98
68) 4-Bromophenyl-phenylether	16.228	248	213494	20.914	ng/ul	94
69) Hexachlorobenzene	16.328	284	262377	22.295	ng/ul	99
70) Atrazine	16.504	200	219442	21.603	ng/ul	100
71) Pentachlorophenol	16.675	266	160406	20.979	ng/ul	98
72) Phenanthrene	17.075	178	1242321	21.759	ng/ul	100
74) Anthracene	17.163	178	1265511	21.903	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.069	216	281797	22.257	ng/uL	98
76) Pentachlorobenzene	14.598	250	281811	21.554	ng/uL	93
77) Carbazole	17.439	167	1210521	22.936	ng/ul	99
78) Di-n-butylphthalate	18.004	149	1402292	21.798	ng/ul	100
80) Fluoranthene	19.098	202	1563785	21.523	ng/ul	99
82) Pyrene	19.463	202	1655794	21.690	ng/ul	99
83) Butylbenzylphthalate	20.374	149	718258	21.323	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.180	252	583030	22.398	ng/ul	92
85) Benzo(a)anthracene	21.245	228	1837782	21.663	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	1075061	21.257	ng/ul	99
87) Chrysene	21.310	228	1750432	21.737	ng/ul	98
89) Di-n-octyl phthalate	22.280	149	1996133	22.152	ng/ul	100
90) Benzo(b)fluoranthene	23.245	252	2006295	22.809	ng/ul	98
91) Benzo(k)fluoranthene	23.310	252	1984220	22.429	ng/ul	98
93) Benzo(a)pyrene	24.033	252	1896471	21.612	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.409	276	966926	8.880	ng/ul	97
95) Dibenzo(a,h)anthracene	27.474	278	1841352	20.603	ng/ul	96
96) Benzo(g,h,i)perylene	28.439	276	1737488	19.815	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 21 22:22:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 8.702 ng/uL

RT: 3.181 min Scan# 31

Delta R.T. 0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion: 88 Resp: 30180

Ion Ratio Lower Upper

88 100

43 27.2 22.2 33.2

58 66.4 50.2 75.2



#5

Pyridine

Concen: 22.430 ng/uL

RT: 3.569 min Scan# 99

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion: 79 Resp: 213321


Ion Ratio Lower Upper

79 100

52 74.1 59.9 89.9

51 34.7 29.0 43.6









#6
Benzaldehyde
Concen: 25.409 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion: 77 Resp: 160914
Ion Ratio Lower Upper
77 100
105 96.1 80.0 120.0
106 95.1 75.0 112.6



#8
Phenol
Concen: 21.450 ng/ul
RT: 6.834 min Scan# 654
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion: 94 Resp: 285229
Ion Ratio Lower Upper
94 100
65 26.9 22.0 33.0
66 35.9 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.374 ng/ul
RT: 7.069 min Scan# 694
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion: 93 Resp: 224259
Ion Ratio Lower Upper
93 100
63 71.1 57.1 85.7
95 30.9 24.9 37.3



#12
2-Chlorophenol
Concen: 21.866 ng/ul
RT: 7.199 min Scan# 716
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:128 Resp: 228255
Ion Ratio Lower Upper
128 100
64 43.4 34.9 52.3
130 33.5 27.0 40.4









#13
2-Methylphenol
Concen: 20.720 ng/ul
RT: 8.075 min Scan# 80
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:108 Resp: 213514
Ion Ratio Lower Upper
108 100
107 90.2 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 21.866 ng/ul
RT: 8.151 min Scan# 878
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion: 45 Resp: 299108
Ion Ratio Lower Upper
45 100
77 15.8 12.5 18.7
79 11.3 9.6 14.4









#16
Acetophenone
Concen: 21.203 ng/ul
RT: 8.451 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:105 Resp: 356338
Ion Ratio Lower Upper
105 100
77 75.4 60.2 90.4
51 25.6 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 20.593 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion: 70 Resp: 172507
Ion Ratio Lower Upper
70 100
42 57.9 45.0 67.6
101 10.0 8.1 12.1
130 23.1 19.2 28.8









#18
4-Methylphenol
Concen: 20.720 ng/ul
RT: 8.399 min Scan# 911
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:108 Resp: 239171
Ion Ratio Lower Upper
108 100
107 114.6 89.8 134.8



#19
Hexachloroethane
Concen: 21.782 ng/ul
RT: 8.699 min Scan# 971
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:117 Resp: 94350
Ion Ratio Lower Upper
117 100
201 92.4 77.4 116.0
199 56.2 49.6 74.4









#22
Nitrobenzene
Concen: 22.212 ng/ul
RT: 8.828 min Scan# 993
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion: 77 Resp: 268980
Ion Ratio Lower Upper
77 100
123 46.1 38.2 57.2
65 13.4 10.7 16.1



#23
Isophorone
Concen: 20.352 ng/ul
RT: 9.351 min Scan# 1082
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion: 82 Resp: 502947
Ion Ratio Lower Upper
82 100
95 7.2 5.9 8.9
138 18.5 14.7 22.1









#25
2-Nitrophenol
Concen: 22.135 ng/ul
RT: 9.534 min Scan# 1113
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:139 Resp: 130613
Ion Ratio Lower Upper
139 100
65 48.3 35.2 52.8
109 30.9 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 20.289 ng/ul
RT: 9.598 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:107 Resp: 244162
Ion Ratio Lower Upper
107 100
121 51.2 40.3 60.5
122 88.5 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.099 ng/ul

RT: 9.840 min Scan# 1164

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion: 93 Resp: 316704

Ion Ratio Lower Upper

93 100

95 31.5 25.8 38.6

123 12.5 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 21.601 ng/ul

RT: 10.063 min Scan# 1203

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion: 162 Resp: 226920


Ion Ratio Lower Upper

162 100

164 63.7 51.0 76.6

98 35.8 28.6 42.8









#30
Naphthalene
Concen: 21.608 ng/ul
RT: 10.463 min Scan# 1271
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:128 Resp: 799688
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.4
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 21.139 ng/ul
RT: 10.581 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:127 Resp: 337583
Ion Ratio Lower Upper
127 100
129 32.7 25.4 38.2









#33
Hexachlorobutadiene
Concen: 21.348 ng/ul
RT: 10.740 min Scan# 1318
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:225 Resp: 136710
Ion Ratio Lower Upper
225 100
223 61.7 49.8 74.6
227 59.5 48.8 73.2



#34
Caprolactam
Concen: 18.961 ng/ul
RT: 11.357 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:113 Resp: 78479
Ion Ratio Lower Upper
113 100
55 149.3 112.4 168.6
56 107.2 88.0 132.0









#35
4-Chloro-3-methylphenol
Concen: 20.674 ng/ul
RT: 11.704 min Scan# 1482
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:107 Resp: 251806
Ion Ratio Lower Upper
107 100
144 25.5 20.6 31.0
142 85.1 67.4 101.2



#36
2-Methylnaphthalene
Concen: 20.789 ng/ul
RT: 12.081 min Scan# 1546
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:142 Resp: 542572
Ion Ratio Lower Upper
142 100
141 89.7 69.6 104.4









#37
1-Methylnaphthalene
Concen: 21.127 ng/ul
RT: 12.304 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:142 Resp: 557917
Ion Ratio Lower Upper
142 100
141 88.6 70.2 105.4
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.283 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:216 Resp: 278260
Ion Ratio Lower Upper
216 100
214 76.2 63.1 94.7
179 21.1 17.0 25.6
108 18.2 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 19.199 ng/ul
RT: 12.428 min Scan# 1605
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:237 Resp: 131697
Ion Ratio Lower Upper
237 100
235 65.6 47.6 71.4
272 14.0 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 21.928 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:196 Resp: 182268
Ion Ratio Lower Upper
196 100
198 97.0 69.9 104.9
200 31.6 23.4 35.2









#42
2,4,5-Trichlorophenol
Concen: 21.988 ng/ul
RT: 12.769 min Scan# 1663
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:196 Resp: 202639
Ion Ratio Lower Upper
196 100
198 92.5 75.9 113.9
200 29.4 24.7 37.1



#43
1,1'-Biphenyl
Concen: 22.230 ng/ul
RT: 13.104 min Scan# 1720
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:154 Resp: 729635
Ion Ratio Lower Upper
154 100
153 40.6 31.8 47.8
76 13.0 10.6 15.8









#44
2-Chloronaphthalene
Concen: 22.134 ng/ul
RT: 13.145 min Scan# 1727
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Instrument :
BNA_N
ClientSampleId :
SSTD020372

Tgt Ion:162 Resp: 580520




Ion	Ratio	Lower	Upper
162	100		
164	31.8	25.8	38.8
127	40.0	33.7	50.5




#45
2-Nitroaniline
Concen: 22.524 ng/ul
RT: 13.363 min Scan# 1764
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion: 65 Resp: 163087

Ion	Ratio	Lower	Upper
65	100		
92	70.4	57.4	86.2
138	109.8	91.6	137.4





#47

Dimethylphthalate

Concen: 21.104 ng/ul

RT: 13.745 min Scan# 1829

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:163 Resp: 723034

Ion Ratio Lower Upper

163 100

194 4.3 3.0 4.6

164 10.0 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 21.944 ng/ul

RT: 13.863 min Scan# 1849

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:165 Resp: 141125


Ion Ratio Lower Upper

165 100

89 60.9 44.7 67.1

121 23.3 17.0 25.4









#50
Acenaphthylene
Concen: 22.138 ng/ul
RT: 13.998 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Instrument : BNA_N
ClientSampleId : SSTD020372

Tgt Ion:152 Resp: 973190




Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.8	23.8
153	12.7	10.5	15.7




#51
3-Nitroaniline
Concen: 22.554 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:138 Resp: 166521

Ion	Ratio	Lower	Upper
138	100		
108	12.0	11.1	16.7
92	112.4	93.9	140.9









#52
 Acenaphthene
 Concen: 21.908 ng/ul
 RT: 14.345 min Scan# 1931
 Delta R.T. -0.000 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020372


Tgt Ion:	153	Resp:	638425
Ion Ratio	Lower	Upper	
153	100		
152	48.2	36.3	54.5
154	87.8	72.4	108.6



#53
 2,4-Dinitrophenol
 Concen: 17.033 ng/ul
 RT: 14.404 min Scan# 1941
 Delta R.T. -0.000 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24

Tgt Ion:	184	Resp:	61948
Ion Ratio	Lower	Upper	
184	100		
63	61.4	50.4	75.6
154	63.5	53.0	79.6









#55
4-Nitrophenol
Concen: 21.658 ng/ul
RT: 14.504 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:109 Resp: 121831
Ion Ratio Lower Upper
109 100
139 136.3 105.6 158.4
65 126.3 98.2 147.2



#56
Dibenzofuran
Concen: 21.704 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:168 Resp: 874477
Ion Ratio Lower Upper
168 100
139 37.4 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 22.371 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:165 Resp: 218612
Ion Ratio Lower Upper
165 100
63 45.0 33.9 50.9
182 3.0 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.227 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:232 Resp: 169066
Ion Ratio Lower Upper
232 100
131 48.3 41.2 61.8
130 2.9 2.2 3.2
166 30.7 27.3 40.9





#59

Diethylphthalate

Concen: 21.353 ng/ul

RT: 15.116 min Scan# 2062

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:149 Resp: 752341

Ion Ratio Lower Upper

149 100

177 21.5 16.8 25.2

150 12.4 10.1 15.1



#61

Fluorene

Concen: 21.530 ng/ul

RT: 15.333 min Scan# 2099

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:166 Resp: 713881


Ion Ratio Lower Upper

166 100

165 95.5 74.6 112.0

167 13.8 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.831 ng/ul
RT: 15.333 min Scan# 2099
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:204 Resp: 344785
Ion Ratio Lower Upper
204 100
206 33.4 26.4 39.6
141 61.8 50.7 76.1



#63
4-Nitroaniline
Concen: 23.404 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:138 Resp: 186302
Ion Ratio Lower Upper
138 100
92 48.6 38.4 57.6
108 78.5 63.8 95.8









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.855 ng/ul
 RT: 15.416 min Scan# 2113
 Delta R.T. -0.000 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SST020372


Tgt Ion:198 Resp: 116105
 Ion Ratio Lower Upper
 198 100
 182 4.8 4.0 6.0
 77 25.6 22.2 33.4



#67
 N-Nitrosodiphenylamine
 Concen: 21.683 ng/ul
 RT: 15.545 min Scan# 2135
 Delta R.T. -0.006 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24

Tgt Ion:169 Resp: 627333
 Ion Ratio Lower Upper
 169 100
 168 64.1 52.6 79.0
 167 35.0 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 20.914 ng/ul
RT: 16.228 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:248 Resp: 213494
Ion Ratio Lower Upper
248 100
250 101.5 78.2 117.2
141 81.7 71.4 107.0



#69
Hexachlorobenzene
Concen: 22.295 ng/ul
RT: 16.328 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:284 Resp: 262377
Ion Ratio Lower Upper
284 100
142 37.2 30.0 45.0
249 30.7 25.5 38.3





#70

Atrazine

Concen: 21.603 ng/ul

RT: 16.504 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:200 Resp: 219442

Ion Ratio Lower Upper

200 100

173 25.8 20.6 31.0

215 46.7 37.6 56.4



#71

Pentachlorophenol

Concen: 20.979 ng/ul

RT: 16.675 min Scan# 2327

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:266 Resp: 160406


Ion Ratio Lower Upper

266 100

264 64.7 49.1 73.7

268 61.4 49.2 73.8









#72
 Phenanthrene
 Concen: 21.759 ng/ul
 RT: 17.075 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24

Instrument :
 BNA_N
 ClientSampleId :
 SST020372

Tgt Ion:178 Resp: 1242321




Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.4	18.6
176	18.5	14.6	21.8




#74
 Anthracene
 Concen: 21.903 ng/ul
 RT: 17.163 min Scan# 2410
 Delta R.T. -0.006 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24

Tgt Ion:178 Resp: 1265511

Ion	Ratio	Lower	Upper
178	100		
179	14.4	12.2	18.4
176	18.5	14.4	21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.257 ng/uL

RT: 13.069 min Scan# 1

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:216 Resp: 281797

Ion Ratio Lower Upper

216 100

214 75.8 63.0 94.6

218 48.4 38.3 57.5



#76

Pentachlorobenzene

Concen: 21.554 ng/uL

RT: 14.598 min Scan# 1974

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:250 Resp: 281811


Ion Ratio Lower Upper

250 100

248 60.0 49.3 73.9

252 57.5 52.9 79.3









#77
 Carbazole
 Concen: 22.936 ng/ul
 RT: 17.439 min Scan# 2457
 Delta R.T. -0.000 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020372


Tgt Ion:167 Resp: 1210521
 Ion Ratio Lower Upper
 167 100
 166 20.5 16.2 24.2
 139 12.6 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 21.798 ng/ul
 RT: 18.004 min Scan# 2553
 Delta R.T. -0.006 min
 Lab File: BN031113.D
 Acq: 21 Apr 2024 18:24

Tgt Ion:149 Resp: 1402292
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.1 10.7
 104 5.1 3.9 5.9









#80
Fluoranthene
Concen: 21.523 ng/ul
RT: 19.098 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:202 Resp: 1563785
Ion Ratio Lower Upper
202 100
101 10.9 9.0 13.6
100 8.3 6.9 10.3



#82
Pyrene
Concen: 21.690 ng/ul
RT: 19.463 min Scan# 2801
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:202 Resp: 1655794
Ion Ratio Lower Upper
202 100
101 13.9 11.1 16.7
100 10.9 9.3 13.9





#83

Butylbenzylphthalate

Concen: 21.323 ng/ul

RT: 20.374 min Scan# 2956

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:149 Resp: 718258

Ion Ratio Lower Upper

149 100

91 72.5 57.3 85.9

206 21.0 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 22.398 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:252 Resp: 583030


Ion Ratio Lower Upper

252 100

254 60.4 54.5 81.7

126 12.8 9.8 14.6





#85

Benzo(a)anthracene

Concen: 21.663 ng/ul

RT: 21.245 min Scan# 3104

Delta R.T. -0.006 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:228 Resp: 1837782

Ion Ratio Lower Upper

228 100

229 20.1 16.7 25.1

226 26.4 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.257 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:149 Resp: 1075061


Ion Ratio Lower Upper

149 100

167 27.4 22.3 33.5

279 4.4 3.0 4.6





#87

Chrysene

Concen: 21.737 ng/ul

RT: 21.310 min Scan# 3114

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:228 Resp: 1750432

Ion Ratio Lower Upper

228 100

226 30.1 22.6 34.0

229 20.3 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 22.152 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Tgt Ion:149 Resp: 1996133





#90

Benzo(b)fluoranthene

Concen: 22.809 ng/ul

RT: 23.245 min Scan# 3444

Delta R.T. -0.012 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:252 Resp: 2006295

Ion Ratio Lower Upper

252 100

253 20.9 17.6 26.4

125 11.2 8.6 12.8



#91

Benzo(k)fluoranthene

Concen: 22.429 ng/ul

RT: 23.310 min Scan# 3455

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24




Tgt Ion:252 Resp: 1984220


Ion Ratio Lower Upper

252 100

253 22.4 17.4 26.0

125 10.8 9.4 14.2









#93
Benzo(a)pyrene
Concen: 21.612 ng/ul
RT: 24.033 min Scan# 31
Delta R.T. -0.000 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24




Instrument :
BNA_N
ClientSampleId :
SSTD020372


Tgt Ion:252 Resp: 1896471
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 12.8 9.9 14.9



#94
Indeno(1,2,3-cd)pyrene
Concen: 8.880 ng/ul
RT: 27.409 min Scan# 4152
Delta R.T. -0.012 min
Lab File: BN031113.D
Acq: 21 Apr 2024 18:24

Tgt Ion:276 Resp: 966926
Ion Ratio Lower Upper
276 100
138 22.8 17.4 26.2
277 24.4 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 20.603 ng/ul

RT: 27.474 min Scan# 4163

Delta R.T. -0.012 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Instrument :

BNA_N

ClientSampleId :

SSTD020372





Tgt Ion:278 Resp: 1841352

Ion Ratio Lower Upper

278 100

139 17.4 13.3 19.9

279 24.0 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 19.815 ng/ul

RT: 28.439 min Scan# 4327

Delta R.T. -0.000 min

Lab File: BN031113.D

Acq: 21 Apr 2024 18:24

Tgt Ion:276 Resp: 1737488

Ion Ratio Lower Upper

276 100

138 23.1 19.0 28.6

277 23.7 19.3 28.9

