

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031058.D
 Acq On : 20 Apr 2024 03:38
 Operator : MA/JU
 Sample : PB160348BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK348

Quant Time: Apr 20 04:13:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	83003	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	373768	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	254822	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.033	188	595505	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	760262	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	1005683	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	10387	5.847	ng/uL	0.00
4) Pyridine-d5	3.552	84	121925	23.830	ng/u1	0.00
7) Phenol-d5	6.810	99	212369	30.330	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	128045	31.432	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	171596	31.475	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	174793	29.684	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	87192	32.468	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	89685	32.128	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	170256	30.928	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	221347	25.730	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	650462	36.682	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	724504	36.055	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	114473	30.929	ng/u1	0.00
60) Fluorene-d10	15.281	176	572349	37.077	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.398	200	73622	24.504	ng/u1	0.00
73) Anthracene-d10	17.133	188	957162	37.083	ng/u1	0.00
81) Pyrene-d10	19.433	212	1272504	34.062	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	1803905	37.826	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
Data File : BN031058.D
Acq On : 20 Apr 2024 03:38
Operator : MA/JU
Sample : PB160348BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK348

Quant Time: Apr 20 04:13:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration

