

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
 Data File : BN031086.D
 Acq On : 20 Apr 2024 23:23
 Operator : MA/JU
 Sample : P2077-12
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AL9

Quant Time: Apr 21 14:13:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	148261	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	679527	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	466499	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.033	188	1079080	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	1264498	20.000	ng/u1	0.00
88) Perylene-d12	24.168	264	1573463	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	12195	3.843	ng/uL	0.00
4) Pyridine-d5	3.558	84	75477	8.259	ng/u1	0.00
7) Phenol-d5	6.810	99	78184	6.251	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	241687	33.214	ng/u1	0.00
11) 2-Chlorophenol-d4	7.169	132	246633	25.327	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	162417	15.442	ng/u1	0.00
21) Nitrobenzene-d5	8.793	128	169491	34.715	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	160554	31.636	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.034	165	299329	29.908	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	456504	29.188	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	1155486	35.594	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	1315146	35.751	ng/u1	0.00
54) 4-Nitrophenol-d4	14.498	143	31935	4.713	ng/u1	0.00
60) Fluorene-d10	15.281	176	1054611	37.319	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	111060	20.400	ng/u1	0.00
73) Anthracene-d10	17.128	188	1760281	37.636	ng/u1	0.00
81) Pyrene-d10	19.433	212	2259333	36.361	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.974	264	2869764	38.461	ng/u1	0.00
Target Compounds						
35) 4-Chloro-3-methylphenol	11.710	107	16707	1.441	ng/u1	92
59) Diethylphthalate	15.116	149	2735865	79.435	ng/u1	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN041924\
Data File : BN031086.D
Acq On : 20 Apr 2024 23:23
Operator : MA/JU
Sample : P2077-12
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
COAL9

Quant Time: Apr 21 14:13:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 19 22:56:25 2024
Response via : Initial Calibration









#35
4-Chloro-3-methylphenol
Concen: 1.441 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. -0.000 min
Lab File: BN031086.D
Acq: 20 Apr 2024 23:23

Instrument :
BNA_N
ClientSampleId :
COAL9

Tgt Ion:107 Resp: 16707
Ion Ratio Lower Upper
107 100
144 24.3 20.6 31.0
142 76.0 67.4 101.2



#59
Diethylphthalate
Concen: 79.435 ng/ul
RT: 15.116 min Scan# 2062
Delta R.T. -0.000 min
Lab File: BN031086.D
Acq: 20 Apr 2024 23:23

Tgt Ion:149 Resp: 2735865
Ion Ratio Lower Upper
149 100
177 21.0 16.8 25.2
150 12.5 10.1 15.1

