

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042023\
 Data File : BN025021.D
 Acq On : 20 Apr 2023 04:10
 Operator : CG/JU
 Sample : SST0.194
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST0.1209

Quant Time: Apr 20 07:46:53 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 20 07:43:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	7198	0.400	ng/ul	0.00
4) Naphthalene-d8	10.742	136	22104	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.575	164	12969	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.326	188	26444	0.400	ng/ul	0.00
17) Chrysene-d12	21.512	240	22332	0.400	ng/ul	0.00
23) Perylene-d12	23.953	264	18107	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	0.000	96	0d	0.000	ng/ul	
6) 2-Methylnaphthalene-d10	12.337	152	2765	0.097	ng/ul	0.00
18) Fluoranthene-d10	19.355	212	6035	0.109	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.792	128	5907	0.105	ng/ul	97
7) 2-Methylnaphthalene	12.409	142	3370	0.096	ng/ul	99
8) 1-Methylnaphthalene	12.623	142	3691	0.099	ng/ul	99
10) Acenaphthylene	14.298	152	4607	0.095	ng/ul	96
11) Acenaphthene	14.640	153	4100	0.098	ng/ul	99
12) Fluorene	15.630	166	4448	0.095	ng/ul	98
15) Phenanthrene	17.369	178	7882	0.106	ng/ul	97
16) Anthracene	17.462	178	5767	0.093	ng/ul	96
19) Fluoranthene	19.383	202	8178	0.106	ng/ul	96
20) Pyrene	19.745	202	8574	0.109	ng/ul	97
21) Benzo(a)anthracene	21.500	228	6519	0.092	ng/ul	98
22) Chrysene	21.550	228	7343	0.096	ng/ul	98
24) Benzo(b)fluoranthene	23.210	252	7248	0.100	ng/ul	77
25) Benzo(k)fluoranthene	23.257	252	6849	0.097	ng/ul#	75
26) Benzo(a)pyrene	23.853	252	6316	0.105	ng/ul#	62
27) Indeno(1,2,3-cd)pyrene	26.518	276	6876	0.097	ng/ul#	99
28) Dibenzo(a,h)anthracene	26.541	278	5240	0.094	ng/ul#	84
29) Benzo(g,h,i)perylene	27.289	276	5974	0.099	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042023\
Data File : BN025021.D
Acq On : 20 Apr 2023 04:10
Operator : CG/JU
Sample : SST00.194
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.1209

Quant Time: Apr 20 07:46:53 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 20 07:43:17 2023
Response via : Initial Calibration

