

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042120\
 Data File : BN010560.D
 Acq On : 21 Apr 2020 11:07
 Operator : CG/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDC00.4

Quant Time: Apr 21 13:14:29 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 13 18:04:22 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2	1,4-Dioxane	0.294	0.287	2.4	75	0.00
3	n-Nitrosodimethylamine	0.191	0.194	-1.6	92	0.00
4 S	2-Fluorophenol	0.876	0.924	-5.5	80	0.00
5 S	Phenol-d6	1.145	1.150	-0.4	76	0.00
6	bis(2-Chloroethyl)ether	0.992	1.034	-4.2	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
8 S	Nitrobenzene-d5	0.280	0.272	2.9	80	0.00
9	Naphthalene	0.993	1.012	-1.9	81	0.00
10	Hexachlorobutadiene	0.239	0.246	-2.9	80	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.755	-14.7	91	0.00
12	2-Methylnaphthalene	0.714	0.731	-2.4	81	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.189	0.150	20.6	67	0.00
15 S	2-Fluorobiphenyl	1.456	1.472	-1.1	81	0.00
16	Acenaphthylene	1.728	1.603	7.2	76	0.00
17	Acenaphthene	1.110	1.116	-0.5	82	0.00
18	Fluorene	1.454	1.456	-0.1	81	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
20	4-Bromophenyl-phenylether	0.222	0.217	2.3	81	0.00
21	Hexachlorobenzene	0.254	0.270	-6.3	85	0.00
22	Pentachlorophenol	0.100	0.071	29.0#	68	0.00
23	Phenanthrene	1.080	1.100	-1.9	82	0.00
24	Anthracene	0.988	0.908	8.1	76	0.00
25 SURR	Fluoranthene-d10	1.213	1.308	-7.8	88	0.00
26	Fluoranthene	1.345	1.321	1.8	79	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
28	Pvrene	1.346	1.427	-6.0	77	0.00
29 S	Terphenyl-d14	0.915	0.972	-6.2	78	0.00
30	Benzo(a)anthracene	1.305	1.179	9.7	65	0.00
31	Chrysene	1.300	1.316	-1.2	74	0.00
32	Bis(2-ethylhexyl)phthalate	0.670	0.456	31.9#	52	0.00
33	Indeno(1,2,3-cd)pyrene	1.149	1.240	-7.9	77	0.00
34 I	Perylene-d12	1.000	1.000	0.0	71	0.00
35	Benzo(b)fluoranthene	1.317	1.341	-1.8	73	0.00
36	Benzo(k)fluoranthene	1.320	1.353	-2.5	74	0.00
37 C	Benzo(a)pyrene	1.166	1.137	2.5	70	0.00
38	Dibenzo(a,h)anthracene	1.066	1.130	-6.0	76	0.00
39	Benzo(a,h,i)perylene	1.073	1.192	-11.1	79	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042120\
Data File : BN010560.D
Acq On : 21 Apr 2020 11:07
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 21 13:14:29 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN041320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 13 18:04:22 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			