

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042123\
 Data File : BN025049.D
 Acq On : 21 Apr 2023 11:05
 Operator : CG/JU
 Sample : 02340-10
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHE10

Quant Time: Apr 22 00:15:54 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 21 04:13:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.930	152	9370	0.400	ng/ul	0.00
4) Naphthalene-d8	10.737	136	30236	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.571	164	18500	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.318	188	37637	0.400	ng/ul	0.00
17) Chrysene-d12	21.515	240	28569	0.400	ng/ul	0.00
23) Perylene-d12	23.956	264	29842	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.306	96	43836	4.518	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.326	152	11597	0.299	ng/ul	0.00
18) Fluoranthene-d10	19.346	212	31773	0.422	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042123\
Data File : BN025049.D
Acq On : 21 Apr 2023 11:05
Operator : CG/JU
Sample : 02340-10
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHE10

Quant Time: Apr 22 00:15:54 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN042023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 21 04:13:07 2023
Response via : Initial Calibration

