

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014350.D
 Acq On : 21 Apr 2021 10:57
 Operator : CG/JU
 Sample : SSTDICC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Quant Time: Apr 21 12:49:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 12:48:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	8172	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	28454	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	14962	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	30873	0.40	ng	0.00
29) Chrysene-d12	21.258	240	29260	0.40	ng	0.00
36) Perylene-d12	23.490	264	24348	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	6863	0.38	ng	0.00
5) Phenol-d6	6.863	99	8254	0.36	ng	0.00
8) Nitrobenzene-d5	8.832	82	9314	0.53	ng	0.00
11) 2-Methylnaphthalene-d10	12.058	152	16958	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1757	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	22766	0.39	ng	0.00
27) Fluoranthene-d10	19.101	212	31420	0.37	ng	0.00
31) Terphenyl-d14	19.707	244	25013	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.255	88	3958	0.38	ng	99
3) n-Nitrosodimethylamine	3.601	42	2136	0.37	ng	# 100
6) bis(2-Chloroethyl)ether	7.121	93	8277	0.40	ng	100
9) Naphthalene	10.518	128	28146	0.39	ng	100
10) Hexachlorobutadiene	10.809	225	6248	0.44	ng	# 100
12) 2-Methylnaphthalene	12.133	142	18290	0.39	ng	100
16) Acenaphthylene	14.042	152	24098	0.42	ng	100
17) Acenaphthene	14.384	154	16128	0.38	ng	100
18) Fluorene	15.374	166	19583	0.38	ng	100
20) 4,6-Dinitro-2-methylph...	15.450	198	1293	0.40	ng	100
21) 4-Bromophenyl-phenylether	16.264	248	6743	0.41	ng	100
22) Hexachlorobenzene	16.373	284	9319	0.44	ng	100
23) Atrazine	16.532	200	3645	0.36	ng	100
24) Pentachlorophenol	16.714	266	1521	0.27	ng	100
25) Phenanthrene	17.104	178	33650	0.39	ng	100
26) Anthracene	17.201	178	25479	0.37	ng	100
28) Fluoranthene	19.131	202	35590	0.37	ng	100
30) Pyrene	19.498	202	34264	0.41	ng	100
32) Benzo(a)anthracene	21.247	228	29993	0.38	ng	100
33) Chrysene	21.300	228	37224	0.40	ng	100
34) Bis(2-ethylhexyl)phtha...	21.184	149	9784	0.45	ng	100
35) Indeno(1,2,3-cd)pyrene	25.725	276	38048	0.33	ng	100
37) Benzo(b)fluoranthene	22.819	252	35155	0.33	ng	100
38) Benzo(k)fluoranthene	22.864	252	35182	0.34	ng	100
39) Benzo(a)pyrene	23.391	252	27552	0.35	ng	100
40) Dibenzo(a,h)anthracene	25.739	278	32033	0.35	ng	100
41) Benzo(g,h,i)perylene	26.410	276	35534	0.36	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014350.D
 Acq On : 21 Apr 2021 10:57
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Apr 21 12:49:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 12:48:18 2021
 Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014350.D
 Acq: 21 Apr 2021 10:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:	152	Resp:	8172
Ion	Ratio	Lower	Upper
152	100		
150	154.3	123.4	185.2
115	59.4	47.5	71.3



1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.255 min Scan# 59
 Delta R.T. 0.000 min
 Lab File: BN014350.D
 Acq: 21 Apr 2021 10:57

Tgt Ion:	88	Resp:	3958
Ion	Ratio	Lower	Upper
88	100		
58	66.2	52.2	78.2
43	25.8	20.6	30.8









n-Nitrosodimethylamine
Concen: 0.37 ng
RT: 3.601 min Scan# 102
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion: 42 Resp: 2136




Ion	Ratio	Lower	Upper
42	100		
74	135.0	108.0	162.0
44	0.0	0.0	0.0




2-Fluorophenol
Concen: 0.38 ng
RT: 5.292 min Scan# 312
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Tgt Ion: 112 Resp: 6863

Ion	Ratio	Lower	Upper
112	100		
64	53.8	43.0	64.6
63	29.2	23.4	35.0









Phenol-d6
 Concen: 0.36 ng
 RT: 6.863 min Scan# 507
 Delta R.T. 0.000 min
 Lab File: BN014350.D
 Acq: 21 Apr 2021 10:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Tgt Ion:	99	Resp:	8254
Ion	Ratio	Lower	Upper
99	100		
42	20.0	16.0	24.0
71	34.8	27.8	41.8



bis(2-Chloroethyl)ether
 Concen: 0.40 ng
 RT: 7.121 min Scan# 539
 Delta R.T. 0.000 min
 Lab File: BN014350.D
 Acq: 21 Apr 2021 10:57


Tgt Ion:	93	Resp:	8277
Ion	Ratio	Lower	Upper
93	100		
63	74.3	59.4	89.2
95	33.7	27.0	40.4






Abundance Scan 857 (10.467 min): BN014350.D\data.ms (-857) (-)

Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4






Tgt Ion:	136	Resp:	28454
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.6
54	5.5	4.4	6.6
68	4.9	3.9	5.9



Abundance Scan 728 (8.832 min): BN014350.D\data.ms (-725) (-)

Nitrobenzene-d5
Concen: 0.53 ng
RT: 8.832 min Scan# 728
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Tgt Ion:	82	Resp:	9314
Ion	Ratio	Lower	Upper
82	100		
128	37.9	30.3	45.5
54	44.8	35.8	53.8






Abundance Scan 861 (10.518 min): BN014350.D\data.ms (-855) (-)

Naphthalene
Concen: 0.39 ng
RT: 10.518 min Scan# 861
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	128	Resp:	28146
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.0	13.6
127	12.9	10.3	15.5






Abundance Scan 884 (10.809 min): BN014350.D\data.ms (-879) (-)

Hexachlorobutadiene
Concen: 0.44 ng
RT: 10.809 min Scan# 884
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57




Tgt Ion:	225	Resp:	6248
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.2	50.6	75.8






Abundance Scan 1063 (12.058 min): BN014350.D\data.ms (-1011) (-)

2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.058 min Scan# 1063
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:152 Resp: 16958
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5






Abundance Scan 1080 (12.133 min): BN014350.D\data.ms (-1010) (-)

2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.133 min Scan# 1080
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57




Tgt Ion:142 Resp: 18290
Ion Ratio Lower Upper
142 100
141 88.2 70.6 105.8
115 36.3 29.0 43.6






Abundance Scan 1307 (14.321 min): BN014350.D\data.ms (-1291) (-)

Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion	Ratio	Lower	Upper
164	100		
162	101.3	81.0	121.6
160	50.9	40.7	61.1






Abundance Scan 1425 (15.818 min): BN014350.D\data.ms (-1417) (-)


2,4,6-Tribromophenol
Concen: 0.32 ng
RT: 15.818 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57



Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	32.7	26.2	39.2





Abundance Scan 1198 (12.938 min): BN014350.D\data.ms (-1198) (-)




2-Fluorobiphenyl
Concen: 0.39 ng
RT: 12.938 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	172	Resp:	22766
Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.2	42.2
170	24.5	19.6	29.4



Abundance Scan 1198 (12.938 min): BN014350.D\data.ms (-1190) (-)





Abundance Scan 1285 (14.042 min): BN014350.D\data.ms (-1216) (-)





Acenaphthylene
Concen: 0.42 ng
RT: 14.042 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Tgt Ion:	152	Resp:	24098
Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.6	23.4
153	13.1	10.5	15.7



Abundance Scan 1285 (14.042 min): BN014350.D\data.ms (-1269) (-)









Acenaphthene
 Concen: 0.38 ng
 RT: 14.384 min Scan# 1312
 Delta R.T. 0.000 min
 Lab File: BN014350.D
 Acq: 21 Apr 2021 10:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4

Tgt Ion:	154	Resp:	16128
Ion Ratio	Lower	Upper	
154	100		
153	111.5	89.2	133.8
152	61.2	49.0	73.4



Fluorene
 Concen: 0.38 ng
 RT: 15.374 min Scan# 1390
 Delta R.T. 0.000 min
 Lab File: BN014350.D
 Acq: 21 Apr 2021 10:57


Tgt Ion:	166	Resp:	19583
Ion Ratio	Lower	Upper	
166	100		
165	98.3	78.6	118.0
167	13.6	10.9	16.3





Abundance Scan 1527 (17.067 min): BN014350.D\data.ms (-1511) (-)

Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.8	6.2	9.4






Abundance Scan 1527 (17.067 min): BN014350.D\data.ms (-1511) (-)




Abundance Scan 1396 (15.450 min): BN014350.D\data.ms (-1320) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.40 ng
RT: 15.450 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Tgt Ion	Ratio	Lower	Upper
198	100		
51	75.0	60.0	90.0
105	39.4	31.5	47.3




Abundance Scan 1396 (15.450 min): BN014350.D\data.ms (-1388) (-)






Abundance Scan 1461 (16.264 min): BN014350.D\data.ms (-1427) (-)

4-Bromophenyl-phenylether
Concen: 0.41 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4







Tgt Ion:	248	Resp:	6743
Ion	Ratio	Lower	Upper
248	100		
250	96.7	77.4	116.0
141	51.7	41.4	62.0



Abundance Scan 1470 (16.373 min): BN014350.D\data.ms (-1429) (-)

Hexachlorobenzene
Concen: 0.44 ng
RT: 16.373 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Tgt Ion:	284	Resp:	9319
Ion	Ratio	Lower	Upper
284	100		
142	33.2	26.5	39.7
249	30.0	24.0	36.0






Abundance Scan 1483 (16.532 min): BN014350.D\data.ms (-1429) (-)

Atrazine
Concen: 0.36 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion:	200	Resp:	3645
Ion Ratio	Lower	Upper	
200	100		
173	26.3	21.0	31.6
215	48.4	38.7	58.1






Abundance Scan 1498 (16.714 min): BN014350.D\data.ms (-1429) (-)


Pentachlorophenol
Concen: 0.27 ng
RT: 16.714 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57



Tgt Ion:	266	Resp:	1521
Ion Ratio	Lower	Upper	
266	100		
264	63.4	50.6	76.0
268	64.4	51.4	77.0





Abundance Scan 1530 (17.104 min): BN014350.D\data.ms (-1525) (-)




Phenanthrene
Concen: 0.39 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	178	Resp:	33650
Ion Ratio	Lower	Upper	
178	100		
176	19.0	15.2	22.8
179	15.3	12.2	18.4



Abundance Scan 1530 (17.104 min): BN014350.D\data.ms (-1505) (-)





Abundance Scan 1538 (17.201 min): BN014350.D\data.ms (-1526) (-)




Anthracene
Concen: 0.37 ng
RT: 17.201 min Scan# 1538
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Tgt Ion:	178	Resp:	25479
Ion Ratio	Lower	Upper	
178	100		
176	18.7	15.0	22.4
179	15.2	12.2	18.2




Abundance Scan 1538 (17.201 min): BN014350.D\data.ms (-1530) (-)





Abundance Scan 1825 (19.101 min): BN014350.D\data.ms (-1827) (-)

Fluoranthene-d10
Concen: 0.37 ng
RT: 19.101 min Scan# 1825
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:	212	Resp:	31420
Ion Ratio	Lower	Upper	
212	100		
106	12.1	9.7	14.5
104	6.7	5.4	8.0





Abundance Scan 1825 (19.101 min): BN014350.D\data.ms (-1785) (-)




Abundance Scan 1831 (19.131 min): BN014350.D\data.ms (-1823) (-)

Fluoranthene
Concen: 0.37 ng
RT: 19.131 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Tgt Ion:	202	Resp:	35590
Ion Ratio	Lower	Upper	
202	100		
101	10.3	8.2	12.4
203	17.3	13.8	20.6



Abundance Scan 1831 (19.131 min): BN014350.D\data.ms (-1791) (-)





Abundance Scan 2125 (21.258 min): BN014350.D\data.ms (-2106) (-)




Chrysene-d12
Concen: 0.40 ng
RT: 21.258 min Scan# 2125
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	240	Resp:	29260
Ion Ratio	Lower	Upper	
240	100		
120	13.4	10.7	16.1
236	26.9	21.5	32.3



Abundance Scan 2125 (21.258 min): BN014350.D\data.ms (-2106) (-)





Abundance Scan 1905 (19.498 min): BN014350.D\data.ms (-1865) (-)




Pyrene
Concen: 0.41 ng
RT: 19.498 min Scan# 1905
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Tgt Ion:	202	Resp:	34264
Ion Ratio	Lower	Upper	
202	100		
200	20.7	16.6	24.8
203	17.6	14.1	21.1



Abundance Scan 1905 (19.498 min): BN014350.D\data.ms (-1865) (-)





Abundance Scan 1947 (19.707 min): BN014350.D\data.ms (-1934) (-)




Terphenyl-d14
Concen: 0.41 ng
RT: 19.707 min Scan# 1947
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.8	6.2	9.4
122	11.6	9.3	13.9



Abundance Scan 1947 (19.707 min): BN014350.D\data.ms (-1917) (-)





Abundance Scan 2124 (21.247 min): BN014350.D\data.ms (-2132) (-)




Benzo(a)anthracene
Concen: 0.38 ng
RT: 21.247 min Scan# 2124
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.6	21.3	31.9
229	19.6	15.7	23.5



Abundance Scan 2124 (21.247 min): BN014350.D\data.ms (-2105) (-)






Abundance Scan 2129 (21.300 min): BN014350.D\data.ms (-2129) (-)




Chrysene
Concen: 0.40 ng
RT: 21.300 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:	228	Resp:	37224
Ion Ratio	Lower	Upper	
228	100		
226	28.9	23.1	34.7
229	19.6	15.7	23.5






Abundance Scan 2118 (21.184 min): BN014350.D\data.ms (-2118) (-)




Bis(2-ethylhexyl)phthalate
Concen: 0.45 ng
RT: 21.184 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Tgt Ion:	149	Resp:	9784
Ion Ratio	Lower	Upper	
149	100		
167	26.6	21.3	31.9
279	3.9	3.1	4.7




Abundance Scan 3203 (25.725 min): BN014350.D\data.ms (-3175) (-)




Indeno(1,2,3-cd)pyrene
Concen: 0.33 ng
RT: 25.725 min Scan# 3203
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4




Tgt Ion	Ratio	Lower	Upper
276	100		
138	23.7	18.9	28.3
277	24.8	19.8	29.8




Abundance Scan 3203 (25.725 min): BN014350.D\data.ms (-3174) (-)




Abundance Scan 2550 (23.490 min): BN014350.D\data.ms (-2536) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.490 min Scan# 2550
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.7	19.8	29.6
265	43.0	34.4	51.6




Abundance Scan 2550 (23.490 min): BN014350.D\data.ms (-2506) (-)





Abundance Scan 2354 (22.819 min): BN014350.D\data.ms (-2354) (-)

Benzo(b)fluoranthene
Concen: 0.33 ng
RT: 22.819 min Scan# 2354
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.8	19.0	28.6
125	12.1	9.7	14.5





Abundance Scan 2354 (22.819 min): BN014350.D\data.ms (-2296) (-)




Abundance Scan 2367 (22.864 min): BN014350.D\data.ms (-2367) (-)

Benzo(k)fluoranthene
Concen: 0.34 ng
RT: 22.864 min Scan# 2367
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.1	19.3	28.9
125	11.8	9.4	14.2




Abundance Scan 2367 (22.864 min): BN014350.D\data.ms (-2279) (-)





Abundance Scan 2521 (23.391 min): BN014350.D\data.ms (-2520) (-)

Benzo(a)pyrene
Concen: 0.35 ng
RT: 23.391 min Scan# 2521
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.9	20.7	31.1
125	14.7	11.8	17.6





Abundance Scan 2521 (23.391 min): BN014350.D\data.ms (-2463) (-)




Abundance Scan 3207 (25.739 min): BN014350.D\data.ms (-3184) (-)

Dibenzo(a,h)anthracene
Concen: 0.35 ng
RT: 25.739 min Scan# 3207
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57


Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.6	13.3	19.9
279	25.2	20.2	30.2



Abundance Scan 3207 (25.739 min): BN014350.D\data.ms (-3184) (-)




Abundance Scan 3403 (26.410 min): BN014350.D\data.ms (-3345) (-)



Benzo(g,h,i)perylene
Concen: 0.36 ng
RT: 26.410 min Scan# 3403
Delta R.T. 0.000 min
Lab File: BN014350.D
Acq: 21 Apr 2021 10:57

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4



Tgt Ion:	276	Resp:	35534
Ion Ratio	Lower	Upper	
276	100		
277	23.9	19.1	28.7
138	20.9	16.7	25.1

