

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014348.D
 Acq On : 21 Apr 2021 09:46
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 21 12:51:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 12:48:18 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	7025	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	24645	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	12894	0.40	ng	0.00
19) Phenanthrene-d10	17.068	188	27806	0.40	ng	0.00
29) Chrysene-d12	21.261	240	24686	0.40	ng	0.00
36) Perylene-d12	23.489	264	23962	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	1537	0.10	ng	0.00
5) Phenol-d6	6.863	99	1820	0.09	ng	0.00
8) Nitrobenzene-d5	8.832	82	1965	0.13	ng	0.00
11) 2-Methylnaphthalene-d10	12.057	152	3713	0.10	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	398	0.09	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	4895	0.10	ng	0.00
27) Fluoranthene-d10	19.104	212	7054	0.09	ng	0.00
31) Terphenyl-d14	19.710	244	5378	0.10	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.255	88	1190	0.13	ng	Qvalue # 82
6) bis(2-Chloroethyl)ether	7.120	93	1795	0.10	ng	99
9) Naphthalene	10.518	128	6223	0.10	ng	97
10) Hexachlorobutadiene	10.809	225	1369	0.11	ng	# 98
12) 2-Methylnaphthalene	12.133	142	4038	0.10	ng	98
16) Acenaphthylene	14.042	152	5129	0.10	ng	99
17) Acenaphthene	14.384	154	3518	0.10	ng	100
18) Fluorene	15.374	166	4292	0.10	ng	97
20) 4,6-Dinitro-2-methylph...	15.450	198	281	0.10	ng	# 1
21) 4-Bromophenyl-phenylether	16.264	248	1489	0.10	ng	96
22) Hexachlorobenzene	16.374	284	2083	0.11	ng	99
23) Atrazine	16.532	200	874	0.10	ng	# 94
25) Phenanthrene	17.104	178	7766	0.10	ng	99
26) Anthracene	17.201	178	5598	0.09	ng	99
28) Fluoranthene	19.134	202	7686	0.09	ng	100
30) Pyrene	19.496	202	7787	0.11	ng	100
32) Benzo(a)anthracene	21.250	228	6280	0.09	ng	98
33) Chrysene	21.303	228	7855	0.10	ng	98
34) Bis(2-ethylhexyl)phtha...	21.186	149	2989	0.16	ng	96
35) Indeno(1,2,3-cd)pyrene	25.724	276	9204	0.10	ng	99
37) Benzo(b)fluoranthene	22.825	252	8179	0.08	ng	# 78
38) Benzo(k)fluoranthene	22.866	252	8179	0.08	ng	# 83
39) Benzo(a)pyrene	23.393	252	6919	0.09	ng	# 61
40) Dibenzo(a,h)anthracene	25.741	278	7493	0.08	ng	# 90
41) Benzo(g,h,i)perylene	26.405	276	8125	0.08	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
Data File : BN014348.D
Acq On : 21 Apr 2021 09:46
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Quant Time: Apr 21 12:51:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 21 12:48:18 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014348.D
 Acq: 21 Apr 2021 09:46




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion: 152 Resp: 7025
 Ion Ratio Lower Upper
 152 100
 150 153.4 123.4 185.2
 115 59.1 47.5 71.3



#2
 1,4-Dioxane
 Concen: 0.13 ng
 RT: 3.255 min Scan# 59
 Delta R.T. -0.000 min
 Lab File: BN014348.D
 Acq: 21 Apr 2021 09:46

Tgt Ion: 88 Resp: 1190
 Ion Ratio Lower Upper
 88 100
 58 50.1 52.2 78.2#
 43 17.8 20.6 30.8#









#4
2-Fluorophenol
Concen: 0.10 ng
RT: 5.292 min Scan# 312
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion: 112 Resp: 1537
Ion Ratio Lower Upper
112 100
64 53.0 43.0 64.6
63 30.4 23.4 35.0



#5
Phenol-d6
Concen: 0.09 ng
RT: 6.863 min Scan# 507
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion: 99 Resp: 1820
Ion Ratio Lower Upper
99 100
42 25.2 16.0 24.0#
71 35.9 27.8 41.8









#6
bis(2-Chloroethyl)ether
Concen: 0.10 ng
RT: 7.120 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampled :
SSTDICC0.1


Tgt Ion: 93	Resp:	1795	
Ion	Ratio	Lower	Upper
93	100		
63	75.6	59.4	89.2
95	33.5	27.0	40.4



#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt	Ion:136	Resp:	24645
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.6
54	5.6	4.4	6.6
68	5.0	3.9	5.9









#8
Nitrobenzene-d5
Concen: 0.13 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	82	Resp:	1965
Ion	Ratio	Lower	Upper
82	100		
128	42.7	30.3	45.5
54	50.2	35.8	53.8



#9
Naphthalene
Concen: 0.10 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:	128	Resp:	6223
Ion	Ratio	Lower	Upper
128	100		
129	12.9	9.0	13.6
127	14.0	10.3	15.5









#10
Hexachlorobutadiene
Concen: 0.11 ng
RT: 10.809 min Scan# 884
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:225 Resp: 1369
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.0 50.6 75.8



#11
2-Methylnaphthalene-d10
Concen: 0.10 ng
RT: 12.057 min Scan# 1063
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:152 Resp: 3713
Ion Ratio Lower Upper
152 100
151 20.0 16.3 24.5









#12
2-Methylnaphthalene
Concen: 0.10 ng
RT: 12.133 min Scan# 1080
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:142 Resp: 4038
Ion Ratio Lower Upper
142 100
141 89.2 70.6 105.8
115 39.7 29.0 43.6



#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:164 Resp: 12894
Ion Ratio Lower Upper
164 100
162 99.7 81.0 121.6
160 50.2 40.7 61.1









#14
2,4,6-Tribromophenol
Concen: 0.09 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:330 Resp: 398
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 39.2 26.2 39.2



#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:172 Resp: 4895
Ion Ratio Lower Upper
172 100
171 35.8 28.2 42.2
170 25.6 19.6 29.4









#16
Acenaphthylene
Concen: 0.10 ng
RT: 14.042 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	152	Resp:	5129
Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.6	23.4
153	13.3	10.5	15.7



#17
Acenaphthene
Concen: 0.10 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:	154	Resp:	3518
Ion	Ratio	Lower	Upper
154	100		
153	111.0	89.2	133.8
152	61.6	49.0	73.4









#18
Fluorene
Concen: 0.10 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:166 Resp: 4292
Ion Ratio Lower Upper
166 100
165 101.5 78.6 118.0
167 13.9 10.9 16.3



#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.068 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:188 Resp: 27806
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.9 6.2 9.4









#20
4,6-Dinitro-2-methylphenol
Concen: 0.10 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampled :
SSTDIC0.1


Tgt Ion:198 Resp: 281
Ion Ratio Lower Upper
198 100
51 238.1 60.0 90.0#
105 89.5 31.5 47.3#



#21
4-Bromophenyl-phenylether
Concen: 0.10 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:248 Resp: 1489
Ion Ratio Lower Upper
248 100
250 97.7 77.4 116.0
141 58.4 41.4 62.0









#22
Hexachlorobenzene
Concen: 0.11 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:284 Resp: 2083
Ion Ratio Lower Upper
284 100
142 31.8 26.5 39.7
249 29.9 24.0 36.0



#23
Atrazine
Concen: 0.10 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:200 Resp: 874
Ion Ratio Lower Upper
200 100
173 31.8 21.0 31.6#
215 50.9 38.7 58.1









#25
Phenanthrene
Concen: 0.10 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 7766
Ion Ratio Lower Upper
178 100
176 18.9 15.2 22.8
179 15.7 12.2 18.4



#26
Anthracene
Concen: 0.09 ng
RT: 17.201 min Scan# 1538
Delta R.T. 0.000 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:178 Resp: 5598
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.9 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.09 ng
RT: 19.104 min Scan# 1824
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 7054
Ion Ratio Lower Upper
212 100
106 12.1 9.7 14.5
104 7.1 5.4 8.0



#28
Fluoranthene
Concen: 0.09 ng
RT: 19.134 min Scan# 1830
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:202 Resp: 7686
Ion Ratio Lower Upper
202 100
101 10.6 8.2 12.4
203 17.3 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.261 min Scan# 2124
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 24686
Ion Ratio Lower Upper
240 100
120 12.1 10.7 16.1
236 26.9 21.5 32.3



#30
Pyrene
Concen: 0.11 ng
RT: 19.496 min Scan# 1903
Delta R.T. -0.002 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46


Tgt Ion:202 Resp: 7787
Ion Ratio Lower Upper
202 100
200 20.7 16.6 24.8
203 17.5 14.1 21.1








#31
Terphenyl-d14
Concen: 0.10 ng
RT: 19.710 min Scan# 1946
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1







Tgt Ion:244 Resp: 5378
Ion Ratio Lower Upper
244 100
212 8.6 6.2 9.4
122 11.9 9.3 13.9



#32
Benzo(a)anthracene
Concen: 0.09 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:228 Resp: 6280
Ion Ratio Lower Upper
228 100
226 27.5 21.3 31.9
229 20.8 15.7 23.5









#33
Chrysene
Concen: 0.10 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	228	Resp:	7855
Ion	Ratio	Lower	Upper
228	100		
226	29.7	23.1	34.7
229	20.5	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.16 ng
RT: 21.186 min Scan# 2117
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:	149	Resp:	2989
Ion	Ratio	Lower	Upper
149	100		
167	24.4	21.3	31.9
279	4.3	3.1	4.7









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.10 ng
RT: 25.724 min Scan# 3201
Delta R.T. -0.001 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:276 Resp: 9204
Ion Ratio Lower Upper
276 100
138 22.8 18.9 28.3
277 24.6 19.8 29.8



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.489 min Scan# 2548
Delta R.T. -0.001 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:264 Resp: 23962
Ion Ratio Lower Upper
264 100
260 24.8 19.8 29.6
265 43.9 34.4 51.6









#37
Benzo(b)fluoranthene
Concen: 0.08 ng
RT: 22.825 min Scan# 2354
Delta R.T. 0.006 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:252 Resp: 8179
Ion Ratio Lower Upper
252 100
253 32.3 19.0 28.6#
125 24.9 9.7 14.5#



#38
Benzo(k)fluoranthene
Concen: 0.08 ng
RT: 22.866 min Scan# 2366
Delta R.T. 0.003 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Tgt Ion:252 Resp: 8179
Ion Ratio Lower Upper
252 100
253 32.0 19.3 28.9#
125 19.5 9.4 14.2#





#39

Benzo(a)pyrene

Concen: 0.09 ng

RT: 23.393 min Scan# 2520

Delta R.T. 0.003 min

Lab File: BN014348.D

Acq: 21 Apr 2021 09:46





Tgt Ion:252	Resp:	6919
Ion Ratio	Lower	Upper
252	100	
253	39.3	20.7 31.1#
125	39.8	11.8 17.6#

Instrument :

BNA_N

ClientSampled :

SSTDIC0.1



#40

Dibenzo(a,h)anthracene

Concen: 0.08 ng




RT: 25.741 min Scan# 3206


Delta R.T. 0.003 min

Lab File: BN014348.D

Acq: 21 Apr 2021 09:46

Tgt Ion:278	Resp:	7493
Ion Ratio	Lower	Upper
278	100	
139	18.6	13.3 19.9
279	31.7	20.2 30.2#





#41
Benzo(g,h,i)perylene
Concen: 0.08 ng
RT: 26.405 min Scan# 3400
Delta R.T. -0.004 min
Lab File: BN014348.D
Acq: 21 Apr 2021 09:46

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:276 Resp: 8125
Ion Ratio Lower Upper
276 100
277 26.1 19.1 28.7
138 23.8 16.7 25.1

