

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014354.D
 Acq On : 21 Apr 2021 13:21
 Operator : CG/JU
 Sample : SSTDIC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC5.0

Quant Time: Apr 21 13:51:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 12:48:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	8350	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	27311	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	14360	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	28867	0.40	ng	0.00
29) Chrysene-d12	21.261	240	32782	0.40	ng	0.00
36) Perylene-d12	23.493	264	24477	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	81088	4.39	ng	0.00
5) Phenol-d6	6.863	99	106762	4.59	ng	0.00
8) Nitrobenzene-d5	8.832	82	114355	6.72	ng	0.00
11) 2-Methylnaphthalene-d10	12.057	152	208681	4.97	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	28651	5.51	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	273781	4.85	ng	0.00
27) Fluoranthene-d10	19.104	212	425678	5.34	ng	0.00
31) Terphenyl-d14	19.710	244	335016	4.88	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.247	88	43768	4.13	ng	99
3) n-Nitrosodimethylamine	3.561	42	31261	5.32	ng	# 95
6) bis(2-Chloroethyl)ether	7.120	93	95475	4.54	ng	98
9) Naphthalene	10.518	128	329163	4.79	ng	99
10) Hexachlorobutadiene	10.809	225	68681	5.02	ng	# 99
12) 2-Methylnaphthalene	12.133	142	222140	4.96	ng	99
16) Acenaphthylene	14.042	152	326622	5.97	ng	100
17) Acenaphthene	14.384	154	209812	5.15	ng	97
18) Fluorene	15.374	166	265590	5.35	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	20788	6.94	ng	# 49
21) 4-Bromophenyl-phenylether	16.264	248	82936	5.42	ng	99
22) Hexachlorobenzene	16.374	284	108520	5.53	ng	99
23) Atrazine	16.532	200	57850	6.08	ng	99
24) Pentachlorophenol	16.715	266	26292	4.94	ng	99
25) Phenanthrene	17.104	178	415366	5.13	ng	100
26) Anthracene	17.201	178	368142	5.69	ng	100
28) Fluoranthene	19.134	202	482878	5.41	ng	100
30) Pyrene	19.496	202	475783	5.11	ng	100
32) Benzo(a)anthracene	21.250	228	477424	5.43	ng	99
33) Chrysene	21.303	228	497344	4.74	ng	99
34) Bis(2-ethylhexyl)phtha...	21.186	149	154289	6.27	ng	99
35) Indeno(1,2,3-cd)pyrene	25.724	276	544280	4.27	ng	100
37) Benzo(b)fluoranthene	22.822	252	479478	4.50	ng	# 95
38) Benzo(k)fluoranthene	22.870	252	486694	4.73	ng	# 95
39) Benzo(a)pyrene	23.393	252	421214	5.37	ng	# 91
40) Dibenzo(a,h)anthracene	25.741	278	456812	4.91	ng	97
41) Benzo(g,h,i)perylene	26.412	276	466704	4.74	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
Data File : BN014354.D
Acq On : 21 Apr 2021 13:21
Operator : CG/JU
Sample : SSTDIC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Apr 21 13:51:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 21 12:48:18 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014354.D
 Acq: 21 Apr 2021 13:21

Instrument :
 BNA_N
 ClientSampled :
 SSTDICC5.0

Tgt Ion: 152 Resp: 8350




Ion	Ratio	Lower	Upper
152	100		
150	154.9	123.4	185.2
115	60.0	47.5	71.3




#2
 1,4-Dioxane
 Concen: 4.13 ng
 RT: 3.247 min Scan# 58
 Delta R.T. -0.008 min
 Lab File: BN014354.D
 Acq: 21 Apr 2021 13:21

Tgt Ion: 88 Resp: 43768

Ion	Ratio	Lower	Upper
88	100		
58	64.1	52.2	78.2
43	25.0	20.6	30.8









#3
n-Nitrosodimethylamine
Concen: 5.32 ng
RT: 3.561 min Scan# 97
Delta R.T. -0.040 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 31261
Ion Ratio Lower Upper
42 100
74 140.8 108.0 162.0
44 14.0 0.0 0.0#



#4
2-Fluorophenol
Concen: 4.39 ng
RT: 5.292 min Scan# 312
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion: 112 Resp: 81088
Ion Ratio Lower Upper
112 100
64 53.9 43.0 64.6
63 29.1 23.4 35.0









#5
Phenol-d6
Concen: 4.59 ng
RT: 6.863 min Scan# 507
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 106762
Ion Ratio Lower Upper
99 100
42 19.2 16.0 24.0
71 34.5 27.8 41.8



#6
bis(2-Chloroethyl)ether
Concen: 4.54 ng
RT: 7.120 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion: 93 Resp: 95475
Ion Ratio Lower Upper
93 100
63 72.7 59.4 89.2
95 32.9 27.0 40.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	27311
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.6
54	5.5	4.4	6.6
68	5.1	3.9	5.9



#8
Nitrobenzene-d5
Concen: 6.72 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:	82	Resp:	114355
Ion	Ratio	Lower	Upper
82	100		
128	41.4	30.3	45.5
54	41.9	35.8	53.8









#9
Naphthalene
Concen: 4.79 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 329163
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 12.6 10.3 15.5



#10
Hexachlorobutadiene
Concen: 5.02 ng
RT: 10.809 min Scan# 884
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:225 Resp: 68681
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 50.6 75.8









#11
2-Methylnaphthalene-d10
Concen: 4.97 ng
RT: 12.057 min Scan# 1063
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 208681
Ion Ratio Lower Upper
152 100
151 20.1 16.3 24.5



#12
2-Methylnaphthalene
Concen: 4.96 ng
RT: 12.133 min Scan# 1080
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:142 Resp: 222140
Ion Ratio Lower Upper
142 100
141 87.9 70.6 105.8
115 35.4 29.0 43.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 14360
Ion Ratio Lower Upper
164 100
162 101.9 81.0 121.6
160 50.5 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 5.51 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:330 Resp: 28651
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 33.0 26.2 39.2









#15
2-Fluorobiphenyl
Concen: 4.85 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 273781
Ion Ratio Lower Upper
172 100
171 34.8 28.2 42.2
170 24.1 19.6 29.4



#16
Acenaphthylene
Concen: 5.97 ng
RT: 14.042 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:152 Resp: 326622
Ion Ratio Lower Upper
152 100
151 19.3 15.6 23.4
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 5.15 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 209812
Ion Ratio Lower Upper
154 100
153 109.5 89.2 133.8
152 56.8 49.0 73.4



#18
Fluorene
Concen: 5.35 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:166 Resp: 265590
Ion Ratio Lower Upper
166 100
165 97.4 78.6 118.0
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:188 Resp: 28867
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.7 6.2 9.4



#20
4,6-Dinitro-2-methylphenol
Concen: 6.94 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:198 Resp: 20788
Ion Ratio Lower Upper
198 100
51 20.2 60.0 90.0#
105 23.7 31.5 47.3#









#21
4-Bromophenyl-phenylether
Concen: 5.42 ng
RT: 16.264 min Scan# 1461
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	248	Resp:	82936
Ion	Ratio	Lower	Upper
248	100		
250	96.7	77.4	116.0
141	49.9	41.4	62.0



#22
Hexachlorobenzene
Concen: 5.53 ng
RT: 16.374 min Scan# 1470
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:	284	Resp:	108520
Ion	Ratio	Lower	Upper
284	100		
142	33.0	26.5	39.7
249	29.4	24.0	36.0









#23
Atrazine
Concen: 6.08 ng
RT: 16.532 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:200 Resp: 57850
Ion Ratio Lower Upper
200 100
173 26.8 21.0 31.6
215 47.2 38.7 58.1



#24
Pentachlorophenol
Concen: 4.94 ng
RT: 16.715 min Scan# 1498
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:266 Resp: 26292
Ion Ratio Lower Upper
266 100
264 62.3 50.6 76.0
268 64.2 51.4 77.0









#25
Phenanthrene
Concen: 5.13 ng
RT: 17.104 min Scan# 1530
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 415366
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.3 12.2 18.4



#26
Anthracene
Concen: 5.69 ng
RT: 17.201 min Scan# 1538
Delta R.T. 0.000 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:178 Resp: 368142
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.4
179 15.3 12.2 18.2









#27
Fluoranthene-d10
Concen: 5.34 ng
RT: 19.104 min Scan# 1824
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:212 Resp: 425678
Ion Ratio Lower Upper
212 100
106 11.9 9.7 14.5
104 6.7 5.4 8.0



#28
Fluoranthene
Concen: 5.41 ng
RT: 19.134 min Scan# 1830
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:202 Resp: 482878
Ion Ratio Lower Upper
202 100
101 10.2 8.2 12.4
203 17.4 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.261 min Scan# 2124
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:240	Resp:	32782
Ion Ratio	Lower	Upper
240	100	
120	12.4	10.7 16.1
236	27.2	21.5 32.3



#30
Pyrene
Concen: 5.11 ng
RT: 19.496 min Scan# 1903
Delta R.T. -0.002 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:202	Resp:	475783
Ion Ratio	Lower	Upper
202	100	
200	20.6	16.6 24.8
203	17.5	14.1 21.1









#31
Terphenyl-d14
Concen: 4.88 ng
RT: 19.710 min Scan# 1946
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:244 Resp: 335016
Ion Ratio Lower Upper
244 100
212 7.4 6.2 9.4
122 10.2 9.3 13.9



#32
Benzo(a)anthracene
Concen: 5.43 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:228 Resp: 477424
Ion Ratio Lower Upper
228 100
226 26.2 21.3 31.9
229 19.8 15.7 23.5









#33
Chrysene
Concen: 4.74 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:228 Resp: 497344




Ion	Ratio	Lower	Upper
228	100		
226	28.6	23.1	34.7
229	19.9	15.7	23.5




#34
Bis(2-ethylhexyl)phthalate
Concen: 6.27 ng
RT: 21.186 min Scan# 2117
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:149 Resp: 154289

Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.3	31.9
279	3.7	3.1	4.7









#35
Indeno(1,2,3-cd)pyrene
Concen: 4.27 ng
RT: 25.724 min Scan# 3201
Delta R.T. -0.001 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:276 Resp: 544280
Ion Ratio Lower Upper
276 100
138 23.3 18.9 28.3
277 24.9 19.8 29.8



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.493 min Scan# 2549
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:264 Resp: 24477
Ion Ratio Lower Upper
264 100
260 24.4 19.8 29.6
265 43.4 34.4 51.6









#37
Benzo(b)fluoranthene
Concen: 4.50 ng
RT: 22.822 min Scan# 2353
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 479478
Ion Ratio Lower Upper
252 100
253 21.9 19.0 28.6
125 9.4 9.7 14.5#



#38
Benzo(k)fluoranthene
Concen: 4.73 ng
RT: 22.870 min Scan# 2367
Delta R.T. 0.006 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:252 Resp: 486694
Ion Ratio Lower Upper
252 100
253 22.0 19.3 28.9
125 9.1 9.4 14.2#









#39
Benzo(a)pyrene
Concen: 5.37 ng
RT: 23.393 min Scan# 2520
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:252 Resp: 421214
Ion Ratio Lower Upper
252 100
253 21.9 20.7 31.1
125 10.6 11.8 17.6#



#40
Dibenzo(a,h)anthracene
Concen: 4.91 ng
RT: 25.741 min Scan# 3206
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Tgt Ion:278 Resp: 456812
Ion Ratio Lower Upper
278 100
139 15.2 13.3 19.9
279 23.5 20.2 30.2





#41
Benzo(g,h,i)perylene
Concen: 4.74 ng
RT: 26.412 min Scan# 3402
Delta R.T. 0.003 min
Lab File: BN014354.D
Acq: 21 Apr 2021 13:21

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:276 Resp: 466704
Ion Ratio Lower Upper
276 100
277 23.7 19.1 28.7
138 19.6 16.7 25.1

