

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014361.D
 Acq On : 21 Apr 2021 18:39
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Apr 21 19:13:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 13:57:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	6890	0.40	ng	0.00
7) Naphthalene-d8	10.467	136	24273	0.40	ng	0.00
13) Acenaphthene-d10	14.321	164	12219	0.40	ng	0.00
19) Phenanthrene-d10	17.067	188	24517	0.40	ng	0.00
29) Chrysene-d12	21.268	240	23257	0.40	ng	# 0.01
36) Perylene-d12	23.490	264	21594	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.292	112	5541	0.39	ng	0.00
5) Phenol-d6	6.863	99	6788	0.39	ng	0.00
8) Nitrobenzene-d5	8.832	82	7604	0.37	ng	0.00
11) 2-Methylnaphthalene-d10	12.057	152	14019	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.818	330	1300	0.33	ng	0.00
15) 2-Fluorobiphenyl	12.938	172	19137	0.40	ng	0.00
27) Fluoranthene-d10	19.101	212	23359	0.36	ng	0.00
31) Terphenyl-d14	19.707	244	18454	0.37	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.255	88	3157	0.38	ng	99
3) n-Nitrosodimethylamine	3.609	42	1333	0.38	ng	# 89
6) bis(2-Chloroethyl)ether	7.120	93	6789	0.40	ng	99
9) Naphthalene	10.518	128	23754	0.39	ng	100
10) Hexachlorobutadiene	10.809	225	5207	0.39	ng	# 99
12) 2-Methylnaphthalene	12.133	142	15342	0.39	ng	100
16) Acenaphthylene	14.042	152	19904	0.38	ng	99
17) Acenaphthene	14.384	154	12931	0.38	ng	97
18) Fluorene	15.374	166	15715	0.37	ng	99
20) 4,6-Dinitro-2-methylph...	15.450	198	864	0.32	ng	92
21) 4-Bromophenyl-phenylether	16.264	248	5322	0.39	ng	99
22) Hexachlorobenzene	16.373	284	7837	0.41	ng	97
23) Atrazine	16.532	200	2592	0.32	ng	99
24) Pentachlorophenol	16.714	266	1084	0.31	ng	99
25) Phenanthrene	17.104	178	26660	0.39	ng	100
26) Anthracene	17.201	178	19783	0.37	ng	100
28) Fluoranthene	19.131	202	27180	0.37	ng	100
30) Pyrene	19.498	202	26104	0.38	ng	100
32) Benzo(a)anthracene	21.247	228	23447	0.38	ng	99
33) Chrysene	21.300	228	29214	0.40	ng	99
34) Bis(2-ethylhexyl)phtha...	21.183	149	7664	0.35	ng	99
35) Indeno(1,2,3-cd)pyrene	25.722	276	34578	0.43	ng	99
37) Benzo(b)fluoranthene	22.822	252	28379	0.36	ng	97
38) Benzo(k)fluoranthene	22.867	252	31492	0.39	ng	100
39) Benzo(a)pyrene	23.394	252	26531	0.39	ng	# 90
40) Dibenzo(a,h)anthracene	25.739	278	27940	0.37	ng	99
41) Benzo(g,h,i)perylene	26.410	276	30131	0.38	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
Data File : BN014361.D
Acq On : 21 Apr 2021 18:39
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Apr 21 19:13:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 21 13:57:02 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.692 min Scan# 610
 Delta R.T. 0.000 min
 Lab File: BN014361.D
 Acq: 21 Apr 2021 18:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 6890
 Ion Ratio Lower Upper
 152 100
 150 152.8 123.4 185.2
 115 58.8 47.5 71.3



#2
 1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.255 min Scan# 59
 Delta R.T. -0.000 min
 Lab File: BN014361.D
 Acq: 21 Apr 2021 18:39

Tgt Ion: 88 Resp: 3157
 Ion Ratio Lower Upper
 88 100
 58 64.9 52.2 78.2
 43 26.5 20.6 30.8









#3
n-Nitrosodimethylamine
Concen: 0.38 ng
RT: 3.609 min Scan# 103
Delta R.T. 0.008 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 1333
Ion Ratio Lower Upper
42 100
74 148.5 108.0 162.0
44 0.0 0.0 0.0



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.292 min Scan# 312
Delta R.T. 0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion: 112 Resp: 5541
Ion Ratio Lower Upper
112 100
64 53.7 43.0 64.6
63 28.6 23.4 35.0









#5
Phenol-d6
Concen: 0.39 ng
RT: 6.863 min Scan# 507
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	99	Resp:	6788
Ion	Ratio	Lower	Upper
99	100		
42	19.5	16.0	24.0
71	33.8	27.8	41.8



#6
bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 7.120 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:	93	Resp:	6789
Ion	Ratio	Lower	Upper
93	100		
63	73.4	59.4	89.2
95	33.5	27.0	40.4









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.467 min Scan# 857
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	24273
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.6
54	5.6	4.4	6.6
68	5.1	3.9	5.9



#8
Nitrobenzene-d5
Concen: 0.37 ng
RT: 8.832 min Scan# 728
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:	82	Resp:	7604
Ion	Ratio	Lower	Upper
82	100		
128	38.4	30.3	45.5
54	45.2	35.8	53.8









#9
Naphthalene
Concen: 0.39 ng
RT: 10.518 min Scan# 861
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 23754
Ion Ratio Lower Upper
128 100
129 11.4 9.0 13.6
127 13.0 10.3 15.5



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.809 min Scan# 884
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:225 Resp: 5207
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 50.6 75.8









#11
2-Methylnaphthalene-d10
Concen: 0.38 ng
RT: 12.057 min Scan# 1063
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 14019
Ion Ratio Lower Upper
152 100
151 20.4 16.3 24.5



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.133 min Scan# 1080
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:142 Resp: 15342
Ion Ratio Lower Upper
142 100
141 88.3 70.6 105.8
115 36.7 29.0 43.6









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.321 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 12219
Ion Ratio Lower Upper
164 100
162 100.8 81.0 121.6
160 52.5 40.7 61.1



#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.818 min Scan# 1425
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:330 Resp: 1300
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.2 26.2 39.2









#15
2-Fluorobiphenyl
Concen: 0.40 ng
RT: 12.938 min Scan# 1198
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	172	Resp:	19137
Ion	Ratio	Lower	Upper
172	100		
171	34.8	28.2	42.2
170	24.0	19.6	29.4



#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.042 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:	152	Resp:	19904
Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.6	23.4
153	12.9	10.5	15.7









#17
Acenaphthene
Concen: 0.38 ng
RT: 14.384 min Scan# 1312
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 12931
Ion Ratio Lower Upper
154 100
153 113.0 89.2 133.8
152 65.3 49.0 73.4



#18
Fluorene
Concen: 0.37 ng
RT: 15.374 min Scan# 1390
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:166 Resp: 15715
Ion Ratio Lower Upper
166 100
165 99.0 78.6 118.0
167 13.7 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.067 min Scan# 1527
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188	Resp:	24517
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	7.7	6.2



#20
4,6-Dinitro-2-methylphenol
Concen: 0.32 ng
RT: 15.450 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:198	Resp:	864
Ion Ratio	Lower	Upper
198	100	
51	80.4	60.0
105	47.1	31.5









#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.264 min Scan# 1461
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	248	Resp:	5322
Ion	Ratio	Lower	Upper
248	100		
250	96.1	77.4	116.0
141	50.0	41.4	62.0



#22
Hexachlorobenzene
Concen: 0.41 ng
RT: 16.373 min Scan# 1470
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:	284	Resp:	7837
Ion	Ratio	Lower	Upper
284	100		
142	36.2	26.5	39.7
249	29.5	24.0	36.0









#23
Atrazine
Concen: 0.32 ng
RT: 16.532 min Scan# 1483
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	200	Resp:	2592
Ion Ratio	Lower	Upper	
200	100		
173	25.9	21.0	31.6
215	49.6	38.7	58.1



#24
Pentachlorophenol
Concen: 0.31 ng
RT: 16.714 min Scan# 1498
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:	266	Resp:	1084
Ion Ratio	Lower	Upper	
266	100		
264	62.8	50.6	76.0
268	62.8	51.4	77.0









#25
Phenanthrene
Concen: 0.39 ng
RT: 17.104 min Scan# 1530
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178	Resp:	26660
Ion Ratio	Lower	Upper
178	100	
176	19.2	15.2 22.8
179	15.3	12.2 18.4



#26
Anthracene
Concen: 0.37 ng
RT: 17.201 min Scan# 1538
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:178	Resp:	19783
Ion Ratio	Lower	Upper
178	100	
176	18.6	15.0 22.4
179	15.2	12.2 18.2









#27
Fluoranthene-d10
Concen: 0.36 ng
RT: 19.101 min Scan# 1825
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 23359
Ion Ratio Lower Upper
212 100
106 12.2 9.7 14.5
104 6.7 5.4 8.0



#28
Fluoranthene
Concen: 0.37 ng
RT: 19.131 min Scan# 1831
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:202 Resp: 27180
Ion Ratio Lower Upper
202 100
101 10.5 8.2 12.4
203 17.3 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.268 min Scan# 2126
Delta R.T. 0.011 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 23257
Ion Ratio Lower Upper
240 100
120 5.6 10.7 16.1#
236 25.5 21.5 32.3



#30
Pyrene
Concen: 0.38 ng
RT: 19.498 min Scan# 1905
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:202 Resp: 26104
Ion Ratio Lower Upper
202 100
200 20.8 16.6 24.8
203 17.4 14.1 21.1









#31
 Terphenyl-d14
 Concen: 0.37 ng
 RT: 19.707 min Scan# 1947
 Delta R.T. -0.000 min
 Lab File: BN014361.D
 Acq: 21 Apr 2021 18:39




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244	Resp:	18454
Ion Ratio	Lower	Upper
244	100	
212	7.8	6.2 9.4
122	11.5	9.3 13.9



#32
 Benzo(a)anthracene
 Concen: 0.38 ng
 RT: 21.247 min Scan# 2124
 Delta R.T. -0.000 min
 Lab File: BN014361.D
 Acq: 21 Apr 2021 18:39

Tgt Ion:228	Resp:	23447
Ion Ratio	Lower	Upper
228	100	
226	27.0	21.3 31.9
229	19.7	15.7 23.5









#33
Chrysene
Concen: 0.40 ng
RT: 21.300 min Scan# 2129
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	228	Resp:	29214
Ion	Ratio	Lower	Upper
228	100		
226	29.6	23.1	34.7
229	19.9	15.7	23.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.35 ng
RT: 21.183 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:	149	Resp:	7664
Ion	Ratio	Lower	Upper
149	100		
167	25.9	21.3	31.9
279	4.1	3.1	4.7









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.43 ng
RT: 25.722 min Scan# 3202
Delta R.T. -0.004 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:276 Resp: 34578
Ion Ratio Lower Upper
276 100
138 22.9 18.9 28.3
277 24.8 19.8 29.8



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.490 min Scan# 2550
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:264 Resp: 21594
Ion Ratio Lower Upper
264 100
260 25.0 19.8 29.6
265 45.9 34.4 51.6









#37
Benzo(b)fluoranthene
Concen: 0.36 ng
RT: 22.822 min Scan# 2355
Delta R.T. 0.003 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 28379
Ion Ratio Lower Upper
252 100
253 24.7 19.0 28.6
125 14.2 9.7 14.5



#38
Benzo(k)fluoranthene
Concen: 0.39 ng
RT: 22.867 min Scan# 2368
Delta R.T. 0.003 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:252 Resp: 31492
Ion Ratio Lower Upper
252 100
253 24.3 19.3 28.9
125 12.1 9.4 14.2









#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.394 min Scan# 2522
Delta R.T. 0.003 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 26531
Ion Ratio Lower Upper
252 100
253 27.6 20.7 31.1
125 23.9 11.8 17.6#



#40
Dibenzo(a,h)anthracene
Concen: 0.37 ng
RT: 25.739 min Scan# 3207
Delta R.T. -0.000 min
Lab File: BN014361.D
Acq: 21 Apr 2021 18:39

Tgt Ion:278 Resp: 27940
Ion Ratio Lower Upper
278 100
139 16.2 13.3 19.9
279 25.6 20.2 30.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.38 ng
 RT: 26.410 min Scan# 3403
 Delta R.T. -0.000 min
 Lab File: BN014361.D
 Acq: 21 Apr 2021 18:39

Tgt Ion:	276	Resp:	30131
Ion Ratio	Lower	Upper	
276	100		
277	24.7	19.1	28.7
138	20.8	16.7	25.1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

