


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014347.D
 Acq On : 21 Apr 2021 09:10
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN042221.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 21 13:57:02 2021






AutoFind: Scans 2453, 2454, 2455; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.5	75755	PASS
68	69	0.00	2	1.7	1521	PASS
69	198	0.00	100	40.2	90691	PASS
70	69	0.00	2	0.6	520	PASS
127	198	10	80	50.5	113968	PASS
197	198	0.00	2	0.3	675	PASS
198	198	100	100	100.0	225856	PASS
199	198	5	9	6.7	15187	PASS
275	198	10	60	26.1	59032	PASS
365	198	1	100	3.6	8135	PASS
441	198	0.01	100	13.1	29608	PASS
442	442	50	100	100.0	190744	PASS
443	442	15	24	21.3	40632	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
 Data File : BN014347.D
 Acq On : 21 Apr 2021 09:10
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Apr 22 05:56:32 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 09:55:03 2021
 Response via : Initial Calibration



TIC: BN014347.D\data.ms

(70) Pentachlorophenol (C)

16.716min (-0.029) 21884.93 ng




response 229798

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	61.33
264.00	61.60	62.73
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042221\
Data File : BN014347.D
Acq On : 21 Apr 2021 09:10
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Apr 22 05:56:32 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 09:55:03 2021
Response via : Initial Calibration



TIC: BN014347.D\data.ms

(77) Benzidine

19.322min (-0.029) 0.00 ng

response 1505166

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.07
183.00	13.20	11.38
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_N
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
4/22/2021	BNA_N	<u>BN014347.D</u>
Compound Name	Response	Retention Time
DDT	764266	20.563
DDD	1628	20.175
DDE	611	19.616
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2239	766505	0.29