

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042223\
 Data File : BN025071.D
 Acq On : 22 Apr 2023 12:09
 Operator : CG/JU
 Sample : PB152281BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK281

Quant Time: Apr 23 00:40:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	264605	20.000	ng/u1	0.00
20) Naphthalene-d8	10.728	136	1187914	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.563	164	773161	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.310	188	1735321	20.000	ng/u1	0.00
79) Chrysene-d12	21.510	240	1787845	20.000	ng/u1	0.00
88) Perylene-d12	23.951	264	2111814	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	41452	5.962	ng/uL	0.00
4) Pyridine-d5	3.716	84	569804	27.371	ng/u1	0.00
7) Phenol-d5	7.075	99	700088	27.138	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	500322	28.996	ng/u1	0.00
11) 2-Chlorophenol-d4	7.446	132	527012	27.798	ng/u1	0.00
15) 4-Methylphenol-d8	8.628	113	562633	27.473	ng/u1	0.00
21) Nitrobenzene-d5	9.087	128	272784	28.361	ng/u1	0.00
24) 2-Nitrophenol-d4	9.810	143	309109	29.670	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.346	165	499221	26.809	ng/u1	0.00
31) 4-Chloroaniline-d4	10.869	131	820369	29.030	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	1872959	31.249	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	1980828	28.840	ng/u1	0.00
54) 4-Nitrophenol-d4	14.757	143	322186	26.223	ng/u1	0.00
60) Fluorene-d10	15.557	176	1515982	29.957	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.675	200	287696	25.475	ng/u1	0.00
73) Anthracene-d10	17.410	188	2412767	29.604	ng/u1	0.00
81) Pyrene-d10	19.704	212	2889572	27.833	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.792	264	3347297	30.314	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042223\
 Data File : BN025071.D
 Acq On : 22 Apr 2023 12:09
 Operator : CG/JU
 Sample : PB152281BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK281

Quant Time: Apr 23 00:40:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:37:14 2023
 Response via : Initial Calibration

