

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042223\
 Data File : BN025073.D
 Acq On : 22 Apr 2023 13:20
 Operator : CG/JU
 Sample : 02399-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COMM7

Quant Time: Apr 23 00:40:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:37:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	213602	20.000	ng/ul	0.00
20) Naphthalene-d8	10.728	136	953490	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.563	164	619934	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.310	188	1387708	20.000	ng/ul	0.00
79) Chrysene-d12	21.504	240	1428307	20.000	ng/ul	0.00
88) Perylene-d12	23.951	264	1680734	20.000	ng/ul	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.293	96	26753	4.767	ng/uL	0.00
4) Pyridine-d5	3.728	84	146994	8.747	ng/ul	0.01
7) Phenol-d5	7.075	99	267382	12.839	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	586379	42.098	ng/ul	0.00
11) 2-Chlorophenol-d4	7.446	132	489110	31.959	ng/ul	0.00
15) 4-Methylphenol-d8	8.628	113	341977	20.685	ng/ul	0.00
21) Nitrobenzene-d5	9.087	128	315249	40.834	ng/ul	0.00
24) 2-Nitrophenol-d4	9.810	143	346431	41.428	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.346	165	577979	38.670	ng/ul	0.00
31) 4-Chloroaniline-d4	10.869	131	315207	13.896	ng/ul	0.00
46) Dimethylphthalate-d6	13.975	166	2172301	45.201	ng/ul	0.00
49) Acenaphthylene-d8	14.257	160	2056297	37.338	ng/ul	0.00
54) 4-Nitrophenol-d4	14.757	143	117993	11.977	ng/ul	0.00
60) Fluorene-d10	15.557	176	1876539	46.247	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.675	200	259606	28.746	ng/ul	0.00
73) Anthracene-d10	17.410	188	2825377	43.350	ng/ul	0.00
81) Pyrene-d10	19.710	212	3570832	43.054	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.792	264	3726560	42.405	ng/ul	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042223\
 Data File : BN025073.D
 Acq On : 22 Apr 2023 13:20
 Operator : CG/JU
 Sample : 02399-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COMM7

Quant Time: Apr 23 00:40:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041023.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Apr 23 00:37:14 2023
 Response via : Initial Calibration

