

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042224\
 Data File : BN031115.D
 Acq On : 22 Apr 2024 10:16
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Quant Time: Apr 22 23:25:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	127655	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.410	136	594915	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.275	164	404334	20.000	ng/u1	0.00
74) Phenanthrene-d10	17.028	188	886136	20.000	ng/u1	0.00
79) Chrysene-d12	21.269	240	1002092	20.000	ng/u1	0.00
88) Perylene-d12	24.163	264	1218351	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	21729	7.953	ng/uL	0.00
4) Pyridine-d5	3.552	84	166463	21.155	ng/u1	0.00
7) Phenol-d5	6.805	99	222314	20.645	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.969	67	135974	21.703	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.163	132	175772	20.964	ng/u1	0.00
15) 4-Methylphenol-d8	8.334	113	182417	20.143	ng/u1	0.00
21) Nitrobenzene-d5	8.787	128	91120	21.318	ng/u1	0.00
24) 2-Nitrophenol-d4	9.499	143	90935	20.466	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.034	165	185377	21.157	ng/u1	0.00
31) 4-Chloroaniline-d4	10.552	131	285306	20.837	ng/u1	-0.01
46) Dimethylphthalate-d6	13.693	166	578121	20.547	ng/u1	0.00
49) Acenaphthylene-d8	13.969	160	686830	21.541	ng/u1	0.00
54) 4-Nitrophenol-d4	14.487	143	113423	19.314	ng/u1	0.00
60) Fluorene-d10	15.275	176	522498	21.332	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.398	200	82032	18.349	ng/u1	0.00
73) Anthracene-d10	17.128	188	847003	22.053	ng/u1	0.00
81) Pyrene-d10	19.428	212	1071431	21.758	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	1276124	22.088	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.181	88	24608	8.498	ng/uL	97
5) Pyridine	3.570	79	174367	21.959	ng/u1	98
6) Benzaldehyde	6.781	77	132756	25.107	ng/u1	99
8) Phenol	6.828	94	235026	21.169	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.064	93	190851	21.787	ng/u1	97
12) 2-Chlorophenol	7.193	128	188904	21.675	ng/u1	99
13) 2-Methylphenol	8.069	108	179776	20.895	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.152	45	252779	22.133	ng/u1	99
16) Acetophenone	8.452	105	305298	21.758	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.434	70	143893	20.574	ng/u1	99
18) 4-Methylphenol	8.399	108	200623	20.817	ng/u1	96
19) Hexachloroethane	8.693	117	77579	21.452	ng/u1	95
22) Nitrobenzene	8.828	77	223368	22.141	ng/u1	98
23) Isophorone	9.346	82	412244	20.024	ng/u1	99
25) 2-Nitrophenol	9.534	139	106071	21.577	ng/u1	98
26) 2,4-Dimethylphenol	9.593	107	210385	20.985	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.834	93	267848	21.419	ng/u1	98
29) 2,4-Dichlorophenol	10.057	162	189104	21.608	ng/u1	98
30) Naphthalene	10.457	128	671483	21.779	ng/u1	99
32) 4-Chloroaniline	10.575	127	280826	21.108	ng/u1	97
33) Hexachlorobutadiene	10.740	225	111905	20.975	ng/u1	96
34) Caprolactam	11.352	113	58179	16.873	ng/u1	96
35) 4-Chloro-3-methylphenol	11.704	107	207758	20.475	ng/u1	97
36) 2-Methylnaphthalene	12.075	142	467248	21.490	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042224\
 Data File : BN031115.D
 Acq On : 22 Apr 2024 10:16
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Quant Time: Apr 22 23:25:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.299	142	472735	21.487	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.446	216	231857	21.914	ng/ul	98
40) Hexachlorocyclopentadiene	12.428	237	107641	18.521	ng/ul	100
41) 2,4,6-Trichlorophenol	12.698	196	149408	21.216	ng/ul	91
42) 2,4,5-Trichlorophenol	12.769	196	165126	21.148	ng/ul	99
43) 1,1'-Biphenyl	13.104	154	617191	22.195	ng/ul	100
44) 2-Chloronaphthalene	13.146	162	488402	21.979	ng/ul	99
45) 2-Nitroaniline	13.357	65	127315	20.754	ng/ul	96
47) Dimethylphthalate	13.740	163	588425	20.272	ng/ul	98
48) 2,6-Dinitrotoluene	13.863	165	112875	20.716	ng/ul	100
50) Acenaphthylene	13.998	152	806466	21.653	ng/ul	99
51) 3-Nitroaniline	14.193	138	129941	20.773	ng/ul	94
52) Acenaphthene	14.340	153	530587	21.490	ng/ul	96
53) 2,4-Dinitrophenol	14.398	184	44147	14.327	ng/ul	95
55) 4-Nitrophenol	14.504	109	92280	19.362	ng/ul	99
56) Dibenzofuran	14.681	168	744113	21.798	ng/ul	98
57) 2,4-Dinitrotoluene	14.651	165	171915	20.765	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.904	232	139356	20.652	ng/ul	98
59) Diethylphthalate	15.110	149	607472	20.350	ng/ul	99
61) Fluorene	15.328	166	599852	21.353	ng/ul	96
62) 4-Chlorophenyl-phenyle...	15.328	204	291016	21.749	ng/ul	99
63) 4-Nitroaniline	15.357	138	148213	21.976	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.410	198	91438	19.048	ng/ul	98
67) N-Nitrosodiphenylamine	15.545	169	513665	21.627	ng/ul	98
68) 4-Bromophenyl-phenylether	16.222	248	179581	21.429	ng/ul	98
69) Hexachlorobenzene	16.328	284	202953	21.007	ng/ul	95
70) Atrazine	16.504	200	171366	20.551	ng/ul	94
71) Pentachlorophenol	16.675	266	124714	19.869	ng/ul	96
72) Phenanthrene	17.069	178	1022297	21.812	ng/ul	98
74) Anthracene	17.157	178	1048697	22.110	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.063	216	238662	22.962	ng/uL	100
76) Pentachlorobenzene	14.592	250	234678	21.865	ng/uL	96
77) Carbazole	17.439	167	967620	22.333	ng/ul	98
78) Di-n-butylphthalate	18.004	149	1116684	21.145	ng/ul	100
80) Fluoranthene	19.092	202	1243516	21.337	ng/ul	100
82) Pyrene	19.457	202	1312989	21.442	ng/ul	99
83) Butylbenzylphthalate	20.375	149	546753	20.236	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.180	252	439080	21.028	ng/ul	95
85) Benzo(a)anthracene	21.251	228	1449820	21.306	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	846617	20.869	ng/ul	98
87) Chrysene	21.310	228	1407550	21.791	ng/ul	98
89) Di-n-octyl phthalate	22.280	149	1489800	21.214	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	1550181	22.613	ng/ul	100
91) Benzo(k)fluoranthene	23.304	252	1529944	22.190	ng/ul	98
93) Benzo(a)pyrene	24.033	252	1492519	21.825	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.409	276	1710753	20.159	ng/ul	96
95) Dibenzo(a,h)anthracene	27.468	278	1454633	20.884	ng/ul	96
96) Benzo(g,h,i)perylene	28.433	276	1379267	20.183	ng/ul	97





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 8.498 ng/uL
 RT: 3.181 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN031115.D
 Acq: 22 Apr 2024 10:16




Instrument :
 BNA_N
 ClientSampleId :
 SST020373


Tgt Ion: 88 Resp: 24608
 Ion Ratio Lower Upper
 88 100
 43 23.0 22.2 33.2
 58 63.1 50.2 75.2



#5
 Pyridine
 Concen: 21.959 ng/uL
 RT: 3.570 min Scan# 99
 Delta R.T. -0.006 min
 Lab File: BN031115.D
 Acq: 22 Apr 2024 10:16

Tgt Ion: 79 Resp: 174367
 Ion Ratio Lower Upper
 79 100
 52 77.1 59.9 89.9
 51 37.0 29.0 43.6









#6
Benzaldehyde
Concen: 25.107 ng/ul
RT: 6.781 min Scan# 646
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion: 77 Resp: 132756
Ion Ratio Lower Upper
77 100
105 101.5 80.0 120.0
106 94.2 75.0 112.6



#8
Phenol
Concen: 21.169 ng/ul
RT: 6.828 min Scan# 653
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion: 94 Resp: 235026
Ion Ratio Lower Upper
94 100
65 27.3 22.0 33.0
66 36.6 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 21.787 ng/ul
RT: 7.064 min Scan# 694
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion: 93 Resp: 190851
Ion Ratio Lower Upper
93 100
63 74.8 57.1 85.7
95 31.8 24.9 37.3



#12
2-Chlorophenol
Concen: 21.675 ng/ul
RT: 7.193 min Scan# 715
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:128 Resp: 188904
Ion Ratio Lower Upper
128 100
64 44.4 34.9 52.3
130 32.9 27.0 40.4





#13

2-Methylphenol

Concen: 20.895 ng/ul

RT: 8.069 min Scan# 8

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :





SSTD020373

Tgt Ion:108 Resp: 179776

Ion Ratio Lower Upper

108 100

107 88.0 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.133 ng/ul

RT: 8.152 min Scan# 878

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion: 45 Resp: 252779


Ion Ratio Lower Upper

45 100

77 15.2 12.5 18.7

79 12.2 9.6 14.4









#16
Acetophenone
Concen: 21.758 ng/ul
RT: 8.452 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:105 Resp: 305298
Ion Ratio Lower Upper
105 100
77 74.0 60.2 90.4
51 24.8 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 20.574 ng/ul
RT: 8.434 min Scan# 926
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion: 70 Resp: 143893
Ion Ratio Lower Upper
70 100
42 57.5 45.0 67.6
101 10.4 8.1 12.1
130 23.6 19.2 28.8









#18
4-Methylphenol
Concen: 20.817 ng/ul
RT: 8.399 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:108 Resp: 200623
Ion Ratio Lower Upper
108 100
107 107.9 89.8 134.8



#19
Hexachloroethane
Concen: 21.452 ng/ul
RT: 8.693 min Scan# 970
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:117 Resp: 77579
Ion Ratio Lower Upper
117 100
201 91.5 77.4 116.0
199 58.5 49.6 74.4









#22
Nitrobenzene
Concen: 22.141 ng/ul
RT: 8.828 min Scan# 993
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion: 77 Resp: 223368
Ion Ratio Lower Upper
77 100
123 46.1 38.2 57.2
65 13.3 10.7 16.1



#23
Isophorone
Concen: 20.024 ng/ul
RT: 9.346 min Scan# 1081
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion: 82 Resp: 412244
Ion Ratio Lower Upper
82 100
95 7.4 5.9 8.9
138 17.9 14.7 22.1









#25
2-Nitrophenol
Concen: 21.577 ng/ul
RT: 9.534 min Scan# 1113
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:139 Resp: 106071
Ion Ratio Lower Upper
139 100
65 45.1 35.2 52.8
109 30.1 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 20.985 ng/ul
RT: 9.593 min Scan# 1123
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:107 Resp: 210385
Ion Ratio Lower Upper
107 100
121 49.6 40.3 60.5
122 87.9 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.419 ng/ul

RT: 9.834 min Scan# 1164

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion: 93 Resp: 267848

Ion Ratio Lower Upper

93 100

95 31.3 25.8 38.6

123 11.8 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 21.608 ng/ul

RT: 10.057 min Scan# 1202

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion: 162 Resp: 189104


Ion Ratio Lower Upper

162 100

164 62.1 51.0 76.6

98 36.6 28.6 42.8









#30
Naphthalene
Concen: 21.779 ng/ul
RT: 10.457 min Scan# 1270
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:128 Resp: 671483
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 12.8 10.5 15.7



#32
4-Chloroaniline
Concen: 21.108 ng/ul
RT: 10.575 min Scan# 1290
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:127 Resp: 280826
Ion Ratio Lower Upper
127 100
129 33.3 25.4 38.2









#33
Hexachlorobutadiene
Concen: 20.975 ng/ul
RT: 10.740 min Scan# 111905
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:225 Resp: 111905
Ion Ratio Lower Upper
225 100
223 62.3 49.8 74.6
227 67.0 48.8 73.2



#34
Caprolactam
Concen: 16.873 ng/ul
RT: 11.352 min Scan# 1422
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:113 Resp: 58179
Ion Ratio Lower Upper
113 100
55 146.8 112.4 168.6
56 113.4 88.0 132.0









#35
4-Chloro-3-methylphenol
Concen: 20.475 ng/ul
RT: 11.704 min Scan# 1482
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:107 Resp: 207758
Ion Ratio Lower Upper
107 100
144 26.3 20.6 31.0
142 80.7 67.4 101.2



#36
2-Methylnaphthalene
Concen: 21.490 ng/ul
RT: 12.075 min Scan# 1545
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:142 Resp: 467248
Ion Ratio Lower Upper
142 100
141 87.0 69.6 104.4





#37

1-Methylnaphthalene

Concen: 21.487 ng/ul

RT: 12.299 min Scan# 1583

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:142 Resp: 472735

Ion Ratio Lower Upper

142 100

141 87.8 70.2 105.4

116 3.6 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.914 ng/ul

RT: 12.446 min Scan# 1608

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Tgt Ion:216 Resp: 231857




Ion Ratio Lower Upper


216 100

214 77.7 63.1 94.7

179 21.6 17.0 25.6

108 19.5 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 18.521 ng/ul
RT: 12.428 min Scan# 1605
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:237 Resp: 107641
Ion Ratio Lower Upper
237 100
235 59.7 47.6 71.4
272 13.2 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 21.216 ng/ul
RT: 12.698 min Scan# 1651
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:196 Resp: 149408
Ion Ratio Lower Upper
196 100
198 98.2 69.9 104.9
200 30.6 23.4 35.2









#42
2,4,5-Trichlorophenol
Concen: 21.148 ng/ul
RT: 12.769 min Scan# 1663
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:196 Resp: 165126
Ion Ratio Lower Upper
196 100
198 94.1 75.9 113.9
200 30.6 24.7 37.1



#43
1,1'-Biphenyl
Concen: 22.195 ng/ul
RT: 13.104 min Scan# 1720
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:154 Resp: 617191
Ion Ratio Lower Upper
154 100
153 39.7 31.8 47.8
76 12.6 10.6 15.8





#44

2-Chloronaphthalene

Concen: 21.979 ng/ul

RT: 13.146 min Scan# 1727

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:162 Resp: 488402

Ion Ratio Lower Upper

162 100

164 31.7 25.8 38.8

127 41.0 33.7 50.5



#45

2-Nitroaniline

Concen: 20.754 ng/ul

RT: 13.357 min Scan# 1763

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion: 65 Resp: 127315


Ion Ratio Lower Upper

65 100

92 70.9 57.4 86.2

138 108.4 91.6 137.4









#47
 Dimethylphthalate
 Concen: 20.272 ng/ul
 RT: 13.740 min Scan# 1828
 Delta R.T. -0.006 min
 Lab File: BN031115.D
 Acq: 22 Apr 2024 10:16




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373


Tgt Ion:	163	Resp:	588425
Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.0	4.6
164	10.8	8.0	12.0



#48
 2,6-Dinitrotoluene
 Concen: 20.716 ng/ul
 RT: 13.863 min Scan# 1849
 Delta R.T. -0.006 min
 Lab File: BN031115.D
 Acq: 22 Apr 2024 10:16

Tgt Ion:	165	Resp:	112875
Ion	Ratio	Lower	Upper
165	100		
89	55.8	44.7	67.1
121	21.4	17.0	25.4









#50
Acenaphthylene
Concen: 21.653 ng/ul
RT: 13.998 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument : BNA_N
ClientSampleId : SST020373


Tgt Ion:	152	Resp:	806466
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.8	23.8
153	12.7	10.5	15.7



#51
3-Nitroaniline
Concen: 20.773 ng/ul
RT: 14.193 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:	138	Resp:	129941
Ion Ratio	Lower	Upper	
138	100		
108	12.4	11.1	16.7
92	110.8	93.9	140.9





#52

Acenaphthene

Concen: 21.490 ng/ul

RT: 14.340 min Scan# 1931

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:153 Resp: 530587

Ion Ratio Lower Upper

153 100

152 45.7 36.3 54.5

154 85.4 72.4 108.6



#53

2,4-Dinitrophenol

Concen: 14.327 ng/ul

RT: 14.398 min Scan# 1940

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:184 Resp: 44147


Ion Ratio Lower Upper

184 100

63 64.0 50.4 75.6

154 60.0 53.0 79.6









#55
4-Nitrophenol
Concen: 19.362 ng/ul
RT: 14.504 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:109 Resp: 92280
Ion Ratio Lower Upper
109 100
139 134.2 105.6 158.4
65 123.8 98.2 147.2



#56
Dibenzofuran
Concen: 21.798 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:168 Resp: 744113
Ion Ratio Lower Upper
168 100
139 37.6 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 20.765 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:165 Resp: 171915
Ion Ratio Lower Upper
165 100
63 43.6 33.9 50.9
182 2.9 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.652 ng/ul
RT: 14.904 min Scan# 2026
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:232 Resp: 139356
Ion Ratio Lower Upper
232 100
131 50.3 41.2 61.8
130 3.0 2.2 3.2
166 32.0 27.3 40.9





#59

Diethylphthalate

Concen: 20.350 ng/ul

RT: 15.110 min Scan# 2061

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:149 Resp: 607472

Ion Ratio Lower Upper

149 100

177 21.6 16.8 25.2

150 12.5 10.1 15.1



#61

Fluorene

Concen: 21.353 ng/ul

RT: 15.328 min Scan# 2098

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:166 Resp: 599852


Ion Ratio Lower Upper

166 100

165 97.8 74.6 112.0


167 13.9 10.6 16.0








#62
4-Chlorophenyl-phenylether
Concen: 21.749 ng/ul
RT: 15.328 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Instrument :
BNA_N
ClientSampleId :
SSTD020373







Tgt Ion:204 Resp: 291016
Ion Ratio Lower Upper
204 100
206 31.4 26.4 39.6
141 63.1 50.7 76.1



#63
4-Nitroaniline
Concen: 21.976 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:138 Resp: 148213
Ion Ratio Lower Upper
138 100
92 50.5 38.4 57.6
108 77.9 63.8 95.8









#66
4,6-Dinitro-2-methylphenol
Concen: 19.048 ng/ul
RT: 15.410 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:198 Resp: 91438
Ion Ratio Lower Upper
198 100
182 5.4 4.0 6.0
77 28.9 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 21.627 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:169 Resp: 513665
Ion Ratio Lower Upper
169 100
168 66.7 52.6 79.0
167 36.6 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 21.429 ng/ul
RT: 16.222 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:248 Resp: 179581
Ion Ratio Lower Upper
248 100
250 97.0 78.2 117.2
141 85.6 71.4 107.0



#69
Hexachlorobenzene
Concen: 21.007 ng/ul
RT: 16.328 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:284 Resp: 202953
Ion Ratio Lower Upper
284 100
142 40.5 30.0 45.0
249 34.5 25.5 38.3





#70

Atrazine

Concen: 20.551 ng/ul

RT: 16.504 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:200 Resp: 171366

Ion Ratio Lower Upper

200 100

173 28.5 20.6 31.0

215 50.8 37.6 56.4



#71

Pentachlorophenol

Concen: 19.869 ng/ul

RT: 16.675 min Scan# 2327

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:266 Resp: 124714


Ion Ratio Lower Upper

266 100

264 60.7 49.1 73.7

268 67.2 49.2 73.8









#72
Phenanthrene
Concen: 21.812 ng/ul
RT: 17.069 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:178 Resp: 1022297
Ion Ratio Lower Upper
178 100
179 14.6 12.4 18.6
176 18.7 14.6 21.8



#74
Anthracene
Concen: 22.110 ng/ul
RT: 17.157 min Scan# 2409
Delta R.T. -0.012 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:178 Resp: 1048697
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 17.9 14.4 21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.962 ng/uL

RT: 13.063 min Scan# 1713

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:216 Resp: 238662

Ion Ratio Lower Upper

216 100

214 78.6 63.0 94.6

218 47.6 38.3 57.5



#76

Pentachlorobenzene

Concen: 21.865 ng/uL

RT: 14.592 min Scan# 1973

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:250 Resp: 234678


Ion Ratio Lower Upper

250 100

248 64.0 49.3 73.9

252 62.1 52.9 79.3









#77
 Carbazole
 Concen: 22.333 ng/ul
 RT: 17.439 min Scan# 2457
 Delta R.T. 0.000 min
 Lab File: BN031115.D
 Acq: 22 Apr 2024 10:16

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020373

Tgt Ion:167 Resp: 967620




Ion	Ratio	Lower	Upper
167	100		
166	21.3	16.2	24.2
139	13.3	10.6	15.8




#78
 Di-n-butylphthalate
 Concen: 21.145 ng/ul
 RT: 18.004 min Scan# 2553
 Delta R.T. -0.006 min
 Lab File: BN031115.D
 Acq: 22 Apr 2024 10:16

Tgt Ion:149 Resp: 1116684

Ion	Ratio	Lower	Upper
149	100		
150	8.9	7.1	10.7
104	5.0	3.9	5.9









#80
Fluoranthene
Concen: 21.337 ng/ul
RT: 19.092 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16




Instrument :
BNA_N
ClientSampleId :
SSTD020373


Tgt Ion:202 Resp: 1243516
Ion Ratio Lower Upper
202 100
101 11.3 9.0 13.6
100 8.6 6.9 10.3



#82
Pyrene
Concen: 21.442 ng/ul
RT: 19.457 min Scan# 2800
Delta R.T. -0.006 min
Lab File: BN031115.D
Acq: 22 Apr 2024 10:16

Tgt Ion:202 Resp: 1312989
Ion Ratio Lower Upper
202 100
101 14.9 11.1 16.7
100 11.6 9.3 13.9





#83

Butylbenzylphthalate

Concen: 20.236 ng/ul

RT: 20.375 min Scan# 2956

Delta R.T. 0.000 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:149 Resp: 546753

Ion Ratio Lower Upper

149 100

91 67.8 57.3 85.9

206 19.8 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 21.028 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:252 Resp: 439080


Ion Ratio Lower Upper

252 100

254 63.5 54.5 81.7

126 13.0 9.8 14.6





#85

Benzo(a)anthracene

Concen: 21.306 ng/ul

RT: 21.251 min Scan# 3104

Delta R.T. 0.000 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:228 Resp: 1449820

Ion Ratio Lower Upper

228 100

229 20.6 16.7 25.1

226 27.0 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.869 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. 0.000 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:149 Resp: 846617


Ion Ratio Lower Upper

149 100

167 26.9 22.3 33.5

279 4.1 3.0 4.6





#87

Chrysene

Concen: 21.791 ng/ul

RT: 21.310 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:228 Resp: 1407550

Ion Ratio Lower Upper

228 100

226 28.7 22.6 34.0

229 18.1 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 21.214 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. 0.000 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Tgt Ion:149 Resp: 1489800





#90

Benzo(b)fluoranthene

Concen: 22.613 ng/ul

RT: 23.251 min Scan# 3445

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:252 Resp: 1550181

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 10.8 8.6 12.8



#91

Benzo(k)fluoranthene

Concen: 22.190 ng/ul

RT: 23.304 min Scan# 3454

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:252 Resp: 1529944


Ion Ratio Lower Upper

252 100

253 23.0 17.4 26.0

125 11.7 9.4 14.2





#93

Benzo(a)pyrene

Concen: 21.825 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. 0.000 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:252 Resp: 1492519

Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 11.7 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.159 ng/ul

RT: 27.409 min Scan# 4152

Delta R.T. -0.012 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16




Tgt Ion:276 Resp: 1710753


Ion Ratio Lower Upper

276 100

138 22.1 17.4 26.2

277 26.1 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 20.884 ng/ul

RT: 27.468 min Scan# 4162

Delta R.T. -0.018 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Instrument :

BNA_N

ClientSampleId :

SSTD020373





Tgt Ion:278 Resp: 1454633

Ion Ratio Lower Upper

278 100

139 17.5 13.3 19.9

279 24.3 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 20.183 ng/ul

RT: 28.433 min Scan# 4326

Delta R.T. -0.006 min

Lab File: BN031115.D

Acq: 22 Apr 2024 10:16

Tgt Ion:276 Resp: 1379267

Ion Ratio Lower Upper

276 100

138 22.2 19.0 28.6

277 22.9 19.3 28.9

