

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042224\
 Data File : BN031131.D
 Acq On : 22 Apr 2024 21:33
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020374

Quant Time: Apr 22 23:30:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	136342	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	618862	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	423324	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.034	188	930553	20.000	ng/u1	0.00
79) Chrysene-d12	21.269	240	1039064	20.000	ng/u1	0.00
88) Perylene-d12	24.169	264	1342968	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	23148	7.932	ng/uL	0.00
4) Pyridine-d5	3.552	84	178231	21.207	ng/u1	0.00
7) Phenol-d5	6.811	99	241727	21.017	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	137739	20.584	ng/u1	0.00
11) 2-Chlorophenol-d4	7.164	132	194446	21.713	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	200662	20.746	ng/u1	0.00
21) Nitrobenzene-d5	8.787	128	98688	22.195	ng/u1	0.00
24) 2-Nitrophenol-d4	9.505	143	105762	22.882	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.040	165	204397	22.425	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	305269	21.432	ng/u1	0.00
46) Dimethylphthalate-d6	13.698	166	613954	20.842	ng/u1	0.00
49) Acenaphthylene-d8	13.975	160	733511	21.974	ng/u1	0.00
54) 4-Nitrophenol-d4	14.493	143	128238	20.857	ng/u1	0.00
60) Fluorene-d10	15.275	176	554347	21.617	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.404	200	95073	20.251	ng/u1	0.00
73) Anthracene-d10	17.128	188	912221	22.617	ng/u1	0.00
81) Pyrene-d10	19.433	212	1148801	22.499	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.969	264	1428716	22.434	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.176	88	24614	7.959	ng/uL	92
5) Pyridine	3.570	79	176272	20.785	ng/u1	99
6) Benzaldehyde	6.787	77	143487	25.408	ng/u1	97
8) Phenol	6.834	94	251907	21.244	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.069	93	197407	21.099	ng/u1	99
12) 2-Chlorophenol	7.199	128	203877	21.903	ng/u1	98
13) 2-Methylphenol	8.075	108	193619	21.070	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.158	45	251869	20.648	ng/u1	98
16) Acetophenone	8.452	105	312881	20.878	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.440	70	153712	20.578	ng/u1	99
18) 4-Methylphenol	8.405	108	212691	20.663	ng/u1	99
19) Hexachloroethane	8.699	117	83035	21.497	ng/u1	97
22) Nitrobenzene	8.834	77	231351	22.045	ng/u1	100
23) Isophorone	9.352	82	454380	21.216	ng/u1	100
25) 2-Nitrophenol	9.534	139	114682	22.426	ng/u1	95
26) 2,4-Dimethylphenol	9.599	107	214821	20.599	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.840	93	275763	21.199	ng/u1	100
29) 2,4-Dichlorophenol	10.063	162	198815	21.839	ng/u1	96
30) Naphthalene	10.463	128	702437	21.901	ng/u1	99
32) 4-Chloroaniline	10.581	127	294781	21.299	ng/u1	97
33) Hexachlorobutadiene	10.740	225	122512	22.075	ng/u1	95
34) Caprolactam	11.357	113	72772	20.288	ng/u1	98
35) 4-Chloro-3-methylphenol	11.710	107	227506	21.554	ng/u1	98
36) 2-Methylnaphthalene	12.081	142	486541	21.512	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042224\
 Data File : BN031131.D
 Acq On : 22 Apr 2024 21:33
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020374

Quant Time: Apr 22 23:30:08 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.304	142	486285	21.248	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.451	216	246825	22.283	ng/ul	99
40) Hexachlorocyclopentadiene	12.428	237	106732	17.541	ng/ul	98
41) 2,4,6-Trichlorophenol	12.704	196	164632	22.329	ng/ul	94
42) 2,4,5-Trichlorophenol	12.775	196	177076	21.662	ng/ul	98
43) 1,1'-Biphenyl	13.110	154	627903	21.567	ng/ul	99
44) 2-Chloronaphthalene	13.151	162	511089	21.968	ng/ul	97
45) 2-Nitroaniline	13.363	65	144055	22.429	ng/ul	99
47) Dimethylphthalate	13.746	163	623039	20.501	ng/ul	100
48) 2,6-Dinitrotoluene	13.869	165	125681	22.031	ng/ul	98
50) Acenaphthylene	13.998	152	846983	21.721	ng/ul	99
51) 3-Nitroaniline	14.198	138	145636	22.237	ng/ul	94
52) Acenaphthene	14.346	153	551064	21.318	ng/ul	97
53) 2,4-Dinitrophenol	14.404	184	56000	17.358	ng/ul	97
55) 4-Nitrophenol	14.510	109	105176	21.078	ng/ul	98
56) Dibenzofuran	14.681	168	761361	21.303	ng/ul	99
57) 2,4-Dinitrotoluene	14.657	165	191626	22.107	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	14.910	232	159319	22.551	ng/ul	95
59) Diethylphthalate	15.116	149	663379	21.225	ng/ul	98
61) Fluorene	15.334	166	631616	21.475	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.334	204	302062	21.562	ng/ul	98
63) 4-Nitroaniline	15.363	138	169055	23.942	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.416	198	103091	20.450	ng/ul	99
67) N-Nitrosodiphenylamine	15.551	169	553831	22.205	ng/ul	99
68) 4-Bromophenyl-phenylether	16.228	248	192959	21.926	ng/ul	93
69) Hexachlorobenzene	16.334	284	218121	21.500	ng/ul	97
70) Atrazine	16.504	200	198889	22.713	ng/ul	96
71) Pentachlorophenol	16.681	266	146179	22.177	ng/ul	93
72) Phenanthrene	17.075	178	1054221	21.419	ng/ul	97
74) Anthracene	17.163	178	1106411	22.213	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.069	216	244426	22.394	ng/uL	99
76) Pentachlorobenzene	14.598	250	248403	22.039	ng/uL	97
77) Carbazole	17.439	167	1050476	23.088	ng/ul	98
78) Di-n-butylphthalate	18.004	149	1276787	23.023	ng/ul	99
80) Fluoranthene	19.098	202	1366267	22.609	ng/ul	99
82) Pyrene	19.463	202	1422678	22.407	ng/ul	98
83) Butylbenzylphthalate	20.375	149	674866	24.088	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.186	252	528566	24.413	ng/ul	92
85) Benzo(a)anthracene	21.251	228	1569455	22.243	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	1009535	23.999	ng/ul	100
87) Chrysene	21.310	228	1445469	21.582	ng/ul	98
89) Di-n-octyl phthalate	22.280	149	1938132	25.037	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	1685705	22.308	ng/ul	97
91) Benzo(k)fluoranthene	23.316	252	1714816	22.564	ng/ul	98
93) Benzo(a)pyrene	24.033	252	1644295	21.813	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.421	276	1633496	17.463	ng/ul	99
95) Dibenzo(a,h)anthracene	27.474	278	1581708	20.601	ng/ul	98
96) Benzo(g,h,i)perylene	28.439	276	1509621	20.040	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA N


ClientSampleId :
SSTD020374

TIC: BN031131.D\data.ms

Abundance

Time-->





1,4-Dioxane-d8,S
Pyridine-d5,S
Bis(2-chlorophenyl) ether-d8,S
1,4-Dichlorobenzene-d4,l
2,2'-oxybis(2-methylphenol)
2-Methylphenol
2-Nitrophenol
Bis(2-chlorophenyl) ether-d8,S
4-Chlorophenol
Hexachlorocyclopentadiene
Caprolactam
4-Chloro-3-methylphenol,C
2-Methylnaphthalene
Hexachlorocyclopentadiene
2,4,5-Trichlorophenol,C
2,4,6-Trichlorophenol,C
2-Nitroaniline
2,6-Dinitrophenol,S
2,4-Dinitrophenol,S
3-Nitroaniline
Acenaphthene
2,3,4,6-Tetrachlorophenol
2,3,4,6-Tetrachlorophenol-d10,S
4-Nitrophenol
4-Nitrophenylamine
4-Nitrophenyl ether
Hexachlorocyclopentadiene
Pentachlorophenol,C
Carbazole
Di-n-butylphthalate
Fluoranthene
Pyrene-d10,S
Butylbenzylphthalate
Chrysene
Benzo(a)anthracene
Benzo(b)fluoranthene
Benzo(k)fluoranthene
Perylene
Indeno(1,2,3-cd)pyrene
Benzo(g,h,i)perylene



#2
1,4-Dioxane
Concen: 7.959 ng/uL
RT: 3.176 min Scan# 31
Delta R.T. -0.005 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion: 88 Resp: 24614
Ion Ratio Lower Upper
88 100
43 32.3 22.2 33.2
58 68.5 50.2 75.2



#5
Pyridine
Concen: 20.785 ng/uL
RT: 3.570 min Scan# 99
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion: 79 Resp: 176272
Ion Ratio Lower Upper
79 100
52 74.5 59.9 89.9
51 34.8 29.0 43.6









#6
Benzaldehyde
Concen: 25.408 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion: 77 Resp: 143487
Ion Ratio Lower Upper
77 100
105 99.8 80.0 120.0
106 99.9 75.0 112.6



#8
Phenol
Concen: 21.244 ng/ul
RT: 6.834 min Scan# 654
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion: 94 Resp: 251907
Ion Ratio Lower Upper
94 100
65 25.8 22.0 33.0
66 36.9 29.8 44.6





#10

Bis(2-Chloroethyl)ether

Concen: 21.099 ng/ul

RT: 7.069 min Scan# 694

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion: 93 Resp: 197407

Ion Ratio Lower Upper

93 100

63 72.1 57.1 85.7

95 31.9 24.9 37.3



#12

2-Chlorophenol

Concen: 21.903 ng/ul

RT: 7.199 min Scan# 716

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:128 Resp: 203877


Ion Ratio Lower Upper

128 100

64 42.8 34.9 52.3

130 31.9 27.0 40.4





#13

2-Methylphenol

Concen: 21.070 ng/ul

RT: 8.075 min Scan# 865

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :





SSTD020374

Tgt Ion:108 Resp: 193619

Ion Ratio Lower Upper

108 100

107 89.4 70.1 105.1



#14

2,2'-oxybis(1-Chloropropane)

Concen: 20.648 ng/ul

RT: 8.158 min Scan# 879

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion: 45 Resp: 251869


Ion Ratio Lower Upper

45 100

77 16.6 12.5 18.7

79 11.4 9.6 14.4









#16
Acetophenone
Concen: 20.878 ng/ul
RT: 8.452 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:105 Resp: 312881
Ion Ratio Lower Upper
105 100
77 76.5 60.2 90.4
51 25.6 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 20.578 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion: 70 Resp: 153712
Ion Ratio Lower Upper
70 100
42 55.9 45.0 67.6
101 11.1 8.1 12.1
130 22.6 19.2 28.8





#18

4-Methylphenol

Concen: 20.663 ng/ul

RT: 8.405 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :





SSTD020374

Tgt Ion:108 Resp: 212691

Ion Ratio Lower Upper

108 100

107 111.4 89.8 134.8



#19

Hexachloroethane

Concen: 21.497 ng/ul

RT: 8.699 min Scan# 971

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:117 Resp: 83035


Ion Ratio Lower Upper

117 100

201 93.6 77.4 116.0

199 58.9 49.6 74.4









#22
Nitrobenzene
Concen: 22.045 ng/ul
RT: 8.834 min Scan# 993
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion: 77 Resp: 231351
Ion Ratio Lower Upper
77 100
123 47.6 38.2 57.2
65 13.7 10.7 16.1



#23
Isophorone
Concen: 21.216 ng/ul
RT: 9.352 min Scan# 1082
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion: 82 Resp: 454380
Ion Ratio Lower Upper
82 100
95 7.6 5.9 8.9
138 18.6 14.7 22.1









#25
2-Nitrophenol
Concen: 22.426 ng/ul
RT: 9.534 min Scan# 1113
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:139 Resp: 114682
Ion Ratio Lower Upper
139 100
65 48.8 35.2 52.8
109 32.4 25.4 38.0



#26
2,4-Dimethylphenol
Concen: 20.599 ng/ul
RT: 9.599 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:107 Resp: 214821
Ion Ratio Lower Upper
107 100
121 53.0 40.3 60.5
122 86.7 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.199 ng/ul

RT: 9.840 min Scan# 1164

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion: 93 Resp: 275763

Ion Ratio Lower Upper

93 100

95 32.2 25.8 38.6

123 11.8 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 21.839 ng/ul

RT: 10.063 min Scan# 1203

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:162 Resp: 198815


Ion Ratio Lower Upper

162 100

164 67.7 51.0 76.6

98 36.6 28.6 42.8









#30
Naphthalene
Concen: 21.901 ng/ul
RT: 10.463 min Scan# 1271
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:128 Resp: 702437
Ion Ratio Lower Upper
128 100
129 10.5 9.0 13.4
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 21.299 ng/ul
RT: 10.581 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:127 Resp: 294781
Ion Ratio Lower Upper
127 100
129 33.5 25.4 38.2





#33

Hexachlorobutadiene

Concen: 22.075 ng/ul

RT: 10.740 min Scan# 11

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:225 Resp: 122512

Ion Ratio Lower Upper

225 100

223 66.1 49.8 74.6

227 65.6 48.8 73.2



#34

Caprolactam

Concen: 20.288 ng/ul

RT: 11.357 min Scan# 1423

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:113 Resp: 72772


Ion Ratio Lower Upper

113 100

55 144.2 112.4 168.6

56 109.1 88.0 132.0









#35
4-Chloro-3-methylphenol
Concen: 21.554 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:107 Resp: 227506
Ion Ratio Lower Upper
107 100
144 27.3 20.6 31.0
142 85.4 67.4 101.2



#36
2-Methylnaphthalene
Concen: 21.512 ng/ul
RT: 12.081 min Scan# 1546
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:142 Resp: 486541
Ion Ratio Lower Upper
142 100
141 87.2 69.6 104.4





#37
1-Methylnaphthalene
Concen: 21.248 ng/ul
RT: 12.304 min Scan# 1584
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33





Instrument :

BNA_N

ClientSampleId :




SSTD020374


Tgt Ion:142 Resp: 486285
Ion Ratio Lower Upper
142 100
141 90.7 70.2 105.4
116 3.8 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 22.283 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:216 Resp: 246825
Ion Ratio Lower Upper
216 100
214 77.3 63.1 94.7
179 21.0 17.0 25.6
108 17.8 14.2 21.4









#40
Hexachlorocyclopentadiene
Concen: 17.541 ng/ul
RT: 12.428 min Scan# 1605
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:237 Resp: 106732
Ion Ratio Lower Upper
237 100
235 61.4 47.6 71.4
272 13.3 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 22.329 ng/ul
RT: 12.704 min Scan# 1652
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:196 Resp: 164632
Ion Ratio Lower Upper
196 100
198 94.0 69.9 104.9
200 30.0 23.4 35.2





#42

2,4,5-Trichlorophenol

Concen: 21.662 ng/ul

RT: 12.775 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:196 Resp: 177076

Ion Ratio Lower Upper

196 100

198 96.1 75.9 113.9

200 28.3 24.7 37.1



#43

1,1'-Biphenyl

Concen: 21.567 ng/ul

RT: 13.110 min Scan# 1721

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:154 Resp: 627903


Ion Ratio Lower Upper

154 100

153 40.5 31.8 47.8

76 12.3 10.6 15.8





#44

2-Chloronaphthalene

Concen: 21.968 ng/ul

RT: 13.151 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:162 Resp: 511089

Ion Ratio Lower Upper

162 100

164 31.6 25.8 38.8

127 39.2 33.7 50.5



#45

2-Nitroaniline

Concen: 22.429 ng/ul

RT: 13.363 min Scan# 1764

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion: 65 Resp: 144055


Ion Ratio Lower Upper

65 100

92 71.6 57.4 86.2

138 112.5 91.6 137.4









#47
Dimethylphthalate
Concen: 20.501 ng/ul
RT: 13.746 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:163 Resp: 623039
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 22.031 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:165 Resp: 125681
Ion Ratio Lower Upper
165 100
89 54.2 44.7 67.1
121 20.1 17.0 25.4









#50
 Acenaphthylene
 Concen: 21.721 ng/ul
 RT: 13.998 min Scan# 1872
 Delta R.T. -0.006 min
 Lab File: BN031131.D
 Acq: 22 Apr 2024 21:33




Instrument : BNA_N
 ClientSampleId : SSTD020374


Tgt Ion:	152	Resp:	846983
Ion Ratio	Lower	Upper	
152	100		
151	20.0	15.8	23.8
153	12.7	10.5	15.7



#51
 3-Nitroaniline
 Concen: 22.237 ng/ul
 RT: 14.198 min Scan# 1906
 Delta R.T. 0.000 min
 Lab File: BN031131.D
 Acq: 22 Apr 2024 21:33

Tgt Ion:	138	Resp:	145636
Ion Ratio	Lower	Upper	
138	100		
108	12.0	11.1	16.7
92	110.7	93.9	140.9





#52

Acenaphthene

Concen: 21.318 ng/ul

RT: 14.346 min Scan# 1931

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:153 Resp: 551064

Ion Ratio Lower Upper

153 100

152 47.3 36.3 54.5

154 87.6 72.4 108.6



#53

2,4-Dinitrophenol

Concen: 17.358 ng/ul

RT: 14.404 min Scan# 1941

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:184 Resp: 56000


Ion Ratio Lower Upper

184 100

63 66.6 50.4 75.6

154 65.5 53.0 79.6





#55

4-Nitrophenol

Concen: 21.078 ng/ul

RT: 14.510 min Scan# 1959

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:109 Resp: 105176

Ion Ratio Lower Upper

109 100

139 134.4 105.6 158.4

65 120.4 98.2 147.2



#56

Dibenzofuran

Concen: 21.303 ng/ul

RT: 14.681 min Scan# 1988

Delta R.T. -0.006 min

Lab File: BN031131.D




Acq: 22 Apr 2024 21:33


Tgt Ion:168 Resp: 761361

Ion Ratio Lower Upper

168 100

139 39.7 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 22.107 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:165 Resp: 191626
Ion Ratio Lower Upper
165 100
63 40.4 33.9 50.9
182 2.8 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.551 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:232 Resp: 159319
Ion Ratio Lower Upper
232 100
131 47.5 41.2 61.8
130 2.5 2.2 3.2
166 31.3 27.3 40.9





#59

Diethylphthalate

Concen: 21.225 ng/ul

RT: 15.116 min Scan# 2062

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:149 Resp: 663379

Ion Ratio Lower Upper

149 100

177 22.2 16.8 25.2

150 12.2 10.1 15.1



#61

Fluorene

Concen: 21.475 ng/ul

RT: 15.334 min Scan# 2099

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:166 Resp: 631616


Ion Ratio Lower Upper

166 100

165 98.4 74.6 112.0

167 13.6 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 21.562 ng/ul
RT: 15.334 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:204 Resp: 302062
Ion Ratio Lower Upper
204 100
206 34.1 26.4 39.6
141 65.5 50.7 76.1



#63
4-Nitroaniline
Concen: 23.942 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:138 Resp: 169055
Ion Ratio Lower Upper
138 100
92 50.8 38.4 57.6
108 80.6 63.8 95.8









#66
4,6-Dinitro-2-methylphenol
Concen: 20.450 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:198 Resp: 103091
Ion Ratio Lower Upper
198 100
182 4.9 4.0 6.0
77 27.5 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 22.205 ng/ul
RT: 15.551 min Scan# 2136
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:169 Resp: 553831
Ion Ratio Lower Upper
169 100
168 66.6 52.6 79.0
167 34.1 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 21.926 ng/ul
RT: 16.228 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:248 Resp: 192959
Ion Ratio Lower Upper
248 100
250 105.5 78.2 117.2
141 84.1 71.4 107.0



#69
Hexachlorobenzene
Concen: 21.500 ng/ul
RT: 16.334 min Scan# 2269
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:284 Resp: 218121
Ion Ratio Lower Upper
284 100
142 36.2 30.0 45.0
249 29.3 25.5 38.3





#70

Atrazine

Concen: 22.713 ng/ul

RT: 16.504 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:200 Resp: 198889

Ion Ratio Lower Upper

200 100

173 27.8 20.6 31.0

215 49.9 37.6 56.4



#71

Pentachlorophenol

Concen: 22.177 ng/ul

RT: 16.681 min Scan# 2328

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:266 Resp: 146179


Ion Ratio Lower Upper

266 100

264 66.6 49.1 73.7

268 67.4 49.2 73.8









#72
Phenanthrene
Concen: 21.419 ng/uI
RT: 17.075 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33




Instrument :
BNA_N
ClientSampleId :
SSTD020374


Tgt Ion:178 Resp: 1054221
Ion Ratio Lower Upper
178 100
179 16.0 12.4 18.6
176 20.3 14.6 21.8



#74
Anthracene
Concen: 22.213 ng/uI
RT: 17.163 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BN031131.D
Acq: 22 Apr 2024 21:33

Tgt Ion:178 Resp: 1106411
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.4
176 18.1 14.4 21.6









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 22.394 ng/uL
 RT: 13.069 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BN031131.D
 Acq: 22 Apr 2024 21:33

Instrument :
 BNA_N
 ClientSampleId :
 SST020374

Tgt Ion:216 Resp: 244426




Ion	Ratio	Lower	Upper
216	100		
214	78.5	63.0	94.6
218	48.6	38.3	57.5




#76
 Pentachlorobenzene
 Concen: 22.039 ng/uL
 RT: 14.598 min Scan# 1974
 Delta R.T. 0.000 min
 Lab File: BN031131.D
 Acq: 22 Apr 2024 21:33

Tgt Ion:250 Resp: 248403

Ion	Ratio	Lower	Upper
250	100		
248	65.8	49.3	73.9
252	66.6	52.9	79.3





#77

Carbazole

Concen: 23.088 ng/ul

RT: 17.439 min Scan# 2457

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:167 Resp: 1050476

Ion Ratio Lower Upper

167 100

166 21.2 16.2 24.2

139 13.6 10.6 15.8



#78

Di-n-butylphthalate

Concen: 23.023 ng/ul

RT: 18.004 min Scan# 2553

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:149 Resp: 1276787


Ion Ratio Lower Upper

149 100

150 9.5 7.1 10.7

104 5.0 3.9 5.9





#80

Fluoranthene

Concen: 22.609 ng/u1

RT: 19.098 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:202 Resp: 1366267

Ion Ratio Lower Upper

202 100

101 10.9 9.0 13.6

100 8.4 6.9 10.3



#82

Pyrene

Concen: 22.407 ng/u1

RT: 19.463 min Scan# 2801

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:202 Resp: 1422678


Ion Ratio Lower Upper

202 100

101 13.2 11.1 16.7

100 10.4 9.3 13.9





#83

Butylbenzylphthalate

Concen: 24.088 ng/ul

RT: 20.375 min Scan# 2956

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:149 Resp: 674866

Ion Ratio Lower Upper

149 100

91 70.3 57.3 85.9

206 18.7 16.2 24.4



#84

3,3'-Dichlorobenzidine

Concen: 24.413 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:252 Resp: 528566


Ion Ratio Lower Upper

252 100

254 61.2 54.5 81.7

126 10.9 9.8 14.6





#85

Benzo(a)anthracene

Concen: 22.243 ng/ul

RT: 21.251 min Scan# 3105

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:228 Resp: 1569455

Ion Ratio Lower Upper

228 100

229 19.8 16.7 25.1

226 26.2 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 23.999 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:149 Resp: 1009535


Ion Ratio Lower Upper

149 100

167 27.6 22.3 33.5

279 4.0 3.0 4.6





#87

Chrysene

Concen: 21.582 ng/ul

RT: 21.310 min Scan# 3114

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:228 Resp: 1445469

Ion Ratio Lower Upper

228 100

226 29.2 22.6 34.0

229 20.4 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 25.037 ng/ul




RT: 22.280 min Scan# 3280


Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Tgt Ion:149 Resp: 1938132





#90

Benzo(b)fluoranthene

Concen: 22.308 ng/ul

RT: 23.251 min Scan# 3445

Delta R.T. -0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:252 Resp: 1685705

Ion Ratio Lower Upper

252 100

253 20.3 17.6 26.4

125 11.4 8.6 12.8



#91

Benzo(k)fluoranthene

Concen: 22.564 ng/ul

RT: 23.316 min Scan# 3456

Delta R.T. 0.006 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:252 Resp: 1714816


Ion Ratio Lower Upper

252 100

253 21.0 17.4 26.0

125 10.3 9.4 14.2





#93

Benzo(a)pyrene

Concen: 21.813 ng/ul

RT: 24.033 min Scan# 3

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:252 Resp: 1644295

Ion Ratio Lower Upper

252 100

253 20.8 17.7 26.5

125 12.7 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 17.463 ng/ul

RT: 27.421 min Scan# 4154

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33




Tgt Ion:276 Resp: 1633496


Ion Ratio Lower Upper

276 100

138 22.1 17.4 26.2

277 23.7 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 20.601 ng/ul

RT: 27.474 min Scan# 4163

Delta R.T. -0.012 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Instrument :

BNA_N

ClientSampleId :

SSTD020374





Tgt Ion:278 Resp: 1581708

Ion Ratio Lower Upper

278 100

139 17.6 13.3 19.9

279 22.1 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 20.040 ng/ul

RT: 28.439 min Scan# 4327

Delta R.T. 0.000 min

Lab File: BN031131.D

Acq: 22 Apr 2024 21:33

Tgt Ion:276 Resp: 1509621

Ion Ratio Lower Upper

276 100

138 22.8 19.0 28.6

277 22.0 19.3 28.9

