

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042224\
 Data File : BN031160.D
 Acq On : 23 Apr 2024 18:57
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Quant Time: Apr 23 22:41:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.628	152	147995	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.410	136	677614	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.281	164	457640	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.028	188	1041357	20.000	ng/u1	0.00
79) Chrysene-d12	21.263	240	1158561	20.000	ng/u1	0.00
88) Perylene-d12	24.162	264	1334950	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	25442	8.032	ng/uL	0.00
4) Pyridine-d5	3.552	84	180841	19.823	ng/u1	0.00
7) Phenol-d5	6.805	99	253164	20.278	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.975	67	146258	20.136	ng/u1	0.00
11) 2-Chlorophenol-d4	7.163	132	200205	20.596	ng/u1	0.00
15) 4-Methylphenol-d8	8.340	113	210591	20.058	ng/u1	0.00
21) Nitrobenzene-d5	8.787	128	103116	21.180	ng/u1	0.00
24) 2-Nitrophenol-d4	9.504	143	109098	21.557	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.034	165	212022	21.245	ng/u1	0.00
31) 4-Chloroaniline-d4	10.557	131	319389	20.479	ng/u1	0.00
46) Dimethylphthalate-d6	13.692	166	647716	20.339	ng/u1	0.00
49) Acenaphthylene-d8	13.969	160	774942	21.474	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	133781	20.127	ng/u1	0.00
60) Fluorene-d10	15.275	176	571874	20.628	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.398	200	99131	18.868	ng/u1	0.00
73) Anthracene-d10	17.128	188	951082	21.071	ng/u1	0.00
81) Pyrene-d10	19.427	212	1186646	20.844	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.968	264	1386579	21.903	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.175	88	26694	7.952	ng/uL	91
5) Pyridine	3.569	79	186595	20.269	ng/u1	97
6) Benzaldehyde	6.781	77	149655	24.413	ng/u1	98
8) Phenol	6.834	94	260023	20.202	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.063	93	209144	20.594	ng/u1	97
12) 2-Chlorophenol	7.193	128	213317	21.112	ng/u1	98
13) 2-Methylphenol	8.075	108	200887	20.140	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.157	45	274166	20.706	ng/u1	100
16) Acetophenone	8.452	105	329651	20.265	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.440	70	159528	19.675	ng/u1	97
18) 4-Methylphenol	8.399	108	222105	19.879	ng/u1	100
19) Hexachloroethane	8.699	117	87657	20.907	ng/u1	99
22) Nitrobenzene	8.828	77	245703	21.383	ng/u1	100
23) Isophorone	9.351	82	473647	20.198	ng/u1	99
25) 2-Nitrophenol	9.534	139	121994	21.788	ng/u1	94
26) 2,4-Dimethylphenol	9.598	107	237339	20.784	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.834	93	293127	20.580	ng/u1	98
29) 2,4-Dichlorophenol	10.063	162	209461	21.013	ng/u1	98
30) Naphthalene	10.463	128	737505	21.001	ng/u1	99
32) 4-Chloroaniline	10.581	127	310435	20.486	ng/u1	100
33) Hexachlorobutadiene	10.740	225	123775	20.369	ng/u1	95
34) Caprolactam	11.357	113	73493	18.713	ng/u1	99
35) 4-Chloro-3-methylphenol	11.710	107	237767	20.573	ng/u1	95
36) 2-Methylnaphthalene	12.081	142	506622	20.457	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN042224\
 Data File : BN031160.D
 Acq On : 23 Apr 2024 18:57
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Quant Time: Apr 23 22:41:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 19 22:56:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.304	142	516032	20.593	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.451	216	259713	21.688	ng/ul	99
40) Hexachlorocyclopentadiene	12.428	237	113353	17.232	ng/ul	97
41) 2,4,6-Trichlorophenol	12.698	196	170287	21.364	ng/ul	94
42) 2,4,5-Trichlorophenol	12.769	196	192298	21.760	ng/ul	99
43) 1,1'-Biphenyl	13.110	154	667714	21.214	ng/ul	99
44) 2-Chloronaphthalene	13.145	162	534047	21.234	ng/ul	97
45) 2-Nitroaniline	13.363	65	151678	21.845	ng/ul	98
47) Dimethylphthalate	13.739	163	657299	20.007	ng/ul	100
48) 2,6-Dinitrotoluene	13.863	165	131126	21.262	ng/ul	95
50) Acenaphthylene	13.998	152	882896	20.944	ng/ul	100
51) 3-Nitroaniline	14.192	138	151079	21.339	ng/ul	97
52) Acenaphthene	14.339	153	583831	20.892	ng/ul	98
53) 2,4-Dinitrophenol	14.404	184	56322	16.149	ng/ul	97
55) 4-Nitrophenol	14.504	109	109215	20.246	ng/ul	99
56) Dibenzofuran	14.681	168	806507	20.874	ng/ul	99
57) 2,4-Dinitrotoluene	14.651	165	200327	21.378	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.910	232	162042	21.217	ng/ul	93
59) Diethylphthalate	15.116	149	689257	20.400	ng/ul	99
61) Fluorene	15.328	166	662863	20.847	ng/ul	95
62) 4-Chlorophenyl-phenyle...	15.328	204	309896	20.462	ng/ul	95
63) 4-Nitroaniline	15.363	138	163655	21.439	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.416	198	107981	19.141	ng/ul	95
67) N-Nitrosodiphenylamine	15.545	169	571306	20.468	ng/ul	97
68) 4-Bromophenyl-phenylether	16.228	248	208259	21.147	ng/ul	89
69) Hexachlorobenzene	16.328	284	239229	21.071	ng/ul	98
70) Atrazine	16.504	200	198341	20.240	ng/ul	96
71) Pentachlorophenol	16.675	266	142866	19.368	ng/ul	97
72) Phenanthrene	17.069	178	1148139	20.845	ng/ul	99
74) Anthracene	17.163	178	1170942	21.007	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.063	216	258905	21.196	ng/ul	99
76) Pentachlorobenzene	14.592	250	260919	20.686	ng/ul	96
77) Carbazole	17.439	167	1083882	21.288	ng/ul	97
78) Di-n-butylphthalate	18.004	149	1298053	20.916	ng/ul	99
80) Fluoranthene	19.092	202	1396657	20.728	ng/ul	97
82) Pyrene	19.457	202	1474778	20.832	ng/ul	100
83) Butylbenzylphthalate	20.374	149	648323	20.754	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.180	252	514277	21.303	ng/ul	97
85) Benzo(a)anthracene	21.245	228	1628443	20.699	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.180	149	989513	21.097	ng/ul	98
87) Chrysene	21.310	228	1514689	20.283	ng/ul	97
89) Di-n-octyl phthalate	22.274	149	1803218	23.434	ng/ul	100
90) Benzo(b)fluoranthene	23.251	252	1712121	22.794	ng/ul	97
91) Benzo(k)fluoranthene	23.310	252	1669507	22.099	ng/ul	99
93) Benzo(a)pyrene	24.027	252	1612607	21.521	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	27.403	276	718366	7.726	ng/ul	96
95) Dibenzo(a,h)anthracene	27.474	278	1485520	19.464	ng/ul	97
96) Benzo(g,h,i)perylene	28.427	276	1391364	18.582	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.952 ng/uL
 RT: 3.175 min Scan# 31
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376


Tgt Ion:	88	Resp:	26694
Ion	Ratio	Lower	Upper
88	100		
43	28.2	22.2	33.2
58	72.1	50.2	75.2



#5
 Pyridine
 Concen: 20.269 ng/uL
 RT: 3.569 min Scan# 99
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Tgt Ion:	79	Resp:	186595
Ion	Ratio	Lower	Upper
79	100		
52	72.9	59.9	89.9
51	34.0	29.0	43.6









#6
Benzaldehyde
Concen: 24.413 ng/ul
RT: 6.781 min Scan# 645
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion: 77 Resp: 149655
Ion Ratio Lower Upper
77 100
105 99.0 80.0 120.0
106 96.4 75.0 112.6



#8
Phenol
Concen: 20.202 ng/ul
RT: 6.834 min Scan# 654
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion: 94 Resp: 260023
Ion Ratio Lower Upper
94 100
65 28.0 22.0 33.0
66 36.5 29.8 44.6









#10
Bis(2-Chloroethyl)ether
Concen: 20.594 ng/ul
RT: 7.063 min Scan# 693
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion: 93 Resp: 209144
Ion Ratio Lower Upper
93 100
63 72.5 57.1 85.7
95 34.1 24.9 37.3



#12
2-Chlorophenol
Concen: 21.112 ng/ul
RT: 7.193 min Scan# 715
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:128 Resp: 213317
Ion Ratio Lower Upper
128 100
64 43.3 34.9 52.3
130 31.8 27.0 40.4









#13
2-Methylphenol
Concen: 20.140 ng/ul
RT: 8.075 min Scan# 865
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:108 Resp: 200887
Ion Ratio Lower Upper
108 100
107 91.2 70.1 105.1



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.706 ng/ul
RT: 8.157 min Scan# 879
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion: 45 Resp: 274166
Ion Ratio Lower Upper
45 100
77 15.4 12.5 18.7
79 11.8 9.6 14.4









#16
 Acetophenone
 Concen: 20.265 ng/ul
 RT: 8.452 min Scan# 91
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57




Instrument :
 BNA_N
 ClientSampleId :
 SST020376


Tgt Ion:105 Resp: 329651
 Ion Ratio Lower Upper
 105 100
 77 78.8 60.2 90.4
 51 26.1 20.6 31.0



#17
 N-Nitroso-di-n-propylamine
 Concen: 19.675 ng/ul
 RT: 8.440 min Scan# 927
 Delta R.T. -0.000 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Tgt Ion: 70 Resp: 159528
 Ion Ratio Lower Upper
 70 100
 42 53.3 45.0 67.6
 101 10.5 8.1 12.1
 130 25.0 19.2 28.8





#18

4-Methylphenol

Concen: 19.879 ng/ul

RT: 8.399 min Scan# 91

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :





SSTD020376

Tgt Ion:108 Resp: 222105

Ion Ratio Lower Upper

108 100

107 112.1 89.8 134.8



#19

Hexachloroethane

Concen: 20.907 ng/ul

RT: 8.699 min Scan# 971

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:117 Resp: 87657


Ion Ratio Lower Upper

117 100

201 98.2 77.4 116.0

199 61.2 49.6 74.4









#22
Nitrobenzene
Concen: 21.383 ng/ul
RT: 8.828 min Scan# 993
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion: 77 Resp: 245703
Ion Ratio Lower Upper
77 100
123 47.5 38.2 57.2
65 13.2 10.7 16.1



#23
Isophorone
Concen: 20.198 ng/ul
RT: 9.351 min Scan# 1082
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion: 82 Resp: 473647
Ion Ratio Lower Upper
82 100
95 7.3 5.9 8.9
138 19.0 14.7 22.1





#25

2-Nitrophenol

Concen: 21.788 ng/ul

RT: 9.534 min Scan# 1113

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:139 Resp: 121994

Ion Ratio Lower Upper

139 100

65 48.3 35.2 52.8

109 29.5 25.4 38.0



#26

2,4-Dimethylphenol

Concen: 20.784 ng/ul

RT: 9.598 min Scan# 1124

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:107 Resp: 237339


Ion Ratio Lower Upper

107 100

121 49.7 40.3 60.5

122 87.3 69.8 104.8





#27

Bis(2-Chloroethoxy)methane

Concen: 20.580 ng/ul

RT: 9.834 min Scan# 1164

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion: 93 Resp: 293127

Ion Ratio Lower Upper

93 100

95 33.1 25.8 38.6

123 11.7 9.8 14.6



#29

2,4-Dichlorophenol

Concen: 21.013 ng/ul

RT: 10.063 min Scan# 1203

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion: 162 Resp: 209461


Ion Ratio Lower Upper

162 100

164 65.7 51.0 76.6

98 35.4 28.6 42.8





#30

Naphthalene

Concen: 21.001 ng/ul

RT: 10.463 min Scan# 1270

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:128 Resp: 737505

Ion Ratio Lower Upper

128 100

129 10.6 9.0 13.4

127 13.3 10.5 15.7



#32

4-Chloroaniline

Concen: 20.486 ng/ul

RT: 10.581 min Scan# 1291

Delta R.T. -0.006 min

Lab File: BN031160.D




Acq: 23 Apr 2024 18:57


Tgt Ion:127 Resp: 310435

Ion Ratio Lower Upper

127 100

129 32.0 25.4 38.2









#33
Hexachlorobutadiene
Concen: 20.369 ng/ul
RT: 10.740 min Scan# 11
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:225 Resp: 123775
Ion Ratio Lower Upper
225 100
223 62.3 49.8 74.6
227 68.2 48.8 73.2



#34
Caprolactam
Concen: 18.713 ng/ul
RT: 11.357 min Scan# 1423
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:113 Resp: 73493
Ion Ratio Lower Upper
113 100
55 138.2 112.4 168.6
56 110.1 88.0 132.0





#35

4-Chloro-3-methylphenol

Concen: 20.573 ng/ul

RT: 11.710 min Scan# 1483

Delta R.T. -0.000 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:107 Resp: 237767

Ion Ratio Lower Upper

107 100

144 27.5 20.6 31.0

142 79.7 67.4 101.2



#36

2-Methylnaphthalene

Concen: 20.457 ng/ul

RT: 12.081 min Scan# 1546

Delta R.T. -0.006 min

Lab File: BN031160.D




Acq: 23 Apr 2024 18:57


Tgt Ion:142 Resp: 506622

Ion Ratio Lower Upper

142 100

141 86.5 69.6 104.4





#37

1-Methylnaphthalene

Concen: 20.593 ng/ul

RT: 12.304 min Scan# 1584

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:142 Resp: 516032

Ion Ratio Lower Upper

142 100

141 91.9 70.2 105.4

116 3.8 3.1 4.7



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.688 ng/ul

RT: 12.451 min Scan# 1609

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Tgt Ion:216 Resp: 259713




Ion Ratio Lower Upper


216 100

214 77.5 63.1 94.7

179 22.3 17.0 25.6

108 17.9 14.2 21.4





#40

Hexachlorocyclopentadiene

Concen: 17.232 ng/ul

RT: 12.428 min Scan# 1605

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:237 Resp: 113353

Ion Ratio Lower Upper

237 100

235 57.5 47.6 71.4

272 12.3 10.8 16.2



#41

2,4,6-Trichlorophenol

Concen: 21.364 ng/ul

RT: 12.698 min Scan# 1651

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:196 Resp: 170287


Ion Ratio Lower Upper

196 100

198 94.8 69.9 104.9

200 30.2 23.4 35.2





#42

2,4,5-Trichlorophenol

Concen: 21.760 ng/ul

RT: 12.769 min Scan# 1663

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:196 Resp: 192298

Ion Ratio Lower Upper

196 100

198 94.3 75.9 113.9

200 30.5 24.7 37.1



#43

1,1'-Biphenyl

Concen: 21.214 ng/ul

RT: 13.110 min Scan# 1721

Delta R.T. -0.000 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:154 Resp: 667714


Ion Ratio Lower Upper

154 100

153 39.3 31.8 47.8

76 12.1 10.6 15.8





#44

2-Chloronaphthalene

Concen: 21.234 ng/ul

RT: 13.145 min Scan# 1727

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:162 Resp: 534047

Ion Ratio Lower Upper

162 100

164 33.0 25.8 38.8

127 39.6 33.7 50.5



#45

2-Nitroaniline

Concen: 21.845 ng/ul

RT: 13.363 min Scan# 1764

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion: 65 Resp: 151678


Ion Ratio Lower Upper

65 100

92 71.3 57.4 86.2

138 111.3 91.6 137.4









#47
Dimethylphthalate
Concen: 20.007 ng/ul
RT: 13.739 min Scan# 1828
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:163 Resp: 657299
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.6
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.262 ng/ul
RT: 13.863 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:165 Resp: 131126
Ion Ratio Lower Upper
165 100
89 60.1 44.7 67.1
121 23.3 17.0 25.4









#50
Acenaphthylene
Concen: 20.944 ng/ul
RT: 13.998 min Scan# 1872
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:152 Resp: 882896
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.2 10.5 15.7



#51
3-Nitroaniline
Concen: 21.339 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:138 Resp: 151079
Ion Ratio Lower Upper
138 100
108 11.9 11.1 16.7
92 114.7 93.9 140.9









#52
 Acenaphthene
 Concen: 20.892 ng/ul
 RT: 14.339 min Scan# 1930
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Tgt Ion:153 Resp: 583831




Ion	Ratio	Lower	Upper
153	100		
152	46.5	36.3	54.5
154	88.7	72.4	108.6




#53
 2,4-Dinitrophenol
 Concen: 16.149 ng/ul
 RT: 14.404 min Scan# 1941
 Delta R.T. -0.000 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Tgt Ion:184 Resp: 56322

Ion	Ratio	Lower	Upper
184	100		
63	63.4	50.4	75.6
154	70.0	53.0	79.6









#55
4-Nitrophenol
Concen: 20.246 ng/ul
RT: 14.504 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:109 Resp: 109215
Ion Ratio Lower Upper
109 100
139 130.8 105.6 158.4
65 122.5 98.2 147.2



#56
Dibenzofuran
Concen: 20.874 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:168 Resp: 806507
Ion Ratio Lower Upper
168 100
139 38.1 31.0 46.6









#57
2,4-Dinitrotoluene
Concen: 21.378 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:165 Resp: 200327
Ion Ratio Lower Upper
165 100
63 44.9 33.9 50.9
182 2.9 2.6 4.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.217 ng/ul
RT: 14.910 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:232 Resp: 162042
Ion Ratio Lower Upper
232 100
131 45.5 41.2 61.8
130 2.8 2.2 3.2
166 31.2 27.3 40.9









#59
Diethylphthalate
Concen: 20.400 ng/ul
RT: 15.116 min Scan# 2062
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:149 Resp: 689257
Ion Ratio Lower Upper
149 100
177 20.8 16.8 25.2
150 12.2 10.1 15.1



#61
Fluorene
Concen: 20.847 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:166 Resp: 662863
Ion Ratio Lower Upper
166 100
165 98.7 74.6 112.0
167 13.4 10.6 16.0









#62
4-Chlorophenyl-phenylether
Concen: 20.462 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:204 Resp: 309896
Ion Ratio Lower Upper
204 100
206 36.1 26.4 39.6
141 67.7 50.7 76.1



#63
4-Nitroaniline
Concen: 21.439 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:138 Resp: 163655
Ion Ratio Lower Upper
138 100
92 49.3 38.4 57.6
108 79.7 63.8 95.8









#66
4,6-Dinitro-2-methylphenol
Concen: 19.141 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:198 Resp: 107981
Ion Ratio Lower Upper
198 100
182 4.8 4.0 6.0
77 24.8 22.2 33.4



#67
N-Nitrosodiphenylamine
Concen: 20.468 ng/ul
RT: 15.545 min Scan# 2135
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:169 Resp: 571306
Ion Ratio Lower Upper
169 100
168 67.5 52.6 79.0
167 36.6 27.8 41.6









#68
4-Bromophenyl-phenylether
Concen: 21.147 ng/ul
RT: 16.228 min Scan# 211
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:248 Resp: 208259
Ion Ratio Lower Upper
248 100
250 89.5 78.2 117.2
141 76.2 71.4 107.0



#69
Hexachlorobenzene
Concen: 21.071 ng/ul
RT: 16.328 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:284 Resp: 239229
Ion Ratio Lower Upper
284 100
142 39.4 30.0 45.0
249 31.6 25.5 38.3





#70

Atrazine

Concen: 20.240 ng/ul

RT: 16.504 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:200 Resp: 198341

Ion Ratio Lower Upper

200 100

173 26.9 20.6 31.0

215 50.1 37.6 56.4



#71

Pentachlorophenol

Concen: 19.368 ng/ul

RT: 16.675 min Scan# 2327

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:266 Resp: 142866


Ion Ratio Lower Upper

266 100

264 63.8 49.1 73.7

268 64.3 49.2 73.8









#72
 Phenanthrene
 Concen: 20.845 ng/ul
 RT: 17.069 min Scan# 2394
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Instrument :
 BNA_N
 ClientSampleId :
 SST020376

Tgt Ion:178 Resp: 1148139




Ion	Ratio	Lower	Upper
178	100		
179	14.9	12.4	18.6
176	18.7	14.6	21.8




#74
 Anthracene
 Concen: 21.007 ng/ul
 RT: 17.163 min Scan# 2410
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Tgt Ion:178 Resp: 1170942

Ion	Ratio	Lower	Upper
178	100		
179	15.4	12.2	18.4
176	18.8	14.4	21.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.196 ng/uL

RT: 13.063 min Scan# 1713

Delta R.T. -0.012 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:216 Resp: 258905

Ion Ratio Lower Upper

216 100

214 80.1 63.0 94.6

218 46.9 38.3 57.5



#76

Pentachlorobenzene

Concen: 20.686 ng/uL

RT: 14.592 min Scan# 1973

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:250 Resp: 260919


Ion Ratio Lower Upper

250 100

248 60.4 49.3 73.9

252 61.1 52.9 79.3









#77
 Carbazole
 Concen: 21.288 ng/ul
 RT: 17.439 min Scan# 2456
 Delta R.T. -0.000 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020376

Tgt Ion:167 Resp: 1083882




Ion	Ratio	Lower	Upper
167	100		
166	22.2	16.2	24.2
139	13.4	10.6	15.8




#78
 Di-n-butylphthalate
 Concen: 20.916 ng/ul
 RT: 18.004 min Scan# 2553
 Delta R.T. -0.006 min
 Lab File: BN031160.D
 Acq: 23 Apr 2024 18:57

Tgt Ion:149 Resp: 1298053

Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.1	10.7
104	4.9	3.9	5.9





#80

Fluoranthene

Concen: 20.728 ng/ul

RT: 19.092 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:202 Resp: 1396657

Ion Ratio Lower Upper

202 100

101 12.3 9.0 13.6

100 9.6 6.9 10.3



#82

Pyrene

Concen: 20.832 ng/ul

RT: 19.457 min Scan# 2800

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:202 Resp: 1474778


Ion Ratio Lower Upper

202 100

101 13.9 11.1 16.7

100 11.5 9.3 13.9









#83
Butylbenzylphthalate
Concen: 20.754 ng/ul
RT: 20.374 min Scan# 2956
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:149 Resp: 648323
Ion Ratio Lower Upper
149 100
91 66.1 57.3 85.9
206 20.7 16.2 24.4



#84
3,3'-Dichlorobenzidine
Concen: 21.303 ng/ul
RT: 21.180 min Scan# 3093
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:252 Resp: 514277
Ion Ratio Lower Upper
252 100
254 65.4 54.5 81.7
126 12.2 9.8 14.6





#85

Benzo(a)anthracene

Concen: 20.699 ng/ul

RT: 21.245 min Scan# 3104

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:228 Resp: 1628443

Ion Ratio Lower Upper

228 100

229 20.0 16.7 25.1

226 26.4 21.2 31.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.097 ng/ul

RT: 21.180 min Scan# 3093

Delta R.T. -0.000 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:149 Resp: 989513


Ion Ratio Lower Upper

149 100

167 27.0 22.3 33.5

279 4.3 3.0 4.6





#87

Chrysene

Concen: 20.283 ng/ul

RT: 21.310 min Scan# 3115

Delta R.T. -0.000 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:228 Resp: 1514689

Ion Ratio Lower Upper

228 100

226 30.5 22.6 34.0

229 18.7 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 23.434 ng/ul




RT: 22.274 min Scan# 3279


Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Tgt Ion:149 Resp: 1803218









#90
Benzo(b)fluoranthene
Concen: 22.794 ng/ul
RT: 23.251 min Scan# 3444
Delta R.T. -0.006 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57




Instrument :
BNA_N
ClientSampleId :
SSTD020376


Tgt Ion:252 Resp: 1712121
Ion Ratio Lower Upper
252 100
253 20.0 17.6 26.4
125 10.3 8.6 12.8



#91
Benzo(k)fluoranthene
Concen: 22.099 ng/ul
RT: 23.310 min Scan# 3455
Delta R.T. -0.000 min
Lab File: BN031160.D
Acq: 23 Apr 2024 18:57

Tgt Ion:252 Resp: 1669507
Ion Ratio Lower Upper
252 100
253 21.3 17.4 26.0
125 11.3 9.4 14.2





#93

Benzo(a)pyrene

Concen: 21.521 ng/ul

RT: 24.027 min Scan# 3

Delta R.T. -0.006 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:252 Resp: 1612607

Ion Ratio Lower Upper

252 100

253 23.2 17.7 26.5

125 13.1 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 7.726 ng/ul

RT: 27.403 min Scan# 4151

Delta R.T. -0.018 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57




Tgt Ion:276 Resp: 718366


Ion Ratio Lower Upper

276 100

138 22.3 17.4 26.2

277 25.8 18.3 27.5





#95

Dibenzo(a,h)anthracene

Concen: 19.464 ng/ul

RT: 27.474 min Scan# 4163

Delta R.T. -0.012 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Instrument :

BNA_N

ClientSampleId :

SSTD020376





Tgt Ion:278 Resp: 1485520

Ion Ratio Lower Upper

278 100

139 17.7 13.3 19.9

279 23.2 17.0 25.6



#96

Benzo(g,h,i)perylene

Concen: 18.582 ng/ul

RT: 28.427 min Scan# 4325

Delta R.T. -0.012 min

Lab File: BN031160.D

Acq: 23 Apr 2024 18:57

Tgt Ion:276 Resp: 1391364

Ion Ratio Lower Upper

276 100

138 24.4 19.0 28.6

277 23.9 19.3 28.9

