

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN042319\
 Data File : BN005216.D
 Acq On : 23 Apr 2019 16:04
 Operator : JU/SJ
 Sample : K2403-20
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A5132

Quant Time: Apr 24 09:05:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 23 03:50:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.60	152	380950	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	1551131	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.23	164	843361	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.99	188	1718956	20.00	ng/ul	0.00
77) Chrysene-d12	21.22	240	1459860	20.00	ng/ul	0.00
85) Perylene-d12	23.43	264	1561145	20.00	ng/ul	0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.22	96	86490	10.96	ng/uL	0.00
5) Phenol-d5	6.77	99	1010122	33.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.95	67	631905	33.86	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	803408	35.09	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	820067	35.06	ng/ul	0.00
19) Nitrobenzene-d5	8.74	128	394427	41.31	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	426733	44.09	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.98	165	800699	36.09	ng/ul	0.00
29) 4-Chloroaniline-d4	10.50	131	1098718	48.02	ng/ul	0.00
43) Dimethylphthalate-d6	13.65	166	2397263	39.16	ng/ul	0.00
46) Acenaphthylene-d8	13.92	160	2933130	39.01	ng/ul	0.00
51) 4-Nitrophenol-d4	14.43	143	280462	29.01	ng/ul	0.00
57) Fluorene-d10	15.23	176	1960501	38.16	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.35	200	126336	16.34	ng/ul	0.00
70) Anthracene-d10	17.09	188	2859977	39.69	ng/ul	0.00
78) Pyrene-d10	19.40	212	2977654	39.87	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.30	264	2804500	41.11	ng/ul	0.02


Target Compounds				Ovalue
44) Dimethylphthalate	13.70	163	472642	7.748 ng/ul 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN042319\
Data File : BN005216.D
Acq On : 23 Apr 2019 16:04
Operator : JU/SJ
Sample : K2403-20
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
A5132

Quant Time: Apr 24 09:05:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 23 03:50:45 2019
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 7.748 ng/ul

RT: 13.70 min Scan# 1821

Delta R.T. -0.01 min

Lab File: BN005216.D

Acq: 23 Apr 2019 16:04

Instrument :

BNA_N

ClientSampleId :

A5132


Tot Ion:163 Resp: 472642

Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.0

164 9.8 8.1 12.1



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E

