

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN042419\
 Data File : BN005282.D
 Acq On : 25 Apr 2019 10:18
 Operator : JU/SJ
 Sample : K2456-23
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GAHP1

Quant Time: Apr 25 13:00:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 25 03:19:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	425019	20.00	ng/ul	0.00
18) Naphthalene-d8	10.36	136	1918163	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.23	164	1235169	20.00	ng/ul	0.00
61) Phenanthrene-d10	16.98	188	2823024	20.00	ng/ul	0.00
77) Chrysene-d12	21.21	240	2057142	20.00	ng/ul	-0.02
85) Perylene-d12	23.42	264	2011328	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	40855	4.64	ng/uL	0.00
5) Phenol-d5	6.78	99	924761	27.70	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.94	67	561168	26.95	ng/ul	0.00
9) 2-Chlorophenol-d4	7.13	132	746662	29.23	ng/ul	0.00
13) 4-Methylphenol-d8	8.29	113	793342	30.40	ng/ul	0.00
19) Nitrobenzene-d5	8.73	128	392600	33.25	ng/ul	0.00
22) 2-Nitrophenol-d4	9.45	143	454166	37.95	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.98	165	834683	30.42	ng/ul	0.00
29) 4-Chloroaniline-d4	10.49	131	1161940	41.07	ng/ul	0.00
43) Dimethylphthalate-d6	13.65	166	2854054	31.83	ng/ul	0.00
46) Acenaphthylene-d8	13.92	160	3434521	31.19	ng/ul	0.00
51) 4-Nitrophenol-d4	14.43	143	362171	25.58	ng/ul	0.00
57) Fluorene-d10	15.23	176	2435757	32.37	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.35	200	289768	22.82	ng/ul	0.00
70) Anthracene-d10	17.08	188	3804802	32.15	ng/ul	0.00
78) Pyrene-d10	19.39	212	3938806	37.43	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.28	264	2950649	33.57	ng/ul	-0.02


Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
44) Dimethylphthalate	13.69	163	546285	6.115	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN042419\
Data File : BN005282.D
Acq On : 25 Apr 2019 10:18
Operator : JU/SJ
Sample : K2456-23
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GAHP1

Quant Time: Apr 25 13:00:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN041919MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Apr 25 03:19:34 2019
Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 6.115 ng/ul

RT: 13.69 min Scan# 1820

Delta R.T. -0.01 min

Lab File: BN005282.D

Acq: 25 Apr 2019 10:18

Instrument :

BNA_N

ClientSampleId :

GAHP1

Tot Ion:163 Resp: 546285

Ion Ratio Lower Upper

163 100

194 4.5 3.4 5.0

164 9.8 8.1 12.1

